

Analysis of urine, oral fluid and fingerprints by liquid extraction surface analysis coupled to high resolution MS and MS/MS - Opportunities for forensic and biomedical science

Bailey, Melanie J.; Randall, Elizabeth C.; Costa, Catia; Salter, Tara L.; Race, Alan M.; De Puit, Marcel; Koeberg, Mattijs; Baumert, Mark; Bunch, Josephine

DOI

[10.1039/c6ay00782a](https://doi.org/10.1039/c6ay00782a)

Publication date

2016

Document Version

Accepted author manuscript

Published in

Analytical Methods

Citation (APA)

Bailey, M. J., Randall, E. C., Costa, C., Salter, T. L., Race, A. M., De Puit, M., Koeberg, M., Baumert, M., & Bunch, J. (2016). Analysis of urine, oral fluid and fingerprints by liquid extraction surface analysis coupled to high resolution MS and MS/MS - Opportunities for forensic and biomedical science. *Analytical Methods*, 8(16), 3373-3382. <https://doi.org/10.1039/c6ay00782a>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

Analysis of Urine, Oral fluid and Fingerprints by Liquid Extraction Surface Analysis Coupled to High Resolution MS and MS/MS – Opportunities for Forensic and Biomedical Science

Melanie Bailey¹, Elizabeth C. Randall^{2,3}, Catia Costa¹, Tara Salter², Alan M. Race², Marcel de Puit^{5,6}, Mattijs Koeberg⁵, Mark Baumert⁷ and Josephine Bunch^{2,3,4}

¹Department of Chemistry, University of Surrey, Guildford, GU2 7XH, UK

²National Physical Laboratory, Teddington, TW11 0LW, UK

³School of Biosciences, University of Birmingham, Birmingham, B15 2TT, UK

⁴School of Pharmacy, University of Nottingham, Nottingham NG7 2RD

⁵Netherlands Forensic Institute, Laan van Ypenburg 6, 2497 GB, The Hague, The Netherlands

⁶Delft University of Technology, Julianalaan 136, 2628 BL, Delft, The Netherlands

⁷Advion Limited Advion Limited, Harlow Enterprise Hub, Edinburgh Way Harlow, Essex CM20 2NQ, UK

Abstract

Liquid Extraction Surface Analysis (LESA) is a new, high throughput tool for ambient mass spectrometry. A solvent droplet is deposited from a pipette tip onto a surface and maintains contact with both the surface and the pipette tip for a few seconds before being re-aspirated. The technique is particularly suited to the analysis of trace materials on surfaces due to its high sensitivity and low volume of sample removal. In this work, we assess the suitability of LESA for obtaining detailed chemical profiles of fingerprints, oral fluid and urine, which may be used in future for rapid medical diagnostics or metabolomics studies. We further show how LESA can be used to detect illicit drugs and their metabolites in urine, oral fluid and fingerprints. This makes LESA a potentially useful tool in the growing field of fingerprint chemical analysis, which is relevant not only to forensics but also to medical diagnostics. Finally, we show how LESA can be used to detect the explosive material RDX in contaminated artificial fingermarks.

1. Introduction

The relatively recent inception of ambient mass spectrometry methods such as desorption electrospray ionisation (DESI) [1, 2], plasma-assisted desorption ionisation (PADI) [3], direct analysis in real time (DART) [4] and paper spray mass spectrometry [5] has provided an exciting new capability for the high throughput analysis of many types of sample. Because the techniques generally do not require any form of sample preparation and can be carried out under ambient conditions, there has been considerable interest in their adoption for the analysis of biological fluids [6-8].

Liquid Extraction Surface Analysis (LESA) [9-13] is a new tool in surface mass spectrometry, which also operates under ambient conditions and has recently become commercially available. The technique involves the deposition of a solvent droplet from a pipette tip onto a surface. The solvent droplet maintains contact with both the surface and the pipette tip for a defined time period (a few seconds) before being re-aspirated. The resulting sample can either be directly electrosprayed or subjected to further manipulation before mass spectrometry analysis. The technique consumes only a very small volume of sample, making it minimally destructive and suitable for the analysis of trace materials on surfaces. Because the analysis takes place under ambient conditions and does not require sample preparation, high throughput screening of samples is possible. In contrast to other ambient analysis techniques mentioned above, the process of surface sampling and ionisation are decoupled, offering opportunities for superior quantification. Whilst LESA has shown promise for the profiling of proteins, lipids, drugs and metabolites in tissue and blood [10-13], its potential application to the chemical profiling of other biological fluids has not yet been considered.

Chemical profiling of biological fluids is a growing research area, with applicability to metabolomics, medical diagnostics, toxicology and roadside drug testing. For these applications a variety of biological matrices may be used or have been proposed for use, with complementary features. In medical diagnostics, urine is widely used for the detection of disease and for therapeutic drug monitoring [14]. In the forensics field, oral fluid is becoming the matrix of choice for roadside testing due to its non-invasive nature, but is rapidly gaining acceptance as a suitable matrix for medical diagnostics [15-16].

Another biological material of interest is a latent fingerprint. The field of fingerprint chemical analysis is also a rapidly growing research area, starting from a recent paradigm shift in the forensic community for using fingerprints for more than just their ridge detail [21-29]. It has been previously shown that the chemical composition relates to the age of a latent fingerprint [28], the sex of the donor [24], their drug habits or medical history [17-20], the way they interact with development reagents [21] and might even be used to confirm their identity using their amino acid profile [29] or to verify their location on questioned documents [23]. This may be useful when a fingerprint is smudged or details of the donor are not contained within the fingerprint database. Many techniques have been proposed for these applications, including spectroscopic techniques [30], which lack the selectivity of mass spectrometry methods; chromatography based approaches [22,29] which require considerable sample preparation and consume at least an entire fingerprint; and vacuum based methods [20-25], which have lower sample throughput and have been shown to degrade fingerprint chemistry [31].

Here, we show how LESA coupled to high resolution mass spectrometry can be used to rapidly detect a wide array of molecules, including ingested drugs and their metabolites, amino acids, fatty acids, peptides and other endogenous compounds in urine, oral fluid and fingerprint. The wide array of compounds detected from a small sample make the technique an attractive approach for diagnostic assays of the future from any of these matrices.

We also show how LESA can detect an explosive material in contaminated fingerprints. Finally, we will demonstrate how fingerprint analysis using LESA can be carried out in conjunction with higher resolution imaging methods (in this case Matrix Assisted Laser Desorption Ionisation Mass Spectrometry (MALDI)), for a more detailed review of compound distribution.

2. Experimental

Sample collection

Latent fingerprints were obtained from an individual attending a drug and alcohol treatment service to receive treatment for drug dependence. "Natural" fingerprints were deposited onto glass microscope slides at a pressure between 400-1000 g and barcoded, before being shipped to the laboratory for analysis. Two corresponding oral fluid samples were collected from the patient using a Quantisal™ collection kit, Alere Toxicology, UK. One oral fluid sample was screened for amphetamines, benzodiazepines, cannabis, cocaine, methadone and opiates, using a standard immunoassay drug screening procedure at LGC Forensics, Teddington, Middlesex. The second was reserved for LESA analysis. A urine sample was also collected from the patient, using an LGC Forensics urine collection kit. Immediately prior to analysis, 0.2 µL drops of the oral fluid solution and the urine were spotted onto separate glass slides. The fingerprints, second oral fluid sample and the urine were stored in a fridge at 2-5°C and then allowed to reach room temperature prior to analysis. Reference spectra were collected from standards of cocaine, methylecgonine (EME), benzoylecgonine (BZE), methadone, 2-ethylidene-1,5-dimethyl-3,3-diphenylpyrrolidine (EDDP), 2-ethyl-5-methyl-3,3-diphenylpyrrolidine (EMDP), heroin and 6-acetylmorphine (6-AM) immediately prior to analysis. Drug and metabolite standards were purchased from Sigma Aldrich Limited, UK. A favourable ethical opinion for sample collection and analysis was granted from the UK National Research Ethics Service, reference 14/LO/0346.

Latent fingerprints were also collected from an additional female and a male volunteer who were not patients of the drug and alcohol treatment service. The hands of the volunteers were not washed prior to deposition of the fingerprints. A second female volunteer with no drug use history provided a sample of urine and oral fluid.

To investigate the possibility of detecting exogenous compounds in fingerprints, artificial fingerprints with controlled contamination were prepared. A polymeric dummy finger was rubbed over forehead and cheek and subsequently loaded with explosives by rubbing the dummy over the explosive loaded tissue paper. Standard solutions (1 ml, 100 µg/ml) of TNT and RDX were absorbed on regular tissue paper, which was placed in a petridish. The tissue paper was wetted with approximately 2 ml of ethanol. The loaded dummy finger was subsequently placed on different microscope slides to make contaminated fingerprints.

LESA

Materials

Methanol, ethanol and acetonitrile used in preparation of electrospray solvents were purchased from Fisher Scientific (Leicestershire, U.K.). Water was purified by an ELGA Option 3 system (Marlow, UK). Formic acid (FA) was purchased from Sigma-Aldrich Company Ltd. (Dorset, U.K.).

Surface Sampling

Automated sample analysis was performed using the LESA Points software (Advion Ithaca, NY), which controls the TriVersa Nanomate. This platform was used to select the sampling location (x and y co-

ordinates) and the z position, relative to the plate height, for sampling routines using the Nanomate probe. The LESA sampling routine involved the collection of a conductive tip from the Advion tip rack before moving to solvent well containing the electrospray solvent solution. The Nanomate probe aspirated a pre-set volume into the conductive tip. The probe relocated to the predetermined location on the surface then descended to 0.2 mm above the surface. The tip dispensed between 0.5 and 3 μL onto the sample forming a liquid micro-junction (LMJ) between the tip and the surface. The LMJ was maintained to allow sufficient time for analytes to be dissolved into the solution (1-3 seconds). The solvent was then reaspirated into the tip (1-3 μL). Finally the tip was rotated and engaged with the back of the ESI chip, and nanospray ionization was initiated. The Triversa Nanomate was coupled with a Thermo Fisher Scientific Orbitrap Velos mass spectrometer.

	Matrix			Analyte	
	Natural fingerprint	saliva	Urine	Proteins	explosives
solvent composition	A	A	A	B	A
solvent volume (μL)	3	0.7	0.7	2	3
dispense (μL)	2	0.5	0.5	1	2
delay (s)	3	2	2	1	2
aspirate (μL)	3	0.8	0.8	2	3
dispensation height (mm)	0.4	0.4	0.4	0.2	0.4
aspiration height (mm)	0.2	0.2	0.2	0	0.2
delivery time (min)	3	3	3	2.5	3
gas pressure (psi)	0.3	0.3	0.3	0.3	0.3
voltage (kV)	1.4	1.4	1.4	1.4	1.4
Fill time	43 ms	9 ms	3 ms	43 ms	43 ms

Table 1 : Optimised LESA methods used for analysis of fingerprints, saliva, explosives and proteins in fingerprints. Solvent A = 70 : 30 Methanol : Water with 0.1% formic acid; Solvent B = 70:30 methanol : water with 2% formic acid;

Mass Spectrometry

Samples were introduced at a flow rate of ~ 80 nL/min with a gas pressure of 0.3 psi, a tip voltage of 1.4 kV and a capillary temperature of 250 °C. MS data were collected in full scan mode at a resolution of 100 000 at m/z 400. Each scan comprised 3 co-added microscans. For all optimisation experiments the Automatic Gain Control (AGC) was used to accumulate sufficient ions for analysis. The AGC target was 1×10^6 with fill times as indicated in Table 1.

Collision Induced Dissociation (CID)

CID was performed in the linear ion trap with nitrogen gas at a normalised collision energy of 30%. Ions were detected in the Orbitrap at a resolution of 100,000 at m/z 400. The AGC target was 5×10^4 with a maximum fill time of 2 s. Each scan comprised 1 microscan. The precursors selected and major product ion m/z values are listed in Table 2.

Data processing

Mass spectrometry data were analyzed using Xcalibur 2.10 software (Thermo Fisher Scientific) where the Xtract program was used to calculate monoisotopic masses. For all other data processing, data in

the Thermo RAW format were converted to mzML using msconvert as part of ProteoWizard [32]. Sequential spectra that corresponded to the same injection (as determined by time from the total ion chromatogram) were summed together and outputted to mzML using a custom MATLAB script. The summed spectra were then converted to imzML using imzMLConverter [33]. All further data processing was performed in MATLAB using in-house software.

Spectra were queried for a compound of interest by extracting the region of the spectra that spanned 20 ppm error either side of the accurate mass. Orthogonal matching pursuit (OMP) was then used to fit peaks to the reduced spectral region [34]. OMP was applied with 7 iterations and a dictionary composed of Gaussian functions (with standard deviation of 1) and an additional constraint that no fitted peaks could be within 1 standard deviation. The fitted peak with the lowest ppm error was assigned as a match.

MALDI MS imaging

Fingerprint samples were coated with CHCA matrix (5 mg/mL in 80% MeOH, 0.1% TFA) using a TM-Sprayer (HTX Technologies, LLC). 8 coatings were deposited. The matrix was sprayed with a flow rate 0.115 mL/min and track speed of 1333 mm/min. The capillary temperature was 90°C and the spacing between tracks was 3mm.

MALDI MSI data were acquired with a QSTAR XL QqToF instrument with an oMALDI 2 ion source using Analyst QS 1.1 with oMALDI server 5.1 (AB Sciex). An Nd:YAG (Elforlight: FQS-200-1-Y-355) diode pumped solid state (DPSS) laser (Elforlight, Daventry, UK). The laser was triggered from the QSTAR XL instrument using the existing TTL trigger signal in conjunction with a function generator (Thurlby Thander Instruments, TG2000 20MHz DDS) triggering the laser at a repetition rate of 1000 Hz. Images were acquired at 100 x 100 μm , using a raster speed of 0.2 mm s^{-1} .

MALDI MS data were converted from proprietary file format to mzML using AB MS Data Converter (AB Sciex version 1.3) and then converted to imzML using imzMLConverter [33] and using in house software.

Results and Discussion

Chemical profiling of fingerprints, oral fluid and urine

Full scan data for fingerprint, oral fluid and urine samples are shown in Figure 1. The urine spectrum is dominated by the creatinine peak at m/z 114. Table 2 shows the compounds searched for (taken from a list of previously detected substances in the review by Girod et al [26]) and the corresponding peaks detected in fingerprint, urine and oral fluid samples. Other peaks are present at higher abundance in the spectra, but these could not be assigned.

In the fingerprint, 17 amino acids were detected, together with 18 fatty acids and 12 miscellaneous substances. It is remarkable to detect this many substances in a single analysis of a fingerprint. The donor used here was female, and female donors are known to give particularly poor fingerprints [22], showing the great potential for this method for fingerprint chemical profiling. Previous reviews of

fingerprint chemical profiling [22], [26] indicate that a range of analytical techniques have previously been required to survey such a wide range of compounds within a fingerprint. Table 2 shows how LESA coupled to mass spectrometry opens up the possibility of doing this within a single analysis. Surveying such a large number of compounds in an individual fingerprint may assist future investigations designed to determine the age of a fingerprint, the identity of a donor from a smudged fingerprint, or indeed may permit medical diagnostics to be carried out from a fingerprint as a convenient sampling matrix.

Analyte	Monoisotopic Mass (M+H)	Chemical Formula	Fingerprint	Oral Fluid	Urine
Amino Acids					
Glycine	76.0399	C2H5NO2	y	ND	ND
Histidine	156.0773	C6H9N3O2	y	√	√
Serine	106.0504	C3H7NO3	y	√	√
Alanine	90.0555	C3H7NO2	y	√	√
Isoleucine / leucine	132.1025	C6H13NO2	y	√	√
Threonine	120.0661	C4H9NO3	y	√	√
Arginine	175.1195	C6H14N4O2	y	√	√
Tyrosine	182.0817	C9H11NO3	y	√	√
Asparagine	133.0613	C4H8N2O3	ND	√	√
Lysine	147.1134	C6H14N2O2	y	√	√
Valine	118.0868	C5H11NO2	y	√	√
Aspartic Acid	134.0453	C4H7NO4	y	√	√
Methionine	150.0589	C5H11NO2S	y	ND	√
Taurine	126.0225	C2H7NO3S	y	ND	√
Citrulline	176.1035	C6H13N3O3	y	√	√
Ornithine	133.0977	C5H12N2O2	y	√	√
Glutamic Acid	148.061	C5H9NO4	y	√	√
Phenylalanine	166.0868	C9H11NO2	y	√	√
Fatty acids					
Octanoic acid (8:0)	145.123	C8H16O2	√	√	√
Nonanoic acid (9:0)	159.1387	C9H18O2	√	√	√
Decanoic acid (10:0)	173.1543	C10H20O2	√	√	√
Dodecanoic acid (12:0)	201.1856	C12H24O2	√	ND	√
Tridecanoic acid (13:0)	215.2013	C13H26O2	ND	ND	ND
Myristoleic acid (14:1)	227.2013	C14H26O2	√	ND	ND
Myristic acid (14:0)	229.2169	C14H28O2	√	ND	√
Pentadecenoic acid (15:1)	241.2169	C15H28O2	√	√	ND
Pentadecanoic acid (15:0)	243.2326	C15H30O2	√	√	√
Palmitoleic acid (16:1)	255.2326	C16H30O2	√	√	√
Palmitic acid (16:0)	257.2482	C16H32O2	√	√	√
Margaric acid (17:0)	271.2639	C17H34O2	√	√	√
Heptadecenoic acid (17:1)	269.2482	C17H32O2	√	ND	√
Linoleic acid (18:2)	281.2482	C18H32O2	√	ND	ND
Oleic acid (18:1)	283.2639	C18H34O2	√	√	√

Stearic acid (18:0)	285.2795	C18H36O2	√		√
Nonadecanoic acid (19:0)	299.2952	C19H38O2	√	√	ND
Eicosanoic acid (20:0)	313.3108	C20H40O2	√	ND	ND
Heneicosanoic acid (21:0)	327.3265	C21H42O2	ND	ND	ND
Docosanoic acid (22:0)	341.3421	C22H44O2	√	ND	ND
Tricosanoic acid (23:0)	355.3578	C23H46O2	ND	ND	ND
Tetracosanoic acid (24:0)	369.3734	C24H48O2	ND	ND	ND
Miscellaneous					
caffeine	195.0882	C8H10N4O2	√	ND	√
caffeine metabolite, Paraxanthine	181.0726	C7H8N4O2	√	ND	√
cholesterol	369.3521	C27H46O	√	ND	ND
creatinine	114.0667	C4H7N3O	√	√	√
squalene	411.3991	C30H50	√	ND	ND
phenol	95.0497	C6H6O	√	√	√
riboflavin	377.1461	C17H20N4O6	y	ND	ND
Uric Acid	169.0362	C5H4N4O3	√	ND	√
Urea	61.0402	CH4N2O	√	ND	√
Lactic Acid	91.0395	C3H6O3	√	√	√
alpha-Tocopherol	431.3891	C29H50O2	y	ND	ND
delta-Tocopherol	403.3578	C27H46O2	y	ND	ND

Table 2: Compounds detected in a single fingerprint, a sample of urine (3 analyses) and a sample of oral fluid (3 analyses) using LESA; list taken from Girod et al [25]. √=detected with counts >3x the blank; ND = not detected

Similarly, Table 2 demonstrates that an array of compounds including lipids, amino acids and other substances can be detected by LESA in urine and oral fluid. This opens up the possibility of using LESA as a tool to rapidly probe an array of urine and oral fluid samples. In biomarker discovery, the ability to survey a wide array of compounds without targeting particular species is key to a successful outcome. Table 2 demonstrates the ability of LESA to do exactly this, at a comparatively high sample throughput compared with the chromatography methods that are conventionally used in these studies. LESA coupled to high resolution mass spectrometry could therefore add significant value to future investigations in this area.

Drugs in Biological Matrices

Spectra acquired from LESA-MS analysis of fingerprint, oral fluid and urine samples collected from a donor without a drug use history did not reveal any peaks of interest for ions relating to cocaine, methadone, heroin and their respective metabolites.

Analysis of Patient Samples

Oral Fluid

The patient's oral fluid screened positive for cocaine and opiates using the LGC immunoassay test. LESA MS analysis of corresponding oral fluid sample indicated presence of cocaine, EME and BZE, corroborating results from immunoassay screening. However, heroin and 6-AM were not detected in oral fluid by LESA. The oral fluid was suspended in a buffer solution provided as part of the collection kit, which has the effect of considerably diluting the oral fluid sample. No extraction from the buffer was carried out prior to LESA, and this, or differing solubilities of analytes in LESA solvents, may explain the reason for the apparent lack of sensitivity of the LESA method to heroin in this matrix. The sensitivity of LESA for oral fluid could be improved by repeated spotting and drying procedures to increase the volume of sample. Methadone and its metabolites were not detected by either method in oral fluids. Methadone was selected for analysis because the patient was on a methadone treatment programme.

Fingerprints

CID spectra revealed detection of expected product ions, consistent with standards of cocaine, BZE, Heroin and 6-AM, in all four regions of the sampled fingerprint, in direct agreement with the LGC immunoassay test. EME was detected in 3 out of the 4 locations analysed. Example CID spectra are given in Figure 2. No ions relating to methadone or its metabolites were detected. Methadone was not detected via immunoassay measurement. There was therefore good agreement between the LESA fingerprint and immunoassay oral fluid screen for all analytes.

Urine

Cocaine, and associated metabolites, in addition to the 6-AM metabolite were successfully detected in the donor urine sample via LESA-MS and immunoassay. Unsuccessful detection of ions of heroin in the urine of the donor can be explained by a difference in the detection window (varying time of excretion post-administration) for the different matrices studied here. Additionally, it has previously been shown that opiates are detected more readily in sweat than in urine [41].

Once again, methadone and metabolites were not detected using either method in the patient's urine. A failure to detect these compounds in any sample, by either method indicates that, perhaps, the patient had not in fact consumed methadone within a detectable window. Further work must be conducted to establish limits of detection and to better understand these issues.

Table 3 summarises the data obtained from LESA-MS of the patient's oral fluid, urine and fingerprints.

Drug and associated metabolites	Parent ion m/z detected (M+H)⁺	Δ ppm from accurate mass	CID transitions: precursor → product ion	Oral fluid immune assay screen	Fingerprint (LESA)	Oral Fluid (LESA)	Urine (LESA)
Cocaine	304.1538	3.5	304→272, 182, 150, 82	Y	Y	Y	Y
EME	200.1272	5.3	200→182, 82		Y	Y	Y
BZE	290.1377	7.3	290→272, 168, 150		Y	Y	Y

Methadone	ND	N/A	ND	N	N	N	N
EMDP	ND		ND		N	N	N
EDDP	ND		ND		N	N	N
Heroin	370.1640	3.9	370→328, 310,	Y (opiates)	Y	N	N
6-AM	328.1531	5.4	211,193 328→310, 286, 211,193,165		Y	N	Y

Table 3 : Drug and metabolite compounds detected in fingerprint, oral fluid and urine samples from the same donor. Four points on a single fingerprint from a donor were chosen at random and analysed using LESA. Three 0.2 μ L spots of oral fluid in a buffer solution and three 0.2 μ L spots of urine were analysed; the table shows the samples in which the relevant parent and precursor ions were detected. ND = Not Detected

A limitation of the LESA method for fingerprint chemical analysis is the lack of spatial resolution. In drug testing cases it may be necessary to take a sample testing positive for a drug using LESA and to verify the location of a substance within a fingerprint ridge, if a higher level of confirmation is required. To investigate the potential of carrying out this type of analysis, MALDI images of the fingerprint which had previously been analysed using LESA are presented in Figure 3. The images (3b and 3c) of ions detected at m/z 638.6 and 550.6 clearly show the three areas sampled from this fingerprint using LESA, and allow the LESA sampling area to be determined. Figure 3d shows a MALDI MS/MS spectrum acquired during CID of m/z 304 for the area of the fingerprint highlighted in Figure 3a, showing the characteristic fragment ion for cocaine at m/z 182. The image of this fragment ion is presented as an inset, showing some ridge detail. Whilst MALDI has been used before in imaging mode on fingerprints spiked with cocaine [20] and in profiling mode to detect cocaine in the fingerprint of a drug user [19], this appears to be the first report of MALDI imaging of drugs in a fingerprint from a drug user. Whilst quantification of substances in fingerprint residue using MALDI is known to be difficult to achieve [22], a two-step process of (a) quantification of a drug residue in a fingerprint using LESA, and then (b) imaging using MALDI would thereby provide quantification and visualisation of a drug metabolite within a fingerprint ridge.

A different LESA method, as described in Table 1 was developed for the detection of peptides and proteins in fingerprints. Application of this method to a fingerprint from a male donor yielded the spectra presented in Figure 4. Several multiply charged ions are detected, indicating the potential of this method for profiling peptides and proteins directly in latent fingerprints. The spectrum presented in figure 4 shows detection of an ion at m/z 1434.68. This +6 charge state ion reveals that molecules of approximately 8600 daltons are detected using this method. Identifications of proteins and peptides in fingerprints have been made previously by others [26], most notably by Francese's group [24] using MALDI. There are several possible advantages of using LESA compared to techniques such as MALDI. These advantages include improved sensitivity afforded by larger sampling regions; opportunities for using multiple fragmentation techniques within a single analysis (e.g. analysis of multiply charged ions can allow electron transfer dissociation (ETD) as well as CID to be used for improved sequence coverage); and long electrospray survey times achieved by using nano-electrospray flow rates for analysis of microliter volumes, which allows for multiple fragmentation

events. Previous research has shown how hundreds of proteins can be detected from digested tissues and blood spots using LESA coupled to LC/MS [35]. Introducing a chromatographic step into LESA-MS analysis of fingerprints, following sample digestion, will likely enable detection and identification of many proteins from a single location of a single fingerprint and will form the topic of further work.

Detection of Explosives in a Contaminated Fingerprint

Figure 5a shows full scan spectra collected during analysis of a standard of RDX. The mass spectrum shows the presence of a peak at m/z 221.0255, which can be assigned to $[M-H]^-$ [36]. Figure 5b shows that this peak is also present in the fingerprint sample, and Figures 5c and d show that the peak is also not present in either the TNT standard (Figure 5c) or the fingerprints that were blank for RDX (but spiked with TNT) (Figure 5d). In contrast, no peaks corresponding to TNT were found in the contaminated fingerprint sample, implying a lack of sensitivity to this compound, or that the compound was unstable in the fingerprint matrix. However, for RDX, these results do demonstrate the possibility of using LESA to detect explosives in contaminated fingerprints. Other mass spectrometry methods including Direct Analyte Probed nanoextraction (DAPNe) [37], Direct Analysis in Real Time [36], desorption electrospray ionisation (DESI) [6] and MALDI [38] have also been demonstrated for this application. LESA offers complementary features compared with MALDI and DESI as described above for proteins. In the case of explosive detection from fingerprints, the long electrospray time that LESA offers allows multiple CIDs to be carried out from a single spot, leaving the rest of the fingerprint untouched. This is significant if a range of explosive or drug substances are to be scanned for. In contrast to DAPNe, LESA probes a larger area (offering enhanced sensitivity) and has a robotic system for high throughput, automated analysis. Future work should explore the limits of detection of LESA for these analytes in a fingerprint matrix, compare with the other techniques that have been previously demonstrated to show promise for this application and explore the integration of LESA into the forensic workflow.

Conclusions

We have demonstrated the potential for LESA to rapidly gain qualitative information from a range of biofluid samples including urine, oral fluid and fingerprints. The detection of drugs and metabolites in the oral fluid, urine and fingerprints of drug users make LESA an interesting candidate for high throughput drug testing, because the absence of sample preparation and chromatography step decreases the time spent on sample handling and analysis. In contrast to other ambient techniques such as DESI, DART, PADI and paper spray, the process of surface sampling and ionisation is more clearly decoupled. This will likely provide an opportunity for superior quantification, and inclusion of multiple reference or internal standards, and will be considered for these analytes and sample types in further work.

The broad array of compounds detected in a single fingerprint makes LESA a very attractive technique for the study of fingerprint chemistry, an area which is rapidly growing in forensic science, due to the possibility of determining fingerprint age as well as the identity of the donor and the development of superior development reagents. We have shown that because the LESA method only consumes a small area of a fingerprint, it could be used in conjunction with imaging mass spectrometry methods

(e.g. MALDI) which have limited quantitative capability, if confirmation is required that a particular substance is located within a fingerprint ridge.

More broadly, this method of sampling a large number of compounds from either oral fluid, urine and fingerprints is highly relevant to medical diagnostics and metabolomics, because an array of samples can be rapidly probed, allowing high throughput analysis to be carried out.

Acknowledgements

The authors would like to acknowledge the National Institute of Health Research for funding through the Clinical Research Network Portfolio ID 17487 and would like to thank Fiona Robinson and Sarah Hamilton at Surrey and Borders NHS Trust as well as Mahado Ismail from the University of Surrey for their help with the collection of fingerprint, oral fluid and urine samples.

References

- [1] Z. Takats, J. Wiseman, B. Gologan and R. G. Cooks, *Mass Spectrometry Sampling Under Ambient Conditions with Desorption Electrospray Ionisation*, *Science* 2004, 306, 471-473
- [2] D. R. Ifa, N. Manicke, A. Dill and R. G. Cooks, *Latent Fingerprint Chemical Imaging by Mass Spectrometry*, *Science*, 321 (5890), 805
- [3] L. Ratcliffe, F. Rutten, D. Barrett, T. Whitmore, D. Seymour, C. Greenwood, Y. Arandan-Gonzalvo, S. Robinson and M. McCourstra, *Surface Analysis Under Ambient Conditions Using Plasma Assisted Desorption Ionisation Mass Spectrometry* *Anal. Chem.* 2007, 49 (16), 6094-6101
- [4] R. B. Cody, J. Alaramse and H Dupont Durst, *Versatile New Ion Source for the Analysis of Materials in Open Air Under Ambient Conditions*, *Anal. Chem.* 2005, 77 (8), 2297-2302
- [5] H. Wang, J. Liu, R. G. Cooks and Z. Ouyang, *Paper Spray for Direct Analysis of Complex Mixtures Using Mass Spectrometry*, *Angewandte Chemie* 2010, 122 (5), 889-892
- [6] D. Ifa, A. Jackson, G. Paglia and R. G. Cooks, *Forensic Applications of Ambient Ionisation Mass Spectrometry*, *Anal. Bioanal. Chem.* 2009, 394, 1995-2008
- [7] F. M. Green, T. L. Salter, P. Stokes, S. Gimore and G. O'Connor, *Ambient Mass Spectrometry: Advances and Applications in Forensics* *Surface and Interface Analysis* 2010, 42 (5), 347-357
- [8] G. A. Harris, A. S. Gathena and F. M. Fernandez, *Ambient Sampling / Ionisation Mass Spectrometry : Applications, Current trends* *Anal. Chem.* 2011, 83 (12), 4508-4538
- [9] Kertesz, V., Van Berkel, G.J. *Fully automated liquid extraction-based surface sampling and ionization using a chip-based robotic nanoelectrospray platform* *J. Mass Spectrom.* 2010, 45, 252-260
- [10] Marshall, P., Toteu-Djomte, V., Bareille, P., Perry, H., Brown, G., Baumert, M., Biggadike, K. *Correlation of skin blanching and percutaneous absorption for glucocorticoid receptor agonists by*

matrix-assisted laser desorption ionization mass spectrometry imaging and liquid extraction surface analysis with nanoelectrospray ionization mass spectrometry Anal. Chem. 2010, 82, 7787–7794

[11] Parson, W.B., Koeniger, S.L., Johnson, R.W., Erickson, J., Tian, Y., Stedman, C., Schwartz, A., Tarcsa, E., Cole, R., Van Berkel, G.J. *Analysis of chloroquine and metabolites directly from whole-body animal tissue sections by liquid extraction surface analysis (LESA) and tandem mass spectrometry*. J. Mass Spectrom. 2012, 47, 1420–1428

[12] Brown, S.H.J., Huxtable, L.H., Willcox, M.D.P., Blanksby, S.J., Mitchell, T.W.: *Automated surface sampling of lipids from worn contact lenses coupled with tandem mass spectrometry*. Analyst 2013 138, 1316–1320

[13] Edwards, R.L., Creese, A.J., Baumert, M., Griffiths, P., Bunch, J., Cooper, H.J. *Hemoglobin variant analysis via direct surface sampling of dried blood spots coupled with high-resolution mass spectrometry* Anal. Chem. 2011, 83, 2265–2270

[14] Negrusz, A. and Cooper, G. *Clarke's Analytical Forensic Toxicology* 2013, London, UK, Pharmaceutical Press

[15] Drummer, Olaf H *Introduction and Review of Collection Techniques and Applications of Drug Testing of Oral Fluid* Therapeutic Drug Monitoring 2008, 30 (2) 203-206

[16] W. M. Bosker and M. A. Huestis *Oral Fluid Testing for Drugs of Abuse* Clinical Chemistry 2009, 55 (11), 1910-1931

[17] Legget, R, Lee-Smith, E. E., Jickells, S. M. and Russell, D. A. *Intelligent Fingerprinting: simultaneous identification of drug metabolites and individuals by using antibody-functionalised nanoparticles*, Angewandte Chemie 2007, 199, 4179-4181

[18] K. Kuwayama, T. Yamamuro, K. Tsjuikawa, H. Miyaguchi, T. Kanamori, Y. Iwata and H. Inoue, Time-course measurements of drugs and metabolites transferred from fingerprints after drug administration: usefulness of fingerprints for drug testing *Forensic Toxicology* 2014, 32, 235-242

[19] Bailey MJ, Bradshaw R, Francese S, Salter TL, Costa C, Ismail M, P Webb R, Bosman I, Wolff K, de Puit M. (2015) *Rapid detection of cocaine, benzoylecgonine and methylecgonine in fingerprints using surface mass spectrometry* Analyst 2015, 140, 6254 - 6259

[20] G. Groeneveld, M. de Puit, S. Bleay, R. Bradshaw & S. Francese, Detection and mapping of illicit drugs and their metabolites in fingermarks by MALDI MS and compatibility with forensic techniques Scientific Reports 2015, 5, Article number: 11716

[21] Bailey MJ, Ismail M, Bright N, Everson D, Costa C, De Puit M, Bleay S, Elad ML, Cohen Y, Geller B, Webb RP, Watts JF. (2013) 'Enhanced imaging of developed fingerprints using mass spectrometry imaging'. *Analyst*, 138 (21), pp. 6246-6250.

[22] Bailey MJ, Bright NJ, Hinder S, Jones BN, Webb RP, Croxton RS, Francese S, Ferguson LS, Wolstenholme R, Jickells S, Jones BJ, Ojeda JJ, Kazarian SG, Bleay S. (2012) 'Chemical characterization of latent fingerprints by matrix-assisted laser desorption ionization, time-of-flight secondary ion mass spectrometry, mega electron volt secondary mass spectrometry, gas chromatography/mass spectrometry, X-ray photoelectron spectroscopy, and attenuated total reflection Fourier transform infrared spectroscopic imaging: An intercomparison'. *Analytical Chemistry*, 84 (20), pp. 8514-8523

- [23] Bright NJ, Webb R, Hinder SJ, Kirkby KJ, Ward NI, Watts JF, Bleay S, Bailey MJ. (2012) 'Determination of the deposition order of overlapping latent fingerprints and inks using Secondary Ion Mass Spectrometry (SIMS)'. *Anal Chem*, 84 (9), pp. 4083-4087
- [24] FERGUSON, Leesa Susanne, WULFERT, Florian, WOLSTENHOLME, Rosalind, FONVILLE, Judith Marlou, CLENCH, Malcolm, CAROLAN, Vikki Amanda and FRANCESE, Simona (2012). Direct detection of peptides and small proteins in fingerprints and determination of sex by MALDI mass spectrometry profiling. *The Analyst*, 137 (20), 4686-4692
- [25] BRADSHAW, Robert, BLEAY, Stephen, WOLSTENHOLME, Rosalind, CLENCH, Malcolm and FRANCESE, Simona (2013). Towards the integration of matrix assisted laser desorption ionisation mass spectrometry imaging into the current fingerprint examination workflow. *Forensic Science International*, 232 (1-3), 111-124
- [26] A. Girod R. Ramotowski, C. Weyermann, Composition of fingerprint residue: A qualitative and quantitative review, *Forensic Science International*, 2012, 223 (1-3), 10-24
- [27] P. Hazarika and D. A. Russell, *Advances in Fingerprint Analysis*, *Angewandte Chemie* 2012, 51 (15), 3524-3531, 2012
- [28] S. Cadd, M. Islam, P. Manson, S. Bleay, 'Fingerprint composition and aging: A literature review' *Science & Justice* 2014, 55 (4) 219-238
- [29] M de Puit, M Ismail and Xioama Xu, *LCMS Analysis of Fingerprints, the Amino Acid Profile of 20 Donors* *Journal of Forensic Sciences* 2013, 59 (2) 364-370
- [30] C. Ricci, P. Phiriyavityopas, N. Curum, K. L. Chan, S. Jickells and S. G. Kazarian, *Chemical Imaging of Latent Fingerprints*, *Appl Spectrosc.* 2007, 61(5), 514-22.
- [31] N. J. Bright, T. R. Willson, D. Driscoll, S.M. Reddy, R. P. Webb, S. Bleay, N. I. Ward, K. J. Kirkby and M. J. Bailey *Chemical changes exhibited by latent fingerprints after exposure to vacuum conditions* *Forensic science international.* 2013, 230 (1-3), 81-6
- [32] M.C. Chambers et al, *Nature Biotechnology*, 2012, 30, 918-920
- [33] A. M. Race, I. B. Styles, J. Bunch, *Journal of Proteomics*, 2012, 75 (16) 5111-5112
- [34] L Denis *et al* 2009 *Inverse Problems* 25 115017
- [35] J Sarbsby, N J Martin, J Bunch and HJ Cooper, *J Am Soc Mass Spectrom* 2014, 25 (11) 1953-61
- [36] H. Chen & R. Zenobi, *Neutral desorption sampling of biological surfaces for rapid chemical characterization by extractive electrospray ionization mass spectrometry* *Nature Protocols* 2008, 3, 1467 - 1475
- [37] K Clemons, J Dake, E Sisco and G Verbek *Forensic Science International* 2013, 231 (1-3) 98-101
- [38] K Kaplan-Sandquist, M A LeBeu, ML Miller, *Forensic Science International* 2014, 235 68-77

[39] P Kintz, A Tracqui, P Mangin, Y Edel, Sweat Testing in Opioid Users with a Sweat Patch, J Anal Toxicol 1996 20 (6) 393-7

Figures




Figure 1 Full scan mass spectra showing Liquid Extraction Surface Analysis (LESA) of (a) blank, (b) natural fingerprint (c) oral fluid (d) urine




Figure 2 Collision induced Dissociation (CID) spectra produced by liquid extraction surface analysis (LESA) analysis of a fingerprint from a donor attending a drug rehabilitation service, showing fragment ion peaks corresponding to (a) 6-AM (b) heroin (c) cocaine (d) benzylecgonine (e) methylecgonine




Figure 3 Matrix Assisted Laser Desorption Ionisation (MALDI) images of a fingerprint previously analysed using liquid extraction surface analysis (LESA). (a) representation of the areas analysed by LESA; images of (b) m/z 638.6 and (c) m/z 550.6; (d) MS/MS spectrum acquired during CID of m/z 304 for the area of the fingerprint highlighted in the inset, showing the characteristic fragment ion for cocaine at m/z 182.




Figure 4 Full scan mass spectrum produced by the “protein” LESA method developed for fingerprints in the range m/z 1300 to 1500




Figure 5 Full scan spectra for standards of (a) RDX (c) TNT and artificial fingerprints contaminated with (b) RDX and (d) TNT