

Utilizing high-energy γ -photons for high-resolution ^{213}Bi SPECT in mice

De Swart, Jan; Chan, Ho Sze; Goorden, Marlies C.; Morgenstern, Alfred; Bruchertseifer, Frank; Beekman, Freek J.; De Jong, Marion; Konijnenberg, Mark W.

DOI

[10.2967/jnumed.115.157685](https://doi.org/10.2967/jnumed.115.157685)

Publication date

2016

Document Version

Accepted author manuscript

Published in

The Journal of Nuclear Medicine

Citation (APA)

De Swart, J., Chan, H. S., Goorden, M. C., Morgenstern, A., Bruchertseifer, F., Beekman, F. J., De Jong, M., & Konijnenberg, M. W. (2016). Utilizing high-energy γ -photons for high-resolution ^{213}Bi SPECT in mice. *The Journal of Nuclear Medicine*, 57(3), 486-492. <https://doi.org/10.2967/jnumed.115.157685>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

Utilizing high-energy gamma photons for high-resolution ^{213}Bi SPECT in mice.

Jan de Swart¹, Ho Sze Chan¹, Marlies C Goorden², Alfred Morgenstern³, Frank Bruchertseifer³, Freek J Beekman^{2,4,5}, Marion de Jong^{1,6}, Mark W Konijnenberg¹.

¹ Department of Nuclear Medicine, Erasmus Medical Centre, Rotterdam, the Netherlands

² Section Radiation, Detection & Medical Imaging, Delft University of Technology, Delft, the Netherlands

³ European Commission, Joint Research Centre, Institute for Transuranium Elements (ITU), Karlsruhe, Germany

⁴ Millabs B.V., Utrecht, the Netherlands

⁵ Department of Translational Neuroscience, Brain Center Rudolf Magnus, the Netherlands

⁶ Department of Radiology, Erasmus Medical Centre, Rotterdam, the Netherlands

Corresponding author:

Jan de Swart

Department of Nuclear Medicine

Erasmus MC

's Gravendijkwal 230

3015 CE Rotterdam

The Netherlands

Phone: +31 10 7043636

E-mail: j.deswart@erasmusmc.nl

Word count: 4999

Running title: Bi-213 microSPECT imaging

ABSTRACT

The combined alpha, gamma, and X-ray emitter ^{213}Bi (half-life 46 min) is very promising for radionuclide therapy. SPECT imaging of ^{213}Bi is challenging, since the majority of emitted photons has a much higher energy (440 keV) than common in SPECT. We assessed ^{213}Bi imaging capabilities of the Versatile Emission Computed Tomograph (VECTor) dedicated to (simultaneous) preclinical imaging of both SPECT and PET isotopes over a wide photon energy range of 25-600 keV.

Methods: VECTor was equipped with a dedicated clustered pinhole collimator. Both the 79 keV X-rays and 440 keV gamma-rays emitted by ^{213}Bi could be imaged. Phantom experiments were performed to determine the maximum resolution, contrast-to-noise ratio and activity recovery coefficient for different energy window settings. Additionally, imaging of [^{213}Bi -DOTA,Tyr³]octreotate and ^{213}Bi -DTPA in mouse models was performed.

Results: Using 440 keV gamma-rays instead of 79 keV X-rays in image reconstruction strongly improved the resolution (0.75 mm) and contrast-to-noise characteristics. Results obtained with a single 440 keV energy window setting were close to those with a combined 79 keV/440 keV window.

We found a reliable activity recovery coefficient down to 0.240 MBq/mL with 30 minutes imaging time. In a tumor-bearing mouse injected with 3 MBq [^{213}Bi -DOTA,Tyr³]octreotate, tumor uptake could be visualized with a one hour post-mortem scan. Imaging a non-tumor mouse at 5 minute frames after injecting 7.4 MBq ^{213}Bi -DTPA showed renal uptake and urinary clearance, visualizing the renal excretion pathway from cortex to ureter. Quantification of the uptake data allowed kinetic modeling and estimation of the absorbed dose to the kidneys.

Conclusion: It is feasible to image ^{213}Bi down to 0.75 mm resolution by using a SPECT system equipped with a dedicated collimator.

Keywords: Bi-213, SPECT, ultra-high-energy.

INTRODUCTION

New opportunities for high linear energy transfer (LET) radionuclide therapy with the alpha particle emitters ^{225}Ac and ^{213}Bi are increasingly being investigated (1-3). The research for peptide receptor radionuclide therapy with alpha particles is mostly focused on labeling peptides with ^{213}Bi . Not only is its short half-life of 46 minutes in good accordance with the rapid targeting to receptor-positive tumors as well as the rapid clearance of peptides, it also raises less concern for detrimental effects because of the absence of non-specific uptake by daughters detached from its peptide or linker due to alpha decay recoil (4).

^{213}Bi offers the best imaging opportunities through its 440 keV gamma-ray and is therefore important for biodistribution and dosimetry studies (5). All other gamma-rays and X-rays emitted by ^{213}Bi and its daughters are either too low in abundance or in energy to be suitable for imaging, possibly with the exception of the X-rays from ^{213}Bi at 77 and 79 keV if appropriate correction methods for down-scatter of the 440 keV gamma-rays are applied (see table 1 in on-line supplemental data (6)). Patient imaging of the uptake pattern of ^{213}Bi labeled antibody HuM195, targeted to CD33 leukemia, and ^{213}Bi -DOTATOC targeting neuroendocrine tumors has been performed by imaging the 440 keV gamma-ray with high-energy collimators (7-9).

Pre-clinical biodistribution studies with ^{213}Bi are challenging due to its short half-life. Typically ^{213}Bi labeled peptide biodistributions have been determined at 1h and 3h in rats and mice, thereby missing essential information on the kinetics in the uptake phase (2). Dynamic imaging of this uptake phase will show the kinetic pattern, but usually lacks good quantification. Imaging of the high-energy (440 keV) photons is severely compromised for most small animal SPECT systems, due to penetration of pinhole edges and the collimator wall. Recently, however, a new dedicated small animal SPECT system based on use of many clustered pinholes has been developed that enables imaging over an energy range from 25 to 600 keV (10). This system has shown to be able to e.g. image SPECT and PET tracers simultaneously at 0.5 mm and 0.75 mm resolution, respectively (11).

The aim of the study was to investigate the capability to (dynamically) image ^{213}Bi in small animals, e.g. using ^{213}Bi labeled peptides. Resolution, contrast-to-noise ratio and activity recovery coefficient for different energy window settings and combinations thereof were optimized in phantom studies. Subsequently imaging of ^{213}Bi -DTPA and [^{213}Bi -DOTA,Tyr³]octreotate in mice was performed.

MATERIALS AND METHODS

Radiochemistry

^{213}Bi was eluted from a standard $^{225}\text{Ac}/^{213}\text{Bi}$ generator (European Commission, Institute for Transuranium Elements (ITU)). For phantom experiments the elution was not chemically altered. ^{213}Bi was labeled with diethylene triamine pentaacetic acid (DTPA) for renal function imaging. ^{213}Bi eluate was added directly into a ready-for-use solution containing, 0.15M TRIS and 64 μM DTPA, at total volume 800 μL and pH 8.5.

For tumor imaging, ^{213}Bi was labeled to [DOTA,Tyr³]octreotate according to the labeling procedure described earlier (12), the incorporation of the radioactivity was >99%, radiochemical purity was >85%. Specific activity was 14.8 MBq/nmol.

Small Animal Imaging System

The VECTor (MILabs B.V.) uses three gamma cameras in a triangular set up. It enables high-energy gamma photon imaging including single 511 keV photons by using a tungsten collimator with clustered pinholes with relatively small opening angles. This reduces the image-degrading effects of pinhole edge penetration by these high-energy photons. The collimator contains 162 pinholes with a diameter of 0.7 mm grouped in clusters of 4. All clustered pinholes together observe a field-of-view which has the shape of an hourglass with a diameter of 44 mm and an average longitudinal length of 33 mm (10). Total body images are obtained by moving the animal through the scanner along a spiral trajectory (13). Data is collected in list mode.

Image Reconstruction

SPECT images were reconstructed by using projections from all bed positions simultaneously (14) using Pixel-based OSEM (15). Three photopeak energy window settings were tested: a window set at the 440 keV photo peak, a window set at 79 keV, or both energy windows simultaneously. Scatter and background were corrected for with the triple-energy window method (16). The 440 and 79 keV

photo peak windows had two adjacent background windows each (figure 1 in supplemental data).

For reconstruction the standard SPECT system matrix was used (17) for the 79 keV energy window, and a 511 keV system matrix for reconstructions using the 440 keV or the combined energy windows.

Phantom Experiments

A 5 mL syringe (internal diameter: 12.06 mm) was filled with 86.2 MBq ^{213}Bi in a volume of 2.0 mL (fill height: 17.7 mm) for determining the ability to recover different amounts of activity. A volume of interest (VOI) was drawn around the activity in the reconstructed image. On a dynamic scan (20 frames of 30 minutes) the activity concentration at the start of the acquisition was 36.2 MBq/mL ^{213}Bi , at the start of the last frame 0.0062 MBq/mL.

A second dynamic scan (90 frames of 5 minutes) of a ^{213}Bi filled syringe was performed. The initial activity concentration in 2.0 mL within the 5 mL syringe was 25.1 MBq/mL ^{213}Bi and 0.0352 MBq/mL at the start of the last frame.

Reconstructions of both experiments were done with 4 subsets and 30 iterations, voxel size was 0.8 mm. A post-reconstruction filter (3D Gaussian) with 0.4 mm Full Width at Half Maximum (FWHM) was applied. All data were corrected for decay. System performance was characterized with the recovery coefficient (RC), defined as the measured apparent radioactivity concentration divided by the true radioactivity concentration. For large objects and sufficient imaging times RC should equal 1.

Spatial resolution was determined by using a Jaszczak resolution phantom (HR-micro phantom, Vanderwilt Techniques) with hollow channels of 0.7, 0.8, 0.9, 1.0, 1.2 and 1.5 mm diameter. It was filled with 119 MBq ^{213}Bi (activity concentration 198 MBq/mL) and scanned for 45 minutes. Images were reconstructed using 32 subsets, 60 iterations and a voxel size of 0.4 mm. No post-reconstruction filters were used. Profiles were determined from five single slices through the measured signal over a 0.4 mm cross-hair line drawn over the rods. Gaussian curves were fitted to these profiles and averaged, and their FWHM values were reported.

To assess the impact of lower numbers of counts on resolution, we also reconstructed this dataset using only 20%, 5% and 1% of the counts from the list-mode data. This emulates scans with shorter scan times or lower activities. We reconstructed the lowest activity scan with 4 subsets and 30 iterations on 0.4 mm voxels which are the same reconstruction settings as used for the mouse scans. For the higher activities more iterations were needed to recover finer details and thus we chose the same reconstruction settings as for the high-count reconstructions. Images were post-filtered with a 3D Gaussian with a 0.6, 0.9 and 1.0 mm FWHM for scans with 20%, 5% and 1% of the counts respectively.

An analysis of contrast-noise characteristics was performed similar to (18). Resolution phantom images were resampled to a fine 0.05 mm grid, and ROIs with a diameter 90% of the rod diameter were placed on top and in between the rods (Figure 2c in supplemental data). This was repeated over 20 slices. The mean activity inside the ROIs placed over the rods (h_d) and in between the rods (b_d) was determined. The contrast C_d for rod size d is then defined to be

$$C_d = \frac{\overline{h_d - b_d}}{\overline{h_d}} \quad (\text{Eq.1})$$

The variability between the ROI mean values was characterized by the noise parameter N_d :

$$N_d = \frac{\sqrt{\sigma_{h_d}^2 + \sigma_{b_d}^2}}{\overline{rois_d}} \quad (\text{Eq.2})$$

The standard deviation σ of h_d and b_d were calculated over all ROIs in 1/3 of the 20 slices to reduce inter-ROI covariance. The denominator $\overline{rois_d}$ indicates the mean value taken over all ROIs. The contrast-to-noise ratio was defined as C_d/N_d .

Quantification calibration procedure and determination of the calibration factors are described in the supplemental data and based on the method in (19).

Ex-vivo Experiment

A nu/nu mouse bearing a CA20948 tumor xenograft was injected **intravenously (iv) in a tail vein** with 3.0 MBq [$^{213}\text{Bi-DOTA,Tyr}^3$]-octreotate. The mouse was euthanized 38 minutes after injection and immediately imaged for 1 hour. **A CT scan was made following SPECT with the integrated CT scanner (acquisition parameters 55kV and 615 μ A and reconstructed with filtered Back Projection), which was used only as an anatomical reference.** SPECT reconstruction parameters comprised 4 subsets, 30 iterations and a voxel size of 0.4 mm. A post-reconstruction filter (3D Gaussian) of 1.5 mm FWHM was applied. Tumor volume was determined **by drawing VOIs** in the CT-images. **The radioactivity uptake was based on the VOIs drawn in the SPECT images. Organs and tumor tissue were counted for 60 s in a Wallac Wizard gamma counter (PerkinElmer). Counting started 2.3 hours after injection.**

In-vivo experiment

In an in-vivo experiment, a balb/c mouse was injected **iv** in a tail vein with 7.4 MBq $^{213}\text{Bi-DTPA}$ under Isoflurane anaesthesia. Dynamic acquisition of the abdominal region was started 3 minutes after injection over a total of nine 5-minute frames. Equal reconstruction parameters were used as for the ex-vivo experiment. **SPECT-based VOIs** were drawn over the urinary bladder and over both kidneys to determine their kinetics. The absorbed dose to the kidneys was determined by using the sphere model within the Olinda/EXM code (20) and calculated for an average single kidney mass of 0.286 mg., **determined from the mouse of the ex-vivo experiment.**

All animal studies were conducted in accordance with the guidelines and after approval of the Animal Welfare Committee of the Erasmus MC.

Statistics

Statistical analysis of the syringe experiments was performed with the Graphpad Prism software (GraphPad Software, Inc.). Average values of the RC **from ^{213}Bi –filled syringes**, normalized to the initial value, were determined for each frame. **Deviations from the horizontal line were analyzed according to the D'Agostini & Pearson normality test. Additionally the Runs test was performed to**

decide whether the residuals followed a random pattern. The cut-off value for the linearity of RC was determined by the 3σ outlier criterion on the moving average.

RESULTS

Quantitative Properties

On the 30-minute dynamic series, the RC was averaged over the linear range (Figure 1). Cut-off values or lower limits for linearity found are indicated in table 1. The RC was comparable for all energy window settings. The single 440 keV window setting showed the largest linearity range (lower limit: 0.24 MBq/ml).

For the 5-minute dynamic scans, the RC for the 79 keV/440 keV combined energy peaks showed the best results as it remained constant down to 0.32 MBq/mL (Figure 2). Below 2 MBq/mL a more scattered pattern of the data points was visible, but this was not a statistical deviation due to the equal variations in the signal above and under the average RC line. The RC for the 79 keV photo peak on the 5-minute dynamic scans did not meet all statistical tests. Therefore the RC line in this graph can only serve as an indication.

In general the quantitative properties for activity recovery for images using the 440 keV peak and the combined peaks are comparable. Especially in the 5-minute scans these two settings have better quantitative properties than the single 79 keV peak setting. There was no difference in results when a 0.4 mm or a 0.8 mm voxel size was applied in the reconstruction settings (data not shown).

Spatial Resolution

Resolution phantom images are shown in Figure 3. Visually, images using the single 440 keV or the combined 79 keV/440 keV energy window setting appear to be much less noisy and have better resolution than the images using the single 79 keV energy window; on the reconstructions 0.7 mm rods could be distinguished for the combined 79 keV/440 keV setting and for the single 440 keV window. Profiles are shown in Figure 4 and the mean FWHMs for the three energy windows are indicated in Table 1. These values showed a significant difference for the combined 79 keV/440 keV setting compared to the single 79 keV setting. Differences between the 440 keV and the combined 79 keV/440 keV setting were small, but the combined setting resulted in a slightly lower variation,

indicating the most stable settings. Contrast and contrast-to-noise curves for all 3 energy window settings are shown in Figure 5 and are in agreement with the visual assessment; both contrast and contrast-to-noise are much better for the 440 keV or the 79 keV/ 440 keV window settings compared to the single 79 keV energy window. All though 0.7 mm rods could be distinguished in reconstructions, profiles reveal that these are really at the resolution limit. The contrast for these rods was found to be $C_{0.7} = 0.34$ for the 79 keV/440 keV combined window setting. To emulate lower activities, we also reconstructed resolution phantom images with only part of the listmode data used (figure 3, supplemental data) . When only 20%, 5% or 1% of the counts were used respectively, rods that could still be distinguished were 0.8 mm, 0.9 mm and 1.2 mm.

Ex-vivo Experiment

Tumor uptake by [^{213}Bi -DOTA,Tyr³]-octreotate in the euthanized mouse was clearly visualized with the 440 keV and the combined 79 keV/440 keV energy window setting, but not with the 79 keV setting (Figure 5), ^{213}Bi uptake in the kidneys was not visible though. The uptake in the xenograft amounted to 38 kBq, corresponding to a concentration of 0.42 MBq/mL for the 90 mm³ volume tumor. The activity in the abdomen below the tumor corresponded to 0.16 MBq/mL. A region within the mouse on the contralateral side of the tumor, not clearly linked to physiological uptake, showed an uptake of 0.12 MBq/mL. The biodistribution assay of this mouse showed a tumor uptake of 5.3% of the injected activity (%IA) and with activity concentration of 0.36 MBq/g. The uptake in the left kidney was 1.9 %IA (0.48 MBq/g).

In-vivo Experiment

In-vivo mouse maximum intensity projection (mip) images are shown in Figure 6 for the combined energy window as this setting was generally found to be optimal. The first frame in Figure 6 shows activity in both kidneys and bladder after injection with 7.4 MBq ^{213}Bi -DTPA. The filling of the urinary bladder was also visible by its enlargement over time and in the first frame already exceeded the maximum displayed intensity. The activity in the kidneys gradually accumulated in the renal medulla

as visualized in the mip image of Figure 6 and the quantitative VOI-based results in Figure 7.

The peak uptake of in the kidneys was 18%IA. The kidney radioactivity uptake ranged between 0.34 and 0.66 MBq (1.2 and 2.3 MBq/mL). These SPECT-based activity concentrations in the kidneys were found to be well above the 0.32 MBq/mL threshold for linear response.

The decay-corrected radioactivity accumulation in the urinary bladder followed a single-exponential build-up pattern with an 11 ± 2 min half-life. The non-decay corrected uptake data of the kidneys could be fitted with a single exponential curve with an effective clearance half-life of 52 min (95% confidence interval: 36-96 min.). A horizontal line was the preferred fit through the decay corrected kidney data. The residence time for the ^{213}Bi DTPA uptake in the kidneys was 11.7 ± 0.4 min, leading to an absorbed dose of 26 ± 2 Gy by 7.4 MBq. The largest part (94%) of this dose was delivered by the ^{213}Po α -particles.

DISCUSSION

Direct imaging of the 440 keV gamma-rays from ^{213}Bi is possible with dedicated high-energy SPECT, despite the low activity injected to avoid an undesirably high dose to the animal. We found that including the 440 keV setting in image reconstruction is essential; only using 79 keV X-rays has a strong negative effect on image quality. Generally, the results for the 440 keV and the combined 440 keV/79keV setting are close. Resolution and contrast-noise properties were slightly better for the combined setting which also showed better quantification properties for short time frames. However, on longer time frames 440 keV alone performed slightly better. We believe that for low-count studies the combined window performs better as it contains more counts, while for higher count levels the down-scatter of 440 keV gammas in the 79 keV photopeak adversely affects quantification.

Quantification sensitivity of the camera in the 5-minute frame setting was high enough to allow kinetic modeling of the kidney uptake and bladder filling, which is apparent when comparing the results from the 5-minute phantom scan and the results of the in-vivo ^{213}Bi -DTPA scan. Tumor and pancreas modeling should also be possible for most DOTA-conjugated somatostatin analogues as their uptake in mice is in the order of approximately 10%IA/mL (2) to even 225%IA/g for Exendin (21). With an injected activity of 5-10 MBq this will lead to activities in these organs that enable 5-minute frame scanning, but results will become uncertain when the activity concentration will drop below 0.32 MBq/mL. In those cases, longer time frames should be used with the risk of missing essential fast kinetic effects.

The maximum resolution reported here was achieved with a static scan and relatively high activity compared to that injected in mice. This was done to test (maximum) system performance for ^{213}Bi imaging and to investigate optimal energy window settings which is easiest on almost noiseless data. As common in SPECT and shown in the supplemental data, the resolution that can be obtained depends on the number of detected counts and thus increasing system sensitivity may be very

beneficial for ^{213}Bi imaging in mice. Such sensitivity improvements have already been realized in practice through increasing crystal thickness which makes the system 2.5 times more efficient. Furthermore, new high-energy collimators with larger sensitivity have been developed. The choice of collimator depends on the imaging task and these higher sensitivity collimators may be very suitable for ^{213}Bi imaging. The VECTor system used in this paper is a first-generation scanner and could not benefit yet from these improvements.

Ex-vivo and in-vivo mouse experiments were performed as proof of principle, to investigate the feasibility to image the tumor and physiological uptake using ^{213}Bi . The reported kidney dose of 26 Gy by ^{213}Bi -DTPA is however high and most probably would lead to renal toxicity at a later stage. Renal toxicity has been observed after scanning ^{111}In labeled peptides with cumulative kidney doses of 20-40 Gy (20). Considering the high LET nature of its radiation 26 Gy by ^{213}Bi will be at least as equitoxic as 20-40 Gy by ^{111}In .

In the biodistribution assay for [^{213}Bi -DOTA,Tyr³]-octreotate showed an absolute tumor uptake of 5.3 %IA and in the kidney of 1.9 %IA. The renal uptake was therefore too low to be detectable on SPECT. These uptake values are higher than found from the SPECT data. It is not clear what caused this discrepancy.

The uptake of ^{111}In -DTPA in rat kidneys has been reported to be 0.9 ± 0.2 %/mL at 38 minutes after injection (22). Initially (at 2-4 min) the peak activity in the renal cortex is 5-7 times higher by perfusion with radioactive blood. The uptake of ^{213}Bi -DTPA in the kidneys seems to be much higher than that observed with ^{111}In -DTPA. This may be the result of the persistent uptake of ^{213}Bi in the kidneys, independent of the compound administered (23). Species-specific differences in DTPA uptake by mice and rats could also cause this difference, but the reported renal uptake of ^{111}In DTPA in dogs was found to be comparable to the rat values (24).

Using significantly higher injected activities of ^{213}Bi labeled peptides for better quantification is not ideal both for the consequentially higher amount of peptide needed, which might partially block receptor-mediated targeting, as well as for increased probability of radiation toxicity.

CONCLUSION

We have shown that it is possible to image ^{213}Bi at sub-mm resolution level with a SPECT system equipped with a dedicated high-energy collimator. We found that the use of the 440 keV gamma-ray peak is essential and produced significantly better images than the 79 keV X-ray peak. Quantification of the ^{213}Bi activity concentration was reliable above 0.240 MBq/mL with 30-minutes image time using the 440 keV energy window setting and above 0.320 MBq/mL with 5-minutes frames using the combined energy window setting. Uptake of [^{213}Bi -DOTA,Tyr³]octreotate in a CA20948 tumor xenograft was well visualized. Dynamic in-vivo imaging of the ^{213}Bi -DTPA distribution in a mouse showed distinct renal uptake patterns, enabling identification of sub-organ distributions (renal cortex). Quantification of the uptake data allowed kinetic modeling and estimation of the absorbed dose to the kidneys, albeit with uncertainties of around 20%.

Acknowledgements

The Dinse Stiftung (Germany) is acknowledged for financial support.

1. Essler M, Gartner FC, Neff F, et al. Therapeutic efficacy and toxicity of ²²⁵Ac-labelled vs. ²¹³Bi-labelled tumour-homing peptides in a preclinical mouse model of peritoneal carcinomatosis. *Eur J Nucl Med Mol Imaging*. 2012;39:602-612.
2. Wild D, Frischknecht M, Zhang H, et al. Alpha- versus beta-particle radiopeptide therapy in a human prostate cancer model (²¹³Bi-DOTA-PESIN and ²¹³Bi-AMBA versus ¹⁷⁷Lu-DOTA-PESIN). *Cancer Res*. 2011;71:1009-1018.
3. Morgenstern A, Bruchertseifer F, Apostolidis C. Targeted alpha therapy with ²¹³Bi. *Curr Radiopharm*. 2011;4:295-305.
4. Jaggi JS, Seshan SV, McDevitt MR, LaPerle K, Sgouros G, Scheinberg DA. Renal tubulointerstitial changes after internal irradiation with alpha-particle-emitting actinium daughters. *J Am Soc Nephrol*. 2005;16:2677-2689.
5. Norenberg JP, Krenning BJ, Konings IR, et al. ²¹³Bi-[DOTA0, Tyr3]octreotide peptide receptor radionuclide therapy of pancreatic tumors in a preclinical animal model. *Clin Cancer Res*. 2006;12:897-903.
6. Eckerman KF, Endo A, . *Medical Internal Radiation Dose Committee : MIRD radionuclide data and decay schemes*. 2nd ed. Reston, VA: Society of Nuclear Medicine; 2007.
7. Sgouros G, Ballangrud AM, Jurcic JG, et al. Pharmacokinetics and dosimetry of an alpha-particle emitter labeled antibody: ²¹³Bi-HuM195 (anti-CD33) in patients with leukemia. *J Nucl Med*. 1999;40:1935-1946.
8. Rosenblat TL, McDevitt MR, Mulford DA, et al. Sequential cytarabine and alpha-particle immunotherapy with bismuth-213-lintuzumab (HuM195) for acute myeloid leukemia. *Clin Cancer Res*. 2010;16:5303-5311.
9. Kratochwil C, Giesel FL, Bruchertseifer F, et al. ²¹³Bi-DOTATOC receptor-targeted alpha-radionuclide therapy induces remission in neuroendocrine tumours refractory to beta radiation: a first-in-human experience. *Eur J Nucl Med Mol Imaging*. 2014;41:2106-2119.
10. Goorden MC, van der Have F, Kreuger R, et al. VECTor: a preclinical imaging system for simultaneous submillimeter SPECT and PET. *J Nucl Med*. 2013;54:306-312.
11. Goorden MC, van der Have F., and Beekman, F. J. . Optimizing image reconstruction for simultaneous sub-mm clustered pinhole PET-SPECT. *Proceedings of the 12th International Meeting on Fully 3D Image Reconstruction in Radiology and Nuclear Medicine* 2013:126-129.
12. Chan HS, de Blois E, Konijnenberg M, et al. Optimizing labeling conditions of ²¹³Bi-somatostatin analogs for receptor-mediated processes in preclinical models. *J NUCL MED MEETING ABSTRACTS*. 2014;55:1179.

13. Vaissier PE, Goorden MC, Vastenhouw B, van der Have F, Ramakers RM, Beekman FJ. Fast spiral SPECT with stationary gamma-cameras and focusing pinholes. *J Nucl Med.* 2012;53:1292-1299.
14. Vastenhouw B, Beekman F. Submillimeter total-body murine imaging with U-SPECT-I. *J Nucl Med.* 2007;48:487-493.
15. Branderhorst W, Vastenhouw B, Beekman FJ. Pixel-based subsets for rapid multi-pinhole SPECT reconstruction. *Phys Med Biol.* 2010;55:2023-2034.
16. Ogawa K, Harata Y, Ichihara T, Kubo A, Hashimoto S. A practical method for position-dependent Compton-scatter correction in single photon emission CT. *IEEE Trans Med Imaging.* 1991;10:408-412.
17. van der Have F, Vastenhouw B, Rentmeester M, Beekman FJ. System calibration and statistical image reconstruction for ultra-high resolution stationary pinhole SPECT. *IEEE Trans Med Imaging.* 2008;27:960-971.
18. Walker MD, Goorden MC, Dinelle K, et al. Performance Assessment of a Preclinical PET Scanner with Pinhole Collimation by Comparison to a Coincidence-Based Small-Animal PET Scanner. *J Nucl Med.* 2014;55:1368-1374.
19. Wu C, van der Have F, Vastenhouw B, Dierckx RA, Paans AM, Beekman FJ. Absolute quantitative total-body small-animal SPECT with focusing pinholes. *Eur J Nucl Med Mol Imaging.* 2010;37:2127-2135.
20. Konijnenberg MW, Breeman WA, de Blois E, et al. Therapeutic application of CCK2R-targeting PP-F11: influence of particle range, activity and peptide amount. *EJNMMI Res.* 2014;4:47.
21. Melis M, Vegt E, Konijnenberg MW, et al. Nephrotoxicity in mice after repeated imaging using ¹¹¹In-labeled peptides. *J Nucl Med.* 2010;51:973-977.
22. Melis M, de Swart J, de Visser M, et al. Dynamic and static small-animal SPECT in rats for monitoring renal function after ¹⁷⁷Lu-labeled Tyr³-octreotate radionuclide therapy. *J Nucl Med.* 2010;51:1962-1968.
23. Slikkerveer A, de Wolff FA. Pharmacokinetics and toxicity of bismuth compounds. *Med Toxicol Adverse Drug Exp.* 1989;4:303-323.
24. McAfee JG, Gagne G, Atkins HL, et al. Biological distribution and excretion of DTPA labeled with Tc-99m and In-111. *J Nucl Med.* 1979;20:1273-1278.




Figure 1. Recovery coefficients for a syringe initially filled with 86.2 MBq ²¹³Bi (in 2.0 ml) scanned in 15 frames of 30 minutes. Results were fitted with a horizontal line when both residuals were normally distributed (D'Agostini-Pearson test) and showed no significant systematic deviation (Runs test). The open markers indicate the excluded RC values. Some data points at low activity concentrations did not fall within the Y-axis boundaries.




Figure 2. Recovery coefficient for a syringe initially filled with 67.69 MBq ^{213}Bi (in 2.0 ml) scanned in 45 frames of 5 minutes. Fitting was performed with the same statistical rules as in figure 2. Data for the first 13 frames were omitted from the graph, but were included in the averaging.




Figure 3. Resolution phantom images of ^{213}Bi SPECT. The phantom has 6 segments containing capillary diameters of 1.5, 1.2, 1.0, 0.9, 0.8 and 0.7 mm. The images show reconstructions for different energy window settings summed over 5 slices (2 mm in total).




Figure 4. Profiles through 0.7, 0.8, 0.9 and 1.2 mm rods for the single 79 and 440 keV and the combined 79 keV/440 keV energy windows (graphs A, B, C and D). Contrast and contrast-to-noise curves for the different rod sizes are shown in graphs E and F.




Figure 5. Ex-vivo image of 3.0 MBq [$^{213}\text{Bi-DOTA, Tyr}^3$]-octreotate injected in a nude mouse. From top to bottom: mip images reconstructed at 79keV, 440 keV and at both energy windows. The numbers in the color table indicate the radioactivity concentration in MBq/mL.




Figure 6. In-vivo mouse **mip** images with 7.4 MBq ^{213}Bi -DTPA. Images were reconstructed using the combined 79 keV/440 keV photo peak setting. In the 9 consecutive 5-minute frames kidneys show up in each frame, initially showing the ureters at the top of the image with gradually a distribution to the renal cortices and filling of the bladder. **The numbers in the color table indicate the radioactivity concentration in MBq/mL.**




Figure 7. Kinetic modeling of ^{213}Bi -DTPA in the urinary bladder and the kidneys. In the decay-corrected data (A) the bladder activity is fit by a single exponential with a $T_{1/2}=11\pm 2$ min ($T_{\text{eff}}=15$ min). The kidneys did not show clearance, whereas in the uncorrected data (B) the renal clearance proceeded with $T_{\text{eff}}=52\pm 10$ min half-life. Images were reconstructed from combined 79 keV/440 keV photo peaks.

Table 1. Results from the measurements of the ^{213}Bi filled phantoms.

Energy window	79 keV	440 keV	79 + 440 keV
RC 30 min frames (mean \pm SE)	1.03 \pm 0.007	1.04 \pm 0.007	1.04 \pm 0.009
Lower limit linearity 30 min (MBq/mL)	0.94	0.24	0.94
RC 5 min frames (mean \pm SE)	0.93 \pm 0.007	1.00 \pm 0.003	1.01 \pm 0.003
Lower limit linearity 5 min (MBq/mL)	2.33	0.90	0.32
FWHM 0.7 mm rods (mm)	1.1 \pm 0.3	1.4 \pm 0.3	1.4 \pm 0.3
FWHM 0.8 mm rods (mm)	1.6 \pm 0.5	1.4 \pm 0.2	1.4 \pm 0.2
FWHM 0.9 mm rods (mm)	1.4 \pm 0.6	1.33 \pm 0.14	1.29 \pm 0.09
FWHM 1.0 mm rods (mm)	1.4 \pm 0.5	1.34 \pm 0.29	1.29 \pm 0.12
FWHM 1.2 mm rods (mm)	1.6 \pm 0.6	1.30 \pm 0.14	1.25 \pm 0.08
FWHM 1.5 mm rods (mm)	2.0 \pm 0.4	1.58 \pm 0.14	1.61 \pm 0.12