

## **Towards a semi-flexible parylene-based platform technology for active implantable medical devices**

Bakhshae Babaroud, Nasim; Kluba, Marta; Dekker, Ronald; Serdijn, Wouter; Giagka, Vasiliki

**Publication date**

2019

**Document Version**

Final published version

**Citation (APA)**

Bakhshae Babaroud, N., Kluba, M., Dekker, R., Serdijn, W., & Giagka, V. (2019). *Towards a semi-flexible parylene-based platform technology for active implantable medical devices*. 1-1. Abstract from 7th Dutch Bio-Medical Engineering Conference, Egmond aan Zee, Netherlands.  
<https://www.bme2019.nl/mobview/presentation/1679>

**Important note**

To cite this publication, please use the final published version (if applicable). Please check the document version above.

**Copyright**

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

**Takedown policy**

Please contact us and provide details if you believe this document breaches copyrights. We will remove access to the work immediately and investigate your claim.

# TOWARDS A SEMI-FLEXIBLE PARYLENE-BASED PLATFORM TECHNOLOGY FOR ACTIVE IMPLANTABLE MEDICAL DEVICES

*Nasim Bakhshae Babaroud<sup>1</sup>, Marta Kluba<sup>2</sup>, Ronald Dekker<sup>2</sup>, Wouter Serdijn<sup>1</sup>, Vasiliki Giagka<sup>1, 3</sup>*

*<sup>1</sup>Bioelectronics Section and <sup>2</sup>Electronic Components, Technology and Materials Section,*

*Department of Microelectronics, Faculty of Electrical Engineering, Mathematics and Computer Science, Delft University of Technology*

*Mekelweg 4, 2628 CD, Delft, The Netherlands*

*<sup>3</sup>Technologies for Bioelectronics Group, Department of System Integration and Interconnection Technologies, Fraunhofer Institute for Reliability and Microintegration IZM*

*Gustav-Meyer-Allee 25, 13355, Berlin, Germany*

*E-mail: N.Bakhshae@tudelft.nl*

## ABSTRACT

Active implantable medical devices have been developed for diagnosis, monitoring and treatment of large variety of neural disorders. Since the mechanical properties of these devices need to be matched to the tissue, soft materials, such as polymers are often preferred as a substrate [1]. Parylene is a good candidate, as it is highly biocompatible and it can be deposited/etched using standard Integrated Circuit (IC) fabrication methods/processes. Further, the implantable devices should be smart, a goal that can be accomplished by including ICs. These ICs, often come in the form of additional pre-packaged components that are assembled on the implant in a heterogenous process. Such a hybrid integration, however, does not allow for size minimization, which is so critical in these applications, as otherwise the implants can cause severe damage to the tissue. On the other hand, it is essential that all components are properly packaged to prevent early failure due to moisture penetration [2].

In this work we use a previously developed semi-flexible platform technology based on a Parylene substrate and Pt metallization, which allows integration of electronic components with a flexible substrate in a monolithic process. We use an IC fabrication-based platform that allows for the fabrication of several rigid regions including Application-Specific Integrated Circuits (ASICs) and other components connected to each other by means of flexible interconnects. We aim to add more functionality to this technology and thereby extend it to a platform for a variety of medical applications. An example of such functionality is integrating Light Emitting Diodes (LEDs) for optogenetic stimulation or integrating Capacitive Micromachined Ultrasound Transducers (CMUTs) for ultrasound stimulation or ultrasound wireless power transfer. Since the long-term reliability is critical for implantable devices, we intend to reinforce our implant with an extra Polydimethylsiloxane (PDMS) encapsulation layer that relies on the low viscosity of the uncured rubber to flow in every detail of the surface to prevent void formation [3]. Therefore, this work also focuses on enhancing the adhesion of PDMS to Parylene, as it must remain strong for the required lifetime of the device.

## REFERENCES

- [1] C. Hassler, T. Boretius, T. Stieglitz, "Polymers for Neural Implants", *Journal of Polymer Science Part B: Polymer Physics*, vol.49, no.1, pp. 18–33, 2010.
- [2] V. Giagka, W. A. Serdijn, "Realizing flexible bioelectronic medicines for accessing the peripheral nerves – technology considerations", *Bioelectronic Medicine*, vol.4, no.1, 2018.
- [3] A. Vanhoestenbergh, N. Donaldson, "Corrosion of silicon integrated circuits and lifetime predictions in implantable electronic devices", *Journal of Neural Engineering*, vol.10, no.3, 2013.