

Higher Order Singular Value Decomposition Filter for Contrast Echocardiography

Wahyulaksana, Gerald; Wei, Luxi; Voorneveld, Jason; Hekkert, Maaïke Te Lintel; Strachinaru, Mihai; Duncker, Dirk J.; De Jong, Nico; Van Der Steen, Antonius F.W.; Vos, Hendrik J.

DOI

[10.1109/TUFFC.2023.3316130](https://doi.org/10.1109/TUFFC.2023.3316130)

Publication date

2023

Document Version

Final published version

Published in

IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control

Citation (APA)

Wahyulaksana, G., Wei, L., Voorneveld, J., Hekkert, M. T. L., Strachinaru, M., Duncker, D. J., De Jong, N., Van Der Steen, A. F. W., & Vos, H. J. (2023). Higher Order Singular Value Decomposition Filter for Contrast Echocardiography. *IEEE Transactions on Ultrasonics, Ferroelectrics, and Frequency Control*, 70(11), 1371-1383. <https://doi.org/10.1109/TUFFC.2023.3316130>

Important note

To cite this publication, please use the final published version (if applicable). Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy


Please contact us and provide details if you believe this document breaches copyrights. We will remove access to the work immediately and investigate your claim.

Higher Order Singular Value Decomposition Filter for Contrast Echocardiography

Geraldi Wahyulaksana¹, Graduate Student Member, IEEE, Luxi Wei², Member, IEEE, Jason Voorneveld³, Maaïke te Lintel Hekkert, Mihai Strachinaru⁴, Dirk J. Duncker⁵, Nico de Jong⁶, Antonius F. W. van der Steen⁷, Fellow, IEEE, and Hendrik J. Vos⁸, Member, IEEE

Abstract—Assessing the coronary circulation with contrast-enhanced echocardiography has high clinical relevance. However, it is not being routinely performed in clinical practice because the current clinical tools generally cannot provide adequate image quality. The contrast agent’s visibility in the myocardium is generally poor, impaired by motion and nonlinear propagation artifacts. The established multipulse contrast schemes (MPCSs) and the more experimental singular value decomposition (SVD) filter also fall short to solve these issues. Here, we propose a scheme to process amplitude modulation/amplitude-modulated pulse inversion (AM/AMPI) echoes with higher order SVD (HOSVD) instead of conventionally summing the complementary pulses. The echoes from the complementary pulses form a separate dimension in the HOSVD algorithm. Then, removing the ranks in that dimension with dominant coherent signals coming from tissue scattering would provide the contrast detection. We performed both in vitro and in vivo experiments to assess the performance of our proposed method in comparison with the current standard methods. A flow phantom study shows that HOSVD on AM pulsing exceeds the contrast-to-background ratio (CBR) of conventional AM and an SVD filter by 10 and 14 dB, respectively. In vivo porcine heart results also demonstrate that, compared to AM, HOSVD improves CBR in open-chest acquisition (up to 19 dB) and contrast ratio (CR) in closed-chest acquisition (3 dB).

Index Terms—Contrast-enhanced ultrasound (CEUS), higher order singular value decomposition (HOSVD), microbubble detection, myocardial perfusion, ultrafast imaging.



I. INTRODUCTION

CONTRAST echocardiography (CE) is a clinically established diagnostic tool for assessing the ejection fraction and regional wall motion when image quality is poor

Manuscript received 10 July 2023; accepted 9 September 2023. Date of publication 18 September 2023; date of current version 9 November 2023. This work was supported in part by the Dutch Research Council (NWO) research programme “Vernieuwingsimpuls – Vidi 2017” under Project QUANTO-16572; and in part by the Medical Delta Scientific Program, The Netherlands, under Project UltraHB. (Corresponding author: Geraldi Wahyulaksana.)

This work involved human subjects or animals in its research. Approval of all ethical and experimental procedures and protocols was granted by the Centrale Commissie Dierproeven (CCD) under Approval No. AVD1010020172411 (SP2100125).

Please see the Acknowledgment section of this article for the author affiliations.

Digital Object Identifier 10.1109/TUFFC.2023.3316130

in regular B-mode imaging [1], [2]. Signal strength of the blood pool and vasculature are enhanced using ultrasound contrast agents (UCA), consisting of micrometer-sized lipid-coated gas bubbles that are injected intravenously [3]. CE can also be used to assess myocardial perfusion. However, accurate and reliable perfusion imaging has been a challenge since its introduction. The shadowing from the ribs and lungs, and attenuation of signal due to the contrast agent in the cardiac cavities, noise, and strong clutter signal originating from stationary and moving tissue are contributing factors that lead to limited image quality [3], [4].

In the past decades, several contrast-specific pulsing schemes have been developed to suppress tissue clutter signal and provide contrast detection. Generally, two categories can be distinguished. One category relies solely on the

Highlights

- We propose a method to detect slowly flowing microbubbles in moving tissue using higher-order singular value decomposition (HOSVD) with spatial, temporal, and pulsing sequences as input dimensions.
- Both in vitro and in vivo results showed that the method provides superior contrast detection compared to the standard SVD filter, and amplitude modulation (AM), applied to high-frame-rate data.
- By combining SVD and contrast-specific pulsing into one algorithm, HOSVD improves microbubble signals detection and reduce motion artifacts in challenging cases such as myocardial perfusion imaging.

generation of harmonic signals due to the nonlinear scattering of ultrasound by the microbubbles. In these schemes, subharmonics or higher harmonics of the originally transmitted ultrasound frequency would be selectively retained to show the contrast agents on screen. Yet, it was found that tissue also generates harmonic frequency content by nonlinear propagation, thus leading to substantial tissue clutter [5]. The other category is contrast pulsing schemes that rely on sending multiple pulses that suppress the linear tissue clutter signal when the echoes are combined by simple subtraction or addition, sometimes after scalar multiplication. The most commonly used multipulse contrast schemes (MPCS) are pulse inversion (PI) [6], amplitude modulation (AM) [7], and their combination, i.e., amplitude-modulated PI (AMPI) [8]. However, as ultrasound also propagates nonlinearly when it travels through tissue or a microbubble cloud, the tissue suppression effectiveness is reduced as well for these schemes. This might be significant for cardiac imaging where the chambers are filled with microbubbles following intravenous administration and the nonlinearity is substantially accumulated due to great imaging depth, which would lead to the unwanted “nonlinear” tissue signal again [9]. Moreover, the rapid myocardial motion throughout the cardiac cycle causes out-of-phase summation of the linear components from multiple transmission, which are therefore not fully suppressed. This again results in tissue clutter through so-called motion artifacts [10], [11].

Another approach to suppress the tissue clutter signal is by using a postprocessing clutter filter. Originally, clutter filtering is performed as frequency filters on a packet of consecutive frames with the assumption that blood flow and tissue motion have distinct spectral characteristics in such ensembles. However, this assumption is not valid when the flow velocity and tissue motion have overlapping spectral content [12], which is generally the case in microvascular myocardial perfusion imaging. More recently, the advancement of high-frame-rate (HFR) ultrasound imaging [13] provides highly coherent data in both space and time that facilitates clutter filtering using spatiotemporal information. HFR contrast-enhanced ultrasound (HFR-CEUS) combined with singular value decomposition (SVD) filtering [14] has been used ubiquitously in research setting to reduce tissue clutter enabling assessment of flow in the vasculature. SVD decomposes the tissue, blood flow, and noise into separate singular vectors, assuming that they have different spatiotemporal coherence. Ideally, the filtering process is then performed by removing tissue and noise singular vectors,

with only the flow signal of the interest remaining. It has been reported that SVD suppresses tissue clutter better than conventional frequency filters and contrast pulsing schemes (AM and AMPI) [14], [15], [16]. However, its performance also degrades with slower flow rates and faster tissue motion since, in that case, the coherence between the contrast agent signal and tissue increases, thus preventing the separation into different components [14], [17], [18]. This is a crucial issue for cardiac imaging where tissue motion can reach up to 9.4 cm/s [19] and blood flow in the microcirculation is generally slower than 1 cm/s. To mitigate this problem, the combination of SVD and AM by applying SVD on AM-processed images was attempted. However, it has been reported that it achieves a lower contrast-to-background ratio (CBR) than an SVD filter alone [16]. In this case, the SVD filter fails to improve contrast detection of the AM-processed images because it operates in the spatiotemporal dimension, where the residual tissue and microbubble signal are highly coherent and of the same order of magnitude in strength.

To achieve better tissue and contrast separation, we propose to expand the filter operating dimensions by using higher order SVD (HOSVD), instead of using “regular” SVD that can only take a 2-D matrix as input. It is a multilinear generalization of SVD, in which HOSVD decomposes an n th-order tensor [20]. Its usage to improve blood flow visualization in medical ultrasound imaging has been proposed by Kim et al. [21]. They demonstrated that using Doppler frames as the 3-D could improve blood flow and perfusion separation from tissue clutter. However, their implementation is aimed at tissue with slow motion and without exploiting the nonlinear property of UCAs.

We expect that HOSVD can improve the MPCS performance to detect UCAs if the underlying structure of the data meets the assumptions attained in the decomposition. In this study, we propose using HOSVD to combine the received echoes from the complementary pulses in the MPCS, instead of the conventional linear signal combinations. In the case of MPCS, it is assumed that contrast agents respond nonlinearly to the ultrasound pulses, whereas tissue response is mostly linear. Thus, we can add the pulsing sequence as a 3-D to the spatial and temporal information. We hypothesize that the nonlinear part of the microbubble signal would be decomposed into different components than the tissue clutter signal, after which data separation can be performed by appropriate component selection. An in vitro measurement with a flow phantom as well as in vivo cardiac measurement were performed to assess the efficacy of our proposed method. The performance of our proposed technique will be evaluated




Fig. 1. Setup of the flow phantom. Probe was positioned inside the chamber to emulate both rigid tissue motion and nonlinear propagation through microbubble cloud.

and compared with the conventionally processed MPCS and the regular spatiotemporal SVD filter.

II. MATERIALS AND METHODS

A. In Vitro Setup

We built an in vitro setup that could emulate rigid tissue motion and the effect of nonlinear propagation through a cloud of bubbles. A tissue-mimicking wall-less flow phantom was made from a suspension of 10% w/v polyvinyl alcohol (PVA) and 1% w/v silicon carbide as background scattering particles, 20% cooling liquid, and 69% distilled water with one completed freeze–thaw thermal cycle, see Fig. 1. To obtain ultrasound images in the presence of nonlinear propagation through a bubble cloud, diluted Definity (Lantheus Medical Imaging, North Billerica, MA, USA) with a concentration of $\sim 0.5\text{--}1.1 \times 10^6$ MB/mL was put into the phantom chamber on top. The same dilution was infused with a perfusion pump through a 1-mm-diameter channel as the microbubbles of interest. Since the minimum flow that this syringe pump could provide was 6 mm/s, we turned off the pump to have negligible flow, thus mimicking capillary perfusion rather than vascular flow. An ultrasound probe (see Fig. 1) was attached to a linear stage. Rigid tissue motion was emulated by moving the probe in the vertical direction during image acquisition. The peak velocity of the probe was 35 mm/s, which emulates the peak velocity of the left main coronary artery environment [22].

B. In Vivo Porcine Myocardial Perfusion Model

Two female Yorkshire \times Norwegian Landrace porcine (35 and 38 kg) experiments followed the European Union and institutional guidelines for the care and use of laboratory animals, with Centrale Commissie Dierproeven (CCD) approval AVD1010020172411 (SP2100125). They were first sedated, put under full anesthesia using pentobarbital (10–15 mg/kg/h) and mechanically vented. Animals were positioned in a supine position. Vital signs were monitored. A diluted (30 \times) Definity¹ solution (Lantheus Medical Imaging Inc.,

¹Registered trademark.

MA, USA) was continuously infused through the jugular vein at 1.5 mL/min. The acquisitions were performed both transthoracically (closed chest) and by direct cardiac access after sternotomy (open chest).

C. Ultrasound Transmission and Beamforming

Radio frequency acquisitions in both in vitro and in vivo experiments were performed with a 64-element phased array transducer (P7-4, Philips ATL, Bothell, WA, USA), connected to a Vantage 256 system (Verasonics Inc., Redmond, WA, USA). The HFR imaging was performed with diverging waves with a pulse repetition frequency of 4.5 kHz, a center frequency of 5.2 MHz (three cycles), and an image depth of 10 cm. Multipulse contrast sequences (MPCS) (AM and AMPI) were adopted on the fundamental frequency (nonlinear fundamental imaging) [8]. AM was used in the in vitro measurement. Both AM and AMPI transmission schemes were used in in vivo experiments. The in vitro HFR transmission sequence only consisted of one transmit aperture (64 elements) with the final frame rate of 1500 Hz. In the in vivo experiment, an acquisition consisted of 1 s of HFR recordings, a high mechanical index (MI) focused beam “FLASH” microbubble destruction sequence, and a subsequent 4 s of HFR recordings. The in vivo HFR transmission sequence consisted of three synthetic transmit apertures (21 elements each) with overlapping beams to increase spatial resolution with coherent compounding [23] and a final frame rate of 500 Hz. The “checkerboard pattern” with even, full, and odd element number transmission was completed per subaperture. The “FLASH” destruction sequence consisted of transmitting 21 cycles focused beam with $MI \approx 1$. Channel data were beamformed offline using the Ultrasound Toolbox [24] in MATLAB (R2022A, Mathworks, Natick, MA, USA, 2022) on a 0.5λ resolution grid.

III. THEORY AND IMPLEMENTATION

A. Data Structure

The acquired beamformed image series (s) in-phase/quadrature (IQ) data can be modeled as a linear mixture of tissue clutter signal (c), microbubble signal (b), and noise (n)

$$s(x, z, t, p) = c(x, z, t, p) + b(x, z, t, p) + n(x, z, t, p) \quad (1)$$

where x stands for the lateral dimension, z stands for the axial dimension, t stands for time, and p stands for transmitted pulse number. The clutter and noise signal need to be removed from the mixture in order to accurately assess the microbubble signal. A further property of the tissue clutter is that it dominates the images, over microbubble signal and noise, which is important as the filter described in Section III-C order the signals partly based on their intensity.

B. Conventional Processing

Conventionally, the MPCS suppresses the linear tissue clutter signal by subtracting (AM) or summing (AMPI)

the signals backscattered from the complementary pulse transmissions (p). Two assumptions are made: 1) the tissue backscatter signal has a linear relationship with the phase and/or amplitude of the transmitted pulse and 2) negligible motion occurs within the time interval of the complementary pulse transmissions. When these assumptions hold, then the residue in the processed signal consists of the nonlinear microbubble signal and noise. However, the CE signals violate the assumptions: nonlinear ultrasound propagation in tissue also generates the second-harmonic frequency content, which will give some residual signal in the linear combination of the complementary pulses that appears as tissue clutter in the contrast-enhanced images [5] and tissue motion induces signal misalignment that causes imperfect linear signal cancellation, leading to the generation of motion artifacts [10], [11].

C. HOSVD Processing Rationale

Instead of linearly combining the complementary pulses, we propose to use HOSVD to retrieve the microbubble signal. HOSVD takes an n th-order tensor as its input and decomposes it into a tensor core and n -orthogonal modes that each has individual ranks. We hypothesize that HOSVD is able to separate the clutter, microbubble, and noise signal, based on their different correlation of spatial, temporal, and the backscatter response (magnitude and phase) to the transmitted pulses. Just like in 2-D SVD, HOSVD decomposes the signal into all-orthogonal singular vectors and sorts them based on the magnitude of the multilinear singular value (MSV).

HOSVD is implemented on an ultrasound image series $s(x, z, t, p)$, arranged as a 3rd-order tensor $\mathcal{T}(x \times z, t, p) \in \mathbb{C}^{x \times z \times t \times p}$, which represents spatial, temporal, and pulsing dimensions as the input. The 3rd-order tensor \mathcal{T} can be decomposed with HOSVD as follows:

$$\begin{aligned} T &= \mathcal{S} \cdot U \cdot V \cdot W, \\ &= \sum_{j_1}^{I_1} \sum_{j_2}^{I_2} \sum_{j_3}^{I_3} \mathcal{S}_{j_1, j_2, j_3} U_{j_1} V_{j_2} W_{j_3} \end{aligned} \quad (2)$$

where \mathcal{S} is the core tensor with dimensions $\mathbb{C}^{x \times z \times t \times p}$, U is the spatial mode singular vectors, V is the temporal mode singular vectors, W is the pulsing mode singular vectors, and I_1 – I_3 are the spatial, temporal, and pulsing mode component ranks, respectively. An example of rearranging AM images and the resulting decomposition is shown in Fig. 2. Assuming that the clutter, microbubble, and noise are decomposed into different ranks, the filtering process is similar to the regular SVD filtering process, which is removing the ranks that mainly consist of clutter and noise. However, the tensor core is not diagonal like in regular SVD, but a 3rd-order nonsparse tensor, which means that every combination from all the modes has a particular tensor core value. As a result, unlike regular SVD that only requires a rank selection that applies to both spatial and temporal singular vectors, the rank selection for HOSVD has to be performed on each dimension separately.

D. Rank Selection for Clutter Filtering

Each of the decomposed dimensions represents a different physical meaning; thus, every dimension needs a specific




Fig. 2. Illustration of the tensor arrangement from stacks of ultrasound images and its decomposition with HOSVD.

rank selection algorithm. It makes the rank selection more flexible and complicated at the same time. After the rank selections are performed, the HOSVD as a clutter filter to retrieve microbubble signal ($\hat{b}(x, z, t)$) can be implemented as follows:

$$\hat{b}(x, z, t) = \sum_{j_1 \in I_1} \sum_{j_2 \in I_2} \sum_{j_3 \in I_3} \mathcal{S}_{j_1, j_2, j_3} U_{j_1} V_{j_2} W_{j_3} \quad (3)$$

where I_1 – I_3 are the selected n -mode ranks for spatial, temporal, and pulsing modes, respectively. We developed different algorithms for the in vitro and in vivo datasets as they have different conditions and complexity.

1) In Vitro Rank Selection:

a) *Pulsing dimension*: The scattered response of both tissue and microbubbles to the transmitted pulse is mostly linear, especially at the fundamental frequency such as used in our experiments. However, microbubbles also generate “nonlinear fundamental” signals in response to the complementary transmitted pulses [4], [8], characterized by either nonlinear amplitude, amplitude-dependent phase response, or both; for our proposed technique, the exact cause is not important. This “nonlinear fundamental” component is different from the linear tissue response and hence will be decomposed into another rank than the linear tissue rank. Since it is assumed that the linear responses will dominate the original data, the ranks with linear response will get the highest singular values,




Fig. 3. (a) B-mode image of the phantom as a reference. (b)–(d) Examples of ultrasound images as a result of pulsing mode rank selection. (e) and (f) Singular vectors that represent the backscatter trend, in response to the transmitted pulses (AM).

which means that they will appear in the first rank or ranks. The residual nonlinear component or components will appear in the last rank since their overall magnitude is lower.

To exemplify the pulsing rank separation, ultrasound images are shown for each pulsing rank in Fig. 3(b)–(d). Here, all the spatiotemporal ranks were kept for illustration. The pulsing mode singular vectors represent the backscatter response of the scatterers to the different transmission pulses ($W_{j3}(p)$), as shown in Fig. 3(e). The first pulsing rank W_1 shows the major components of the magnitude and phase in response to the AM transmission. In this rank, the first (even) and third (odd) transmitted pulse each contain approximately half energy, and indeed, the magnitude of the second response (full aperture) has twice the magnitude of the others and the phase is almost constant. The residual phase shift in the three responses originates from the overall motion of the probe. As a comparison, the phase of the first pulsing rank is constant when the probe is static [Fig. 3(f)]. The reconstructed image of the first pulsing rank W_1 corresponds closely to the B-mode/fundamental imaging [Fig. 3(a)]. The second pulsing rank W_2 mostly has energy on the even and odd transmission. This can be explained by its correspondence to the residual difference between the even- and odd-element transmissions:




Fig. 4. (a) Spatial mode singular vectors, reshaped as ultrasound spatial images ($U(x,z)$). (b) Tensor core magnitude of the spatial singular vectors when only $W_3(p)$ is chosen.

although the wavefields do overlap, there remain minor differences in their wavefronts since the elements by which the wavefield is transmitted are their mutual adjuncts [Fig. 3(c)]. Finally, the third rank shows the total residual signal, which is mainly originating from nonlinear responses of microbubbles deviating from the general tissue and (linear) bubble responses but can still contain noise. Indeed, the reconstructed image mainly shows the microbubble signal inside the target channel [Fig. 3(d)]. Thus, the clutter filtering for the pulsing dimension was performed by removing the first and second pulsing ranks as they predominantly consist of unwanted clutter signal.

b) Spatial dimension: Like 2-D-SVD, the spatial singular vectors $U^{x \times z \times t \times p}$ can be reshaped into ultrasonic spatial images $U^{x \times z \times t \times p}$. The manual distinction between the spatial mode singular vectors ($U_{j1}(x,z)$) that consist of clutter, microbubble, and noise signal is straightforward because they can simply be visualized. As shown in Fig. 4(a), $U_7(x,z)$ and $U_9(x,z)$ visibly consist of microbubble signal. Accordingly, the clutter filtering should be performed by keeping those two ranks. A comparison between Fig. 4(a) and (b) shows that the spatial ranks with the highest magnitude in pulsing rank 3 correspond to the images with the highest visibility of the microbubble-containing channel. Thus, the automatic rank selection was performed by choosing the spatial ranks that have the highest magnitude. Thus, the automatic rank selection was performed by choosing the spatial ranks that have the highest magnitude, as well as those above the threshold of 0.4.

The microbubbles had negligible flow speed compared to the probe velocity, so we do not expect that the microbubble and tissue signal are decomposed into different temporal ranks. The temporal singular vectors ($V_{j2}(t)$) that represent the spectral variation of the signal and the MSV of the temporal mode are shown in Fig. 5. Since there was no separation between tissue and microbubble signal in the




Fig. 5. (a) Modulus of the first four temporal singular vectors ($V_1(t)$ – $V_4(t)$). (b) Normalized magnitude of the temporal mode MSV.




Fig. 6. (a) Singular vectors that represent the decomposed backscatter response to AMPI transmission. (b)–(d) Examples of ultrasound images as a result of pulsing mode rank selection.

temporal dimension, all temporal singular vectors were kept for the in vitro data processing.

2) In Vivo Rank Selection:

a) *Pulsing dimension:* Similar to the in vitro data, we expect the “nonlinear fundamental” microbubble signal to be decomposed into rank 3 of the pulsing mode vector ($W_3(p)$). As shown in Fig. 6, the pulsing mode singular vectors followed the same trend as the in vitro dataset. Yet, since AMPI was used in this specific example, the phase of the rank shifts by π between full and even/odd transmission. Moreover, the reconstructed in vivo images show the tissue responses for ranks 1 and 2, and contrast signal in rank 3. Thus, the same rank selection was applied to the in vivo dataset: keeping the 3rd pulsing rank.

b) *Spatial dimension:* The spatial structure, as well as nonrigid motion, of the in vivo dataset is substantially more complex than the in vitro dataset. Unlike the in vitro spatial mode decomposition, the tensor core energy is widely spread across spatial singular vectors and there are no singular vectors that have significantly higher magnitude than the rest. As a




Fig. 7. (a) Representative ranks of the reshaped spatial mode singular vectors ($U_{j1}(x, z)$). (b) Normalized magnitude of the spatial MSV. (c) Gradient of the spatial MSV, the peak (rank 31) is chosen as the start of the selected ranks.

result, the in vitro rank selection algorithm could not be applied straightforwardly.

Analogous to the regular 2-D SVD, the ranks were sorted based on the total energy and the subspaces that contain residual tissue clutter signal mostly are allocated in the lower ranks, and those that contain microbubble signal are allocated mostly in the middle-to-high ranks. Since the spatial mode MSV has gradual changes, we expect that there is a detectable trend on the slope of the MSV when the rank subspaces change their dominant contents, similar to the regular SVD structure [25]. The gradient of the MSV is shown in Fig. 7. The local peaks of the derivative (rank 31) coincide with the transition of the subspaces, so it was chosen as the start of the selected ranks. Rank 15 was chosen as the start of the peak searching because, empirically, we observed that the tissue subspace occupied more than the first 15 ranks. This value of 15 may depend on ensemble length and tissue motion, although we found that this value worked for both closed-chest data and open-chest data. All the subsequent ranks are selected with the assumption that microbubble signal, especially the “nonlinear fundamental” part, has lower energy than the tissue signal and the spatial decomposition does not have ranks that only contain noise.

c) *Temporal dimension:* Different from the in vitro case, where the only flow of interest is within the channel, microbubble flow in the chambers of the heart is not negligible and might influence the signal decomposition when filtering for perfusion in the heart walls. The microbubble flow peak velocity in the chambers is faster than the tissue motion and the coronary circulation also has a wide range of flow velocity. Cardiac motion typically is nonrigid. All of these factors will




Fig. 8. (a) Frequency content of the temporal mode singular vectors. (b) Dominant frequency of each temporal mode singular vector and the selected low-frequency ranks.

cause the microbubble signal to appear in different ranks in the temporal dimension. Since we are interested in the slow flow in the myocardium and not the fast flow in the chamber and the fast tissue motion, rank selection can be performed on the temporal dimension.

The temporal singular vectors and their frequency contents are shown in Fig. 8(a). It can be seen that the singular vectors corresponding with higher temporal ranks have increasing dominant frequency, with the first rank representing almost static motion and the last ranks representing mostly noise and aliased motion around the highest Doppler velocities. Since our interest mainly lies in detecting slow flow in the microcirculation, with velocities slower than 1 mm/s, we need to retain the lower ranks and discard the higher ones. We calculated the dominant (maximum) frequency of each rank [as shown in Fig. 8(b)] and selected the ranks with a dominant frequency below 50 Hz, and the effect of this choice will also be demonstrated.

E. Postprocessing and Analysis

1) *In Vitro Image Quality Quantification*: HOSVD processing was performed on the beamformed quadrature demodulated data using the Tensorlab toolbox [26]. To quantify the efficacy of our proposed method, we implemented HOSVD with the rank selection method as mentioned above and compared this with regular B-mode, conventional AM, and SVD spatiotemporal filter [14] on both the full-aperture transmit pulse and conventional AM. We performed our analysis during two scenarios: when the probe was static and when it was moving at a constant velocity of 35 mm/s. Each scenario consisted of a time interval of 200 ms (comprising 300 temporal frames). For both scenarios, HOSVD was implemented on an ensemble of 20 temporal frames (13.3 ms) with ten samples of temporal overlap. For the moving probe scenario, SVD was also implemented on an ensemble of 20 temporal frames, while all 300 frames were used for the static probe scenario to optimize tissue and flow separation [25]. The example of SVD rank selection is provided in the appendix. The regions of interest (ROIs) were drawn on the channel (contrast) and PVA (background) to perform the quantitative analysis in the first frame (Fig. 9). The ROIs then automatically followed (using a 2-D cross correlation-based global motion estimator) the channel




Fig. 9. Regions used for calculating CBR. Blue and white regions are used for contrast and background signal, respectively. Yellow region indicates the field of view to display the result of different filters.




Fig. 10. (a) Regions used for calculating CBR in the cardiac measurements. Blue (myocardium) and white (epicardium) regions were used for contrast and background signal, respectively. (b) Blue region (myocardium) used for calculating contrast signal strength. LV and RV indicate the left and right ventricles, respectively.

and PVA, while the probe was moving in the subsequent frames. CBR was then calculated to evaluate the filters' performances to suppress clutter signal, defined as $CBR_a = 20\log_{10}(\overline{RMS}_{contrast}/\overline{RMS}_{background})$, where \overline{RMS} is the time-averaged root-mean-square signal strength in the time interval (200 ms). Overview box plots were calculated with: whiskers cover approximately 99% of the data; the bottom and top of each box denote the 1st and 3rd quartile of the data, respectively; and solid lines in the middle of each box signify the data median.

2) *In Vivo Image Quality Quantification*: The three synthetic-aperture transmissions were coherently compounded on each pulse before HOSVD is implemented. The ensemble length of 20 temporal samples (40 ms) with ten samples overlap was used for HOSVD clutter filtering. The performance of our proposed HOSVD filter was compared with the conventionally processed AM. Power Doppler was computed using $PD(z, x) = \sum_{t=1}^T |S(z, x, t)|^2$, where $S(z, x, t)$ are the filtered ultrasound images. The tissue suppression efficacy during motion was assessed on the open-chest acquisitions, using $CBR_p = 10\log_{10}(\overline{RMS}_{contrast}/\overline{RMS}_{background})$, in 75 temporal frames (150 ms) in the ECG signal in Fig. 12(e). We use a factor of 10 in this equation as the contrast and background signal are power Doppler signals, i.e., already squared values of the magnitude. In order to quantify this tissue suppression, we drew ROIs on the myocardium where no tissue clutter was visible and hence only showed contrast or noise signal and on the region where the residual tissue clutter due to motion was clearly observed—acting as background. The areas are shown in Fig. 10(a).

The contrast detection efficacy on the closed-chest acquisition was assessed during end-diastole by calculating the




Fig. 11. (a)–(d) Representative contrast images of different processing during moving probe. (e) CBR values overview, obtained with different processing filters and probe motion. Box plots represent median, 50%, and 99% ranges.

contrast ratio (CR) before and after the FLASH sequence, calculated by $CR = 10 \log_{10}(\text{RMS}_{\text{pre-FLASH}}/\text{RMS}_{\text{post-FLASH}})$. The ROI on the myocardium to compare the pre-FLASH and post-FLASH destruction sequence is shown in Fig. 10(b).

IV. RESULTS

A. In Vitro

Representative phantom contrast images during motion (35 mm/s) are shown in Fig. 11(a)–(d). The flow channel is visible in all images, yet the background is most suppressed when using AM + HOSVD thus giving the best visibility of the channel. When quantifying [Fig. 11(f)], the CBR of HOSVD on AM exceeds that of the conventionally processed AM by 10 dB, AM + SVD by 11 dB, that of both SVD-filtered, and unfiltered by 14 dB. During the time interval where the probe was not moving (images not shown for brevity), HOSVD achieved 5-, 4.5-, 12-, and 17-dB CBR improvement over conventional AM, AM + SVD, SVD filtered, and unfiltered images, respectively. See Fig. 11(f) for these quantified results, including their spread. The SVD-filtered images have worse CBR than the AM images or the AM + HOSVD. Note that SVD-based contrast filtering assumes that the contrast agent signal is highly decorrelated with the tissue signal, i.e., has distinct motion. In this experiment that mimics perfusion and slow flow in relatively fast-moving cardiac tissue, such an assumption is largely




Fig. 12. Comparison of open-chest apical view AM images, processed conventionally and with HOSVD. Conventionally processed AMPI images (a) and (c) can be compared with the HOSVD-processed images (b) and (d), respectively. Tissue motion artifacts at the apex of the heart are more suppressed and microbubble signals could be seen better (cyan boxes) with HOSVD than the conventional processing.

violated since the contrast agents have very low flow velocity in the channel and, on top, they move with the tissue in the “moving probe” experiment.

B. In Vivo

Representative power Doppler images of the open-chest acquisition (apical view) during early diastole are shown in Fig. 12. Qualitatively, it can be observed that the HOSVD + AM images [Fig. 12(b) and (d)] show more microbubble signal in the myocardium (cyan arrow) and more suppressed tissue clutter signal (white arrow), compared to the conventionally processed AM images [Fig. 12(a) and (c)]. Quantitatively, HOSVD + AM provided 19-dB CBR improvement over the conventionally processed AM images. The ECG signal of the image period (40 ms) is shown in Fig. 12(e).

The example images from the closed-chest acquisition (parasternal short axis view) pre- and post-FLASH destruction sequence processed with different schemes are shown in Fig. 13. Compared to the open-chest apical view, the nonlinear propagation artifact was more consequential in this view because the right ventricle (that contains a high concentration of microbubbles) was in the ultrasound propagation path. On the other hand, the motion artifacts might be expected to be less severe. These images show that the tissue signal is suppressed and the microbubble signal can be seen inside the myocardium boundaries.




Fig. 13. (a) and (c)–(f) Comparison of closed-chest parasternal short axis AM images, processed conventionally and with HOSVD. (b) Frequency contents of the temporal mode singular vectors and the rank selection. Image pre-FLASH destruction processed with (a) conventional AM shows less microbubble signal in the interventricular septum (cyan boxes) with (c) low-rank HOSVD-processed image. (d) High-rank HOSVD-processed HOSVD does not show microbubble signal inside the myocardium and more signal in the chamber because it does not have low-frequency components in slow time. Post-FLASH destruction of both processing (e) and (f) has much less microbubble signal in the interventricular septum.

In the pre-FLASH-destruction images, the results of conventional AM processing and HOSVD filtering with two temporal mode rank selections are shown in Fig. 13. To demonstrate the effect of the temporal mode rank selection [based on Fig. 13(b)], images that represent low-rank signals [Fig. 13(c)] and high-rank signals [Fig. 13(d)] are displayed. We could identify a myocardial perfusion signal after low-rank HOSVD processing [Fig. 13(c)] and visibly more compared to the conventional AM [Fig. 13(a)]. On the contrary, the high-rank HOSVD [Fig. 13(d)] processed image does not show a microbubble signal inside the myocardium but more signal inside the left ventricle. These results show that HOSVD could separate the slow- and fast-moving microbubble signal. The images post-FLASH destruction sequence, processed with conventional AM and low-rank HOSVD [Fig. 13(e) and (f)], displays significantly less signal inside the myocardium; only a feeding vessel is observed, which can be expected right after the FLASH destruction pulses. It signifies that the detected signal on pre-FLASH images originates from microbubbles

rather than tissue residuals. The signal strength ratio inside the myocardium region pre- and post-FLASH destruction sequence of the low-rank HOSVD processed images is 3 dB higher than the conventional AM processing. It indicates that the HOSVD processing provides better contrast detection inside the myocardium.

V. DISCUSSION

In this study, we proposed to use HOSVD as a method to process the complementary pulses of MPCS instead of conventionally summing or subtracting them. We expect HOSVD to improve contrast detection in case in the presence of motion, nonlinear propagation, and slow microbubble flow, where the commonly used conventional MPCS and spatiotemporal SVD are less effective. The results of both our in vitro and in vivo porcine heart model show that HOSVD provides superior contrast detection. In our flow phantom experiment, HOSVD exceeds conventional MPCS by 10 dB, AM + SVD by 11 dB, and spatiotemporal SVD and unfiltered by 14 dB each. In our in vivo open-chest implementation, HOSVD provides better motion artifact suppression than conventional MPCS by 19 dB and improves contrast detection in a transthoracic (closed chest) experiment by 3 dB.

HOSVD contrast detection improvement over conventional MPCS and SVD is more powerful in the case when motion artifacts exist. As shown by the results of our in vitro experiment, HOSVD provides 10-dB CBR improvement over AM when the probe is moving, compared to 5-dB improvement when the probe is static (where AM is generally working well in suppressing tissue signal). In the case of SVD, it has been reported that the spatiotemporal SVD filter's capability to detect flow is deteriorated when the flow speed is slower than the tissue motion velocity [14], [17], [18]. Hence, the nearly static microbubble flow in moving probe in vitro experiment and the in vivo and the myocardial perfusion are difficult cases for spatiotemporal SVD filtering because of the high spatiotemporal correlation between the microbubble and tissue signals when they are moving along. The improvement in contrast signal separation achieved by HOSVD over SVD becomes clear by comparing the spatial mode singular vectors of the two methods. As seen in Section II, HOSVD was able to separate contrast and tissue signal into individual spatial components. On the other hand, SVD was not able to discriminate since all component still consists of a mixture of contrast and tissue signal (see the Appendix), which make the filtering process unfeasible since no component or set of components could be selected that contains contrast signal only. Consequently, the in vitro results show that SVD completely fails to separate the microbubble signal when the probe is moving since it provides the same CBR as in the unfiltered images, while HOSVD improved the CBR up to 17 dB, and the CBR improvement in the open-chest in vivo results is due to the fact that HOSVD could better overcome the motion artifacts, compared to conventional AM processing. The tissue displacement (especially at the apex of the heart) caused signal misalignment when the pulses were summed (conventionally processed) and thus impaired

tissue suppression. The SVD filter was also tested (not shown here) on the in vivo dataset, but it only showed improvement of contrast visibility over AM during the brief moments where tissue was mostly still, and therefore, using the SVD filtering technique would have limited usability in the clinical application of myocardial perfusion imaging.

Analogous to the regular 2-D SVD filtering, HOSVD filter outcome is determined by the rank selection. Due to the different complexity of the images, we used different rank selection methods for spatial and temporal modes to process the in vitro and in vivo data. The rank selection for the pulsing ranks was straightforward in both our experiments: the last rank contained most of the nonlinear bubble signals, for both AM and AMPI. In an additional test (not shown here), we applied HOSVD filtering on a PI dataset and found again that taking the last rank—rank 2, in that case—showed most contrast. It is also worth mentioning that the performance of HOSVD by rank selection on the pulsing mode is already better than the conventional pulse summation in the in vitro dataset and yields at least similar results for the in vivo dataset.

The spatial selection based on the tensor core magnitude pulsing mode rank 3 only works in the in vitro dataset. The tissue-mimicking material (PVA) has a homogeneous and much simpler structure than the structure of a porcine heart. Furthermore, the motion emulated by the probe is rigid, while the cardiac motion field is complex. In the in vitro case, the contrast-signal ranks were clearly visible both from the images as well as the rank magnitudes, thus easily facilitating the rank selection. In the in vivo case, the separation between tissue and contrast signal was not done into unique ranks, yet a global crossover point between the signals could be identified. The temporal selection that we adapted for the in vivo dataset worked well, yet it did continue on a fairly good separation of the microbubble and tissue response from selecting the last pulsing rank. We expect that the physiological tissue motion of the organ will, in the end, determine the algorithm: with limited and relatively uniform motion, the in vitro algorithm should be tested first, and with complex motion field, the in vivo algorithm. We performed temporal selection for the in vivo data to select perfusion, and by doing so, we removed some of the fast-moving bubble signals. In vitro temporal selection is not possible nor necessary because the microbubble inside the channel did not flow. Thus, its only motion was solely modulated by the rigid global motion. It should be noted that we adopted an empirical threshold for choosing spatial ranks of both in vitro and in vivo methods. Although these values work for our data, they should be checked and adjusted for different ensemble length and motion.

The decomposition of HOSVD highly depends on the ensemble length of the temporal samples. The temporal length needs to be optimized according to the motion that occurs in the images. For in vitro data during probe motion, increasing the ensemble length from five to 20 temporal samples increased the resulting CBR by 1.5 dB. However, based on our empirical results, increasing the ensemble length even more does not improve the CBR further, so 20 temporal samples were chosen. A longer ensemble is potentially better when

the probe is static; yet, 20 temporal samples already provide sufficient CBR. For the in vivo data, 20 temporal samples (180 total frames with “checkerboard pattern” transmission and three synthetic subaperture) provided satisfactory results.

CBR is an appropriate parameter to quantitatively assess the filter’s performance in the in vitro data because the contrast and background regions are clearly located in different pixels. On the other hand, it is less obvious to determine contrast and background regions in our in vivo data since it can be expected that all tissue is well perfused in these healthy animals. For cardiac application, normally, CBR is calculated by drawing the contrast region on the chamber [11] or a big vessel [16] that only contain a microbubble signal and the background region is drawn on the myocardium. However, for assessing myocardial perfusion, the contrast is located inside the myocardium itself. Even after filtering, the myocardium still consists of both microbubble and tissue signals. This issue is alleviated for our open-chest images because some part of the myocardium predominantly consist of tissue signal motion artifact. Thus, an ROI that represents tissue region can be drawn. However, it measures the filter performance to suppress tissue clutter instead of quantitatively assess the amount of myocardial perfusion. For closed-chest images, there was not any region that only consisted of tissue signal, but it can be observed that the myocardium was predominantly consisting of only tissue signal, right after the FLASH destruction sequence. Thus, we drew one region on the myocardium and the contrast detection between HOSVD and MPCS was assessed by comparing the ratio between pre- and post-FLASH destruction pulses. There are several possible reasons why the quantitative improvement in the open-chest data (19 dB) is significantly greater than that in the closed-chest data (3 dB). First, it is important to note that the parameters used to assess the two are different, with CBR being used to assess tissue suppression and CR being used to assess contrast detection. Second, in the open-chest acquisition, the unobstructed myocardium is located directly under the probe, whereas in the closed-chest acquisition, the myocardium is located 3–4 cm deeper and obstructed by layers of fat and nonlinear propagation of microbubbles in the right ventricle. Finally, tissue motion in the open-chest (unbounded) heart in the apical view is faster than in the parasternal short axis, resulting in the AM being less effective in suppressing tissue signals and allowing for greater improvement.

It should be noted that our HOSVD clutter filtering has some shortcomings. It is possible that some parts of the microbubble signal may be removed due to imperfect signal decomposition. In addition, the computation time for our HOSVD clutter filtering method is longer than that for 2-D SVD. In our in vivo data processing, it took 5.5 s to process an ensemble of 20 frames, as opposed to 0.25 s for 2-D SVD. However, it is important to mention that real-time processing is not an essential necessity for now, and therefore, the longer computation time is not a crucial issue and might later be achieved by code optimization and/or parallelization. Furthermore, our current implementation can still be improved. First, the spatial singular vectors that consist of microbubble signal mostly are not located in consecutive ranks. Thus,

selecting a subspace of the spatial singular vectors means that some ranks that consist of tissue signal mostly would be kept, thus reducing contrast to background. Clustering the ranks instead of selecting a range could be a solution. Second, the optimal ensemble length depends on the amount of motion that occurs in the data. Adaptive ensemble length tailored to the amount of motion (cardiac phase) could improve contrast detection. Third, the inclusion of more input dimensions, such as imaging frequency or coded pulsing, might be investigated in the future, yet this would add even more complexity to the rank selection. Fourth, when using a probe that would have sufficient bandwidth to also allow second harmonic imaging, the data might be further filtered to isolate the second harmonic signal. Since our probe had limited bandwidth, we used the full fundamental band only. Fifth, anatomical cropping of the ROI from the overall cardiac images may result in better clutter suppression since fewer data need to be decomposed into the limited number of ranks. However, we found that coarse manual segmentation led to loss of details in the images such as the exact border delineation, which reduced our confidence in the interpretation of the contrast images. Automated segmentation may overcome this issue. Finally, HOSVD could benefit from an implementation of prefilter tissue motion compensation because the tissue signal would be more coherent and longer ensemble lengths could be adopted. We expect that the *in vivo*-data decomposition will resemble more the *in vitro*-data decomposition and associated simpler and better contrast signal isolation. To reach that aim, the motion should be corrected while preserving the local tissue phase information such as achieved by, e.g., the Lagrangian beamformer by Cormier et al. [27].

One further step of using the HFR imaging might be actual tracking of the contrast agent microbubbles, e.g., during the first filling of the vessels after FLASH destruction, which would provide further insight in flow velocities in the vessels independent of their flow direction. Such tracking might be applied using the recent ultrasound localization techniques; the proposed filter can be used to initially suppress the tissue clutter. In the current clinical use of CE, time-intensity curves of FLASH-replenishment tests are regularly used to assess local perfusion. We showed that at least in *in vitro* experiments, our technique is more robust to motion than the other methods. Yet, we expect that the robustness to motion *in vivo* is not sufficient to fully avoid cyclic patterns in the TICs, and cardiac gating might still be preferred while analyzing FLASH-replenishment tests.


VI. CONCLUSION

We have presented the application of HOSVD on multipulse contrast sequences to improve contrast detection in ultrafast ultrasound imaging. Both our *in vitro* and *in vivo* assessments have demonstrated that our proposed implementation provides superior contrast detection over the conventional processing scheme in MPCS and the more recent spatiotemporal SVD filter. The improvement could give more confidence in detecting microbubble even in the case of mild tissue motion and visualizing flow in the coronary circulation.


APPENDIX

Examples of SVD rank selection when spatiotemporal SVD was implemented on moving probe images are shown in the figures below. Low ranks with high spatial correlation were selected. However, SVD was unable to separate tissue and microbubble signals in both the implementation of full aperture transmission and AM images.

SVD on full aperture



SVD on AM images



ACKNOWLEDGMENT

The authors thank Daniel J. Bowen and Bernardo Raposo Loff Barreto, Erasmus MC, The Netherlands, for performing the sonography in some of our *in vivo* acquisition. They would also like to thank Robert Beurskens, Erasmus MC, The Netherlands, for the support on building the *in vitro* setup.

Geraldi Wahyulaksana, Luxi Wei, Jason Voorneveld, and Mihai Strachinaru are with the Biomedical Engineering, Department of Cardiology, Erasmus University Medical Center, 3015 GD Rotterdam, The Netherlands (e-mail: g.wahyulaksana@erasmusmc.nl; l.wei@erasmusmc.nl; j.voorneveld@erasmusmc.nl, m.strachinaru@erasmusmc.nl).

Maaïke te Lintel Hekkert and Dirk J. Duncker are with the Thoraxcenter, Department of Experimental Cardiology, Erasmus University Medical Center, 3015 GD Rotterdam, The Netherlands (e-mail: m.telintelhekkert@erasmusmc.nl; d.duncker@erasmusmc.nl).

Nico de Jong, Antonius F. W. van der Steen, and Hendrik J. Vos are with the Biomedical Engineering, Department of Cardiology, Erasmus University Medical Center, 3015 GD Rotterdam, The Netherlands, and also with the Medical Imaging Group, Department of Imaging Physics, Faculty of Applied Sciences, Delft University of Technology, 2600 AA Delft, The Netherlands (e-mail: n.dejong@erasmusmc.nl; a.vandersteen@erasmusmc.nl; h.vos@erasmusmc.nl).

REFERENCES

- [1] R. Senior et al., "Clinical practice of contrast echocardiography: Recommendation by the European association of cardiovascular imaging (EACVI) 2017," *Eur. Heart J.-Cardiovascular Imag.*, vol. 18, no. 11, p. 1205, Nov. 2017, doi: [10.1093/ehjci/jex182](https://doi.org/10.1093/ehjci/jex182).
- [2] M. Strachinaru and F. J. T. Cate, "Microbubble enhanced echocardiography in current cardiology practice," *Rev. Cardiovascular Med.*, vol. 23, no. 6, p. 202, May 2022.
- [3] M. Dewey et al., "Clinical quantitative cardiac imaging for the assessment of myocardial ischaemia," *Nature Rev. Cardiol.*, vol. 17, no. 7, pp. 427–450, Jul. 2020, doi: [10.1038/s41569-020-0341-8](https://doi.org/10.1038/s41569-020-0341-8).
- [4] M. A. Averkiou, M. F. Bruce, J. E. Powers, P. S. Sheeran, and P. N. Burns, "Imaging methods for ultrasound contrast agents," *Ultrasound Med. Biol.*, vol. 46, no. 3, pp. 498–517, Mar. 2020, doi: [10.1016/j.ultrasmedbio.2019.11.004](https://doi.org/10.1016/j.ultrasmedbio.2019.11.004).
- [5] M. A. Averkiou, "Tissue harmonic ultrasonic imaging," *Comptes Rendus L'Académie Sci.-Phys.*, vol. 2, no. 8, pp. 1139–1151, 2001, doi: [10.1016/S1296-2147\(01\)01259-8](https://doi.org/10.1016/S1296-2147(01)01259-8).
- [6] D. H. Simpson, C. T. Chin, and P. N. Burns, "Pulse inversion Doppler: A new method for detecting nonlinear echoes from microbubble contrast agents," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 46, no. 2, pp. 372–382, 1999.
- [7] G. A. Brock-Fisher, "Means for increasing sensitivity in non-linear ultrasound imaging systems," *J. Acoust. Soc. Amer.*, vol. 101, no. 6, p. 3240, Jun. 1997, doi: [10.1121/1.418339](https://doi.org/10.1121/1.418339).
- [8] P. J. Phillips, "Contrast pulse sequences (CPS): Imaging nonlinear microbubbles," in *Proc. IEEE Ultrason. Symp.*, vol. 2, 2001, pp. 1739–1745, doi: [10.1109/ULTSYM.2001.992057](https://doi.org/10.1109/ULTSYM.2001.992057).
- [9] M.-X. Tang, N. Kamiyama, and R. J. Eckersley, "Effects of nonlinear propagation in ultrasound contrast agent imaging," *Ultrasound Med. Biol.*, vol. 36, no. 3, pp. 459–466, Mar. 2010, doi: [10.1016/J.ULTRASMEDBIO.2009.11.011](https://doi.org/10.1016/J.ULTRASMEDBIO.2009.11.011).
- [10] C. Tremblay-Darveau et al., "The role of microbubble echo phase lag in multipulse contrast-enhanced ultrasound imaging," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 65, no. 8, pp. 1389–1401, Aug. 2018, doi: [10.1109/TUFFC.2018.2841848](https://doi.org/10.1109/TUFFC.2018.2841848).
- [11] A. Stanzola et al., "Motion artifacts and correction in multipulse high-frame rate contrast-enhanced ultrasound," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 66, no. 2, pp. 417–420, Feb. 2019, doi: [10.1109/TUFFC.2018.2887164](https://doi.org/10.1109/TUFFC.2018.2887164).
- [12] A. Heimdal and H. Torp, "Ultrasound Doppler measurements of low velocity blood flow: Limitations due to clutter signals from vibrating muscles," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 44, no. 4, pp. 873–881, Jul. 1997.
- [13] O. Couture, M. Fink, and M. Tanter, "Ultrasound contrast plane wave imaging," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 59, no. 12, pp. 2676–2683, Dec. 2012, doi: [10.1109/TUFFC.2012.2508](https://doi.org/10.1109/TUFFC.2012.2508).
- [14] C. Demeñe et al., "Spatiotemporal clutter filtering of ultrafast ultrasound data highly increases Doppler and fulltrasound sensitivity," *IEEE Trans. Med. Imag.*, vol. 34, no. 11, pp. 2271–2285, Nov. 2015, doi: [10.1109/TMI.2015.2428634](https://doi.org/10.1109/TMI.2015.2428634).
- [15] Y. Desailly, A.-M. Tissier, J.-M. Correas, F. Wintzenrieth, M. Tanter, and O. Couture, "Contrast enhanced ultrasound by real-time spatiotemporal filtering of ultrafast images," *Phys. Med. Biol.*, vol. 62, no. 1, pp. 31–42, Jan. 2017, doi: [10.1088/1361-6560/62/1/31](https://doi.org/10.1088/1361-6560/62/1/31).
- [16] J. Voorneveld et al., "High-Frame-Rate contrast-enhanced ultrasound for velocimetry in the human abdominal aorta," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 65, no. 12, pp. 2245–2254, Dec. 2018, doi: [10.1109/TUFFC.2018.2846416](https://doi.org/10.1109/TUFFC.2018.2846416).
- [17] J. Zhu et al., "High frame rate contrast-enhanced ultrasound imaging for slow lymphatic flow: Influence of ultrasound pressure and flow rate on bubble disruption and image persistence," *Ultrasound Med. Biol.*, vol. 45, no. 9, pp. 2456–2470, Sep. 2019, doi: [10.1016/j.ultrasmedbio.2019.05.016](https://doi.org/10.1016/j.ultrasmedbio.2019.05.016).
- [18] G. Wahyulaksana et al., "Independent component analysis filter for small vessel contrast imaging during fast tissue motion," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 69, no. 7, pp. 2282–2292, Jul. 2022, doi: [10.1109/TUFFC.2022.3176742](https://doi.org/10.1109/TUFFC.2022.3176742).
- [19] A. Thorstensen, H. Dalen, B. H. Amundsen, and A. Støylen, "Peak systolic velocity indices are more sensitive than end-systolic indices in detecting contraction changes assessed by echocardiography in young healthy humans," *Eur. Heart J.-Cardiovascular Imag.*, vol. 12, no. 12, pp. 924–930, Dec. 2011, doi: [10.1093/EJCHOCARD/JER178](https://doi.org/10.1093/EJCHOCARD/JER178).
- [20] L. De Lathauwer, B. De Moor, and J. Vandewalle, "A multilinear singular value decomposition," *SIAM J. Matrix Anal. Appl.*, vol. 21, no. 4, pp. 1253–1278, Jan. 2000, doi: [10.1137/S0895479896305696](https://doi.org/10.1137/S0895479896305696).
- [21] M. Kim, C. K. Abbey, J. Hedhli, L. W. Dobrucki, and M. F. Insana, "Expanding acquisition and clutter filter dimensions for improved perfusion sensitivity," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 64, no. 10, pp. 1429–1438, Oct. 2017, doi: [10.1109/TUFFC.2017.2719942](https://doi.org/10.1109/TUFFC.2017.2719942).
- [22] G. Shechter, J. R. Resar, and E. R. McVeigh, "Displacement and velocity of the coronary arteries: Cardiac and respiratory motion," *IEEE Trans. Med. Imag.*, vol. 25, no. 3, pp. 369–375, Mar. 2006, doi: [10.1109/TMI.2005.862752](https://doi.org/10.1109/TMI.2005.862752).
- [23] C. Papadacci, M. Pernot, M. Couade, M. Fink, and M. Tanter, "High-contrast ultrafast imaging of the heart," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 61, no. 2, pp. 288–301, Feb. 2014, doi: [10.1109/TUFFC.2014.6722614](https://doi.org/10.1109/TUFFC.2014.6722614).
- [24] A. Rodriguez-Molares et al., "The UltraSound ToolBox," in *Proc. IEEE Int. Ultrason. Symp. (IUS)* Oct. 2017, pp. 1–4, doi: [10.1109/ULTSYM.2017.8092389](https://doi.org/10.1109/ULTSYM.2017.8092389).
- [25] D. Maresca, M. Correia, M. Tanter, B. Ghaleb, and M. Pernot, "Adaptive spatiotemporal filtering for coronary ultrafast Doppler angiography," *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 65, no. 11, pp. 2201–2204, Nov. 2018, doi: [10.1109/TUFFC.2018.2870083](https://doi.org/10.1109/TUFFC.2018.2870083).
- [26] N. Vervliet, O. Debals, L. Sorber, M. Van Barel, and L. De Lathauwer. (2016). *TensorLab 3.0*. [Online]. Available: <http://www.tensorlab.net>
- [27] P. Cormier, J. Porée, C. Bourquin, and J. Provost, "Dynamic myocardial ultrasound localization angiography," *IEEE Trans. Med. Imag.*, vol. 40, no. 12, pp. 3379–3388, Dec. 2021, doi: [10.1109/TMI.2021.3086115](https://doi.org/10.1109/TMI.2021.3086115).



Geraldi Wahyulaksana (Graduate Student Member, IEEE) was born in Jakarta, Indonesia, in 1992. He received the M.S. degree in electrical engineering from the Eindhoven University of Technology, Eindhoven, The Netherlands, in 2017. He is currently pursuing the Ph.D. degree with the Department of Biomedical Engineering, Erasmus Medical Center, Rotterdam, The Netherlands.

After working on contrast agent modeling for MRI, from 2017 to 2019, he worked as a

Software Designer at Son and developed an object detection system for a wireless power transfer system with air-coupled ultrasound. His research interests include methods to detect microbubbles with high-frame-rate ultrasound imaging.



Luxi Wei (Member, IEEE) received the B.Sc. degree in honors biophysics from the University of British Columbia, Vancouver, BC, Canada, and the M.Sc. degree in medical biophysics from the University of Toronto, Toronto, ON, Canada, in 2016 and 2019, respectively. She is currently pursuing the Ph.D. degree with the Department of Biomedical Engineering, Erasmus Medical Center, Rotterdam, The Netherlands.

Her research interest includes high-frame-rate ultrasound, contrast agent imaging, volumetric imaging, and transducer development.



Jason Voorneveld was born in Johannesburg, South Africa, in 1987. He received the B.Sc. degree in electromechanical engineering and the M.Sc. degree in biomedical engineering from the University of Cape Town, Cape Town, South Africa, in 2009 and 2014, respectively, and the Ph.D. degree in biomedical engineering from the Erasmus University Medical Center, Rotterdam, The Netherlands, in 2019.

He is currently working as a Postdoctoral Fellow with the Erasmus University Medical Center, Rotterdam. His research interests include high-frame-rate ultrasound imaging, blood-flow quantification, and ultrasound contrast agents.



Maaike te Lintel Hekkert received the bachelor's degree in applied science, Rotterdam, The Netherlands, in 2003.

She has been working as a Research Technician with the Division of Experimental Cardiology, Department of Cardiology, Erasmus University Medical Center, Rotterdam. She has performed all kinds of cardiovascular experiments in pigs, rats, and mice, but her main focus has been on ischemia/reperfusion experiments, with open- and closed-chest conditions. She is

adept in all aspects of the experiments, including sedation, intubation, anesthesia, execution of the experimental protocols, monitoring of hemodynamics, and postoperative care. In addition, she has many years of experience with coronary catheterization as well as complex surgical procedures.



Mihai Strachinaru received the M.D. degree and the cardiology specialization from the "Carol Davila" University of Medicine and Pharmacy, Bucharest, Romania, in 2001 and 2008, respectively, and the Ph.D. degree in shear wave echocardiography from the Department of Biomedical Engineering, Erasmus University Medical Center, Rotterdam, The Netherlands, in 2020.

He then moved on to work as a Cardiologist in France from 2008 to 2009, Belgium from 2009 to 2016, and The Netherlands from 2016 to 2019. During this period, he developed a keen interest in cardiac noninvasive imaging, especially echocardiography. He currently holds Romanian, French, and European certifications in transthoracic and transesophageal echocardiography. He is currently a Postdoctoral Researcher with the Erasmus University Medical Center and an Intensive Care Cardiologist with the CHIREC Hospital, Brussels, Belgium. His research interests include high-frame-rate ultrasound imaging in all its aspects: 2-D and 3-D, cardiac shear waves, ultrafast Doppler, contrast imaging, and related preclinical and clinical applications.



Dirk J. Duncker received the Ph.D. and M.D. degrees from Erasmus University, Rotterdam, The Netherlands, in 1988 and 1989, respectively.

He worked as a Postdoctoral Research Fellow with the University of Minnesota, Minneapolis, MN, USA, from 1991 to 1994, after which he returned to the Department of Cardiology (Thoraxcenter), Erasmus University Medical Center, Rotterdam. In 1995, he received a five-year Research Fellowship Stipend from

the Royal Netherlands Academy of Arts and Sciences, Amsterdam, The Netherlands, and in 2000, he became an Established Investigator of the Netherlands Heart Foundation. He was appointed as a Professor of Experimental Cardiology in 2003. His research interests include the regulation of coronary blood flow and coronary microvascular tone in health as well as in a variety of cardiovascular disorders, including metabolic disorders, ischemic heart disease, cardiac hypertrophy, and heart failure.



Nico de Jong graduated from the Delft University of Technology, Delft, The Netherlands, in 1978. He received the M.Sc. degree in applied physics in pattern recognition and the Ph.D. degree in acoustic properties of ultrasound contrast agents in 1993.

From 2003 to 2011, he was a part-time Professor with the University of Twente, Enschede, The Netherlands, in the Group Physics of Fluids headed by Prof. Detlef Lohse. He is currently the Vice Head of Biomedical Engineering with the Thoraxcenter, Erasmus University Medical Center, Rotterdam, The Netherlands, headed by Prof. Ton van der Steen. He is the Head of the Medical Imaging Group, Technical University, Delft. He is the Founder and an Organizer of the Annual European Symposium (this year for the 25th time, see <http://www.echocontrast.nl>) on ultrasound contrast imaging, held in Rotterdam and attended by approximately 175 scientists from universities and industries all over the world. He is on the Safety Committee of the World Federation of Ultrasound in Medicine and Biology (WFUMB).

Dr. de Jong is an Associate Editor of *Ultrasound in Medicine and Biology* and the IEEE TRANSACTIONS ON ULTRASONICS, FERROELECTRICS, AND FREQUENCY CONTROL. He has been a guest editor for special issues of several journals. He teaches at technical universities and the Erasmus University Medical Center, Rotterdam. He has graduated 38 Ph.D. students and is currently supervising more than 12 Ph.D. students.



Antonius F. W. van der Steen (Fellow, IEEE) received the M.Sc. degree in applied physics from the Technical University Delft, Delft, The Netherlands, and the Ph.D. degree in medical science from the Catholic University Nijmegen, Nijmegen, The Netherlands, in 1989 and 1994, respectively.

He is currently the Head of Biomedical Engineering with the Thorax Center, Erasmus University Medical Center, Rotterdam, The Netherlands. He is an expert in ultrasound, cardiovascular imaging, and cardiovascular biomechanics. He has a career at the crossroads of engineering, health care, and industry. He has experience in running large consortia as the Co-Founder and former Chairman of the Medical Delta, which comprises over 280 scientists working on technical solutions for sustainable health. He was also the Co-PI of ParisK, one of the large CTMM projects (16 MEuro). His international profile is high, with more than 200 invited lectures all over the world, and Guest Professorship/Guest Researcher in Canada, Japan, and China.

Dr. van der Steen is a fellow of the European Society of Cardiology. He is a member of the Netherlands Academy of Technology (AcT) and a Board Member of the Royal Netherlands Academy of Sciences (KNAW). He was a recipient of the Simon Stevin Master Award and the NWO PIONIER Award in Technical Sciences.



Hendrik J. Vos (Member, IEEE) received the M.Sc. degree in applied physics from Delft University of Technology, Delft, The Netherlands, and the Ph.D. degree from the Department of Biomedical Engineering, Erasmus University Medical Center, Rotterdam, The Netherlands, in 2004 and 2010, respectively.

He worked as a Postmaster Researcher with the University of Florence, Florence, Italy, and a Contract Researcher for the petrochemical industry on cutting-edge ultrasonic solutions. He currently is an Associate Professor with the Erasmus University Medical Center, Rotterdam, and the Delft University of Technology, Delft. He received a Dutch NWO-TTW-VIDI personal grant in 2018. His research interests include acoustical array technology for biomedical imaging in all its aspects: transducers, 2-D and 3-D beamforming, cardiac shear waves, ultrafast Doppler, contrast imaging, and related subclinical and clinical studies.