# Spectral Tailoring for Boron Neutron Capture Therapy

The research described in this thesis was performed within a cooperation of the Institute for Energy of the Joint Research Centre (JRC) of the European Commission, P.O. Box 2, 1755 ZG, Petten, The Netherlands and the Section Physics of Nuclear Reactors (PNR), of the Department of Radiation, Radionuclides & Reactors ( $R^3$ ) of the Faculty of Applied Sciences of the Delft University of Technology, Mekelweg 15, 2629 JB, Delft, The Netherlands.

## Spectral Tailoring for Boron Neutron Capture Therapy

Proefschrift

ter verkrijging van de graad van doctor aan de Technische Universiteit Delft, op gezag van de Rector Magnificus prof. dr. ir. J.T. Fokkema, voorzitter van het College voor Promoties, in het openbaar te verdedigen op dinsdag 26 juni 2007 om 15:00 uur

> door: Victor Alexander NIEVAART natuurkundig ingenieur, geboren te Apeldoorn

Dit proefschrift is goedgekeurd door de promotoren: Prof.[em] dr. ir. H. van Dam Prof. dr. ir. T.H.J.J. van der Hagen Toegevoegd promotor: Dr. ir. J.L. Kloosterman

Samenstelling promotiecommissie: Rector Magnificus, voorzitter Prof.[em] dr. ir. H. van Dam Prof. dr. ir. T.H.J.J. van der Hagen Dr. ir. J.L. Kloosterman Prof. dr. med. W. Sauerwein Prof. dr. ir. M. Hendriks-de Jong Prof. dr. H.T. Wolterbeek Dr. R.L. Moss

Technische Universiteit Delft Technische Universiteit Delft Technische Universiteit Delft University Hospital Essen Erasmus Medisch Centrum Rotterdam Technische Universiteit Delft Joint Research Centre Petten

© 2007 V.A. Nievaart and IOS Press

All rights reserved. No part of this book may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, without prior permission from the publisher.

ISBN 978-1-58603-762-8

Keywords: Boron Neutron Capture Therapy, Optimal Source Neutrons, Adjoint Monte Carlo

Published and distributed by IOS Press under the imprint Delft University Press

Publisher IOS Press Nieuwe Hemweg 6b 1013 BG Amsterdam The Netherlands tel: +31-20-688 3355 fax: +31-20-687 0019 email: info@iospress.nl www.iospress.nl www.dupress.nl

LEGAL NOTICE The publisher is not responsible for the use which might be made of the following information.

PRINTED IN THE NETHERLANDS

This thesis consists of two parts. The first part provides a complete overview of the research performed and the second part consists of a collection of key papers:

- V.A. Nievaart, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen and H. van Dam, "A parameter study to determine the optimal source neutron energy in Boron Neutron Capture Therapy of brain tumours," Phys.Med.Biol. 49, 4277-4292 (2004).
  © 2004 IOP Publishing Ltd ; www.iop.org/journals/pmb
- II. V.A. Nievaart, D. Légràdy, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen and H. van Dam, "Application of adjoint Monte Carlo to accelerate simulations of mono-directional beams in treatment planning for Boron Neutron Capture Therapy," Med.Phys. 34 (4), 1321-1335 (2007).
   © 2007 Am. Assoc. Phys. Med.; www.medphys.org
- III. V.A. Nievaart, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen, H. van Dam, A. Wittig, M. Malago and W. Sauerwein, "Design of a Rotating Facility for Extracorporal Treatment of an Explanted Liver with Disseminated Metastases by Boron Neutron Capture Therapy with an Epithermal Neutron Beam," Rad.Res. 166, 81-88 (2006). © 2006 by Radiation Research Society ; www.radres.org

<b>CONTENTS</b>	
-----------------	--

SUMMARY					
SAME	SAMENVATTING				
PUBL	ICATIONS	XV			
СНАР	TER 1 INTRODUCTION	1			
1.	GENERAL INTRODUCTION	1			
1.1	THE BASICS OF BNCT	1			
1.1.1	The four maior BNCT dose components in tissue	2			
1.1.2	Biologically weighted doses and $^{10}B$ compounds	6			
1.1.3	Brief history of BNCT	7			
1.1.4	Present status of BNCT	8			
1.2	SPECTRAL TAILORING FOR BNCT	9			
1.2.1	Search for the optimal source neutron energy for BNCT	10			
1.2.2	The scope of this thesis	10			
СНАР	TER 2 OPTIMAL NEUTRONS AND DOSIMETRY	13			
2.	OPTIMAL SOURCE NEUTRONS WITH REGARD TO	13			
	DOSIMETRY IN BNCT				
2.1	BACKGROUND ON BNCT DOSIMETRY AND	13			
	RADIOBIOLOGY				
2.2	A PARAMETER STUDY FOR BNCT OF THE BRAIN	14			
2.2.1	Set-up and chosen parameter ranges	14			
2.2.2	The influence of parameter biasing on the results	16			
2.2.3	Results: the optimal source neutron energy	17			
2.2.4	Results: the influencing parameters	19			
2.2.5	Results: Improvements when using the optimal source neutrons	21			
2.3	CONCLUSIONS	23			
СНАР	TER 3 ADJOINT TECHNIQUES IN BNCT	25			
3.	APPLICATION OF ADJOINT MONTE CARLO TECHNIQUES	25			
	IN BNCT				
3.1	BACKGROUND, THEORY AND MCNP	25			
3.2	GENERAL ADJOINT SET-UP FOR BNCT	26			
3.3	ADJOINT TREATMENT PLANNING WITH A MONO-	27			
	DIRECTIONAL BEAM				
3.3.1	Adjoint Point Detector Technique (APDT)	28			
3.3.2	Legendre EXpansion Technique (LEXT)	29			
3.4	EXAMPLE: OPTIMUM CHARACTERISTICS IN THE	31			
	IRRADIATION OF A HUMAN HEAD				
3.4.1	Calculation times of forward MC, APDT and LEXT	32			
3.4.2	Optimum irradiation locations and directions	35			
3.4.3	Optimum source neutron energy group	36			
3.5	CONCLUSIONS	38			
СНАР	TER 4 THERMAL NEUTRON FIELD FACILITY	39			
4.	DESIGN OF A HOMOGENEOUS THERMAL NEUTRON	39			
	FIELD FACILITY FOR BNCT				
4.1	DESIGN OF AN IRRADIATION FACILITY FOR THE	39			

#### EXTRA-CORPORAL TREATMENT OF LIVER CANCER WITH AN EXISTING EPITHERMAL BEAM 4.1.1 Design parameters and selected materials 40 42 4.1.2 Simulating rotation with the Monte Carlo code Results of the designed liver facility at the HFR Petten 4.1.3 42 OPTIMUM NEUTRON BEAM DESIGN AND HOLDERS FOR 44 4.2 EXTRA-CORPORAL BNCT IRRADIATIONS 4.2.1 Set-up of the holder models 44 4.2.2 Expected contributions of the source neutrons inside the models 45 Optimum source neutron energies 48 4.2.3 52 CONCLUSIONS 4.3 **CHAPTER 5 CONCLUSIONS** 53 57 APPENDIX NOMENCLATURE 61 **ABBREVIATIONS** 63 REFERENCES 65 ACKNOWLEDGEMENT 73 **CURRICULUM VITAE** 75 PAPERS I-III 77

#### Spectral Tailoring for Boron Neutron Capture Therapy

Since the first clinical trials on Boron Neutron Capture Therapy in the 1950s, BNCT research has been mainly focussed on the treatment of (deep-seated) brain tumours, in particular, glioblastoma multiforme. Promising work to treat other cancers at other locations and even other diseases are in progress. Therefore, the chemists, medical doctors, physicists and biologists involved in BNCT are not only continuing to investigate and improve the (brain) clinical results, but are also investigating the new applications in BNCT. The work presented in this thesis is in the field of physics and deals, from three different viewpoints, with obtaining the optimal source neutron energy to optimise BNCT. The optimal source neutron energy is defined such as to obtain as many as possible ( $n,\alpha$ )-absorptions due to <sup>10</sup>B in the tumours and as low as possible total neutron dose in the healthy tissues and organs at risk.

Firstly, the relation between the optimal source neutron energy and the radiation biology of brain BNCT was investigated. The biological weighting factors of the four major BNCT dose components, the skin and cranium thickness, the tolerance dose in skin and brain, the <sup>10</sup>B concentration and the number of beam gammas per source neutron were varied in a theoretical study. The parameter value ranges are bounded by unexpected and/or unrealistic values. It was investigated as to what is the optimal source neutron energy for four tumours at different depths, in each of the 136 million configurations for all combinations of parameter values. By far, the modality of the optimal source neutron energies is between 1 keV and 10 keV. However, depending on where the tolerance dose is reached first, in the skin or brain, low values for <sup>10</sup>B and fast neutron related parameters in this limiting tissue result in lower or higher than modal source neutron energies.

Secondly, adjoint Monte Carlo (MC) techniques are developed to find more quickly the optimal source neutrons' location, direction and energy. The adjoint MC is very suitable for this task because the adjoint particles fly mainly towards regions that from the statistics point of view are the best directions to irradiate from. Until now it was impossible to gather acceptable statistics of adjoint MC particles which traverse the adjoint detector perpendicularly, rendering the adjoint method inapplicable for mono-directional beams. The BNCT beam available in Petten can be regarded as mono-directional. This problem is solved with the use of next event estimators or with the application of a Legendre expansion technique. In the first case, adjoint particles are transported deterministically through a beam shaped channel to a point detector far away from the geometric model. The particles will traverse the disk shaped entrance of this tube (the beam exit in the actual geometry) perpendicularly. This method is slow when many events are involved that are not contributing to the point detector, e.g. neutrons in a scattering medium. In a second approach, adjoint particles that traverse an adjoint shaped detector plane are used to estimate the Legendre coefficients for expansion of the angular adjoint function. This provides an estimate of the adjoint function for the direction normal to the detector plane. In a realistic head phantom with 10 organs at risk and 10 tumours, the two adjoint techniques are 1.8 to 3.3 times faster than the forward MC calculations when 1020 different orientations of a gamma beam with a diameter larger than 5 cm are simulated. In case of a neutron beam, only the adjoint technique based on Legendre expansion is faster, 6.6 up to 20 times, than forward MC. In general, in case of small diameter beams adjoint MC calculations are only preferable for a large number of beams and a small number of regions of interest. For larger beam sizes, fewer beams and/or many regions of interest makes the adjoint favourable over the forward calculations. As well as being able to obtain the optimal locations to irradiate from, the optimal source neutron energy at every location around the head can also be obtained with adjoint MC. Compared with the Petten beam spectrum, it is found that only thermal and low-energy epithermal source neutrons can give significant improvements to the ratio of the thermal neutron flux in tumours to organs at risk.

Thirdly, the optimal source neutron energies are determined in order to obtain a homogeneous thermal neutron fluence in a prescribed volume. Ideally, when the homogeneity, defined as the ratio of minimum to maximum thermal neutron flux, is unity, the same thermal neutron related dose can be given in every part of this volume. When using the Petten beam with its current neutron spectrum for the extracorporeal BNCT treatment of liver cancer, the best homogeneity obtained is 0.68 in a volume of 2.4 litres. This volume is spheroidal shaped and rotating. The rotation is simulated in MCNP by averaging the particle tracks in tori shaped tally volumes. With a combination of source neutrons of 30% around 0.1 eV and 70% around 10 keV, a homogeneity of 0.95 can be reached in a cuboid model. This result was obtained after calculating the detector response functions for thermal neutrons in different volume shapes (i.e. cuboid, cylinder and sphere) as a function of source neutron energy. By applying linear programming, the detector response functions of the source neutron energies were combined such that the homogeneity in each volume shape is optimised.

The outcome of the three parts of this thesis shows that 3 neutron energy regimes should be prescribed in BNCT. As well as the 10 keV epithermal source neutrons, low epithermal source neutrons of around 1 eV and thermal source neutrons with energies of 0.1 eV must be used.

Petten, May 2007, V.A. Nievaart

#### **Spectrum Optimalisatie voor Boron Neutron Capture Therapy**

Al vanaf de eerste klinische studies naar Boron Neutron Capture Therapy<sup>1</sup> (BNCT) in de jaren 50 ligt de nadruk van het BNCT onderzoek voornamelijk op de behandeling van diep gesitueerde hersentumoren en wel in het bijzonder de glioblastoma multiforme. Pas de laatste decennia zijn veelbelovende ontwikkelingen gaande om andere vormen en locaties van kanker en zelfs niet kankerzijnde ziektes te gaan behandelen. Dit betekent dat de bij BNCT betrokken chemici, artsen, fysici en biologen niet alleen proberen de hersenresultaten te verbeteren maar ook onderzoek te doen naar de nieuwe BNCT toepassingen. Dit proefschrift heeft een fysische grondslag en behandelt vanuit 3 invalshoeken het verkrijgen van de optimale bronneutronenenergie om zodoende BNCT te verbeteren. De optimale bronneutronen worden gekenmerkt door een energie die zoveel mogelijk (n, $\alpha$ )-absorpties tengevolge van <sup>10</sup>B in de tumor genereert en tegelijkertijd een zo laag mogelijke dosis geeft in de gezonde weefsels en stralingsgevoelige organen.

De eerste invalshoek van dit proefschrift is de relatie tussen de optimale bronneutronenenergie en de radiobiologie van BNCT toegepast voor hersenen. Hiervoor zijn in deze theoretische studie de biologische weegfactoren van de vier belangrijkste BNCT dosiscomponenten, de huid- en schedeldiktes, de tolerantiedosis in huid en hersenen, de <sup>10</sup>B concentratie en het aantal brongamma's per bronneutron in de bundel gevarieerd. Voor al deze parameters zijn intervallen gekozen die zijn begrensd door onrealistische en/of niet meer te verwachten waarden. Onderzocht is wat de optimale bronneutronenenergie is voor tumoren op 4 verschillende dieptes in elk van de 136 miljoen configuraties tengevolge van alle mogelijke combinaties van parameterwaarden. Voor veruit de meeste configuraties blijken bronneutronen met een energie tussen de 1 keV en 10 keV optimaal te zijn. Alleen voor lage waarden van de <sup>10</sup>B en snelle neutronen dosis gerelateerde parameters zijn er soms afwijkende (lagere of hogere) optimale bronneutronenenergieën. Dit geldt voor deze parameters in het weefsel (huid of hersenen) waar de tolerantiedosis als eerste wordt bereikt.

Ten tweede zijn er adjoint Monte Carlo (MC) technieken ontwikkeld om sneller de optimale locatie, richting en energie van de bronneutronen te vinden. Adjoint MC is zeer geschikt omdat de adjoint deeltjes zich voornamelijk bewegen naar die plekken die statistisch gezien het gunstigste zijn om vanuit te bestralen. Tot nu toe was het onmogelijk om genoeg adjoint MC deeltjes te verzamelen die loodrecht door het detectievlak gaan. Het verkrijgen van een goede statistiek voor de adjoint in geval van een eenrichtingsbundel was dus onmogelijk. De BNCT bundel in Petten (NL) kan worden beschouwd als een eenrichtingsbundel. Het probleem is opgelost met het gebruik van 'volgende-gebeurtenis-schatters' en door toepassing van een 'Legendre-ontwikkelings' techniek. In het eerste geval worden adjointdeeltjes

<sup>&</sup>lt;sup>1</sup> In het Nederlands is BNCT vertaald als Borium Neutronenvangst Therapie.

deterministisch door een kanaal getransporteerd, die is gevormd zoals de bundel, naar een punt ver weg van de ingang. Zodoende zullen de adjointdeeltjes deze ingang (de bundelopening in werkelijkheid) loodrecht passeren. Deze methode is nadelig als veel deeltjes moeten worden gevolgd die het detectiepunt niet kunnen bereiken zoals het geval is bij neutronen in een verstrooiend medium. Bij de tweede methode worden de adjointdeeltjes die een adjointdetectorvlak passeren gebruikt voor het schatten van de Legendre coëfficiënten om zodoende de hoekafhankelijke adjointfunctie te kunnen ontwikkelen. Dit resulteert in een schatting voor de adjointfunctie in de richting loodrecht op het adjoint detectievlak. In een realistisch scenario, een hoofdfantoom met 10 tumoren in de hersenen en 10 stralingsgevoelige organen, zijn de twee adjoint technieken 1,8 tot 3,3 keer sneller dan normale voorwaartse MC berekeningen als 1020 verschillende posities van een gammabundel met een diameter groter dan 5 cm moeten worden gesimuleerd. In het geval van een neutronenbundel is alleen de Legendre techniek sneller dan normale voorwaartse MC berekeningen en wel 6,6 tot 20 keer. Voor kleine bundeldiameters kan worden geconcludeerd dat adjoint MC berekeningen voordelig zijn als er relatief veel bundelposities en weinig tumoren en/of stralingsgevoelige organen in het spel zijn. Voor grotere bundeldiameters is de adjoint methode al aantrekkelijk voor minder bundelposities en/of meer gebieden waarin de dosis berekend moet worden. Buiten de optimale bundelposities kan de adjoint techniek ook worden gebruikt om de optimale bronneutronenenergie overal rondom het hoofd te bepalen. In vergelijking met het spectrum van de Pettenbundel blijkt dat toepassing van de ene keer alleen thermische en de andere keer alleen laagepithermische bronneutronen significante verbeteringen geven in de verhouding tussen het thermische neutronenfluentietempo in de tumoren en stralingsgevoelige organen.

Ten derde is de optimale bronneutronenenergie onderzocht om een homogeen thermische neutronenfluentie te verkrijgen in een bepaald volume. Deze homogeniteit is gedefinieerd als de verhouding tussen de minimale en de maximale thermische neutronenfluentietempi in een volume. Idealiter heeft deze verhouding een waarde 1 wat betekent dat de thermische neutronen gerelateerde dosis overal in het volume hetzelfde kan zijn. In geval van behandeling van leverkanker met BNCT in Petten, waarbij de lever buiten het lichaam wordt gebracht, is de homogeniteit 0,68. Deze waarde wordt bereikt in een roterende sferoïde met een volume van 2,4 liter en gebruikmakend van het bestaande spectrum van de Pettenbundel. Met de MCNP code kan rotatie worden gesimuleerd door de geregistreerde padlengtes van de deelties in een torus te middelen over het torusvolume. Een mix van 30% 0.1 eV bronneutronen en 70% 10 keV bronneutronen resulteert in een homogeniteit van 0.95 in een kubusvormig model. Dit resultaat is verkregen door als functie van de bronneutronenenergie de detectorresponsiefuncties te berekenen voor thermische neutronen in verschillende modellen (te weten: kubus-, cilinder- en bolvormig). Met behulp van lineair programmeren zijn de detectorresponsiefuncties zo gecombineerd dat de homogeniteit in elke volumevorm is geoptimaliseerd.

De uitkomst van deze drie delen van het promotieonderzoek is dat 3 bronneutronenenergieregimes beschikbaar zouden moeten zijn in BNCT. Buiten de 10 keV epithermische bronneutronen moeten ook laagepithermische bronneutronen van rond de 1 eV en thermische bronneutronen met energieën van 0,1 eV beschikbaar zijn.

Petten, Mei 2007, V.A. Nievaart XIV PUBLICATIONS

#### PUBLICATIONS

#### FIRST AUTHOR:

V.A. Nievaart, R.L. Moss, J.L. Kloosterman and T.H.J.J. van der Hagen, "Calculating the tumour-specific optimal source neutron energy for Boron Neutron Capture Therapy with particle production and adjoint Monte Carlo techniques," PHYSOR 2002, Seoul Korea, October 7-10, (2002).

V.A. Nievaart, J.L. Kloosterman, T.H.J.J. van der Hagen, R.L. Moss, "Forward and adjoint Monte Carlo calculations for investigating the use of an adjustable filter in BNCT," In: W. Sauerwein, R. Moss, A. Wittig, editors Research and Development Neutron Capture Therapy. Bologna: Monduzzi Editore, 1065-72 (2002).

V.A. Nievaart , "Maximising the  $\alpha$ -production in brain tumours under different equivalent dose configurations in BNCT," ISNCT/YMNM3 meeting, Pisa Italy, Nov 27 - Dec 2, (2003).

V.A. Nievaart, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen and H. van Dam, "A parameter study to determine the optimal source neutron energy in Boron Neutron Capture Therapy of brain tumours," Phys.Med.Biol. 49, 4277-4292 (2004).

V.A. Nievaart, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen, H. van Dam, "Feasibility study for the extra-corporal treatment of liver cancer by BNCT at the HFR Petten," ISNCT-11, October 11-15, Boston, Conference proceeding, USA (2004).

V. Nievaart, A. Wittig, R. Moss, J. Rassow, W. Sauerwein, "Optimisation of treatment planning for multi-beam Boron Neutron Capture Therapy (BNCT) using epithermal neutron beams in patients with multiple metastases to the brain from malignant melanoma," Strahlentherpie und Onkologie, Abstractband zum DEGRO-Kongress, Sondernr.1, May (2006).

V.A. Nievaart, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen, H. van Dam, A. Wittig, M. Malago, W. Sauerwein, "Design of a rotating facility for extracorporal treatment of an explanted liver with disseminated metastases by BNCT with an epithermal neutron beam at the HFR in Petten", Radiation Research 166, 81-88 (2006).

V.Nievaart, R Moss, A Wittig, and W Sauerwein, "Extra-corporal treatment of Liver Metastases by BNCT at the HFR Petten", Journal of Physics: Conference Series of the Institute of Physics 41, 369-374 (2006).

Sander Nievaart, Ray Moss, Wolfgang Sauerwein, Massimo Malago, Jan Leen Kloosterman, Tim van der Hagen and Hugo van Dam, "An optimum source neutron spectrum and holder shape for extra-corporal treatment of liver cancer by BNCT,"

12th International Congress on Neutron Capture Therapy Kagawa International Conference Hall, Sunport Takamatsu, Japan, Oct 9-13 (2006).

Sander Nievaart, Ray Moss, Wolfgang Sauerwein and Andrea Wittig, "Use of linear programming to obtain an optimum, multi-beam treatment plan in BNCT," 12th International Congress on Neutron Capture Therapy Kagawa International Conference Hall, Sunport Takamatsu, Japan, Oct 9-13 (2006).

V.A. Nievaart, D. Légràdy, R.L. Moss, J.L. Kloosterman, T.H.J.J. van der Hagen and H. van Dam, "Application of adjoint Monte Carlo to accelerate simulations of mono-directional beams in treatment planning for Boron Neutron Capture Therapy," Med.Phys. 34 (4), 1321-1335 (2007).

V.A. Nievaart, G.G. Daquino and R.L. Moss, "Monte Carlo based treatment planning systems for Boron Neutron Capture Therapy in Petten The Netherlands," submitted to Journal of Physics: Conference Series of the Institute of Physics (2007).

#### **CO-AUTHOR:**

Mudde, RF, Deutz, L, Nievaart, VA, & Maanen, HRE van, "LDA-measurements of the turbulence in and around a venture," In W Rodi & M Mulas (Eds.), Proceeding of Ercoftac internationa symposium of engineering turbulence modelling and measurements, Ercoftac, Sardinia, Italy, pp. 80-91 (2005).

W. Sauerwein, M. Malago, R. Moss, S. Altieri, G. Hampel, A. Wittig, V. Nievaart, L. Collette, P. Mauri, R. Huiskamp, J. Michel, G. Daquino, G. Gerken, N. Bornfeld, C.E. Broelsch, "Bor-Neutroneneinfangtherapie (BNCT) zur Behandlung diffuser, nicht resektabler Lebermetastasen," Strahlentherpie und Onkologie, Abstractband zum DEGRO-Kongress, Sondernr.1, May (2006).

A. Wittig, L. Collette, J. Heimans, P. Paquis, K. Haseslsberger, V. Barsegian, C. Loquai, G. Kaiser, R. Moss, J. Rassow, F. Stecher-Rasmussen, R. Huiskamp, V. Nievaart, S. Bührmann, P. Bet, B. Hahn, K. Hideghéty, H. Arlinghaus, W. Sauerwein, "A strategy to introduce Boron Neutron Capture Therapy (BNCT), a novel radiotherapy modality, into clinical practice," The clinical trials of the EORTC BNCT Group; Strahlentherpie und Onkologie, Abstractband zum DEGRO-Kongress, Sondernr.1, May (2006).

G.Gambarini, G.G.Daquino, R.L.Moss, M.Carrara, V.A.Nievaart, E.Vanossi, "Gel dosimetry in the BNCT facility for extra-corporal treatment of corporal of liver cancer at the HFR Petten," 10th Symposium on Neutron Dosimetry. Progress in dosimetry of neutrons and light nuclei Uppsala (SV), 12-16 June (2006).

A. Roca, G.G. Daquino, R.L. Moss, S. Nievaart, R. Schmidt F. Stecher-Rasmussen, "Determination of the neutron capture sensitivity of ionisation chambers based on

neutron capture reaction rates," 12th International Congress on Neutron Capture Therapy Kagawa International Conference Hall, Sunport Takamatsu, Japan, Oct 9-13 (2006).

Ray Moss, Sander Nievaart, Lucien Pott, Andrea Wittig, Wolfgang Sauerwein, "Design and testing of a rotating, cooled device for extra-corporal treatment of liver cancer by BNCT in the epithermal neutron beam at the HFR Petten," 12th International Congress on Neutron Capture Therapy Kagawa International Conference Hall, Sunport Takamatsu, Japan, Oct 9-13 (2006).

G.G.Daquino, G.Gambarini, S.Nievaart, M.Carrara, E.Vanossi, "Use of Gel dosimetry to characterize the dose distribution in the spheroidal holder for liver treatment at the HFR Petten," 12th International Congress on Neutron Capture Therapy Kagawa International Conference Hall, Sunport Takamatsu, Japan, Oct 9-13 (2006).

P. van Lent, A. Sloetjes, A. Blom, G. Krijger, G. Koning, S. Nievaart, W. van den Berg and R. Moss, "Selective elimination of synovial macrophages by boron neutron capture therapy prevents onset of murine experimental arthritis," Journal of Controlled Release, Volume 116, Issue 2, 28, pp e106-e107 (2006).

Peter van Lent, Gerard Krijger, Wouter Hofkens, Sander Nievaart, Annet Sloetjes, Gerben Koning, Ray Moss, Wim van den Berg, "Selective Elimination of Synovial Macrophages by Boron Neutron Capture Therapy prevents onset of Joint Inflammation and Cartilage Destruction during Experimental Arthritis," submitted to Annals of Rheumatic diseases (2007).

Andrea Wittig, Massimo Malago, Laurence Collette, René Huiskamp, Sandra Bührmann, Victor Nievaart, Gernot M. Kaiser, Karl-Heinz Jöckel, Kurt Werner Schmid, Uta Ortmann, Wolfgang A. Sauerwein, "Uptake of two 10B-compounds in liver metastases of colorectal adenocarcinoma for extracorporeal irradiation with BNCT (EORTC trial 11001)," in preparation.

M.M. Fretz, G.C. Krijger, U.D. Woroniecka, V.A. Nievaart, W. Jiskoot, R. Moss, G.A. Koning and G. Storm, "Targeted liposomes for boron neutron capture therapy (BNCT): importance of target receptor density," in preparation.

A. Roca, S. Nievaart, R.L. Moss, F. Stecher-Rasmussen, "Validating a MCNPX model of Mg(Ar) and TE(TE) ionisation chambers exposed to 60Co gamma-rays," in preparation.

## Chapter 1

## Introduction

#### 1. General introduction

Boron Neutron Capture Therapy (BNCT) is a form of radiotherapy using neutrons for treating various types of cancer and some other non-malignant diseases. The basics of BNCT will be discussed in the next section followed by two sections describing theory that is important for the three successive chapters. To put the research presented in this thesis in context, a brief description of the history and present status of BNCT is given. This thesis is based on three articles in the field of BNCT published in journals covering the combined field of physics and medicine. Chapters 2, 3 and 4 are written as extended summaries of the articles with some novelties added. The full articles are printed in part II of this thesis. The motives for performing this research are explained in Section 1.2 which ends with the description of the scope of this thesis.

#### 1.1 The basics of BNCT

The basic concept of BNCT is that cancer or other 'bad' cells are loaded with the isotope boron-10 (<sup>10</sup>B) after which the site containing these cells is irradiated with neutrons. <sup>10</sup>B and the neutrons are non-toxic. After <sup>10</sup>B has captured a neutron, a nuclear reaction takes place and releases two heavy particles, being an alpha particle (<sup>4</sup>He) and lithium ion (<sup>7</sup>Li). This so-called (n, $\alpha$ ) absorption reaction is illustrated in Figure 1.1. According to the energies, the alphas and <sup>7</sup>Li nuclei can be regarded as short ranged particles since they travel less than 10  $\mu$ m in tissue. This range is similar to the size of a human cell and implies that the heavy particles have a high probability to kill or damage the cancer cell.



The cell is killed when the alpha or Li particle causes a double-strand break of the DNA. This occurs when a heavy particle travels through the cell nucleus. The probability of this event together with the probability of having a neutron reacting with a  $^{10}$ B in the first place requires that for successful BNCT the cell is loaded with around 10<sup>9</sup> of  $^{10}$ B atoms [1,2].

In the field of BNCT, often the energy spectrum of the neutrons is classified in 3 parts: Thermal neutrons below 0.5 eV, epithermal neutrons between 0.5 eV and 10 keV and fast neutrons above 10 keV and below 20 MeV. It is for thermal neutrons,

indicated with  $n_{th}$  in Figure 1.1, that the probability to react with <sup>10</sup>B is high. For thermal neutrons, this probability, known as the microscopic nuclear absorption cross section<sup>1</sup> ( $\sigma_a$ ) of <sup>10</sup>B is proportional with 1/ $\nu$ , where  $\nu$  is the velocity of the incoming neutron. For example, the absorption cross section of <sup>10</sup>B for 0.025 eV neutrons is 3837 barn and only 6 barn for 10 keV neutrons. However, the neutrons slow down due to interactions with tissue. This means that the starting neutron energy, coming from the source, has to be epithermal or fast in order to become thermal in a deep seated tumour after slowing down.

BNCT is a disease targeted therapy as the neutrons will only kill the cells which are labelled with <sup>10</sup>B. Unfortunately, with the presently available <sup>10</sup>B administrating compounds, also healthy cells will contain some <sup>10</sup>B. Besides this, human tissue contains certain isotopes that react with neutrons as well. These reactions result in a dose given to the healthy tissue which should not exceed a certain limit, called the tolerance dose. These extra dose components will be further explained in the two following sections.

Although the concept of BNCT might look quite simple and despite the fact that it is seven decades after its first proposal, BNCT is still under investigation. So far, as a maximum achievement, only phase I/II<sup>2</sup> clinical trials are performed with only a relatively low number (a few hundred) of patients involved. After all these years, researchers of various disciplines are still challenged by the two key issues of BNCT: Finding a non-toxic <sup>10</sup>B administrating compound, which brings the isotope into the 'bad' cells only or at least significantly more than in the healthy tissue and designing a treatment beam which is developed to deliver the optimal neutrons at the right location while minimizing the dose to healthy tissue. This latter issue about finding the BNCT source neutrons with the optimal energy and direction is studied in this thesis. The research took place at the BNCT facility of the Institute for Energy, Joint Research Centre (JRC) of the European Commission in Petten, The Netherlands. The centre's main target is to lead and participate in so-called scientific networks. For the optimisation of the neutronics as performed in this work, the JRC worked together with the section Physics of Nuclear Reactors (PNR) of the Delft University of Technology in The Netherlands.

#### 1.1.1 The four major BNCT dose components in tissue

Human tissue consists mainly of hydrogen, oxygen, carbon and nitrogen. By far, most reactions of neutrons are with H and N. Together with the presence of boron in

<sup>&</sup>lt;sup>1</sup> The microscopic cross section is a measure for the probability of a nuclear reaction for the nucleus and is expressed in *barn* which is  $10^{-24}$  cm<sup>2</sup> [3].

<sup>&</sup>lt;sup>2</sup> Most often clinical studies consist of four separate stages. They start from studying the effects of the treatment on healthy tissues (phase I), after which the focus is shifting towards treating the disease (phase II) and come to a scheme to treat the disease optimally (phase III) and ends with registering the treatment (phase IV).

the tissue, the majority of the total physical dose<sup>3</sup> considered in BNCT is delivered by H, N and <sup>10</sup>B. Because H has two types of reactions giving a physical dose, the total physical dose in BNCT consists of four components. These two H-related physical doses are described in the last two rows of Table 1.1. Rows 1 and 2 of this table describe the physical doses due to the isotopes <sup>10</sup>B and <sup>14</sup>N, respectively. In the remaining text the 'physical dose' is often shortened by writing 'dose'. As indicated in Table 1.1 in light and dark grey, the first three dose components are related to 'thermal' neutron reactions while the fourth dose component is due to reactions with 'fast' neutrons.

Physical dose		Dose* symbol	Reaction type	Scheme	Remarks
		~~	· <b>J</b> P ·		
la	Boron dose	$D_B$	n,α		Biological effects in tumour and normal tissue are related to <sup>10</sup> B micro-distribution
Therma	Thermal neutron dose	$D_p$	n,p		Induced proton 620 keV
	Induced gamma-ray dose	$D_{\gamma}$	n,y	H C	Induced γ-rays 2.2 MeV
Fast	Fast neutron dose	$D_n$	n,n		Energy of recoiling proton is on average half the neutron energy

Table 1.1. Overview of the four major dose components in BNCT.

\* According to IAEA Techdoc 1223 [4].

In a block-shaped model of H<sub>2</sub>O, see Figure 1.2, in which realistic amounts of N and <sup>10</sup>B are added, is the behaviour of the four dose components as a function of source neutron energy and depth in the model illustrated (see Figures 1.3 to 1.7). Light water with a few mass percent of nitrogen makes a good material to simulate average human tissue. As drawn in Figure 1.2, a spherical tumour ( $\emptyset$  4 cm) is positioned at 4 cm depth and contains 30 ppm of <sup>10</sup>B which is uniformly distributed. The assumption of having a three times higher <sup>10</sup>B-concentration in the tumour than in the healthy surroundings is a realistic 'average' [5,6]. The *D<sub>B</sub>* (see definition Table 1.1) as a function of the tumour is at all energies clearly visible because of

<sup>&</sup>lt;sup>3</sup> Physical dose is defined as the specific energy deposited around a certain point in a medium due to ionising radiation. It is written in unit Gray with symbol Gy. 1 Gy = 1 J/kg.



the three times higher concentration of <sup>10</sup>B and consequently more  $(n,\alpha)$ -reactions. This results in the horizontal darker grey band at the full width of the figure. For source neutron energies above 1 eV and below 3 keV, the  $D_B$  is significantly higher in the first 2 cm of the tumour. Between 10 keV and 100 keV the boron dose is more uniformly distributed in the whole tumour and is still in high contrast to the healthy surroundings. This is desired. Figures 1.4 and 1.5 show the  $D_p$  and  $D_y$  respectively, which look quite similar due to the fact that they are related to thermal neutron reactions as well. At low source neutron energies there are many thermal reactions at shallow depths. For increasing source neutron energy the majority of the thermal reactions are



Figure 1.3. Physical boron dose per source neutron as a function of source neutron energy and depth in the light water model with tumour at the centreline.



neutron.



Figure 1.5. As in Figure 1.3 but for the physical gamma dose per source neutron (actually kerma).



Figure 1.6. As in Figure 1.3 but for the physical fast neutron dose per source neutron (actually kerma).

'smeared out' stretching deeper into the phantom. The stripes and 'folding' behaviour around 1 MeV is due to resonances in the oxygen neutron cross section data. The typical 'thermal-reaction' look is also visible in Figure 1.3 but affected by the presence of the tumour. Figure 1.6 shows the  $D_n$  which is completely different from the described thermal dose figures. According to this figure, only for source neutrons above 30 keV, there is a physical fast neutron dose which increases rapidly with neutron energy. It is not visible in Figure 1.6 that  $D_n$  is already significant for source neutrons between 1 keV and 30 keV. This dose contribution can not be neglected, compared with the thermal dose components but is only significant superficially at the first 5 to 10 mm of the phantom.

The  $D_B$  and  $D_p$  presented here are determined by calculating with the Monte Carlo  $(MC)^4$  code MCNP4C2 from Los Alamos Laboratories [7], the alpha and proton productions, due to the <sup>10</sup>B and N respectively. Since these alphas and protons deposit their energy locally, the physical doses result from multiplying the calculated particle production densities with the released energy. The  $D_y$  and  $D_n$  are

<sup>&</sup>lt;sup>4</sup> Monte Carlo method: Weights of simulated particles are followed when travelling through the geometry. These weights can change due to interactions with the materials. These interactions occur probabilistically and are based on the nuclear cross section data of the materials.

calculated by multiplying the photon and neutron fluxes in MCNP with energy dependent tables of kerma<sup>5</sup> factors for water taken from ICRU46 [8]. Therefore, the  $D_{\gamma}$  and  $D_n$  are kermas rather then physical doses. Nevertheless, kerma and physical dose are equal in case of a charged particle equilibrium<sup>6</sup> which is supposed here. Furthermore, a background dose exists due to gamma rays in the BNCT neutron beam. Figure 1.7 depicts the  $D_{\gamma}$  from these 'beam' gammas which will be further indicated as  $D_{b\gamma}$ . For mono-directional source gammas having 24 discrete energies, chosen at logarithmically equal intervals between 1 keV and 20 MeV, the plot indicates a very slowly decreasing dose as a function of depth that is hardly visible. Note that the scale is logarithmic. A typical ratio of source neutrons to source gammas is 20 which is the case at the BNCT facility in Petten.



Figure 1.7. Physical gamma dose per source gamma due to (unwanted) gammas already present in the beam as a function of source gamma energy and depth in the light water model.

### 1.1.2 Biologically weighted doses and <sup>10</sup>B compounds

The secondary particles in BNCT, i.e. the alpha particles, protons, recoiling protons and electrons accompanying the  $D_B$ ,  $D_p$ ,  $D_n$  and  $D_\gamma$  as presented in Table 1.1, deposit their energy differently in the tissue. For example, the energy deposition along the short track of an alpha particle is very dense in comparison with that of an electron of which the track is longer. As a result, the human cells respond biologically differently when irradiated with 1 MeV alpha particles or 1 MeV gammas. This makes that the sum of the different particle energy depositions per unit of mass, the physical doses, has no biological meaning.

Many BNCT investigations aim to establish that the biological effects of  $D_B$ ,  $D_p$  and  $D_n$  can be translated into gamma dose equivalents. In this way the four dose components in BNCT can be added and the (total) dose given to the patient at each point in tissue can be described. The reason to translate the doses into gamma dose

<sup>&</sup>lt;sup>5</sup> Kerma is defined as the Kinetic Energy Released per unit Mass and consists of the energy that is transferred after the first collision.

<sup>&</sup>lt;sup>6</sup> There is charged particle equilibrium when for every charged particle leaving a certain volume in an irradiated medium, another charged particle of the same type, having the same energy and direction, enters the volume.

equivalents comes from the fact that a lot of experience in conventional radiotherapy (using mainly gammas) is gathered in the last century. The translation is performed by multiplying each physical dose component by a biologically weighted factor also known as the Relative Biological Effectiveness (RBE) factor. After addition, the total dose is called the 'total biologically weighted dose'<sup>7</sup> having the symbol  $D_w$  [4]. Currently, however, after many years of research, the values of these factors are still under discussion. As will be seen in Chapter 2, the RBE factor to 'translate' the  $D_B$ is replaced by a Compound related Biologically Effectiveness (CBE) factor. This factor embodies the 'normal' boron dose related RBE but is corrected for the applied boron compound [5,9]. Up to now, only two boron compounds are approved to be given to patients in clinical BNCT trials. The first drug is Borocaptate Sodium

(BSH) and the second Boronophenylalanine (BPA) of which the chemical structures are shown in Figure 1.8 [10-13]. Due to their respective natures, the distribution of the <sup>10</sup>B over the cells and the positioning in the cells, with respect to the cell nucleus, are different. This is taken into account by the CBE factor.



Figure 1.8. Chemical structures of BSH and BPA

#### 1.1.3 Brief history of BNCT

(This section is partly a brief outline taken from section 1.3.2 by Philipp [14]). The existence of neutrons was proven by Chadwick [15] twelve years after Rutherford had already postulated the existence of these particles in 1920 [16]. In 1936, after <sup>7</sup>Li and alpha particles were detected when <sup>10</sup>B reacts with thermal neutrons [17], Locher [18] suggested to apply this phenomenon in radiation therapy. In the early 1940s, Kruger and Zahl et al. [19-22] made some first promising radiobiological experiments in cell-cultures and mice, proving that BNCT worked as suggested. In 1941, it was Zahl et al. [22] who first proposed the use of epithermal neutrons instead of thermal neutrons, whenever humans would be treated in clinical trials. This remark of Zahl et al. about the usage of thermal and/or epithermal neutrons will be shown to play a very important role throughout this thesis.

It was not until 1950s that the first clinical trial on BNCT was started in the United States using thermal source neutrons [23]. At Brookhaven National Laboratory (BNL) and the Massachusetts Institute of Technology (MIT), 10 patients suffering from glioblastoma multiforme (a malignant type of brain tumour), were irradiated after a <sup>10</sup>B-enriched borax solution was intravenously given. The overall result, including a further 18 patients treated at MIT in a second protocol in which the skin, cranium and dura were removed, was unsatisfactory and consequently BNCT was

<sup>&</sup>lt;sup>7</sup> In Paper I the biological weighted dose, as defined for BNCT according to IAEA-TECDOC-1223, is assumed to be similar to the equivalent dose throughout the whole manuscript.

halted in the US in 1961. It was concluded that thermal neutrons do not penetrate deep enough and the boron compounds used were not very tumour selective [24]. From 1968, in Japan, clinical trials continued with the open-craniotomy procedure, where several intracerebral malignancies were treated. The Japanese professor Hatanaka who had been already involved in a clinical trial in the US, can be regarded as the catalyst of BNCT at that time. Together with other researchers, the Japanese investigated new boron compounds and treated more than 120 patients of whom some survived for a long term [25,26]. This outcome encouraged BNL and MIT to start new BNCT trials in the US in 1994 and 1996, respectively. From then onwards the focus was merely on the use of epithermal neutron beams which are able to penetrate skin and cranium and make removal of these unnecessary. In 1997, a European trial at the High Flux Reactor (HFR) in Petten (the Netherlands) started under the medical supervision of the university hospitals in Amsterdam (NL) and Essen (D), involving glioma patients [27-29]. In 1999 and 2000 respectively, similar clinical trials were started in Finland and Sweden [30,31]. All the trials mentioned so far were performed at nuclear reactors of which the 'medical' reactor at BNL in the US closed in 2000 and the Studsvik reactor in Sweden halted in 2005. In Italy (Pavia), in 2001, a very promising BNCT experiment was initiated by irradiating an explanted liver suffering from (inoperable) diffuse metastases. In 2003, a phase I/II clinical trial on skin melanoma started at a reactor in Argentina. In the same year, in Petten, a new protocol was approved and a phase I/II clinical trial started for patients suffering from melanoma metastases in the brain. This trial is performed in cooperation with the MIT. Reactor based clinical trials are also under investigation in Czech Republic, South-Korea, Taiwan and Russia. Accelerator-based BNCT is being investigated in Birmingham (UK), Italy, Argentina, Russia and in the US.

#### **1.1.4 Present status of BNCT**

The overall opinion of BNCT researchers and its critics is that *the* major improvement in BNCT is to be expected from new boron compounds that bring more <sup>10</sup>B in the tumour cells. In this respect, the application of liposomes is studied [32,33] which can be seen as bags ( $\sigma$  50-200nm) carrying a medicine. The liposomes can be programmed to connect only to 'bad' cells after which the medicine is transferred. The Petten BNCT group, in cooperation with the Delft University of Technology and the Universities of Utrecht and Nijmegen, are investigating the treatment of ovarian-carcinoma [34,35] and rheumatoid arthritis (after earlier studies [36,37]) using liposomes filled with BPA [38]. Other ongoing medical and biological studies concern the possibility of mixing the presently registered boron-carriers BSH and BPA [39,40] and the visualisation of the <sup>10</sup>B-uptake in cells by special microscopy (e.g. EELS [41]).

After many investigations, a major challenge in BNCT is still the translation of the several dose components into biological equivalents as discussed in section 1.1.2. It is impossible to compare the results among BNCT centres because different weighting factors have been applied or a different concept of dose-description is followed. In this respect, Riley and Binns [42,43] have started inter-comparison measurements using their own detectors and techniques at many BNCT facilities in order to standardize the used beams and enable the comparison of the results of the

treated patients. Promising and related to this issue is the development of a typical BNCT beam dependent radiobiological number which characterises the response of the cells when irradiating with this beam [44,45].

To predict the BNCT dose components as well as the neutron and gamma fluxes in patient treatment planning and experiments, MC based computer simulations are performed. The advantage of MC is the 3D-modelling capability and often the obtained precision in the results when compared with measurements whilst the disadvantage is the rather long calculation time. It is for the first reason (precision) that also physicists in conventional radiotherapy are interested and 'variance reduction' investigations to reduce the calculation time are in progress to overcome the MC disadvantage.

The last interesting trend in BNCT to be reported here concerns the disease targeted nature of BNCT which is most valuable when dealing with metastasised diseases. As already mentioned in Section 1.1.3, the treatment of an explanted liver full of metastases resulted in the survival of one of the patients for almost 4 years. This result encouraged many BNCT groups and also the group of Petten/Essen and Delft, to investigate the feasibility of such a project at the HFR in Petten.

#### **1.2 Spectral tailoring for BNCT**

As written in section 1.1.3, Zahl et al. was the first to propose the use of epithermal instead of thermal source neutrons. The use of epithermal source neutrons was extensively studied by Fairchild [46,47] at Brookhaven. Mainly epithermal neutrons in the range of 1 eV to 20 keV were shown to be useful for treating deep-seated brain tumours through the intact skull. The many succeeding publications discuss the characteristics of newly designed epithermal beams at the BNCT research centres [48-53]. The focus in these articles is on the application of filter and moderator materials, the shape of the beam assemblies and the quality of the resulting beam. Furthermore, these publications have in common that an existing neutron source is filtered and moderated such that it delivers neutrons in the energy range recommended by Fairchild. Actually, the spectral tailoring for BNCT consists of two parts:

- 1. Defining the source neutron energies of the BNCT treatment beam in order to obtain the most <sup>10</sup>B absorption reactions in the tumour. In addition, the location, direction and dimensions of the BNCT treatment beam need to be optimised for every individual tumour size and location.
- 2. Developing and constructing the filter with the appropriate materials and obtain the from 1. resulting energies starting with an available source.

The first part became the main issue of this PhD-research as will be further explained in Section 1.2.2.

As an example, the neutron filter in Petten has been designed and installed to let pass through only epithermal neutrons, whilst at the same time it has to reduce the unwanted photons, coming from the HFR reactor core. The design and used filter materials together with its main 'treatment' characteristics are given in Figure 1.9.



Figure 1.9. The current neutron filter for BNCT at the High Flux Reactor in Petten.

#### 1.2.1 Search for the optimal source neutron energy for BNCT

To the knowledge of the author, there are four publications (Yanch et al. [54,55], Bisceglie et al. [56] and Bleuel et al. [57]) that deal with the tailoring of the source neutrons as described in the last section. These more fundamental investigations are mainly initiated by the development of accelerator-based neutron sources which have the ability to obtain narrow neutron energy spectra that can be varied. The publications have in common that mono-energetic and mono-directional neutrons are simulated using MC and describe the doses and fluxes realised in a phantom. Succinctly, in all these investigations the focus is on deep-seated brain tumours whilst the doses are calculated with a fixed set of CBE/RBE factors. In all these investigations, the simulated neutrons and gammas start from the source after which the resulting effects are calculated in the tumour and healthy tissues.

The publications of Yanch et al. describe the dosimetric properties as a function of discrete neutron energies, beam size, collimation and different phantom shapes. It is reported that only the geometric differences show an influence on the contribution of the individual dose components to the tumour dose. The optimal source neutron energies to treat a tumour at 7 cm depth in tissue are in the range of 4.0 eV to 40.0 keV. Bisceglie et al. conclude that an optimal source neutron energy for BNCT is in the order of a few keV's for deep-seated tumours at 5 cm. Bleuel et al. conclude that neutrons between 2 keV to 20 keV are the most desirable in BNCT.

#### 1.2.2 The scope of this thesis

The Petten BNCT group together with the PNR department in Delft initiated a study to optimise the neutronics component of BNCT. In this respect, an investigation started to search for the optimal source neutron energies and continue the work already performed and described in Section 1.2.1. When taking into account some related challenges (see Section 1.1.4) the BNCT-physicists and medical physicists in conventional radiotherapy are currently facing, this thesis deals with the following issues:

- The influence of the dose weighting factors, and other dose related issues such as the tolerance doses in the tissues and the <sup>10</sup>B concentrations, on the optimal source neutron energy in BNCT. This issue is discussed in Chapter 2 that is based on *Paper I*.
- The development and application of adjoint MC calculation techniques. In adjoint MC the simulated particles travel 'backwards' which means from the tumour to the source which is the exit of the treatment beam. In realistic scenarios, this approach should be able to provide much faster the information on the optimal origin, direction and energy of the source particles compared with 'normal' MC. Chapter 3 and *Paper II* discuss the outcomes in all details.
- The analysis of having the optimal source neutrons not only for deep seated tumours, but also for mid-range and shallow positioned tumours. This is relevant when applying BNCT to cancer metastases which are spread throughout the brain or other organ, such as the liver. Chapter 4 describes a set-up to obtain a homogeneous (to a certain degree) thermal neutron field in a volume that contains a liver, given an epithermal neutron beam. This subject is published in *Paper III*. Chapter 4 describes also a method to obtain with adjoint-like techniques the optimal shape of the liver container and source neutron energy spectrum whenever this spectrum is free to choose.

From the knowledge obtained by studying these issues we may conclude that a 'variable' or set of different neutron filter(s) in BNCT would give, in every specific case, an optimal treatment from the neutrons point of view. This is essential knowledge, for designing a 'new' filter at the HFR in Petten or elsewhere.

It is often stated that the future of BNCT as a serious treatment relies on the availability of new boron carriers. However, this thesis shows that a better understanding of the BNCT *neutronics* gives a considerable improvement as well.

## Chapter 2

### **Optimal neutrons and dosimetry**

#### 2. Optimal source neutrons with regard to dosimetry in BNCT

In this chapter, the optimal source neutron energy in BNCT for brain tumours is investigated as a function of several biological and physical parameters. When investigating the optimal source neutron energy for BNCT one is looking for those source neutrons which cause maximum damage to the cancer cells and ideally none, but practically only a tolerable damage to the healthy cells. For the case concerning irradiation of the human head, different types of cells (tissues) with different tolerance doses are involved. Therefore, in principle, the best source neutrons are not simply the neutrons producing the highest <sup>10</sup>B absorption rate in the tumour before reaching the tolerance dose in one of the healthy tissues. This requires insight into the 'dosimetry' that is based on results coming from radiobiology.

#### 2.1 Background on BNCT dosimetry and radiobiology

The values of the tolerance doses that are currently being used in conventional radiotherapy have been mainly determined empirically by observing the levels of early and late side effects which develop in patients who underwent radiotherapy. The vast amount of data that has been collected during the time since the introduction of radiotherapy stems from treating patients with different fractionation schemes and doses of gamma rays and megavoltage X-rays [58]. The value of the tolerance dose strongly depends on the number of fractions into which the total dose delivered to the patient was divided. Mathematical methods have been developed that allow adjusting the value of the tolerance dose depending on the number of fractions [59]. Another factor which has a profound impact on the level of the tolerance dose is the quality of radiation used for treating the patient. It is well known that the biological effect per unit dose is higher for high LET<sup>8</sup> radiation as compared to low LET radiation [60]. This is due to the differences in the density of ionisation events inside a cell. While the cellular DNA repair mechanisms can cope with DNA damage that is evenly distributed inside a cell nucleus, multiple damaged sites produced by high LET radiation pose a more serious problem. Hence, at the same level of dose, high LET radiation is more effective in killing cells than low LET radiation. The need for comparing the doses of radiations of different qualities that induced the same level of biological damage triggered the introduction of the relative biological effectiveness – RBE (defined in Chapter 1). RBE values can be determined experimentally in *in vivo* and *in vitro* experiments.

<sup>&</sup>lt;sup>8</sup> In radiobiology LET stands for Linear Energy Transfer and is defined as the energy lost by charged particles due to interactions per unit of distance.

The problem with the RBE values is that their level depends on the cell system used for the experiment and the analysed endpoint [61]. Hence, when a new radiotherapy modality is developed, during which high LET radiation will be applied, it is not possible to simply recalculate the tolerance doses for irradiated organs on the basis of RBE values that were determined under laboratory conditions. Hence, no validated methods exist with the help of which the tolerance doses for high LET radiation could be calculated based on the clinical experience with low LET radiation. For ethical reasons it is not possible to determine the tolerance dose by exposing patients to various doses of high LET radiation as was done with X-rays and photons during the early days of radiotherapy [58]. This problem is especially pertinent to such complex radiotherapy modalities as BNCT, where the organs at risk are exposed to a mixed beam of both high and low LET radiations. Despite numerous radiobiological investigations it is not clear whether the effects of both radiation qualities are additive or synergistic [62,63]. In order to circumvent the radiobiological and medical issues described above, the optimal source neutron energy can be determined for a range of RBE values. This approach, extended by varying other parameters, such as the <sup>10</sup>B concentration, is applied in this chapter. Such an approach is at least a strong indication, as to whether there is a significant influence of the biological and clinical values on the optimal source neutron energy in BNCT and if so, what are these influencing parameters.

#### 2.2 A parameter study for BNCT of the brain

The biological and physical parameters that are varied enclose the *RBE* factors, the <sup>10</sup>B concentration, the intensity of the gammas present in the beam and the thickness of skin and cranium. Another parameter that is chosen to vary is the tolerance dose set in skin and brain. All parameter value ranges are discussed in the next section. After setting ranges for these parameters, for every configuration, the optimal source neutron energy is calculated. The optimal source neutron energy allows most of the neutrons to react with <sup>10</sup>B present at certain tumour positions under the constraint of not exceeding a pre-set dose limit in healthy tissue.

#### 2.2.1 Set-up and chosen parameter ranges

A cubic phantom is irradiated with neutrons from a 120 mm diameter disc shaped source with 22 discrete neutron energies, logarithmically chosen between 0.1 eV and 1 MeV. The calculations are carried out with the Monte Carlo code MCNP4C2 [7]. The MCNP geometry is shown in Figure 2.1. The neutrons are mono-directional. They first hit a layer of skin, then a layer of cranium before reaching the brain in which the tumours are located at four separate depths. All tissue compositions and densities are as defined in the ICRU46 report [8]. Small MCNP tallies (volume 78.5 mm<sup>3</sup>) are located along the beam centre line, at every millimetre, in order to calculate the dose as a function of depth in the phantom. As will be seen in the next section, the dose is also averaged over the whole volume (1.366 litres) of the brain which is represented by the drawn hemisphere in Figure 2.1.



Figure 2.1. Cross section of the MCNP geometry.

The biologically weighted dose<sup>9</sup>  $D_w$  in every tally *i* in the phantom is determined, as follows:

 $D_{w} = \left(C_{B} \cdot CBE_{B} \cdot D_{B}\right)_{i} + RBE_{p} \cdot \left(D_{p}\right)_{i} + \left(RBE_{n} \cdot D_{n}\right)_{i} + RBE_{\gamma} \cdot \left(D_{\gamma} + \lambda_{b\gamma} \cdot D_{b\gamma}\right)_{i}$ (2.1)

The  $CBE_B$  in this equation is the compound adjusted RBE as explained in Section 1.1.2 and the  $C_B$  is the concentration of <sup>10</sup>B. The *D*'s represent the absorbed doses for the thermal neutrons (p), the fast neutrons (n) and induced gammas  $(\gamma)$ . The absorbed dose<sup>10</sup> for the beam gammas  $(b\gamma)$  is given per source gamma and therefore has to be corrected with the term  $\lambda_{b\gamma}$  which is defined as the ratio of source gammas to source neutrons. The ranges of these parameters are shown in Table 2.1. The ranges are mainly based on the BPA [64] and BSH [65] related treatment protocols used in Petten and on a boron uptake study [66]. Since these protocols are based on current literature, it is of no surprise that these ranges practically include all values used in BNCT literature (see Nigg [67]).

During the MCNP calculation, 10 ppm of <sup>10</sup>B is assumed in all tissues. The influence and implications of the boron concentration in the tissues is discussed in section 3.1 of Paper I. After the MCNP calculations, a post-processing program calculates all the different configurations. When combining all parameters in categories III and IV in Table 2.1, there are almost 25 million possible configurations. By ignoring double occurrences, mostly zero values, and obviously deleting physically impossible combinations, the number of possible configurations could be reduced significantly and becomes 4,527,600.

<sup>&</sup>lt;sup>9</sup> Notation here is according to IAEA-TECDOC-1223 [4]. In Paper I the used indices are different, e.g.  $D_w = H$  (see Footnote 1),  $D_B = D_{10B}$ ,  $D_p = D_N$  and  $D_n = D_H$ .

<sup>&</sup>lt;sup>10</sup> The absorbed dose is also known as the physical dose.

Table 2.1. All varied parameters in this study categorized with their values, ranges and/or step sizes.

I. Thickness of skin and cranium								
Skin–Cranium [mm]: 3-3, 5-5 and 7-7								
II. Toleranc	II. Tolerance dose ratios							
$\frac{\text{skin(point}^{*})}{\text{brain(point}^{*})} = \frac{1}{5}, \frac{1}{3}, 1, 3 \text{ or } 5$								
$\frac{\text{skin(point}^*)}{\text{brain(volume)}} = \frac{1}{5}, \frac{1}{3}, 1, 3 \text{ or } 5$								
III. Relative biological effectiveness factors and boron concentrations								
	$C_B$ [ppm]	$CBE_B[-]$	$RBE_p[-]$	$RBE_{n}[-]^{**}$	$RBE_{\gamma}[-]$			
Skin	0-80 step 10	0-4 step 1		1-6 step 1				
Cranium	0-20 step 10	0-2 step 1	1-5 step 1	1-6 step 1	0.5-1 step 0.5			
Brain	0-30 step 10	0-2 step 1		1-6 step 1				
IV. Beam gammas								
Source gamma energy [MeV]: 1, 5 and 10								
Number of source gammas on every source neutron $(\lambda_{b\gamma})$ [-]: 0, 1/20 and 1/10								
*) Point is defined as tally volume of 78.5 $\text{mm}^3$								

\*) Point is defined as tally volume of 78.5 mm<sup>3</sup>

\*\*) The *RBE<sub>n</sub>* cannot exceed its previous tissue value [68].

#### 2.2.2 The influence of parameter biasing on the results

One of the major problems when interpreting the results of this study is the presence of some parameter values a reader regards as unrealistic. The problem is that the optimal source neutron energies resulting from the 'unwanted' parameter value(s) cannot be recognized. It is simply impossible to produce a graph that shows the outcome of every single configuration, as every parameter needs a dimension. A solution to this problem can only be obtained interactively: first record the outcomes of all 4,527,600 configurations into a data file and after the user has selected the parameter value ranges of interest, the results are collected, processed and presented. Furthermore, due to physical and other criteria by which the parameter values are selected, not every parameter value is equally represented in the total result. This can be regarded as biasing. For example (see Table 3 in Paper I) a  $RBE_n$  (= $RBE_H$ ) value in skin of 6 is present in 38% of all configurations. Again the outcomes of such a parameter value cannot be 'recognized' in the presented results. The interactive solution proposed above is necessary to accomplish this.

To re-cap, the main interest of this study is to investigate the role of each parameter and its value for all settings.

#### 2.2.3 Results: the optimal source neutron energy

The percentage of parameter configurations resulting in a certain optimal source neutron energy when treating tumours between 20 mm and 80 mm from the skin, under the constraint that the allowed tolerance dose in skin is three times higher than in the volume of the brain is represented in Figure 2.2. This tolerance dose ratio is comparable with the ratio as described in the EORTC protocol on 'metastatic malignant melanoma in the brain' [64] that prescribes not to exceed a biologically weighted dose of 22 Gy in a point in the skin and 7 Gy averaged in the brain. In Figure 2.2, the skin and cranium thicknesses are 3 mm (further referred to as the '3 mm phantom').



Figure 2.2. Percentage of parameter configurations resulting in a certain optimal source neutron energy as function of one tolerance dose ratio in the 3 mm phantom. The vertical axis displays the location of the tumour.

Note that Figure 2.2 is an interpolated contour plot over 22 x 4 grid points; 22 discrete energies on the horizontal axis times 4 depths of the tumour. It can be seen that for tumours at 20 mm depth, roughly 30% (yellow) of the configurations prescribe 1 keV source neutrons and another 40% result in 2 keV.

Although it is difficult to see in Figure 2.2, for the same depth, approximately 20% (orange) of the configurations result in 0.2 eV source neutrons. Overall, for all tumour depths the majority of the configurations prefer source neutron energies between 1 keV and 10 keV. For all ten tolerance dose ratios studied, for the 3 mm phantom, the percentage configurations with certain optimal source neutron energies are presented in Figure 2.3. The graph is an interpolation over  $22 \times 10 \times 4$  grid points (energies x tolerance ratios x tumour depths). The results for the 5 mm and 7 mm phantoms are shown in Figures 2.4 and 2.5, respectively. Some obvious trends noticeable in Figures 2.3 to 2.5 are mentioned here and are explained further in the next section.

- The differences between the Figures 2.3 to 2.5 show that there is an influence of the skin and cranium thicknesses on the optimal source neutron results.
- The tolerance dose ratios show an influence. For every phantom, whenever the tolerance dose in skin is below or equal to the tolerance dose in brain, the same 'image' is obtained. The other tolerance ratios can be further classified according to 'point' brain or 'volume' brain.
- Despite the skin and cranium thicknesses, the tumours at 40 mm and 60 mm depth result mostly in 2 keV and 4 keV source neutrons, whatever phantom, whilst the tumour at 80 mm 'needs' source neutrons with higher energies when the skin and cranium becomes thicker.

• For the tumour at 20 mm depth, the majority of the configurations prefer lower source neutron energies whenever the skin and cranium become thicker.



Figure 2.3. As in Figure 2.2 but for all the ten studied tolerance dose ratios (3 mm phantom).



Figure 2.4. As in Figure 2.3, but for the 5 mm phantom.


Figure 2.5. As in Figure 2.3, but for the 7 mm phantom.

#### 2.2.4 Results: the influencing parameters

In the following, the influencing parameters are discussed in the order as presented in Table 2.1.

#### Category I: Thickness of skin and cranium

The thickness of the skin and cranium has a significant influence on the optimal source neutron energy results for the shallowest and deepest seated tumours. A transition can be seen when comparing the Figures 2.3 to 2.5.

In the case of the shallow seated tumour (at 20 mm depth), with increasing skin and cranium thickness, the majority of the optimal source neutron energies 'moves' from 2 keV and some 100 eV towards 1 keV and finally to 500 eV for the 7 mm phantom. The explanation for this transition is given with the help of Figure 2.6 which shows the ratios of the thermal neutron flux in the tumour located at 20 mm to the maximum flux in skin for the three studied thicknesses. The lowest curve represents the ratio of the flux in the tumour to the maximum in brain. The higher the value of the ratios, the more thermal neutrons are in the tumour than in skin or brain. This is preferable. At first sight, source neutrons greater than 100 keV are optimal but, regarding the fast neutron dose due to recoiling protons as shown in Figure 1.6 in Chapter 1, appear unusable. Furthermore, the brain-result in Figure 2.6 is slightly curved between 0.1 eV and 10 keV with a maximum around 100 eV. The skin curves become steeper with decreasing skin thickness. As a result, it seems that the skin curves prescribe 'higher' source neutron energies whilst the brain curve prescribes 'lower' energies. Consequently, there is more profit using source neutrons with higher energies with decreasing skin thickness. For the deepest seated tumour



Figure 2.6. This graph belongs to the explanation given in the text for the relation between skin-cranium thicknesses and the optimal source neutron energy for shallow seated tumours.



Figure 2.7. This graph belongs to the explanation given in the text for the relation between skin-cranium thicknesses and the optimal source neutron energy for deep seated tumours.

at 80 mm, a transition towards higher optimal source neutron energies is observed from 5 keV via 10 keV towards 20 keV (see Figures 2.3 to 2.5). The basic mechanisms causing this dependence on skin thickness is explained with Figure 2.7. Figure 2.7 shows the ratios of the fast neutron flux to the thermal neutron flux for the three skin thicknesses at the border between skin and cranium. The lower this ratio the better and thus lower source neutron energies are preferable. The opposite effect is caused by that the number of thermal neutrons in the deepest seated tumour increases with increasing source neutron energy (see Figure 1.3 in Chapter 1). Although, the higher the energy of the source neutrons the better, the 3 mm skin curve is limited by a more quickly increasing fast neutron component in the skin than the other skin thicknesses. This explains the relation between the increasing energy of the optimal source neutrons and increasing skin thickness.

#### Category II: Tolerance dose ratios

An important outcome of this parameter study is that the optimal source neutron energy as a function of parameter variations shows clear dependence on the chosen tolerance dose in skin and brain (see Figures 2.3 to 2.5). Three groups, distinguished by having similar characteristics in the results, can be identified according to the tolerance ratios:

\*Group 1, in which the tolerance dose in a point in the skin is lower than or equal to the tolerance dose in a point in the brain and also over the total volume of the brain. For this group the skin turns out to be the treatment limiting tissue.

\*Group 2, which has a higher tolerance dose in a point in the skin than in a point in the brain. The brain turns out to be the treatment limiting tissue in the majority of the cases.

\*Group 3, is the same as Group 2, except that the tolerance dose in brain is set over the whole volume of the brain, then both the skin as well as the brain can be the treatment limiting tissue.

## Category III: Relative biological effectiveness factors and boron concentrations

According to Table 5 in Paper I, it is clear which parameters are of direct influence on the optimal source neutron energy; i.e.  $C_B$ ,  $CBE_B$  and the  $RBE_n$ . For Group 1, by far, all dose limits are reached in skin at the interface with cranium. Furthermore, in the case of Group 2, the  $C_B$  and  $CBE_B$  for brain tissue and the  $RBE_n$  for all tissues are the influencing parameters. In recording the location of where the tolerance dose is exceeded, most positions are at the thermal neutron fluence peak between 20 mm and 40 mm. Finally, in the case of Group 3, a mixture of  $C_B$  and  $CBE_B$  for both skin and brain tissue and the  $RBE_n$  for all tissues, are the influencing parameters.

#### Category IV: Beam gammas

For all phantom dimensions, there is an influence notable for tumours at 20 mm and 40 mm for Groups 2 and 3. For the higher values of the gamma related parameters, the optimal source energies tend to lower energies.

#### 2.2.5 Results: Improvements when using the optimal source neutrons

Following the above results concerning which source neutron energy ensures the maximum alpha production in the tumour, a logical follow-up question is: is it necessary to provide all these 22 source neutron energies? To investigate this, the number of alphas generated by the optimal neutrons is compared with the number of alphas as produced by each of the 22 source neutron energies as described in section 2.2.1. To clarify this, as an example, the alpha productions in the tumours obtained

with the optimal source neutron energies are compared with the alpha productions resulting from using only 10 keV source neutrons. This is done for all the 4,527,600 configurations. The improvements in alpha production when one could select in every configuration the optimal source neutron energy instead of only 10 keV source neutrons, is shown in Figure 2.8 for the 5 mm phantom. The percentage of alphas in the tumours that are produced *more* with the optimal source neutrons than with the 10 keV source neutrons, goes up to 400%. In other words: for a specific parameter setting (configuration) the optimal source neutron energy is 5 times better.



Figure 2.8. Maximum percentage improvement of the number of alphas produced with the optimal source neutron energy in comparison with the number of alphas produced with 10 keV source neutrons for one tolerance ratio (5 mm phantom).

All the percentages presented in Figure 2.8 are taken for the configurations where this improvement in alpha production is a maximum. As mentioned in Paper I, graphs like Figure 2.8 show a dependency with the tolerance ratio but then there is no significant difference concerning these graphs for the 3 mm and 7 mm phantoms. In Figure 2.8, for this particular '400% improvement' case, the use of 0.2 eV source neutrons for superficially located tumours becomes a reality when dealing with very low <sup>10</sup>B related parameter values in skin (see further Table 5 in Paper I). The same article predicts that besides using 10 keV source neutrons, the availability of source neutrons with energies in the order of tens of eV will cover all configurations and assure the best alpha production possible in BNCT. In how far this prediction is true was investigated after the publication of Paper I. The results are presented here.

All possible pair-configurations out of the above mentioned 22 source neutron energy results are investigated. For example, the outcomes of the maximum improvement when using the optimal source neutron energies instead of only 0.1 eV source neutrons is combined with the results of the 2 keV source neutrons. For every pair of source neutron energies the smallest improvement values are collected. It turns out that the availability of 500 eV and 10 keV source neutrons provide the smallest difference in alpha production compared with the optimal source neutrons. However, the 'maximum' improvement when these 500 eV and 10 keV source neutrons are available is still 50%, for a certain parameter setting. The outcome of an attempt to minimise the maximum improvement even further is shown Figure 2.9. When 3 beams with three different source neutron energies can be chosen out of the 22 studied source neutron energies, it turns out that the best set consists of 5 eV, 500 eV and again 10 keV. With this trio the maximum improvement of having all 22 source neutron energies available is only 28%.



Figure 2.9. With the availability of source neutrons of 5 eV, 500 eV and 10 keV only a maximum improvement of 28% in alpha production would be achieved when using the optimal source neutron energy.

# 2.3 Conclusions

The optimal source neutron energy delivers a maximum of alphas in the tumour, while not exceeding tolerance dose constraints. It turns out that the definition of these tissue dependent tolerance doses greatly influence the source energy making a neutron 'optimal'. The results presented in Figures 2.3 to 2.5 indicate that according to the ratio of the tolerance dose set in skin to brain three groups can be distinguished. The thickness of skin and cranium affects also the choice of which source neutrons are better to use. The parameters of influence are the <sup>10</sup>B concentration ( $C_B$ ), the boron related *CBE* and the *RBE* for fast neutrons. It turns out that the tissue in which the tolerance dose is reached first determines the parameters causing the major deviation in the source neutron energies. Future studies will involve the field size of the beam and the influence of the shape of the phantom. Also the possibility to process user-defined ranges of the parameter values on-line, by using an interactive program, is necessary to continue and control the discussed parameter biasing.

Another important conclusion from the alpha production improvement results (see Figure 2.9), is that the availability of 5 eV, 500 eV and 10 keV source neutrons would improve the treatment plans (in this study) enormously. In this studied brain case, the nuclear physicist and/or treatment planner in BNCT could deliver, no matter what the circumstances or parameter settings, most of the alphas at the tumour location.

# Chapter 3

# Adjoint techniques in BNCT

## 3. Application of adjoint Monte Carlo techniques in BNCT

In specific cases, adjoint Monte Carlo (MC) calculations can provide the optimum setting to the treatment planner giving shorter calculation times compared with 'normal' forward MC. However, the adjoint method in its standard form is not applicable when dealing with mono-directional treatment beams as in BNCT. In this chapter, techniques are developed to overcome this problem and the method is demonstrated in a specific example.

## 3.1 Background, theory and MCNP

In reactor physics, an equation can be defined that is adjoint to the neutron transport equation. With the proper adjoint source the solutions of this adjoint equation can be physically defined as a measure of the "importance" of a neutron in contributing to the response of the detector [69]. In the field of irradiating materials, e.g. phantoms and patients, it can simply provide at defined locations (points, areas, volumes) the expected detector contribution of source particles, as a function of energy, position and starting angle. These 'detectors' can be tumours or organs at risk (OAR), which together, will be further mentioned as regions of interest (ROI).

In MCNP [7] the adjoint equation is solved by tracing histories 'backward'; the normal MC method is reversed and the adjoint particle obtains after an event the energy and angle that a forward particle would normally have before this event. As stated in the manual, MCNP in forward (=normal or standard) mode is preferable when the detector volume or area is large. The adjoint mode is interesting when the detector volume or area is small but the source volume or area is large. Because the histories start in this small detector area and are traced 'backward', they will have a higher probability to contribute to the larger phase space region of the source<sup>11</sup>. In principle, regarding the statistics, the more particles contribute to the estimate under investigation, the smaller the variance becomes. A good description of adjoint MC is given in Wagner et al. [70] and Hoogenboom [71].

An important result, as derived in Bell and Glasstone [69], is the relation between the 'forward' flux and the adjoint function,  $\phi$  and  $\phi^+$  respectively. The total detector response, depending on the detector response function  $\Sigma_d$  is the integral at the right hand side of

$$\int Q(r,\Omega,E)\phi^+(r,\Omega,E)dVd\Omega dE = \int \Sigma_d(r,\Omega,E)\phi(r,\Omega,E)dVd\Omega dE$$
(3.1)

<sup>&</sup>lt;sup>11</sup> In adjoint MC the forward source becomes the adjoint detector and forward detector becomes the adjoint source

in which r is the position,  $\Omega$  the solid angle, E the energy and V the volume. This integral is taken over all phase-space variables and provides the same result as the integral of  $\phi^+$  multiplied with the source function Q. In other words: knowing the adjoint function and source function of a treatment beam enables to determine e.g. the total detector response of the flux, dose or reaction rate inside a certain area, as similarly obtained with a forward calculation.

As written in the manual, MCNP has an option called 'SCX' which enables to register at the detector the initial energy of the particle emitted by the source. Regardless of the mode in which MCNP is running (forward or adjoint), four possible outcomes can be gathered. These, sometimes confusing possibilities, are listed in Table 3.1.

Table 3.1. Overview MCNP modes

MCNP mode / Particles	Obtained at detector:
start at source:	
Forward	Total detector response
Forward+SCX	Expected contribution of forward source
	particles to the forward detector response
	(=numerically proportional to the adjoint
	function)
Adjoint	Adjoint function
Adjoint+SCX	Expected contribution of adjoint source
	particles to the adjoint detector response
	(=numerically proportional to the total detector
	response)

Because MCNP normalizes every source description automatically, this means in BNCT that the results of the adjoint calculation are normalized to the tumour (adjoint source) instead of the neutron beam (forward source). A proper normalization is necessary as one is interested in 'per-particle-results' related to the real source. Although the re-normalization of the adjoint results can be derived analytically, the 'SCX' option provides an additional check. The normalization of adjoint results is discussed in the report by Wagner et al. [70] and also Difilippo [72].

#### 3.2 General adjoint set-up for BNCT

In BNCT, to know the contribution of source neutrons to a reaction rate in the tumour, the adjoint source spectrum should be similar to the  $(n,\alpha)$  capture cross section of <sup>10</sup>B. Most of the organs at risk, as defined in, for example, the melanoma metastasis of the brain protocol, are in deeper lying structures which will only suffer from the dose given by the thermal reactions. In theory, a well defined adjoint source energy spectrum should be similar to the  $(n,\alpha)$ , (n,p) and  $(n,\gamma)$  cross sections of <sup>10</sup>B, <sup>14</sup>N and <sup>1</sup>H, respectively, depending on the reaction rate of interest. These cross sections have in common that the probability to have a reaction at thermal neutron energies is orders of magnitude higher than for high energies. In practice, in

BNCT, calculating the thermal neutron flux will provide a good first estimate of how the thermal neutron related doses in the phantom will behave (see Chapter 1). In the following, the adjoint source spectrum as well as the forward detector response function are taken to be thermal and defined uniform up to 0.5 eV.

Most of the fast neutron dose is delivered at the outside in the first few centimetres of the tissue. Because the interaction of the beam-particles always starts in the skin, the fast neutron dose component can be regarded as uniform and needs only to be estimated at one location. Note that when dealing with the fast neutrons in the adjoint approach, it is difficult or maybe even impossible, as the skin will form a large adjoint source.

For all calculations presented in this chapter, a new 172 group neutron cross section library is made, according to the XMAS energy structure. All cross section tables, for nuclides present in human tissue used in BNCT, are based on JEF2.2 evaluations for  $37^{\circ}$ C and with S( $\alpha$ , $\beta$ ) thermal treatment.

# 3.3 Adjoint treatment planning with a mono-directional beam

Nowadays, most investigations in treatment planning focus on decreasing the computational time to obtain the doses without making concessions on the accuracy [73-75]. When less time is needed for the dose calculations, simply more time is available for optimization of a treatment plan. There are mainly three algorithms used in treatment planning to calculate the dose which are the convolution/superposition method, the pencil beam method and the MC method. The first two mentioned methods are widely applied in treatment planning and can be characterized as being fast but less accurate compared with the MC method which is accurate but needs more calculation time. Acceleration of Monte Carlo calculations by variance reduction techniques is also of great interest and is investigated in a wide area of research. Especially in treatment planning, Monte Carlo provides more accurately calculated dose rates in the heterogeneous human tissues [76,77] than with the other methods. However, the current long calculation times prevent that the MC methods can be applied for every single beam configuration for every patient.

In comparison with the above mentioned investigations, far less articles have been published in the field of adjoint MC as used in treatment planning. The existing articles [72, 78-81] all deal with gammas and it is only in the work of Lilly [82] in which the adjoint method is related to BNCT. However, Lilly uses a discrete ordinate radiation transport code to optimize a neutron filter for BNCT.

Since BNCT is worldwide still in a clinical trial phase, relatively, not many patients are treated. As a consequence, little experience exists in positioning neutron beams and making treatment plans for various locations and positions of the cancer. In this chapter, we will show that the adjoint MC technique can be an improvement for BNCT and can help the treatment planner in selecting optimum beam settings.

In the published adjoint articles mentioned above, two approaches concerning the adjoint detectors can be distinguished:

1. The adjoint detectors are formed by segmentation of a sphere or cylinder surrounding the irradiated geometry. All adjoint particles are tallied in the angular and energy bins for each segment.

The adjoint detectors are shaped according to the beam exit of the treatment beam. Only adjoint particles resembling the characteristics of the treatment beam, e.g. energy and angle, are tallied in the appropriate bins.

If the interest is to know the delivered dose or reaction rate of a treatment beam to a certain ROI, the first approach would be ideal if the segments are very small and the bin structure very fine. Knowing this (almost) continuous adjoint function for the phase space coordinates around the geometry, the treatment beam could be positioned everywhere and directed freely. However, to obtain reasonable statistics the segments and binning have to be coarse, as such the response of a treatment beam cannot be obtained. Nevertheless, this is useful as a way to obtain a first impression of the directions that are interesting to irradiate from [78].

The second approach delivers, together with the appropriate source function Q (see equation 3.1), the same total detector response as a forward MC calculation. However, a major drawback of this approach is that only adjoint particles at the adjoint detector can contribute with flight directions and energies in the ranges of the treatment beam; e.g. a 2 degree angular divergence requires adjoint particles arriving at the detector within an angle of 2 degrees with the normal of the adjoint detector plane. The probability that adjoint MC particles arrive at the adjoint detector within certain (narrow) boundary conditions is small and it will take a long calculation time before good statistics are obtained. As a result, the adjoint MC approach seems inapplicable for performing BNCT with a mono-directional neutron beam. To overcome this, two techniques are developed which will be discussed in the next sections.

## 3.3.1 Adjoint Point Detector Technique (APDT)

Since the probability that particles will traverse a plane detector perpendicularly within a narrow solid angle is very small, one has to 'force' them. In MCNP this 'forcing' technique is called DXTRAN; at every event the contribution a particle will have to a certain specified region is calculated deterministically. DXTRAN can be used in forward and adjoint mode. This specified region can be defined as a point and by positioning this point far away from the geometry and surrounding it by a beam-shaped body, a situation as depicted in Figure 3.1 arises.



Figure 3.1. Adjoint Point Detector Technique (APDT).

For a disc shaped beam opening, the adjoint point detector is surrounded by a cylinder. As can be seen in Figure 3.1, all contributions to the adjoint point detector pass the entrance of the cylinder (almost) perpendicularly, while all others are "killed". This entrance behaves as a virtual disc shaped adjoint detector. The further

away the adjoint point detector is positioned the more perpendicular the contributions traverse the entrance. In order to prevent angular spreading in the example discussed in section 3.4, the distance from the centre of the irradiated phantom to the beam exit (=adjoint detector disc) is chosen to be  $10^4$  cm. This value depends on the size of the phantom and the size of the beam exit. The normalisation factor for this adjoint technique is provided in Appendix A. As there is a possibility for a score at the adjoint point detector for every event, the relative error decreases more quickly, as compared with the analogue MCNP. A disadvantage of the technique is when a lot of particle events are involved of which many will not contribute to the point detector because these contributions 'travel' through the region outside the beam-shaped body (see Figure 3.1) and a lot of CPU time is wasted. Many particle events are caused by e.g. highly scattering materials and large geometric dimensions.

# 3.3.2 Legendre EXpansion Technique (LEXT)

The second technique to determine the adjoint function for the direction perpendicular to the adjoint detector is by use of the Functional Expansion Technique (FET). Suppose that the angular adjoint function at a certain adjoint detector for a certain energy group looks like the 2D function  $f(\alpha,\beta)$  plotted in Figure 3.2. When this function behaves well, it can be expressed by

$$f(\alpha,\beta) = \sum_{l=0}^{\infty} \sum_{m=0}^{\infty} d_l d_m \psi_l(\alpha) \psi_m(\beta)$$
(3.2)

in which  $d_{l,m}$  are coefficients and  $\psi_{l,m}$  are orthonormal basis functions. It is in the work of Beers and Pine [83] that this so-called FET is applied in MC. They accomplished that all samples contribute to the estimates of the coefficients  $d_{l,m}$ . This can be proven, for example, for one dimension  $(d_l)$  by combining the definition



Figure 3.2. Example of an angular adjoint function in 2D at an adjoint detector for one energy group.

for the total detector response and the scalar product describing the coefficients (equation 8 and 9 in Paper II),

$$d_{l} = \int f(\alpha)\psi_{l}(\alpha)d\alpha \approx \frac{1}{N} \sum_{i=1}^{N} w_{i}\psi_{l}(\alpha)$$
(3.3)

in which N is the number of samples and w is the statistical weight. For the two dimensions  $(\alpha,\beta)$ , both accompanying coefficients  $d_{l,m}$  can be evaluated from N. The indices l and m have to be truncated by values of L and M such that the function f is still well approximated by the sums of the products of d and  $\psi$ . In any case, L and M should be high enough for a converged result but not too high because of the increasing relative error for every extra coefficient used. Note that there is a direct relation among the number of samples, number of used coefficients, the convergence of the result and its relative error. This relation should be investigated once for every geometric model and problem set-up (e.g. detector size, number of energy bins, material characteristics). This may be regarded as a disadvantage of the technique. In this work it is chosen that the orthonormal basis functions are the Legendre polynomials: Legendre EXpansion Technique (LEXT). At this stage it is not well known how angular adjoint functions in different circumstances (e.g. other geometries and particle types) will look like. Therefore, investigating other base functions which describe the adjoint functions better in particular cases is expected to be subject of future studies. The Legendre polynomials are orthogonal at (-1,1)and have the best pointwise convergence properties near the centre of this interval [84,85]. As a consequence, since the interest is in the normal direction,  $\alpha$  and  $\beta$  have to be chosen such that the normal falls in the centre of their ranges (see the arrow in Figure 3.2). The parameterisation in Figure 3.2 is as chosen and described in Paper II. Actually,  $\alpha$  and  $\beta$  are the angles which describe the directions of the adjoint particles traversing the adjoint detector and can be defined in many ways as long as the normal directions fall at the centres of the intervals. Note that the adjoint particle can cross the adjoint detector from one side only. In this parameterisation the ranges of  $[-\pi/2, \pi/2]$  and  $[0, \pi]$  will be linearly scaled into the Legendre range of [-1,1]. Functions f with steep gradients and/or zero values demand extra coefficients to describe these characteristics properly. The use of too many coefficients should be prevented and can be accomplished by truncating the angles (see Section II.F of Paper II).

Before the Legendre technique is applied, all adjoint particles will be recorded in a so-called PTRAC file written by MCNP, on a sphere surrounding the geometry. With the recorded position and flight direction information it can be determined whether an adjoint particle traversed a certain area. In a post processing program these areas can be mathematically described with a shape resembling the neutron beam exit (=adjoint detector). The number, positions and orientations of these adjoint detectors are free to choose in the post-processing program and no new MCNP calculation has to be done. As described in Paper II the weight of every adjoint particle traversing an adjoint detector is recalculated directly with the LEXT to provide its weight if it was flying in the normal direction. After this 'weight-

adaptation' the average and relative error are calculated like in a normal tally result of MCNP.

## 3.4 Example: Optimum characteristics in the irradiation of a human head

The APDT and LEXT are demonstrated by calculating the total detector responses of the fluxes due to thermal neutrons and gammas<sup>12</sup> in 10 randomly distributed tumours and 10 OAR (in total 20 adjoint sources) in a patient suffering from brain cancer. The MCNP geometry of the patient's head is shown in Figure 3.3, as created by the program Sabrina [86]. The OAR are defined as written in the protocol in Petten used to treat metastatic malignant melanoma in the brain [64]. CT-images of the head of a patient are translated, using Scan2MCNP [87], into a MCNP geometry consisting of 47520 voxels. Each voxel can be filled with either brain tissue, soft tissue, cranium or mixtures of these.



Figure 3.3. Phantom head with 10 OAR and 10 tumours which is used as an example throughout this section.

The model is surrounded by air. Figure 3.4a shows the head phantom surrounded by 60 centre points where all adjoint detector discs are positioned. These centre points are described in azimuthal and polar angles as shown for one adjoint detector disc in Figure 3.4b. At each of the 60 positions, 17 discs with different, systematically chosen orientations of their outer normals (pointing away from the phantom) are positioned. See Figure 3.4c.

<sup>&</sup>lt;sup>12</sup> In the remaining text shortened as 'thermal neutron flux' and 'gamma flux'



Figure 3.4. (a) MCNP head model surrounded by 60 adjoint detector disc centre points. (b) The adjoint detector centre points are described in polar and azimuthal angles. The initial disc normal points at the centre of the phantom. (c) At each of the 60 centre points, 17 adjoint detector discs with different orientations are defined.

This implies that in total 1020 beams around the head are modelled. These 1020 adjoint detector discs are also the virtual adjoint disc detectors in the APDT and the beam exits in the forward MC method. All calculations, for all 1020 beams, are performed for beam diameters of 5 cm, 10 cm and 15 cm. The source energy spectrum is taken from the BNCT treatment beam called HB11 as available in Petten. It consists of mainly 10 keV source neutrons and has a 2 degree divergence which can be regarded as mono-directional. The gamma source spectrum is taken uniform between the limits 0.01 MeV and 20.0 MeV.

In the next subsections the results will be discussed, starting with the calculation times of the forward MC and the two adjoint MC techniques. Thereafter, the results for this phantom head itself will be discussed by showing the optimum positions and orientations for the Petten neutron beam to irradiate from. The other feature coming out of the adjoint calculation, the optimum source neutron energies, is discussed last.

#### 3.4.1 Calculation times of forward MC, APDT and LEXT

The calculation times of the forward MC, APDT and LEXT can be compared after setting the allowed relative error in the 20 ROI for the three methods to <5% (95% confidence interval). These relative errors are averaged values from the results of a certain number of beams. To save time, 255 out of the 1020 beams around the head are chosen and their relative errors in the tumours and OAR are averaged. Since the relative error of the results in Monte Carlo is 'scaled' with the square root of the number of tracked particles, it can be determined how many particles need to be run and consequently how much time it takes to get a certain relative error. All averaged relative errors are, before scaling, up to 15% in the 95% confidence interval. In case of the LEXT, the post-processing time is linear proportional with the MCNP time to obtain the PTRAC file. The resulting times for the three methods are summarized in Table 3.2. In the column containing the forward MC results, the absolute times are given in days. The calculation times of the adjoint techniques are normalised towards these forward MC times and written in bold face. It is clear from Table 3.2 that in the case of the head phantom example, the smallest beams (5 cm) are

calculated most quickly with forward MC for both neutrons and gammas. In the case of neutrons, the 10 cm and 15 cm beams are 6 and 20 times faster calculated with the LEXT respectively. The APDT lacks from the fact that the adjoint thermal neutrons scatter a lot due to the hydrogen in the tissues, for which each time, a contribution to the adjoint point detector is determined. In many cases the contribution cannot be made for the reason described at the end of section 3.3.1.

5.2. The total calculation times for the three methods.						
Diameter of adjoint detector/	Time to calculate 1020 beams					
beam exit [cm]	Normalised to forward					
NEUTRONS	FORWARD	APDT	LEXT			
5	<b>1.00</b> (=140 days*)	27.40	1.52			
10	<b>1.00</b> (=113 days*)	18.42	0.15			
15	<b>1.00</b> (=88 days*)	12.94	0.05			
GAMMAS						
5	<b>1.00</b> (=53 days*)	1.57	1.47			
10	<b>1.00</b> (=42 days*)	0.52	0.55			
15	<b>1.00</b> (=27days*)	0.33	0.30			

Table 3.2. The total calculation times for the three methods.

\*) Results gathered on a Pentium IV with a 3 GHz processor and 512 Mb of memory.

For gammas, for the 10 cm and 15 cm beam exits, the calculation times for the APDT improve because gammas interact far less with the head phantom materials. The calculation times for these large beam exits are similar to the LEXT in case of 20 ROI. One can imagine that when less ROI and/or less gamma beams are involved, less APDT calculations are needed and the APDT becomes preferable to the LEXT. An overview of when forward MC calculations are faster than the two presented adjoint MC techniques, as a function of the number of ROI against the number of beams, is given in Figures 3.5 and 3.6, for neutrons and gammas respectively. Every figure contains 3 lines belonging to each beam exit diameter. Above the line of a certain beam diameter, the calculations are faster using forward MC, below the line by adjoint MC. Figures 3.5 and 3.6 are based on scaling the calculation times of the forward MC, APDT and LEXT head phantom results according to:

#### Forward $MC \propto$ number of beams

#### APDT $\propto$ number of beams and number of ROI

# LEXT MCNP $\propto$ number of ROI LEXT Post Proc. $\propto$ number of beams and number of ROI

The increase of MCNP calculation time due to the increasing complexity of the geometry when more ROI are involved is not taken into account. Due to this, in reality, the LEXT results will be somewhat better than presented because of the post processing time which does not depend on the complexity of the geometry.



Figure 3.5. Number of neutron beams with different diameters vs. number of tumours/OAR for which adjoint or forward MC is preferable. For a given beam diameter, the region below the line is the area where adjoint MC is preferable.



Figure 3.6. Like Figure 3.5 but for gamma beams with different diameters.

The exercise presented here can also be performed for a fixed beam diameter but as a function of ROI size/volume. The same kind of curves would be expected, as presented in Figures 3.5 and 3.6, although the largest beam diameters will represent the smallest ROI and vice versa.

#### 3.4.2 Optimum irradiation locations and directions

The following results are obtained with the LEXT for the Petten neutron beam with a diameter of 15 cm. The differences between the LEXT outcomes and 255 controlling forward calculations are within the statistical uncertainties. From this point, the thermal neutron fluxes used are chosen to be the average fluxes over all tumours and the average over all OAR. Due to this averaging, the statistics improve and therefore an optimum treatment plan can be obtained more quickly. It is subject for further research to see if such an approach can hold in comparison with a treatment plan which is optimised taking into account every single tumour and OAR. Figure 3.7 shows the optimum orientation out of the 17 discrete orientations the Petten beam can have at each of the 60 positions around the phantom head as shown in Figure 3.4.





This optimum is defined as to have the highest ratio of thermal neutron flux in the tumours to OAR. As a constraint, it is chosen that the thermal neutron flux in the tumours has to be >75% of the maximum attained thermal neutron flux in the tumours for that beam position. This constraint prevents the situation that the flux in the tumours is low but the flux in the OAR is close to zero which gives anyhow a high ratio. However, the time the patient needs to be irradiated becomes too long. The optimum ratios around the head are also given in the grey histogram in Figure 3.8; the 60 beam positions are ordered with the azimuthal angle. The polar angles

and orientation numbers information is omitted. For instance at beam position 13 in Figure 3.8, around 40 times more thermal neutrons are delivered in the tumours then in the OAR. It is clear that the best locations to irradiate the tumours are from the back of the head. Therefore, with the LEXT, 20 times faster than with normal forward MC, the whole head has been 'scanned' by 1020 beams and the treatment planner can obtain the best positions and orientations for irradiation. It is nearly impossible to have 'missed' a good beam configuration. This is therefore a good starting point to combine the best beams and optimise the treatment plan.

#### 3.4.3 Optimum source neutron energy group

The optimum source particle energy can also be obtained with the adjoint technique. This is simply the source particle energy giving the highest contribution of interest in the tumour and the lowest in the OAR. In terms of adjoint MC, this means that only a few (or ideally none) adjoint source particles originating from the OAR will reach the adjoint detector. This results in a large accompanying relative error. Of course one has to be confident that enough histories have been run in order to assure that the result has converged. It is expected that obtaining acceptable relative errors within various bins at several adjoint detectors around the phantom when using an isotropic adjoint source, is a good indicator that the MC run has converged. This problem of having large errors in some ROI will not occur rapidly in the results with the Petten beam, as many energy groups are involved by which the statistical uncertainty decreases after integration. In the case of single source neutron energy groups, it is for now impossible to draw conclusions on the absolute ratio values of the thermal neutron flux in the tumours to OAR. It is expected that MCNP variance reduction techniques are needed to overcome this problem of the large relative errors in the OAR. This is subject for further investigations. Another approach would be the use of deterministic codes.

For our phantom head example, irradiated with a 15 cm diameter neutron beam, the maximum ratios of the thermal neutron flux in the tumours to OAR are displayed in Figure 3.9. The grey levels indicate the maximum ratios out of 17 orientations, for 17 energy groups as a function of the 60 positions around the head. For this purpose, the 172 energy groups are condensed to 17 groups (the first 12 groups and 16 times 10 groups) in order to improve the statistics and be more realistic regarding obtaining a certain energy group; it is impossible to obtain neutrons in a very narrow energy range apart from having them in a great number. In Figure 3.9, the accompanying optimum orientation is not indicated. The maximum ratios have to meet two additional constraints concerning the thermal neutron flux in the tumours: (1) It has to be >75% of the highest possible thermal neutron flux at that position for a certain orientation and energy group. (2) The relative error has to be <5% in the 95% confidence interval. When these constraints are not met the ratio is displayed as zero. Constraint 1, in particular, truncates the results in Figure 3.9 by which no results are obtained for the lowest and highest energy groups; the values are shown in white. The order of the 60 beam positions in Figure 3.9 is similar to Figure 3.8 and it can be seen that the beams irradiating the back of the head prefer lower source neutron energies. The white histogram in Figure 3.8 shows the ratios obtained similar to those discussed earlier for the Petten beam results, but for the optimum



Figure 3.8. Maximum ratios thermal neutron flux in tumours to OAR of the Petten beam in 60 positions around the phantom head example.



Figure 3.9. Maximum ratios of the thermal neutron flux in the tumours to OAR out of 17 orientations for 17 source neutron energy groups around the phantom head.

energy groups. Because of the high relative errors for the OAR, these ratio values are not trustworthy but nevertheless presented to indicate the kind of improvements that can be expected. In any case, less thermal neutrons in the OAR means that the tumours can be irradiated longer and more boron reactions can be realised.

# **3.5 Conclusions**

Application of the adjoint MC for mono-directional neutron beams enables the optimum irradiation directions and positions for a BNCT treatment plan to be obtained more quickly. The next step would be investigating how to optimise the treatment plan by selecting the best beams out of all information provided by the adjoint. The adjoint also provides information on which source energies give high contributions to the tumours and small to the OAR. It needs investigating further how to deal with the accompanying large relative error in the OAR since this makes it difficult to judge the absolute improvements when using single energy groups. Nevertheless, the adjoint outcome suggests using low source neutron energies, in case of the head phantom example. This has to be further investigated with respect to the skin dose and the dose given to the deepest seated tumours.

# Chapter 4

# Thermal neutron field facility

## 4. Design of a homogeneous thermal neutron field facility for BNCT

The BNCT treatment of organs suffering from cancer which are irradiated extracorporally becomes more and more interesting after the promising results obtained by researchers in Italy [88-90]. Furthermore, researchers in the field of drug development for BNCT desire a facility containing a homogeneous thermal neutron field for in-vitro studies. In this chapter, a measure for homogeneity is defined as the ratio of minimum thermal neutron flux to maximum thermal neutron flux in a volume. As explained in Chapter 1, the thermal neutron flux is a perfect indicator for the <sup>10</sup>B absorptions.

In section 4.1, the description of a facility especially designed for liver treatment by an epithermal neutron beam is given, while section 4.2 discusses the best design for a facility if a new neutron beam would be constructed.

# 4.1 Design of an irradiation facility for the extra-corporal treatment of liver cancer with an existing epithermal beam

In Pavia (Italy), two livers of patients suffering from haematogenous metastasis of colorectal cancer, were treated extra-corporally with BNCT. The first patient lived for almost 4 years (communication from A. Zonta, Pavia Italy) while the second patient died a month after the operation due to heart failure. The success of the first patient, which implies undeniable evidence that BNCT can be a promising treatment, encouraged us to investigate whether extra-corporal liver irradiations can be performed with the existing epithermal neutron beam at the High Flux Reactor in Petten (The Netherlands). The study has the aim to design a facility in which a homogeneous thermal neutron field can be obtained with epithermal source neutrons; the BNCT treatment beam in Petten itself cannot be tailored due to the ongoing clinical trials concerning BNCT of brain tumours {REF 5,6}. As the treatment will be extra-corporal, there is a limit which the patient may be anhepatic, which when corrected for the loss of transportation time, according to the liver surgeon, allows an irradiation time of at most 180 minutes. The study has the additional goal to obtain the same thermal neutron fluence of  $4x10^{12}$  (+/- 20%) cm<sup>-2</sup> as was applied in Pavia. Another requirement (according to the Pavia researchers) is that the liver should not exceed a total weighted dose (as defined in [4] of 15 Gy [89]. The weighting factors (explained in Chapter 1) for the different dose components are those as applied for BPA: RBE<sub>B</sub>=1.3 or 3.8 (healthy or tumour tissue);  $RBE_p = 3.2$ ;  $RBE_p=3.2$ ;  $RBE_y=1.0$ . An important issue that should be investigated thoroughly after a facility has been built is the dose limit for healthy liver. In this chapter, the focus is mainly on obtaining the homogeneous thermal neutron field. This means finding the parameters with the greatest influence, obtaining their optimum values and describing the choice of the materials. More details about the project to treat extra-corporal livers in Petten are given in Paper III.

## 4.1.1 Design parameters and selected materials

The BNCT neutron beam at Petten has an epithermal neutron flux of  $3.3 \times 10^8$  cm<sup>-2</sup>s<sup>-1</sup> with an average energy of 10 keV [29]. The skin, cranium and brain tissues are used to thermalise the neutrons by the elastic scattering properties of the prevailing hydrogen to E<0.5eV. This thermalisation effect is indicated in Figure 4.1: The epithermal neutron flux decreases while the thermal neutron flux increases when the neutrons travel into a cube or sphere of H<sub>2</sub>O which has similar properties to human tissue. The thermal neutron flux has a maximum at around 2.5 cm depth, the so-called thermal maximum. The thermal neutron flux in the sphere (with diameter of 25.0 cm) is higher than in the cube (sides of 20.0 cm) at deeper positions; the thermal flux is distributed over a larger area.



Figure 4.1. The epithermal and thermal neutron flux along the beam centre-line as a function of distance in a cube and sphere of  $H_2O$ .

The fixed parameters in the existing Petten set-up are all source related: Energy and spatially dependent source intensity and a maximum radius  $R_{src}$  of the irradiation beam of 8.0 cm. At the edges of the Petten beam, the neutron intensity drops rapidly. An average liver has a weight of around 1.7 kg [91] and would fit in an imaginary cube with sides of 12.0 cm. From Figure 4.1 it can be seen that the liver would need to be irradiated from more than one side in order to get a flat neutron fluence distribution.

By extrapolating the idea of irradiating a body from more than one side, the step to irradiate a rotating body which is rotational-symmetric seems straightforward. The simulation of the rotation is discussed in subsection 4.1.2. A spheroid shaped liver holder with a polar axis  $R_{pol}$  and an azimuthal axis  $R_{azi}$  is chosen, see Figure 4.2. The advantage of a spherical design is that the intensity fall off near the edges of the beam is compensated by the presence of less liver material to irradiate near the

bottom and top of the sphere.  $R_{pol}$  can be chosen to be the same as  $R_{src}$ . The  $R_{azi}$  should be chosen to be a compromise between large enough to create space for the liver and still provide enough thermal neutrons at the centre of the liver. In fact the thermal neutron profile, as shown in Figure 4.1, is distributed over a volume resembling a torus due to rotation of the spheroidal liver holder.

At the edges of the liver holder the tori-volumes are larger, which will decrease even more the already low thermal neutron flux density since this flux is 'smeared out' over this larger volume. This has to be compensated by surrounding the holder by neutron- moderating and reflecting materials in order to increase the number of thermal neutrons at the edges. In addition to selecting appropriate materials, the dimensions of the materials are also variable. Typical dimensions that can be varied in order to optimize are q1,q2,q3,q4 and q5 as indicated in Figure 4.2.



Figure 4.2. Schematic overview of the liver irradiation set-up with all the design parameters and some 3D drawn tori for tallying neutrons simulating rotation.

For moderation of the epithermal source neutrons, a scattering medium containing a lot of hydrogen should be selected. This requirement is met by the plastic polymethylmethacrylate (PMMA) which is solid and stiff. PMMA has a mass fraction for hydrogen which is around 20% less compared with liver tissue. This difference is reduced to 10% due to the higher density of the plastic. PMMA is non-toxic and transparent and therefore suitable for the liver holder. A disadvantage of hydrogen is the production due to neutron capture of 2.2 MeV gammas. For this reason the rotating liver holder is surrounded by sufficient PMMA to create the necessary thermal neutron build-up and supplemented by graphite for more scattering of the neutrons with less production of gammas. In Figure 4.2 these two materials are defined as build-up and reflector respectively. After selecting these materials, the Monte Carlo code MCNP [7] is used to simulate the neutrons and gammas through the geometry. By varying the parameters q1 to q5, one at a time

and observing the changes, the optimum set-up can be approximated. During this phase the thermal neutron flux is monitored at certain characteristic positions, i.e. at the centre, the thermal maximum, the equator, top and bottom, to obtain a homogeneous neutron flux density. At the same time, it is desired to have as low as possible gamma production. When these criteria are closely met, the thermal neutron flux is calculated everywhere in the liver by simply programming many tori evenly distributed inside the liver. After calculating the entire thermal neutron field inside the liver holder for several combinations of promising parameter settings, the most homogeneous result is chosen. In the calculations the liver tissue composition is taken from ICRU 46 [8]).

#### 4.1.2 Simulating rotation with the Monte Carlo code

By programming tori, only one MCNP calculation is needed in which the beam irradiates the liver from one direction. MCNP adds all the track lengths of the neutrons or gammas inside the torus volume and gives, when dividing by the volume of the torus  $(V_{torus})$ , the neutron and/or gamma flux. From another point of view, the flux can be seen as a time-averaged flux inside the rotating liver. Mathematically, the above explanation of the calculation of a flux in a rotating torus is as follows: Imagine the torus to be fixed and that the beam is rotating around it with a

revolution time T. This means that the angular flux  $\phi(r, \Omega, E, t)$ , as defined in Bell

and Glasstone [69] at r within the torus becomes a function of time t. The angular flux is also depending on the direction  $\Omega$  and energy E of the neutrons. Integrating over time T and dividing by T gives a time averaged angular flux, as shown between the brackets in equation (4.1). In fact:

$$\overline{\phi}_{V_{torus}}^{rev} = \frac{1}{V_{torus}} \int_{V_{torus}} dV \int_{\Omega} d\Omega \int_{E} dE \left\{ \frac{1}{T} \int_{T} dt \phi(\vec{r}, \Omega, E, t) \right\} = \overline{\phi}_{V_{torus}}$$
(4.1)

gives the total flux averaged over  $V_{torus}$  per revolution (rev). This is in fact what MCNP calculates with the so called F4 tally [92].

#### 4.1.3 Results of the designed liver facility at the HFR Petten

Two liver irradiation facility designs are discussed which represent the two extremes: As large as possible irradiated volume with acceptable homogeneity and the smallest acceptable volume with a high homogeneity. The first design has a volume of 2.4 litres and is restricted by the homogeneity demand of +/-20% in the thermal neutron fluence distribution. The second design has a volume of 1.6 litres and is restricted by the minimum volume of liver (together with the conservation liquid) that can be expected. After several MCNP calculations, the optimum parameter settings are obtained which are presented in Table 4.1.

	R <sub>pol</sub>	<b>R</b> <sub>azi</sub>	q1	$q^2$	q3	<i>q4</i>	q5
	[cm]						
1.6ℓ holder	8.0	7.0	1.0	3.5	3.5	15.0	1.0
2.4ℓ holder	8.0	8.5	0.5	3.0	3.5	15.0	1.0

Table 4.1. Optimal parameter settings for liver irradiation set-ups.

Parameters q1, q2, q3 are of direct influence in the Petten set-up while q4 and q5 are of little importance; as long as the graphite thickness q4 > 15.0 cm. The graphite increases by up to 15% of the thermal neutron flux at the edge of the holder volume. It appeared that the build up PMMA around the holders almost doubles the number of thermal neutrons inside the holder. Figure 4.3 shows the cross sections of the thermal neutron flux distributions in the two liver holders and Table 4.2 contains an overview of the characteristic thermal neutron flux, the maximum weighted dose rates and the resulting irradiation times.



Figure 4.3. Left: The thermal neutron flux in the 2.4 litres liver holder. Right: same for 1.6 litres holder.

Description:	2.4ℓ holder	1.6l holder
Average thermal neutron flux	$3.8 \times 10^{8} \text{ cm}^{-2} \text{s}^{-1}$	$4.7 \text{x} 10^8 \text{ cm}^{-2} \text{s}^{-1}$
	-20% minimum	-12% minimum
	+17% maximum	+9% maximum
Time to deliver $4 \times 10^{12}$ cm <sup>-2</sup>	175 min	142 min
Maximum weighted dose rate (point) 8 ppm <sup>10</sup> B	5.8 Gy h <sup>-1</sup>	6.5 Gy h <sup>-1</sup>
Time to deliver 15 Gy maximum dose	155 min	138 min

Table 4.2. Characteristic fluxes and weighted dose rates in both liver holders.

It is due to the smaller  $R_{azi}$  in the 1.6 litres holder that a better homogeneity can be obtained. In both holders the minimum flux is at the edge and centre of the liver volume and the maximum is as to be expected at the position near to the thermal maximum (see Figure 4.3). The thermal maximum is still present and recognizable in both holders. However in the largest holder, the ratio of maximum over minimum within 8.5 cm is 1.5 instead of almost 4 as seen in Figure 4.1. The 2.4 litres holder is

sufficient to hold livers up to 2.1 kg, which do exist (private communication M. Malago and W. Sauerwein, University Hospital Essen, Germany ).

In the 1.6 litres holder, the time to deliver the thermal neutron fluence and the time after which the tolerance dose is exceeded is roughly in agreement. For the largest holder, there seems to be no improvement in homogeneity over the Pavia thermal neutron flux distribution; it has to be recalled that the Petten beam is rather small for such a large organ. Although the required fluence of  $4 \times 10^{12}$  cm<sup>-2</sup> can be given within 180 minutes, the maximum prescribed dose will be exceeded when applying this fluence. This is partly caused by the higher photon dose rate, compared with Pavia, due to the beam photons. Nevertheless, this result is promising enough to proceed first with dosimetry and after that animal studies.

# 4.2 Optimum neutron beam design and holders for extra-corporal BNCT irradiations

The facility described in section 4.1 for irradiating an extra-corporal liver with epithermal source neutrons, tailored specifically for the HFR in Petten, is designed around a given source description. The 'Forward SCX technique' (see Chapter 3) that gives the expected detector contribution of the source particles which is numerically the same as the result of the adjoint method, can be applied for the inverse exercise: To find a neutron source distribution and spectrum that provides a homogeneous thermal neutron field in a given volume.

Actually, the best shape of the holder, i.e. cuboid, cylindrical or spherical (see Figure 4.4) has a direct dependence on the design of a new optimum neutron beam. It is chosen to study these models with volumes of 2, 4 and 6 litres. Despite of the previous sections, in all models, it is chosen to surround the liver only by PMMA and omit the graphite in order to simplify the set-up at this stage. The holder thicknesses are 0.5 cm while a surrounding PMMA block for build up and scattering of neutrons, is 50x50 cm<sup>2</sup> (width x height). In case of the cylinder and sphere, the block starts 2 cm before the rotation axis.

The new beam exit is chosen to be shaped according to the cross-sectional area of the model: this means a rectangular shape for the cuboid and cylinder and circular for the spherical holder. In the calculations, the dimensions of the beam exit (width and height or diameter) are chosen such that there is always sufficient overlap to each side of the cross-sectional area of the maximum investigated liver volume;  $30x40 \text{ cm}^2$  for the cuboid,  $25x39 \text{ cm}^2$  for the cylinder and a diameter of 32 cm for the sphere. In fact, these dimensions ensure that the beam area is two times the cross-sectional area of the largest holder (see dimensions of the 6 litres holders in Figure 4.4). The new beam will have a zero divergence and the desired neutron energy in the liver volume is <0.5 eV and the liver tissue contains 15 ppm of <sup>10</sup>B homogeneously distributed in the liver.

#### 4.2.1 Set-up of the holder models

In the simulation, every holder model contains smaller cells which are the MCNP detectors. In these cells, the expected detector contributions of the source neutrons are obtained. This is necessary to monitor the homogeneity in the entire model during optimisation as will be further explained in section 4.2.3. Due to symmetry

only a quarter of the cuboid model and half of the cylindrical and spherical models have to be filled by these cells. Figure 4.4 shows the holder models and the surrounding PMMA with on the left the defined neutron sources.



Figure 4.4. Cuboid shaped, cylindrical and spherical holder models (all 6 litres) in which the expected detector contributions are calculated together with the neutron beam exits.

The cell structure per model is organized in the following way:

\* Cuboid: A quarter of the 6 litre cuboid model is filled with 300 rectangular parallelopiped shaped cells; width x height x depth= $5 \times 6 \times 10$  cells. The cuboid holder is irradiated from two sides (see arrow in Figure 4.4). In reality, the holder is turned half-way the irradiation time. For the 2 and 4 litre holders, the optimisation is performed over fewer cells; the outside cells are not taken into account whereby the thickness remains always 10 cells=10 cm. Separate MC calculations for these smaller volumes are not performed since the effect of the presence of 10% more hydrogen in liver tissue compared with PMMA is regarded to be insignificant compared with other influences such as the actual environment and beam characteristics.

\* Cylinder: Half of the cylinder consists of 48 ring-shaped cells; diameter x height=8 x 6 rings. The cylindrical holder is rotating along a vertical axis. Similar to the cuboid model: the volume of the cylindrical holder is decreased by not taking into account the lowest 16 and 32 rings for the 4 and 2 litre holders, respectively, during the optimisation process.

\* Sphere: The top half of the sphere is filled (not drawn) with 13 tori-shaped cells. The 4 'tori' along the vertical rotation axis are actually spheres. In every volume, the tori fill the space evenly. The straight arrow, drawn in the spherical holder model, in Figure 4.4 indicates that the diameter of the holder can be decreased, whereby the tori will be redistributed.

#### **4.2.2** Expected contributions of the source neutrons inside the models

The forward source is chosen to have two uniform spectra; of 33 and 43 energy groups between the limits  $1 \times 10^{-5}$  eV to 27.4 keV and  $1 \times 10^{-5}$  eV to 19.6 MeV, respectively. These values of the upper limits come from the XMAS energy structure. By investigating these two spectra, the optimum source spectrum is obtained with and without the influence of the damaging fast neutrons. The energy groups are condensed groups of the 172 energy groups library described in Chapter 3. Everywhere, the statistical uncertainties in all the calculations presented are <5%

(95% confidence interval). In order to keep the graphs surveyable, only 10 cells of the cuboid model and 8 rings of the cylinder are considered. The results for the 6 litre holder models are presented in the Figures 4.5 to 4.7.



Figure 4.5. Expected detector contributions of the source neutrons of 10 cells along and nearest to the beam centreline inside the cuboid model when irradiated from one side.



Figure 4.6. Expected detector contributions of the source neutrons of 8 rings nearest to the beam centreline inside the cylindrical model.



Figure 4.7. Expected detector contributions of the source neutrons of 9 tori and 4 spheres.

In Figure 4.5, 10 cells from the cuboid model, indicated at the upper right, are the nearest to the beam centre line and their expected detector contributions of the source neutrons are shown. Cell 1 is defined here as closest to the beam exit which explains the peaked preference at around 0.5 eV source neutrons. The accompanying importance value of 8.0x10<sup>-5</sup> implies that for every 12,500 source neutrons of around 0.5 eV leaving the 1200 cm<sup>2</sup> source with 43 energy groups, only 1 neutron contributes to neutrons <0.5 eV in this cell 1. Cell 10, farthest away from the beam exit, has a preference for 1 MeV source neutrons. In Figure 4.6, the expected detector contributions of the source neutrons of 8 coaxial rings nearest to the beam centre line are shown. The outer ring has a preference for thermal source neutrons while the thermal neutrons in the centre ring come mainly from fast source neutrons. The expected detector contributions of the source neutrons of the rings in-between are shaped according to a transition from thermal to fast with plateau-shapes for ring 3 and 4. All the expected detector contributions of the source neutrons of the 13 tori of the spherical model (diameter of 22.6 cm) are given in Figure 4.7. There is a great similarity with the expected detector contributions of the source neutrons obtained in the cylindrical model, e.g. the curves are sharply peaked at the fast source neutrons in the tori (actually spheres) at the centre of the models.

Since the cuboid holder will be irradiated from two sides, the curves shown in Figure 4.5 are added such that cell 1 and cell 10 are added, cell 2 and cell 9, etc. The result is presented in Figure 4.8 for cells 1 to 5. The other 5 cells are similar due to symmetry.



Figure 4.8. Expected detector contributions of the source neutrons when 10 cells are irradiated from two sides. Due to the symmetry only 5 cells are shown. The cells are along and nearest to the beam centreline inside the cuboid model.

# 4.2.3 Optimum source neutron energies

In order to realise homogeneity in the holder models, it is desirable to produce the same number of thermal neutrons (E<0.5 eV) in every cell. It is <u>not</u> to be expected that a high homogeneity can be obtained by using a single source neutron energy group. Therefore, it is shown here that by combining the contributions of source neutrons of different energy groups k with weight  $a_k$ , a better homogeneity is attained. The contributions are labelled as the 'importance' (*imp*) and the sum of the weighed importances over all energy groups in every cell j is called  $WI_j$ . Due to the direct relation between the number of source neutrons and importance, homogeneity can be defined as:

Homogeneity = 
$$\frac{\min_{j=1,J}(WI_j)}{\max_{j=1,J}(WI_j)}$$
, where  $WI_j = \sum_{k=1}^G a_k \cdot imp(k,j)$ 

The closer to unity the better is the homogeneity. The minimum and maximum are selected after summing the weighed importances up to *G* energy groups in every cell *j* (*J* cells in total); *G* is 33 or 43 energy groups. The weighs  $a_k$  of the source energy groups are determined by using a linear optimisation scheme. The so-called Simplex method (described in [93-95]) can maximise an objective function *Z* under the constraints provided, whenever there is a solution. The scheme to solve is:

## **Objective function:**

#### **3 constraints:**

$$Z = \max\left(\sum_{j=1}^{J} WI_{j}\right) \qquad \qquad WI_{j} < L \\ WI_{j} \ge Homogeneity \cdot L \text{ with } 0 < Homogeneity < 1 \\ \sum_{k=1}^{G} a_{k} = 1$$

That is: the optimization process will provide the maximum number of thermal neutrons in the liver under the constraint that the ratio of the minimum and maximum  $WI_j$ , among all cells, is between the *Homogeneity* and unity. Similar to the  $a_k$ 's, L is a variable that will be solved during the optimisation. It is needed to prevent setting an absolute constraint which is already done with constraint 3; this keeps the total number of source neutrons constant. The system is solved each time with a given *Homogeneity* that increases in small steps from 0 towards 1. The maximum homogeneity is found for the last value of the *Homogeneity*, for which the Simplex method is able to provide a solution; if the *Homogeneity* increases further, the constraints cannot be satisfied.

In Figures 4.9 and 4.10, the Simplex solution vectors are shown containing the weights  $a_k$  of the energy groups, for the three holder models and three volumes. Figure 4.9 shows the result when source neutrons up to 19.6 MeV may be used while Figure 4.10 is limited up to 27.4 keV.

It is clear from both figures that the maximum homogeneity is reached when using the 'lowest' in combination with the 'highest' neutron energy groups, regardless of the holder model or volume; one exception is the 6 litre spherical holder when no fast neutrons may be used (see Figure 4.10).

For the results of all the "homogeneity vs. neutron source spectrum", the calculations are summarised in Table 4.3. The last two columns of Table 4.3 give the obtained homogeneities for the spectra shown in Figures 4.9 and 4.10.

Model	Volume	Petten	Uniform source spectrum			Optimum source	
	(litres)	spectrum				spectrum (Simplex)	
Г		0eV-	0eV-	<1eV	>1keV	0eV-	0eV-
Energy range		19.6MeV	27.4keV		<27.4keV	27.4keV	19.6MeV
CUBE	2	0.63	0.77	0.48	0.54	0.96	0.98
	4	0.63	0.77	0.48	0.54	0.95	0.97
	6	0.62	0.76	0.47	0.52	0.93	0.95
CYLINDER	2	0.67	0.46	0.26	0.73	0.82	0.99
	4	0.67	0.46	0.26	0.72	0.81	0.99
	6	0.67	0.46	0.26	0.65	0.78	0.88
SPHERE	2	0.62	0.45	0.31	0.60	0.72	0.87
	4	0.44	0.30	0.19	0.55	0.65	0.90
	6	0.36	0.23	0.12	0.44	0.52	0.91

Table 4.3. Obtained optimal homogeneities in the models for different neutron source spectra.



Figure 4.9. The optimum source neutron spectra with a 19.6 MeV upper limit for the 3 geometrical models (3 volumes) to obtain the maximum homogeneity.



Figure 4.10. The optimum source neutron spectra with a 27.4 keV upper limit for the 3 geometrical models (3 volumes) to obtain the maximum homogeneity.

It can be seen that the cylindrical and spherical holders improve significantly whenever fast source neutrons are allowed. The cuboid liver holder provides the best homogeneous thermal neutron field. In studying all Simplex outcomes, it is concluded that the homogeneities will not change significantly compared to the optimum solution when another nearby energy group is selected or the weights are slightly changed. The numbers presented in the last two columns of Table 4.3 are the arithmetical optima. The table also shows the maximum homogeneities when using source neutrons with either a uniform spectrum up to 27.4 keV, or only thermal or only epithermal energy groups. For all spectra in Table 4.3, the fact that the homogeneity of the cuboid and cylindrical models does not vary greatly with decreasing volume suggests that the dimensions of the beam exit can be reduced. Furthermore, Table 4.3 shows that the Petten design presented in section 4.1, might have been improved by choosing a cylindrical liver holder instead of the spherical one. Unfortunately, this is not possible due to the rather small beam opening in Petten. The right choice for the spherical design in the Petten set-up is also conferred when taking into account the minimum summed weighed importance  $(WI_i)$  obtained in the holder model; this means a minimum number of delivered thermal neutrons. In Figure 4.11, this minimum  $W_{I_i}$  is shown for every solution of the Simplex method (27.4 keV upper limit case) somewhere in the liver as a function of Homogeneity.



Figure 4.11. The minimum reached importance in a certain cell inside the 3 geometrical models, for 3 volumes each, as a function of the homogeneity.

Normalised to per source area, this minimum  $WI_j$  is highest in the 2 and 4 litre spherical holder models. In these models, the source neutrons can be said to be at their most effective. Therefore, given the small existing beam opening and intensity, as in Petten, the spherical holder seems from this graph to be the best choice. The curves in Figure 4.11 are based on the source areas as defined in section 4.2. In order to have the same ratio of beam area to cross-sectional area for all models with the

same volume, the curves for the spherical holders of 2 and 4 litres have to be multiplied with 0.67 and 0.84, respectively. It can be seen that even then, the contributions of the source neutrons in these spherical models are still higher.

Describing Figure 4.11 in terms of the Simplex method: For low values of the *Homogeneity* in the constraints, the objective function is maximised by having large  $WI_j$ 's in many cells at the cost of low  $WI_j$ 's in a small number of cells. For the 2 litres spherical and for all the cylindrical and cuboid shaped holders, it can be seen that the minimum  $WI_j$  somewhere in the holders, first increase then decrease towards the maximum homogeneities. This means that for the maximum homogeneities, the source neutrons are less effective and more source neutrons are needed to reach a certain effect. This does <u>not</u> mean immediately that it will take longer to irradiate since that depends on the available source neutron energy spectrum in combination with the source strength.

# 4.3 Conclusions

To conclude, it seems that the cuboid shaped model would be the best option when a new neutron beam is to be designed and constructed. The investigated thickness of 10 cm is sufficient for both livers and other similarly shaped volumes, such as for example the irradiation of cell cultures, for which also a homogenous thermal neutron field is required. The disadvantage, namely the rather large beam exit, can be circumvented by transporting the cuboid holder up and down (or left and right) through a smaller neutron beam with a high intensity. Furthermore, for neutron sources that may have relatively low strength, the 2 and 4 litre spherical holder models would be the preferred configuration which is practically the case with the present epithermal neutron beam in Petten. Note that the obtained homogeneity in the 2.4 litre spheroid liver holder (0.68) of section 4.1 is close to the optimum, as can be concluded from the spherical results in Table 4.2; the spheroid shape improves the result even further.

# Chapter 5

# Conclusions

The quest for the optimal source neutrons for BNCT was investigated from three viewpoints:

\* In the first point of view, as there is no consensus among radiobiologists about the proper values of the parameters influencing the biologically weighted dose, e.g. RBE values and tolerance doses, almost all of these parameters were varied. For brain tumours located between 20 mm up to 80 mm from the skin it was found that for most parameter configurations, epithermal source neutrons, between 1 keV and 10 keV, are optimal. Only the parameters related to <sup>10</sup>B and fast neutrons result in diverging optimal source neutron energies. For example, low values for the <sup>10</sup>B concentrations and the accompanying CBE for the boron dose result often in lower than modal source neutron energies to be optimal. Smaller values for the RBE of the fast neutron dose enable higher source neutron energies to be optimal and thus more destructive in the tumour. These parameters are of influence in the tissue in which the tolerance dose is reached first. This is depending on the ratio of the tolerance dose set for skin to brain. It is likely that in the future, boron compounds will be available which give <sup>10</sup>B concentrations close to zero in skin and other healthy tissues. Furthermore, it is possible that in the future BNCT treatment modalities (e.g. extra-corporeal, at low temperatures) the recoil protons are found to be less damaging for the cells. Of course this needs to be investigated.

This parameter study shows that having three neutron beams, one nearly thermal (5 eV) and two epithermal (500 eV and 10 keV), assures an optimal treatment, no matter what the influencing biologically weighted dose parameter values may be.

\* Two newly developed adjoint Monte Carlo (MC) techniques, the Legendre Expansion Technique (LEXT) and the Adjoint Point Detector Technique (APDT), enable the simulation of mono-directional gamma, as well as neutron, beams. When the beam diameters are larger than 5 cm, the results for 1020 beams are calculated 1.8 up to 20 times faster than when compared with results from forward MC. Overall, for small diameter neutron and gamma beams (around 5 cm), the adjoint MC techniques are preferred when thousands of different locations and orientations of a mono-directional beam need to be calculated and when there are no more than ten tumours and/or organs at risk (OAR). For larger beam diameters is adjoint MC preferable up to hundreds of regions of interest whenever even a 'few' hundred of mono-directional beams are investigated. Overall, in order to take advantage of the adjoint technique, the user has to be interested in beam positions at many locations around the irradiated patient or phantom. Apart from BNCT, the LEXT and APDT are of value in conventional radiotherapy since most of the treatment beams are mono-directional or have just a small divergence. Thereby, the conventional radiotherapy community would be interested to implement MC for dose

calculations, since the results are closer to the measurements than the results of other dose calculation techniques. For the LEXT, it is expected that other base functions than Legendre polynomials, will give improvements whenever they are more appropriate to describe the angular adjoint functions in a certain problem (e.g. particle type, geometry). At present, the divergence of a beam is easy to simulate with the APDT, just by changing the position of the adjoint point detector. However, a problem occurs when the optimal source neutron energy is investigated with adjoint MC because by definition the optimal source neutrons give a high contribution to the tumours and a low contribution to the OAR. As a consequence, the values of interest in the OAR have a large relative error, and no judgement can be given on the outcome. Nevertheless, preliminary results for a model with ten tumours in the brain showed that only thermal and low epithermal source neutrons can give significant improvements to the ratio of the thermal neutron flux in tumours to the thermal neutron flux in the OAR, when compared with the ratio resulting from the Petten BNCT beam. Further analysis in the field of treatment planning optimisation is obviously the next step.

\* The last point in this thesis is the investigation to obtain the optimal source neutrons for future BNCT applications such as the extra-corporeal irradiation of organs (e.g. liver). The first part of this research shows that it is possible to irradiate a large liver, with a volume of 2.4 litres, at the current BNCT facility in Petten. This requires rotation of the liver in a spheroid holder in order to obtain thermal neutron fluence as homogeneous as possible with the rather small epithermal beam. The homogeneity is defined as the ratio of the minimum to the maximum thermal neutron flux in the liver holder. Simulation of the doses and fluxes in the rotating liver was performed with MCNP by programming torus shaped tallies in the liver and irradiating from one side. Averaging the flux and/or dose over the volume of the torus provides the answer as if the torus is rotating around the symmetry axis. The homogeneity in the designed and constructed 2.4 litre liver holder is 0.68. Since more researchers are interested in a volume with a homogenous thermal neutron field, e.g. for cell experiments, a systematic investigation was started in which also the neutron source could vary in shape and energy. In cubic volumes a homogeneity of around 0.95 can be realised when the source neutrons are mixed in a composition of around 30% thermal (around 0.1 eV) and 70% epithermal (around 10 keV). In spherical volumes the homogeneities cannot exceed 0.72, but these spherical shapes are prefered when high effectiveness per source neutron is required. This is necessary in case of a low source strength in combination with a limited irradiation time.

To conclude: in 1941, Zahl *et al.* {REF} proposed the use of epithermal instead of thermal source neutrons in BNCT. The outcome of the three parts of this thesis has shown that 3 neutron energy regimes should be prescribed. Besides the 10 keV epithermal source neutrons, low epithermal source neutrons of around 1 eV and thermal source neutrons with energies of 0.1 eV must also be used. With these source neutrons available, BNCT treatment planning can be optimised, whatever the biological factors may be, wherever the tumours are located, or wherever the patient
or the patient's organ is situated. This work shows that significant improvements in the delivery of a therapeutic radiation dose can be given if a new neutron filter is designed in order to provide a variable neutron spectrum as described above.

56 APPENDIX

## APPENDIX

## Normalisation of adjoint MCNP results for mono-directional detector contributions

In Monte Carlo calculations, the probability that adjoint particles cross a certain detector with a certain angle is nearly zero. Knowing even that more of these particles are necessary to improve the statistics of the result indicates that it will be impossible to calculate such tasks with standard Monte Carlo. Therefore, a forward calculation concerning mono-directional source particles cannot be derived with the adjoint method by means of 'basic' MCNP. Figure A.1 shows schematically this forward set-up with a disc shaped source and spherical tumour as a detector. This figure and also the Figures A.2, A.3 and A.4 contain the parameter descriptions as used throughout this appendix and the characteristics as will be later used in Table A.1

In fact, a surface crossing estimator  $(EST_{sc})$  should have been used for monodirectional adjoint particles. So the detector response function should be like

$$\frac{1}{A_{src}}\delta(\Omega + \Omega_{src})\delta(y - y_{src})$$
 with (x, z) within the detector area

Although it can be performed with 'basic' MCNP, this is not a feasible estimator for a Monte Carlo calculation, as no particle (like the grey arrow in Figure A.2) will have exactly the direction  $-\Omega_{src}$ . Therefore, a next-event estimator would be required which calculates from a given collision site the probability of scattering  $p(-\Omega_{src})$  in the right direction and the attenuation  $e^{-\xi}$  to get to the source disc as illustrated in Figure A.3.  $\xi$  is the number of mean free paths from the collision point to the detector (in fact to the source plane). Hence, the required estimator should be

$$EST_{sc} = p(-\Omega_{src})\frac{e^{-\zeta}}{A_{src}}$$
(a.1)

As such an estimator is not present in MCNP, this estimator is replaced by the nextevent point detector estimator of MCNP, as illustrated in Figure A.4. The point detector estimator ( $EST_{pd}$ ) gives for a particle entering a collision

$$EST_{pd} = p(\Omega_{det}) \frac{e^{-\xi}}{s^2}$$
(a.2)

where  $p(\Omega_{det})$  is the probability of scattering into the direction  $\Omega_{det}$  towards the point detector and *s* is the distance between the collision site and the detector. This is provided with the F5 tally in MCNP [7]. If the position of the point detector is sufficiently far away,  $\Omega_{det}$  will be sufficiently equal to the (opposite) source direction  $\Omega_{src}$ . If the system dimensions are small compared to the



Figure A.2. Adjoint 'basic' set-up.



Figure A.3. Adjoint 'required' set-up.



Figure A.4. Adjoint 'point detector' set-up.

distance to the detector, *s* can be considered constant and equal to the distance between the centre of the system and the point detector.

According to the methodology introduced by Wagner *et al* [70], Table A.1 can be composed for the geometry drawn in the Figures A.1 up to A.4 with the volume averaged flux in the tumour as the forward detector. Practically, after filling in the source and detector densities in Table A.1, the normalization factor is the product of the entries in the forward problem column divided by the product of the entries in the adjoint problem column.

The 3 columns for an adjoint calculation are (by approximation) equivalent with respect to the intention of the estimator. Comparing the results from the columns for the forward and the 'basic' adjoint equation (for a flat source spectrum and a flat detector energy response) a multiplicative normalization factor in the adjoint equation should be  $4\pi$ . The estimator suggested for the adjoint 'required' calculation is equivalent to adjoint 'basic' and here also the normalization factor  $4\pi$  results. By approximation this estimator is also equivalent to the adjoint point detector, but from comparison of the relevant columns, the factor  $s^2$  must be compensated, as well as a factor  $A_{src}$ . Hence, for the actually calculated adjoint point detector the normalization factor  $F_{norm}$ , with which the result of the adjoint calculation must be multiplied becomes

$$F_{norm} = \frac{4\pi}{A_{src}} s^2 = \frac{4}{R_{src}^2} s^2$$
(a.3)

Note that here the situation is given for the total response over energy. If the response for one group is required, the adjoint source is limited to this group and the source density in energy is 1 instead of 1/G where G is the number of energy groups. This gives another factor 1/G in the normalization.

; 'point
or'
$r \in V_{det}$
pic)
$R_{g'}$
$\vec{j}$ for
ponse
*
)
flat
pic R F fc por flat

Table A.1. Normalisation form for the sample problem illustrated in the Figures A.1 up till A.4

## Acknowledgement

The exact derivation of the normalisation factors presented here is performed by J.E. Hoogenboom.

## NOMENCLATURE

А	surface area	$m^2$
а	weigh of source energy group	-
С	concentration	ppm
CBE	compound related biological effectiveness factor	-
d	coefficient of orthonormal basis functions	-
D	physical dose	Gy
Е	energy	eV
EST <sub>pd</sub>	next-event point detector estimator	$m^{-2}$
EST <sub>sc</sub>	surface crossing estimator	m <sup>-2</sup>
f	arbitrary function	-
Fnorm	normalization factor for adjoint results	-
G	number of energy groups	-
J	number of cells	-
L	variable	-
Ν	number of samples	-
р	direction probability function	-
Q	source function	$cm^{-3} s^{-1}$
q1:q5	dimensions (thickness, distance)	m
r	position	-
R	radius	m
RBE	relative biologically effectiveness factor	-
R <sub>g</sub>	group wise response function	-
S	distance between collision site and detector	m
$S_g$	group wise source function	-
Т	revolution time	S
t	time	S
V	volume	m'
W	weight of Monte Carlo particle	-
WI	sum of weighed importances	-
Х	spatial coordinate	m
У	spatial coordinate	m
Z	objective function	-
Z	spatial coordinate	m
Greek s	ymbols	
$\phi^+$	adjoint function	-
$\phi$	forward flux	cm <sup>-2</sup> s <sup>-1</sup>
$\overline{\phi}_{V}$	volume averaged forward flux	cm <sup>-2</sup> s <sup>-1</sup>
α	arbitrary angle	rad
ß	arbitrary angle	rad
δ	Dirac-function	-
λ	ratio source gammas to source neutrons	-

ξ	number of mean free paths to detector
$\Sigma_{\rm d}$	detector response function

cm<sup>-1</sup>

\_

ster

- $\psi$  orthonormal basis functions
- ψ orthonorr Ω direction

## Subscripts

azi	azimuthal
В	boron-10
bγ	beam gamma related
det	detector
i	index for samples
j	index for cells
k	index for energy groups
1	index for coefficients
m	index for coefficients
n	fast neutron related
р	thermal neutron related
pol	polar
rev	revolution
src	source
W	biologically weighted
γ	induced gamma related

## **ABBREVIATIONS**

APDT	Adjoint Point Detector Technique
BNCT	Boron Neutron Capture Therapy
BNL	Brookhaven National Laboratory
BPA	Borono-phenylalanine
BSH	Borocaptate Sodium
CBE	Compound related Biological Effectiveness
DNA	Deoxyribonucleic acid
EELS	Electron Energy Loss Spectroscopy
EORTC	European Organisation for Research and Treatment of Cancer
EST	Estimator
FET	Functional Expansion Technique
HFR	High Flux Reactor
IAEA	International Atomic Energy Agency
ICRU	International Commission on Radiation Units & measurements
JRC	Joint Research Centre
KERMA	Kinetic Energy Released per unit Mass
LET	Linear Energy Transfer
LEXT	Legendre EXpansion Technique
MC	Monte Carlo
MCNP	Monte Carlo N-Particle transport code
MIT	Massachusetts Institute of Technology
OAR	Organs At Risk
PMMA	Polymethyl Methacrylate)
PNR	Physics of Nuclear Reactors
RBE	Relative Biological Effectiveness
ROI	Regio Of Interest

### REFERENCES

- T. HARTMAN, J. CARLSSON, "Radiation dose heterogeneity in receptor and antigen mediated boron neutron capture therapy," Radiother Oncol. 31(1), 61-75 (1994).
- J. A. CODERRE, J. C. TURCOTTE, K. J. RILEY, P. J. BINNS, O. K. HARLING, W. S. KIGER, III, "Boron Neutron Capture Therapy: Cellular Targeting of High Linear Energy Transfer Radiation," Technology in Cancer Research & Treatment Volume 2, Number 5, October (2003).
- 3. J.J. DUDERSTADT, L.J. HAMILTON, Nuclear reactor analysis, John Wiley & Sons Inc, New York (1976).
- 4. International Atomic Energy Agency IAEA: Current status of neutron capture therapy. IAEA-TECDOC-1223. Vienna (2001).
- J.A. CODERRE, G.M. MORRIS, "The radiation biology of boron neutron capture therapy," Rad. Res., 151, 1-18, (1999).
- E.H. ELOWITZ, R.M. BERGLAND, J.A. CODERRE, D.D. JOEL, M. CHADHA, A.D. CHANANA, "Biodistribution of pboronophenylalanine in patients with glioblastoma multiforme for use in boron neutron capture therapy," Neurosurgery 42(3), 463-8 (1998).
- 7. J.F. BRIESMEISTER, MCNP A General Monte Carlo N-Particle Transport Code, Version 4C. LA-13709-M (2000).
- 8. ICRU 46. Photon, electron, proton and neutron interaction data for body tissues. International Commission on Radiation Units and Measurements. Bethesda. MD (1992).
- G. M. MORRIS, J. A.CODERRE, J. W. HOPEWELL, P. L.MICCA, M. REZVANI, "Response of Rat Skin to Boron Neutron Capture Therapy with p-boronophenylalanine or Borocaptate Sodium," Radiother. Oncol. 32, 144-153 (1994).
- A.H. SOLOWAY, H. HATANAKA, M.A. DAVIS, "Penetration of Brain and Brain Tumor. VII. Tumor-binding Sulfhydryl Boron Compounds," J. Med. Chem., 10:714-717, (1967).
- H. C. MILLER, N. E. MILLER, E. L. MUETTERTIES, "Synthesis of Polyhedral Boranes," J. Am. Chem. Soc. 85(23), 3885 – 3886 (1963).
- H.R. SNYDER, A. H. REEDY, W.LENNARZ, "Synthesis of Aromatic Boronic Acids. Aldehydo Boronic Acids and a Boronic Acid Analog of Tyrosine," J. Am. Chem. Soc. 80, 835-838 (1958).
- Y. MISHIMA, C. HONDA, M. ICHIHASHI, H. OBARA, J. HIRATSUKA, H. FUKUDA, H. KARASHIMA, T. KOBAYASHI, K. KANDA, K.YOSHINO, "Treatment of Malignant Melanoma by Single Thermal Neutron Capture Therapy with Melanoma-seeking 10B-compound," Lancet II, 388-389 (1989).
- 14. K.H.I. PHILIPP, Tolerance of normal brain tissue to Boron Neutron Capture Irradiation, PhD Thesis, Katholieke Universiteit Nijmegen (2004).
- 15. J. CHADWICK, "The Existence of a Neutron," Proc. Roy. Soc., A,

136, 692-708 (1932).

- 16. E. RUTHERFORD, "Nuclear Constitution of Atoms," Proc. Roy. Soc. A97, 374-400 (1920).
- H.J. TAYLOR, M. GOLDHABER, "Detection of nuclear disintegration in a photographic emulsion," Nature (Lond.) 135, 341-348 (1935).
- G.L. LOCHER, "Biological effects and therapeutic possibilities of neutrons," Am. J. Roentgenol. Radium. Ther. 36, 1-13 (1936).
- P. G. KRUGER, "Some Biological Effects of Nuclear Disintegration Products on Neoplastic Tissue," Proc.Natl.Acad.Sci.26, 181-192 (1940).
- P.A. ZAHL, F.S. COOPER, "Physical and biological considerations in the use of slow neutrons for cancer therapy," Radiology 37, 673-682 (1940).
- P. A. ZAHL, F. S. COOPER, J. R. DUNNING, "Some in vivo Effects of Localized Nuclear Disintegration Products on a Transplantable Mouse Sarcoma," Proc.Natl.Acad.Sci. 26, 589-598 (1940).
- 22. P. A. ZAHL, F. S. COOPER, "Localization of Lithium in Tumor Tissue as a Basis for Slow Neutron Therapy," Science, 64-65 (1941).
- L.E. FARR, W.H. SWEET, J.S. ROBERTSON, S.G. FOSTER, H.B. LOCKSLEY, D.L.SUTHERLAND, M.L. MENDELSOHN, E.E. STICKEY, "Neutron Capture Therapy with Boron in the Treatment of Glioblastoma Multiforme," Am. J. Roentgenol., 71, 279-291 (1954).
- W. H. SWEET, "Medical aspects of boron slow-neutron capture therapy," In: Workshop on Neutron Capture Therapy, Report BNL-51994, eds. R.G. Fairchild, V.P. Bond (Brookhaven Natl. Lab., Upton, NY), 173-177 (1986).
- H. HATANAKA, Y. NAKAGAWA, "Clinical results of longsurviving Brain Tumor Patients who underwent boron neutron capture therapy," Int. J.Radiat.Oncol.Biol.Phys. 28, 1061-1066 (1994).
- H. HATANAKA, "Clinical results of boron neutron capture therapy," Basic Life Sci. 54, 15-21 (1990).
- 27. R.L. MOSS, "Progress towards boron neutron capture therapy at the High Flux Reactor Petten," Basic Life Sci. 54, 169-83 (1990).
- G. DETLEF, R. MOSS, editors, Boron neutron capture therapy: toward clinical trials of glioma treatment (HB), Plenum Press, New York (1992).
- R.L. MOSS, O. AIZAWA, D. BEYNON, R. BRUGGER, G. CONSTANTINE, O. HARLING, H.B. LIU, P. WATKINS, "The requirements and development of neutron beams for neutron capture therapy of brain cancer," J. Neuro-Onc. vol 33, 27-40 (1997).
- 30. L. KANKAANRANTA, T. SEPPÄLÄ, M. KALLIO, J. KARILA, C. ASCHAN, T. SEREN, M. KORTESNIEMI, P. KOTILUOTO, E. JÄRVILUOMA, M. KULVIK, J. LAAKSO, A. BRANDER, J. JÄÄSKELÄINEN, J. VÄHÄTALO, I. AUTERINEN, S. SAVOLAINEN, M. FÄRKKILÄ, H. JOENSUU, "First clinical

results on the Finnish study on BPA-mediated BNCT in glioblastoma," - NCT Osaka 2000. 9th International Symposium on Neutron Capture Therapy for Cancer, Program & Abstracts. Osaka, JP, 2 - 6 Oct., 31-32 (2000).

- J. CAPALA, B.H. STENSTAM, K. SKOLD, P.M. AF ROSENSCHOLD, V. GIUSTI, C. PERSSON, E. WALLIN, A. BRUN, L. FRANZEN, J. CARLSSON, L. SALFORD, C. CEBERG, B. PERSSON, L. PELLETTIERI, R. HENRIKSSON, "Boron neutron capture therapy for glioblastoma multiforme: clinical studies in Sweden," J.Neurooncol. 62(1-2), 135-44 (2003).
- 32. B.T.S. THIRUMAMAGAL, X.B. ZHAO, A.K. BANDYOPADHY, S. NARAYANASAMY, J. SAMUEL, R. TIWARI, D.W. GOLIGHTLY, V. PATEL, B.T. JEHNING, M.V. BACKER, R.F. BARTH, R.J. LEE, J.M. BACKER, W. TJARKS, "Receptor-Targeted Liposomal Delivery of Boron-Containing Cholesterol Mimics for Boron Neutron Capture Therapy (BNCT)," Bioconjugate Chem. 17 (5), 1141 -1150 (2006).
- J. CARLSSON, E.B. KULLBERG, J. CAPALA, S. SJÖBERG, K. EDWARDS, L. GEDDA, "Ligand Liposomes and Boron Neutron Capture Therapy," J.Neuro-Oncology 62, Number 1, 47-59 (2003).
- G.C. KRIJGER, M.M. FRETZ, R. CARVALHO, O. STEINEBACH, W. JISKOOT, G. STORM, G.A. KONING, "Tumor targeted lipsomes for neutron capture therapy," Radiochim. Acta 93, 589-593 (2005).
- G.A. KONING, M.M. FRETZ, U. WORONIECKA, G. STORM, G.C. KRIJGER, "Targeting liposomes to tumor endothelial cells for neutron capture therapy," Appl. Radiat. Isot. 61 (5), 963-967 (2004).
- R.A. WATSON-CLARK, M.L. BANQUERIGO, K. SHELLY, M.F. HAWTHORNE, E. BRAHN, "Model studies directed toward the application of boron neutron capture therapy to rheumatoid arthritis: Boron delivery by liposomes in rat collagen-induced arthritis," Proc.Natl.Acad.Sci. 95, 2531–2534 (1998).
- J.C. YANCH, S. SHORTKROFF, R.E. SHEFER, S. JOHNSON, E. BINELLO, D. GIERGA, A.G. JONES, G. YOUNG, C. VIVIEROS, A. DAVISON, C. SLEDGE," Boron Neutron Capture Synovectomy: Treatment of Rheumatoid Arthritis Based on the 10B(n,alpha)7Li Nuclear Reaction," Medical Physics (26)3, 364-375 (1999).
- P. VAN LENT, "Selective Elimination of Synovial Macrophages by Boron Neutron Capture Therapy prevents onset of Murine Experimental Arthritis," 12th International Congress on Neutron Capture Therapy, Proceedings of ICNCT-12 Takamatsu, JP, 9 – 13 Oct. (2006).
- 39. Y. SAKURAI, K. ONO, S. MIYATAKE, A. MARUHASHI, "Improvement effect on the depth-dose distribution by CSF drainage and air infusion of a tumour-removed cavity in boron neutron capture therapy for malignant brain tumours," Phys. Med. Biol. 51, 1173-

1183 (2006).

- K. YOKOYAMA, S. MIYATAKE, Y. KAJIMOTO, S. KAWABATA, A. DOI, T. YOSHIDA, T. ASANO, M. KIRIHATA, K. ONO, T. KUROIWA, "Pharmacokinetic study of BSH and BPA in simultaneous use for BNCT," J. Neuro-Oncology 78 Number 3, 227-232 (2006).
- J. MICHEL, G. BALOSSIER, A. WITTIG, W. SAUERWEIN, K. ZIEROLD, "EELS spectrum-imaging for boron detection in biological cryofixed tissues," Instr.sci.technol. 33 no6, 631-644 (2005).
- K.J. RILEY, P.J. BINNS, D.D. GREENBERG, O.K. HARLING, "A Physical Dosimetry Intercomparison for BNCT," Med.Phys. 29 (5), May (2002).
- 43. P.J. BINNS, K.J. RILEY, O.K. HARLING, W.S. KIGER 3RD, P.M. MUNCK AF ROSENSCHOLD, V. GIUSTI, J. CAPALA, K. SKOLD, I. AUTERINEN, T. SEREN, P. KOTILUOTO, J. UUSI-SIMOLA, M. MAREK, L. VIERERBL, F.SPURNY, "An international dosimetry exchange for boron neutron capture therapy. Part I: Absorbed dose measurements," Med.Phys. 32(12), 3729-36 (2005).
- 44. J. BENCZIK, T. SEPPÄLÄ, M. SNELLMAN, H. JOENSUU, G. M. MORRIS, J. W. HOPEWELL, "Evaluation of the Relative Biological Effectiveness of a Clinical Epithermal Neutron Beam Using Dog Brain," Radiat. Res. 159, 199–209 (2003).
- 45. J. GUEULETTE, P.J. BINNS, B.M. DE COSTER, X-Q. LU, S.A. ROBERTS, K.J. RILEY, "RBE of the MIT fission converter beam for crypt cell regeneration in mice," Radiat. Res. 164, 805-809 (2005).
- R.G. FAIRCHILD, "Development and Dosimetry of an Epithermal Neutron Beam for Possible Use in Neutron Capture Therapy, I. 'Epithermal' Neutron Beam Development," Phys.Med.Biol. 10 No. 4, 491-504 (1965).
- R.G. FAIRCHILD, "Development and Dosimetry of an Epithermal Neutron Beam for Possible Use in Neutron Capture Therapy, II. Absorbed Dose Measurements in a Phantom Man," Phys.Med.Biol. 11 No. 1, 15-30 (1966).
- 48. O.AIZAWA, "Research on neutron beam design for BNCT at the Musashi Reactor," Basic Life Sci. 54, 109-24 (1990).
- G. TRACZ, L. DABKOWSKI, D. DWORAK, K. PYTEL, U. WOZNICKA, "The filter/moderator arrangement-optimisation for the boron-neutron capture therapy (BNCT)," Radiat. prot. dosim.110 no1-4, 827-831 (2004).
- J.M. VERBEKE, A.S. CHEN, J.L. VUJIC, K.N. LEUNG, "Optimization of Beam-Shaping Assemblies for BNCS Using the High-Energy Neutron Sources D-D and D-T," Rad.Biol.Med. 134(3), 278-293 (2001).
- 51. T. KOBAYASHI, Y. SAKURAI, K. KANDA, Y. FUJITA, K. ONO,

"The Remodeling and Basic Characteristics of the Heavy Water Neutron Irradiation Facility of the Kyoto Univesity Research Reactor, Mainly for Neutron Capture Therapy," Rad.Biol.Med. 131(3), 354-378 (2000).

- T. MATSUMOTO, H.B.LIU, R.M.BRUGGER, "Design studies of an epithermal neutron beam for neutron capture therapy at the Musashi reactor," J. Nucl. Sci. Technol. 32 no2, 87-94 (1995).
- 53. H.B. LIU, R.M. BRUGGER, D.C. RORER, P.R. TICHLER, J.P. HU, "Design of a high-flux epithermal neutron beam using 235U fission plates at the Brookhaven Medical Research Reactor," Med Phys. 21(10), 1627-31 (1994).
- J.C. YANCH, X.L. ZHOU, G.L.BROWNELL, "A Monte Carlo Investigation of the dosimetric properties of monoenergetic neutron beams for neutron capture therapy," Radiat.Res.126, 1-20 (1991).
- J.C. YANCH, O.K. HARLING, "Dosimetric Effects of Beam size and collimation of epithermal neutrons for Boron Neutron Capture Therapy," Rad.Res. 135, 131-145 (1993).
- E. BISCEGLIE, P. COLANGELO, N. COLONNA, P. SANTORELLI, V. VARIALE, "On the optimal energy of epithermal neutron beams for BNCT," Phys.Med.Biol. 45, 49-58 (2000).
- 57. D.L. BLEUEL, R.J. DONAHUE, B.A. LUDEWIGT, J. VUJIC, "Development of a Neutron Energy-Biased In-Air Figure of Merit for Predicting In-Phantom BNCT Neutron Beam Characteristics," In: Proceedings of the 8th International Symposium on Neutron Capture Therapy for Cancer, La Jolla, CA, Sep. 13-18, (1999).
- 58. H.D. THAMES, J.H. HENDRY, Fractionation in Radiotherapy, Taylor & Francis Books Ltd., London (1987).
- 59. J.F. FOWLER, "Biological factors influencing optimum fractionation in radiation therapy," Acta Oncologica 40(6), 712-717 (2001).
- D.T. GOODHEAD, J. THACKER, R. COX, "Effects of radiations of different qualities on cells: molecular mechanisms of damage and repair," Int.J.Radiat.Biol. 63, 543-556 (1993).
- 61. E.J. HALL, A. GIACCIA, Radiobiology for the Radiologist, Lippincott Williams and Wilkins, Philadelphia, San Francisco, Sixth Edition (2005).
- A.L. BROOKS, G.J. NEWTON, L.J. SHYR, F.A. SEILER, B.R. SCOTT, "The combined effects of alpha-particles and X-rays on cell killing and micronuclei induction in lung epithelial cells," Int.J.Radiat. Biol. 58, 799-811 (1990).
- 63. Y. FURUSAWA, M. AOKI, M. DURANTE, "Simultaneous exposure of mammalian cells to heavy ions and X-rays, Advances in Space Research," 30, 877-884 (2002).
- 64. A. WITTIG ET AL., Early phase II study on BNCT in metastatic malignant melanoma using the boron carrier BPA, EORTC protocol 11011 (2003).
- 65. W. SAUERWEIN ET AL., Postoperative Treatment of glioblastoma

with BNCT at the Petten Irradiation Facility, EORTC Protocol 11961 (1999).

- 66. K. HIDEGHETY, W. SAUERWEIN, A. WITTIG, C. GOTZ, P. PAQUIS, F. GROCHULLA, K. HASELSBERGER, J. WOLBERS, R. MOSS, R. HUISKAMP, H. FANKHAUSER, M. DE VRIES, D. GABEL, "Tissue uptake of BSH in patients with glioblastoma in the EORTC 11961 phase I BNCT trial," J.Neurooncol. 62(1-2), 145-56 (2003).
- D.W. NIGG, "Some recent trends and progress in the physics and biophysics of neutron capture therapy," Progr. In Nuclear Energy 35 No 1, 79-127 (1999).
- C. VROEGINDEWEIJ, F. WHEELER, F. STECHER-RASMUSSEN, R. HUISKAMP, "Microdosimetry Model for Boron Neutron Capture Therapy. II. Theoretical Estimation of the Effectiveness Function and Surviving Fractions," Rad. Res. 155, 498-502 (2001).
- 69. G. I. BELL, S. GLASSTONE, Nuclear Reactor Theory, Van Nostrand Reinhold Company, New York, NY (1970).
- J.C. WAGNER, E.L. REDMOND II, S.P. PALMTAG, J.S. HENDRICKS, "MCNP: Multigroup/Adjoint Capabilities," LA-12704, Los Alamos National Laboratory (1994).
- J.E. HOOGENBOOM, "Methodology of Continuous-Energy Adjoint Monte Carlo for Neutron, Photon, and Coupled Neutron-Photon Transport," Nucl. Sci. Eng. 143, 99-120 (2003).
- 72. F.C. DIFILIPPO, "Forward and adjoint methods for radiotherapy planning," Med. Phys. 25(9), 1702-10 (1998).
- 73. J.V. SIEBERS, M. LAUTERBACH, S. TONG, Q. WU, R. MOHAN, "Reducing dose calculation time for accurate iterative IMRT planning," Med.Phys. 29(2), 231-237 (2002).
- W. ULMER, J. PYYRY, W. KAISSL, "A 3D photon superposition/convolution algorithm and its foundation on results of Monte Carlo calculations," Phys.Med.Biol. 50, 1767-1790 (2005).
- 75. Q. WU, D. DJAJAPUTRA, M. LAUTERBACH, Y. WU, R. MOHAN, "A fast dose calculation method based on table lookup for IMRT optimization," Phys.Med.Biol. 48, N159-N166 (2003).
- C.M. MA, E. MOK, A. KAPUR, T. PAWLICKI, D. FINDLEY, S. BRAIN, K. FORSTER, A. L. BOYER, "Clinical implementation of a Monte Carlo treatment planning system," Med.Phys. 26(10), 2133-2143 (1999).
- J.J. DEMARCO, I.J. CHETTY, T.D. SOLBERG, "A Monte Carlo tutorial and the application for radiotherapy treatment planning," Med.Dosim. 27(1), 43-50 (2002).
- 78. R. JERAJ AND P. KEALL, "Monte Carlo-based inverse treatment planning," Phys. Med. Biol. 44,1885-96 (1999).
- 79. M. GOLDSTEIN, "The adjoint Monte Carlo a viable option for efficient radiotherapy treatment planning," Proc. of the 19th Conf. of

the Israeli Nucl. Soc. 11-2 (1996).

- S. YOO, M.E. KOWALOK, B.R. THOMADSEN, D.L. HENDERSON, "Treatment planning for prostate brachytherapy using region of interest adjoint functions and greedy heuristic," Phys.Med.Biol. 48, 4077-90 (2003).
- B. WANG, M. GOLDSTEIN, X.G. XU, N. SAHOO, "Adjoint Monte Carlo method for prostate external gamma beam treatment planning: an application to 3D patient anatomy," Phys.Med.Biol. 50, 923-935 (2005).
- R.A. LILLIE, "BNCT filter optimisation using two-and threedimensional discrete ordinates," In: Proceedings of the 3-D Deterministic Radiation Transport Computer Programs: Features, Applications and Perspectives, vol. 1, Paris, France, 379 (1996).
- B.L. BEERS, V.W. PINE, "Functional Expansion Technique for Monte Carlo Electron Transport Calculations" IEEE Trans. Nucl. Sci. 23, No. 6 (1976).
- D.P. GRIESHEIMER, W.R. MARTIN, "Monte Carlo Based Angular Flux Distribution with Orthogonal Function Expansion," Trans. Am. Nucl. Soc. 89 (2003).
- 85. J.P. BOYD, Chebyshev and Fourier spectral method, Dover Publications, New York (2000).
- J.T. WEST III, "SABRINA, an Interactive Three-Dimensional Geometry-Modeling Program for MCNP," Los Alamos National Laboratory Report LA10688M (1986).
- K.A. VAN RIPER, "A CT and MRI scan to MCNP input conversion program," Rad.Prot.Dos. 115 No. 1–4, 513–516 (2005).
- 88. T. PINELLI, A. ZONTA, S. ALTIERI, S.BARNI, A. BRAGHIERI, P. PEDRONI, P. BRUSCHI, P. CHIARI, C. FERRARI, C. ZONTA, "TAOrMINA: From the first idea to the application to the human liver," In: W. Sauerwein, R. Moss, A. Wittig, eds Research and Development Neutron Capture Therapy. Bologna: Monduzzi Editore, 1065-72 (2002).
- 89. T. PINELLI, S. ALTIERI, F. FOSSATI, A. ZONTA, C. FERRARI, U. PRATI, L. ROVEDA, S. NGNITEJEU TATA, S. BARNI, D.M. FERGUSON, "Operative modalities and effects of BNCT on liver metastases of colon adenocarcinoma," In: M.F. Hawthorne, K. Shelly, R.J. Wiersema, eds. Frontiers in Neutron Capture Therapy. New York: Kluwer Academic, Plenum Publishers, 1427-40 (2001).
- 90. R. NANO, S. BARNI, P. CHIARI, T. PINELLI, F. FOSSATI, S. ALTIERI, C. ZONTA, U. PRATI, L. ROVEDA, A. ZONTA, "Efficacy of boron neutron capture therapy on liver metastases of colon adenocarcinoma: optical and ultrastructural study in the rat," Oncol.Rep. 11, 149-153 (2004).
- H. GRAY, Anatomy of the Human Body. Philadelphia: Lea & Febiger; 1918. Thoroughly revised and re-edited by WH. Lewis New York: Bartleby.com (2000).

- 92. J.A. FAVORITE, Memorandum: Proposed Changes to the Tallies Section of the MCNP4C Manual, Chapter 2, Secs. V.A-V.C, X-5:02-017(U), LA-UR-02-1892 (2002).
- 93. G.B. DANTZIG, Linear Programming and Extensions. Princeton, NJ Princeton University Press (1963).
- 94. Numerical Recipes in Fortran 77: The art of scientific computing. Copyright (C) 1986-1992 by Cambridge University Press.
- 95. T. COLEMAN, M. A. BRANCH, A. GRACE, Optimization Toolbox For Use with MATLAB, User's Guide Version 2, The MathWorks, Inc. (1999).

## ACKNOWLEDGEMENT

I am willing to defend the proposition that says that without any of you mentioned in the acknowledgement, there would have been no thesis and list of propositions at all! One could argue that some people are not directly involved but I would reply that these people enjoy(ed) my life which is most important when trying to achieve something like a Ph.D. While writing this acknowledgement at the last moment (of course), I am afraid that I might forget to thank someone which would destroy my latest proposition immediately. Therefore, as a solution:

Many thanks to .....! How could I ever forget you? Sorry.

As you can imagine, it is impossible for me to start thanking all the persons in an order of 'importance'. There are the people who coached and helped me scientifically and there are the people who supported me privately in the dark moments I went through during this Ph.D. adventure. To be honest, I was at one moment close to giving up. I choose the 'proven' and therefore 'safe' order of thanking you which is roughly arranged from work-related friends to private friends to closest family.

In Delft, I want to thank the Professors Hugo van Dam and Tim van der Hagen for their trust, advice and support over the years. I learned a lot in the monthly meetings in which often one of you attacked my work while the other one protected me. Fortunately, Dr.ir. Jan Leen Kloosterman was always at my side in these meetings. However, outside the meetings Jan Leen pushed me hard to keep track of an imaginary schedule. Thank you for that, I know now that it is absolutely necessary. Always interested in my work was Dr.ir. Danny Lathouwers; the man with whom I baked bitterballen and vlammetjes during our duty in the IRI-bar (thank you Ruben for teaching us)! Let me continue with thanking Dr.ir. Eduard Hoogenboom, Ine, Riny, Piet, Camiel, Dick, August and Martin. Of course, we started more or less together, and finished it: Alexander, Marek and the Húngàrian guy I forgot the name of (C15?). All students from PNR thank you for letting me feel welcome during my monthly returns to Delft.

In Petten, I want to thank our director Professor Kari Törrönen, Dr. Roberto May and Dr. Marc Becquet. Thank you Dr. Ray Moss. Of course, very important, Gerda, thank you for all your support. Mentioning Gerda, with tears in my eyes: where would I have been without TIM, TAS, Conge, WAB and Sysper2! Thank you all people of the JRC's administration and IT department for making impossible things possible. All my fellow colleagues and agents: Antoaneta, Jim, Arthur, Josselin, Lucien, Giuseppe, Günther, Andrzej, Yuan-DHL, Bart, Alain, Finn, Klaas, Anton and Francois. I really appreciate the friendliness and interest of all the people from NRG working around the reactor. Thanks und grüsse to the Essen crew in the persons of Professor Wolfgang Sauerwein, Dr. Andrea Wittig and Dr. Reinhard Hentschel. Thank you all BNCT colleagues and friends worldwide: Super Cecile, Silva, Andrea, Stead, Jingli, Stuart, Kent, Ray, Hiroaki, Saverio, Grazia, Simone, The Elisabettas, Jerome, Marjan, Gerard and Gerben. Thanks to Wilko and Corine for your Dutch BNCT support. Not to forget the people who encouraged me during my graduation work and motivated me to continue: Professor Rob Mudde, Dr. Hans van Maanen and Jaap Beekman. Also related to this is Hanneke who encouraged me to go for it. Many thanks!

Okkebas, Het Eiland (Allan, Jan and Maikel) and Thijs G thanks for the great time. Please let me repeat it once again: no Cris and Haipo, no coffee, no thesis. Harold and Zoe thanks for also helping me with the linear programming. I will never forget the man who gave me a lot of support and who will make me feel very welcome in Hungary, whenever I go there: David!

Let me now thank my parents and family. Joop and Hedy, you are always there for me and this was once again proven as you helped us after the birth of Anna and Sofie. This was a very hectic period. Let me immediately add the name of Joke Wilkes. Many many thanks! Side support was given by Bert, Jorine, Eveline and Yvonne. I appreciated the support and understanding I received from my supervisors in Delft and Petten in this time.

My dear beloved Sabine, Anna and Sofie. I am very grateful to live my life with you every day. You gave me all the support possible to finish this thing. I am really looking forward to continue with you without having a thesis on my mind. This is for you!



Finally, I include here the first school-pictures of our daughters of whom we can be proud. With these pictures they have also something to look at in my thesis!

Petten, 10<sup>th</sup> of May 2007, Sander

## **CURRICULUM VITAE**

Victor Alexander Nievaart

Call name: Sander Date of birth: June 19, 1973 Place of birth: Apeldoorn, The Netherlands

1985-1992:	High school "Veluws College" in Apeldoorn; HAVO and VWO.
1992-2000:	M.Sc. in applied physics at the Delft University of Technology, The Netherlands. Majored in the Transport Phenomena group on Laser Doppler Anemometry in multiphase flows.
2000-2004:	Ph.D. student within a cooperation of the Section Physics of Nuclear Reactors of the Faculty of Applied Sciences of the Delft University of Technology and the Institute for Energy of the Joint Research Centre of the European Commission in Petten, The Netherlands. Ph.D. topic is 'Spectral tailoring for Boron Neutron Capture Therapy'.
2004-2007:	Continuing Ph.D. and involved in treatment planning in the BNCT clinical trials in Petten, investigating BNCT treatments of extra- corporeal livers and rheumatoid arthritis.

PUBLICATIONS 77

PART II

**PAPERS I-III** 

INSTITUTE OF PHYSICS PUBLISHING Phys. Med. Biol. 49 (2004) 4277–4292

## A parameter study to determine the optimal source neutron energy in boron neutron capture therapy of brain tumours

#### V A Nievaart<sup>1,2</sup>, R L Moss<sup>2</sup>, J L Kloosterman<sup>1</sup>, T H J J van der Hagen<sup>1</sup> and H van Dam<sup>1</sup>

 Reactor Physics Department, Delft University of Technology, Mekelweg 15, 2629JB Delft, The Netherlands
 Joint Research Centre of the European Commission, Postbus 2, 1755ZG Petten, The Netherlands

E-mail: victor.nievaart@jrc.nl

Received 3 February 2004 Published 3 September 2004 Online at stacks.iop.org/PMB/49/4277 doi:10.1088/0031-9155/49/18/006

#### Abstract

The values of the parameters used in boron neutron capture therapy (BNCT) to calculate a given dose to human tissue vary with patients due to different physical, biological and/or medical circumstances. Parameters include the tissue dimensions, the <sup>10</sup>B concentration and the relative biological effectiveness (RBE) factors for the different dose components associated with BNCT. Because there is still no worldwide agreement on RBE values, more often than not, average values for these parameters are used. It turns out that the RBE-problem can be circumvented by taking into account all imaginable parameter values. Approaching this quest from another angle: the outcome will also provide the parameters (and values) which influence the optimal source neutron energy. For brain tumours it turns out that the <sup>10</sup>B concentration, the RBE factors for <sup>10</sup>B as well as fast neutrons, together with the dose limit set for healthy tissue, affect the optimal BNCT source neutron energy. By using source neutrons of a few keV together with neutrons of a few eV, it ensures that, under all imaginable circumstances, a maximum of alpha (and lithium) particles can be delivered in the tumour.

#### 1. Introduction

Boron neutron capture therapy (BNCT) is a promising treatment for various types of cancer. Basically, BNCT treatment consists of two parts: firstly the tumour is loaded with the isotope boron-10 ( $^{10}$ B) and secondly the tumour is irradiated with thermal neutrons. When a thermal

0031-9155/04/184277+16\$30.00 © 2004 IOP Publishing Ltd Printed in the UK

4278 V A Nievaart et al " Troduces an alpha neutron is captured 1 and a lith? , principle, can break the principle - cell, thus th. BNCT is distinct from conventional radiation ⁺he C2 <sup>^</sup> presence of more <sup>10</sup>B resulting in a higher loca. .ation. In u. cells can be spared. recent years, in seven. 'wide, clinical trials have been performed ly deal with the treatment o1 <sup>1</sup>ife threatening brain tumours, na mising BNCT results have lastoma multiforme. Although, quite . 'd in the treatment of other types of cancers of the body (e.c atment of liver cancer reported by Pinelli et al -1ts are for the glioblastoma trials, which seem to suffer inc <sup>10</sup>R to the right location inside the brain and i chanc .. nead without remu. studies (recent related publication T timal when as many neur .... with the boron atoms inside the ..... ame ....e damage as possible to the vital parts of the 1 time In (skin, eyes, ears, glands, etc). The tissue de .vfn In equivalent dose which, when a preset equiv-. colerance do. indicates when treatment should be hat ....ng the equivalent dose be considered that neutrons ent mainly consisting of hydroge عد nd nitrogen) introduce c ... In BNCT, the equivalent dose is usua ' to be made un ...outions, namely the <sup>10</sup>B reaction producing itre sotons. The resulting total equivalent dose follow ano the absoruwe been weighted with particle-der ....ıve biological effective... .....anage caused by a proton is different from . no particle with the same V gammas, energy. The PPT of which ...wn from conventional gamma therapy. ve to be no worldwide agreement among the BNC. RBE factors or what should be the right proc. the con. s them. Discussio. using them at all, to providing values to tw cimal places. These unceru and for a nuclear physicist who is look optimal energy of the neutrons, com. > to irradiate a patient by BN NCT literature, when dealing with the  $o_{\rm P}$ tron energies, the rebased on one set of RBE factors, according w Vanch et c' 2000, Bleuel et al 1998). As such the question ... the n. PE factors on the optimal source neutron energy variation of 5. -ab as the thickness of un as the values of the allowea was .....isive overview on energy. It takes the correlation between into conc? rsearchers solution that allowed us  $e^{1}$ · in ...ed manner. This means that there is an interest ;--.rameter configuration. 1 the next section (section 2) we describe source neutron energy ...s section also gives more deta raints of a tolerance dose in tissue hysical or absorbed dose : . J deals with the calculation set-up, sy metry and the mr ...s section the chosen parameter ranges are pr

A parameter study to deter	mine the optimal source neutron energy in BNCT of brain	n tumours	4279
and how the post refor both " C	m section 5.	<sup>1</sup> provides th	energies.

#### *:*ory

optimal source	e neutron energy in B1	、 、	rrs is determined	by calculati <sup>,</sup>
lent dose at o	every location along a	beam	roduction	at every l
'umour is	expected. In this para	ameter study, unc	Ч	at ever
n is c	alculated using			
$H_i = \mathbf{C}$	$\sim \mathbf{R}\mathbf{R}\mathbf{E}_{\mathbf{N}}$ .	$(D_{\rm N})_i + ({\rm RBE}_{^{\rm LI}} \; \cdot \;$	D 1	$_{\mathcal{A}}$ $(1)$
where CBE <sub>10B</sub> is th	6			ducts of <sup>10</sup> B,
which is 1;1	Luu for th	e characteristics of	the comp	<sup>-</sup> <sup>10</sup> B
ator	or healthy cells. T	he parameters $C_{10F}$	$_3$ and $D_{10B}$ are	
ed 10	<sup>b</sup> B dose, respectively.	$D_{\rm N}$ is the absorbe	d dec	Jan
of ther	mal neutrons with nitr	ogen and RBF.	-12	factor to trai.
into an ec	quivalent dose. The t	wo othe	Jai effectiv	veness factors
or the prot	on recoil of hydroger	- -	anmas. $D_{\rm H}$ is the	absorbed prot
se, also kr	nown as fast r	, repres	sents the induced	absorbed gam
n gamm	as, pro	e reactor core	(as in Petten repo	orted by Ver1
1-pr	. ine	$D_{h\nu}$ term. The ind	ex <i>i</i> shows that all	element
the	on the positi	tion in tissue. The	RBE <sub>N</sub> translates o	m <sup>1-</sup>
effect of us	~~ nroto	ns from the nitroge	n reaction	Jame
in all tissues. The	ðu.			JuleV gammas
from the hydrogen	reaction		heam g	gammas does not
exceed 10 Mey				nce, it is
choser	t with a loca	tion-independent	$RBE_{\nu}$ . The $i$	וויר
<u>þ-</u>	<sup>•</sup> he energy	of the incoming r	eutrons and depe	nd u.
. Conse	در <sup>-</sup> h	osen to be location	dependent. Due f	to the stru
dissues, su	uch as 10.	rtion of bloo	d vessels, the bor	on concentra
as the CBE	10B. differs from		,	
culation of	the optimal source n	eu.	ally searching	for the maxim
vduction u	nder limiting condition	ons, which	1 as	
	( / xxtolerance )			

n.  $\gamma x \left\{ \left( \frac{H^{\text{tolerance}}}{H(E)} \right)_{\text{somewhere healthy tissue}} \cdot \alpha(E)_{\text{turnour}} \right\}$ 

with the dimensions

[total number -'

Ŀ

n۲

""tron].

(*L*) represents the alpha production in the tumour fraction on the right-hand side is the source neutron source neutron each the tolerance dose ( $H^{\text{tolerance}}$ ) at a produced in source neutron source neutron source neutron source neutron source neutron source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron source neutron source neutron source dose in the healthy tissue ( $H^{\text{to}}$  number of alpha source neutron so

4280



2.1. Physica.

By studying the equival masqueraded by the multiplication ' logically or me wie quantities. beam gammas  $D_{b\gamma}$ , although slightly decreasing sidered as four major absorbe.  $D_{\rm H}$  and  $D_{\gamma}$  which arise in the head due i neutrons, are presented in 115 on of depth for the case when skin n are both 5 mm. This figure shows ~ neutron case, with 10 ppr 'I tissues. The data in figure 1 are extracted  $\sim$  with the Monte r<sup>o</sup> (Briesmeister 2000) in a cubic head model to h is disc. <sup>1</sup> lines give the ratios of the fast:epithermal and the ual 1. -----Juetined in energy groups, Sauerwein et al (1999). ... une phantom and most of them would he -ties of water with which <sup>L</sup> ..... compared. The diffusion iong. ∽ and the r , settined as the root mean square distance a neutron ° a ., to its capture as a thermal neutron, is 58.4 mm (7 **.**aı. s clear that all dose contributors, except for pr .e to thermal Jugh the probability of a neutron scatter: . decreases for higher en nergy, and likewise the absorbed ? ...g protons will be significantly h results in a proton recoil ...gh in the skin and decreases rapidly phantom.



#### 3. Сал

The calculation of the  $v_{\rm F}$  anterent parameter configurations consist post-processing calculation addermined with changing parameters such as bo. factors and parameters affecting the influence o. authermote, assing, a ratio concerning the tolerance dose i. an is varied using prese.

#### **CNP** calculation

med phantom, filled with brain tissue, was mou-'rns of 24' mm (width  $\times$  depth  $\times$  height) using MCNP4c2. figure 'w air. A disc-shaped neutron source, with 60 \* ..cd this same centrally, 150 are layers .... over three MCNP of skin and cranium. calculations using " mm. These values ... of the head of the reference are extwhich it j. ....ad possesses the greatest variety of tissue thicknow hat ...... and density of skin are very similar to those - t L Justion of the .ie unlikely skin thickness of 7 mm should h ing the cranium. All tissue composition ~ taken from ICRU 46 ( During the MCNP transport calc-, cranium and brain possess 10 p anere is an effect when <sup>10</sup>B is homogen vhich is homogeneously ? rted in the tissue . considered if this would affect this study sign?

4282			V	/ A Nievaart et al
The 7 mm phant	-		the max	ximum values
of 80	SKIII, Cran	ium and Diam,		rce neutron
	turns out that at 50 r	nm from the skin ther	e is a	rce of
	fux between	a phantom really loade	ed with the ac.	۶'
and the a	ne. N	Note that it affects onl	v the proportion	n
rmal effects a	nd the	Tortunately, for low s	ource neutron e	nerg
e deviation bec	omes larger, tus	very small f	ast component.	Toge
e presence of s	tatistical uncertainties.	to th	e conclusion the	atac
Id complicate	e the investigation unner	essati.	e concrasion and	
will be se	en later the preset dose	limit in bram.	rs an c	
in be se	the brain. Therefore, ha	alf a sphere is modely.	uii	100
b.		an a sphere is modelles		ugo
The is			91	re 22 discrete
even logarithmic	_ <b>1</b> `			o 1 MeV This
results :	, u⊥eV (	) 215 eV and () 464 r		V and
1'	att is shown that only h	vdrogen and nitrogen	deliv	* and
ur m	ajor BNCT dose contrib	utors Source neutr	uej.	0.
we negli	gible dose contributors d	lue to thresh	لار ther nu	clides
ticcuca A	a described in Blauel a		f neutron sou	reacon
ussues. A	toway' by recordin	tu, the e	algulated mono	anorgati
directional	regulte This	JI the C	actuated mono-	energen
	ifted d	nonants in the speet	This offect i	a kont
Sue IS SIL		aponents in the spectru		is kep <sup>,</sup>
10.	J. Notimed ladden	shamad as amatery shar	wa tha lagati	
4-11:		-snaped geometry snow	ws the locar	کړ. نه
taines.	samed regio	ons in which the coord		actions.
In order to inves.	ъ 0			0 mm, 40 mm,
60 mm and 80 mi	m Irom **		aned as brai	n tissue). The
MCNP tam?				individual
VOD	. or MCNP4c2 calc	culations, the standard	ENL.	hrary
	>t chlorine wh	ere the ENDF/6.8 eval	uation is used,	
Lions U				

#### Post processing

' ranium thickness 'ery output of the three MCNP calculation. n) a data file containing the four absorbed dosc. ron as a f hantom and source neutron energy, is created. 11. fun、 ree neutron energies, all absorbed doses per sor Jrain - haadhaad d .. energies are averaged u. discussed later). Since .... ..., the  $D_{b\gamma}$  as given in equation (1) h ` A Fortran code is ..... me and multiply every value ... writt arameter(s) -1 uation (1). Table 1 presents all the parameters.  $+^1$ ' vant s, as varied in this study. Jased on a bu All the choices concerning the parameter re-.dy (Hideghéty et al 2003), the current P\* used in Petten (Sauer) 999, Wittig et al 2001) combined values that can be found in th ...eter ranges are extended so that the br erature (e.g. Nigg 1999) vioreover, all RBEs can have the value 1 w' ain 'highly une

	Table 1	* **			and st	ep sizes.		
	_						-• a	
		$CBE_{10B} \ (-)$	$RBE_{N}\left( -\right)$	$RBE_{H}{}^{a}\left( -\right)$	$\text{RBE}_{\gamma}$ (-,			` •., (–)
	U	-n 1	1-5 step 1	1-6 step 1	0.5-1	1, 5 ano		-
۰m	0-20 ster			1-6 step 1	step 0.5	10		
n	0-30 step 10	U		1-6 step 1				

 $he\ RBE_H\ cannot\ exceed\ its\ previous\ tissue$  .

 Table 2. Number of parameter configurations.

_	- <u>-</u>	RBE <sub>N</sub>	RBE <sub>H</sub>	$RBE_{\gamma}$	Beam gamma energy · <sup>3</sup>		
						000	_

. resence of each parameter value in the 4 527 600 configur-

	$ ightarrow CBE_1$	0 <b>B</b>	$RBE_N$			$RBE_{\rm H}$				-6У	
	Cranium	Brain	All tissues		e.			tissues	A t	All issues	All
						10	38%		1	43%	0
	٦ry	Ever			,0	18%	27%	Every			
				3	11%	21%	18%	value	5	29%	
5%			. 5%	4	18%	21%	11%	50%			
				5	27%	18%	5%				

# the idea of ant difference whether the compound carrying a RSH is *et al* 2001).

ine sake . be post-processing calculation, the  $C_{10B}$  and the are combined to avo. The  $RBE_{H}$  is chosen to be at least adjust. e since the recoiling proton c. the neutron energy which decreases eutron travels deeper into the phane. roindeweij et al 2000). The  $\Im E_H$  value for deeper lying tissue cannot  $e_A$ . tissue value. · hove, je  $RBE_{\gamma}$ , defined to be unity for gamma energies ter setting of 0.5 as well, since it is possible that. foı. mas (see Kuni 1998). The beam gammas in th-. with 1 10 14-17 14 Juakel 2002) discrete enc. show a dose rate of 1 U, ... at the thermal neutron peak. A MCNT anse rate, there is one  $\gamma$ treme  $\lambda_{b\gamma}$ bear Jource neutrons. Taking this result 1/10 are selected. Again, to avoid duplicate value rgies are combined. Table 2 shows the resulting total m ...1 the range of the parameter values combined ....ns to speed scessing affects directly the number of rarameter value is prese e total number of configurations se present percentages (rounded rameter value. For insta- $\perp_{\gamma}$  can be either 0.5 or 1, it is immedia ach value will ? . of all the 4 527 600 configurations.

4284					V A Nievaart et al
D	Due to the rest.			~ the	e presence of the
RBE <sub>H</sub>	l q;w		s tilat no.		resent in
th	an be re	garded as a bias. Al	lthough this i	ssue w	the the
		in that the main inte	erest of this s	tudy is to in	Ves.
	parameter a.	∼ all settings.			
	he final part of the H	<sup>1</sup> culati	on is the cho	pice of the t	olerance 、
e	. This study follows the	Hti.	the sense th	at limits are	set for th
וי	rain: a point tolerance dos	e for skm .	<sup>11</sup> as	a volume to	olerance de
	The latter is actually the	same as allowing .		vints. Fo	or the rat:
	lose in point skin:poi	nt brain, as well as	the po.	• ••	ain, t <sup>1</sup>
	en: 1/5, 1/3, 1	, 3 and 5. Note that	t it is actual	1,	11
dept.	ource s	trength, which is car	ncelled out in	th-	dose د.
ratio a	approa				interest in the
absolu	ute equivalent				' ' <sup>1</sup> can also be
interp	rete	Joers on how the dos	e component	S	To To
obtr	in reasona	ble time, the 'point	dose' in this	paper .	
	.ume (78.5 mm <sup>3</sup> ).				
	J discussing the resu	lts, it should be po	oint <sup>*</sup>	JN	ly the folle
	ers were varied or not	varied in this s+		Lers varied	were the tis
	s, the RBEs, the <sup>10</sup> B co	ncentrati	, ine to	lerance dose	e and the tume
	n principle, this study ;	ال	a on the disc	rete source	neutron energ
	vare shape of *'	source i	ntensity distr	ibution and	the field si7
	'ari		-		

## 4. Results

In figures 3–7 ° for the source neutron (2000) will be rounded to integer values. "The statistical uncertainties are omitted in. "The statistical uncertainties are omitted in. "The statistical uncertainties are omitted in." "The statistical uncertainties are omitted in."

? optimal source neutron energy

	ws the optimal source neutron energy	gy for a pna.	∽f skin ∂r	
ι	function of the four tumour dept	ns and according w.		
skin:p	The colour in this figur	e is a measure of the r	1er	er
configura	tions	stans -		
Figu	re 3 is the result of a		uation of the dis	screte
values be	elonging		mour depths.	This
graph °'	July of the parame	ter configuration	<u>`</u> ne	utron
P	eV, 5 keV and 10 keV for tu ۔	mours at 20, 40, 60 -		۰in,
	J. Because of the limited colour reso	lution, whick	5 L	
rte	ved in redness, the regions between 0%	and T	th grey com	
xam	pple, at a tumour depth of 20 mm. <sup>+1-</sup>	.ui a fe	w per cent of pai	
igura	tions having an optimal sour	J of 5 eV (betwee	en the grey lines a	
10 e'	V). At the other end	tumours at 80 mm	are optimally irr?	
ി	keV source r			



After studying the c \* - the other tolerance oroups. dose ratios 2 .... identice Tatios Studies ... As sr presentative of the ratios of point skin:point br-~**t** *,* ' un, of 1/5, 1/3 and 1. In this group, the majori+ 1 all tumour positions lie between 1 keV and 10 Pattern ex. .. oe split into two gro in figure 3. The remaining four tolerance ng to whether they are defined ov volume brain. For the phant nm skin and 5 mm cranium .se ratios, representing these two grou ו in figure 4.



nce dose which is higher in point sking and (bottom graph of finger filler) and (bottom graph of filler) and (bottom graph of filler) and of the already described grounce ratios. For tumour source and source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for this is that lower source and the explanation for the

PAPER I 89

Table 4. The the			tolerance dose ratios.
_		1	_
	tolerance de	ose in point skin is <i>lower than or equal</i> dose in <i>point</i> brain or <i>volume</i> brain	<i>l</i> 1/5, 1/5
Group 2	than .	in point skin is <i>higher</i> oint brain	3/1 and 5/1
Group 3	The tolerance than the tolerand	' inher	3/1 and 5/1

` best to react with <sup>10</sup>B situated at shallow ' energie J.J.m. For tolerance 1a. ....an brain, it will most likely be that ? is averaged over a volum? ∿at une dose limit in brain. Decu. extrem ...utron energies can produce, under certain condit: nv in the tumour. discussion accompanying figures 3 and 4. \*\* . that the resu studied tolerance ratios can be divided ? , which is summarized in

' to that the optimal source neutron of Nevertus . 11 parameter variations ... ... and brain. For the other phantom calculation mo to the ten tolerance hut now dose ratios, is also for the 3 skin and cranium thicknesses, respectively • raphs of figures 3 and 5, it is of no doubt that the ant influence on the optimal source neutron e. J Cram. <sup>(1)</sup> mm from the skin, the differences are 1. tumours position. icant but for the shallow. ted tumours a transition can be seen e of the 20 mm tumour, with incimium thickness, the majority nal source neutron energies 'move' from 100 eV towards 1 ke 500~eV for the 7 mm phantom. For the  $\delta \upsilon$  . rsition tow rce neutron energies is observed. In fact, huge un his in the c. in results. The basic mechanism causing this decomponent: the thicker +1 thickness origin. Juons of the  $D_{\rm H}$  within the skin  $\sim$ , which are tissue  $D_{\rm H}$  value at the thickness independent surface of th

#### ing parameters

e clear indications presented above, it is the source of the most significant deviation of the three phantoms, end to be ratio and end to the source of the three phantoms, end to be ratio and end to the three phantoms, end to be ratio and end to the three phantoms, end to be ratio and end to the three phantoms, end to the three phantoms are the

#### 90 INCLUDED PUBLICATIONS

4288			V A Nievaart et al
Tab	e 5 D		i source neutron energies in a 5 mm
	- ussue that th	ne parameter is set for.)	ine square brackets
	Phar	ntom: 5 mm skin 5 mm crani	um
our energy		-mV	Deviating higher energy
rance dose: point sl	$\frac{1}{5 \times \text{point}}$ brain	(1-,	
חm 1 keV 2.15 ki 4.64 ki יח keV	$\begin{array}{c} 21 \text{ eV}: C_1 \\ \text{eV} \\ 464 \text{ eV}: C \\ \text{eV} \\ 464 \text{ eV}: C \\ 464 \text{ eV}: C \\ 464 \text{ eV}: C \end{array}$	$C_{10B} \cdot CBE_{10L}$ $C_{10B} \cdot CBE_{10B} [skm.]$ $C_{10B} \cdot CBE_{10B} [skin] < 20$ $C_{10B} \cdot CBE_{10B} [skin] < 20$	ev: RBE <sub>H</sub> [sk: ٦٩٤,
Toleran	ist brain (r	epresenting group 2)	
20 mm 2. 40 mm 2 <sup>15</sup>			$_{\rm OB} \cdot {\rm CBE}_{10B} [{\rm brain}] < 20$ ${\rm CBE}_{10B} [{\rm brain}] >$
60			21.5
10 keV	7		$_{-10B} \cdot CBE_{10L}$ , $_{\kappa}BE_{H} [all] < 2$
e dose: point sh 1 keV 2.15 k 4.64 k of occurrence It we	cin 5 × volume brain 0.215 eV eV eV eV eV present a rositions.	(represented to a series of the second series of the second second series of the second secon	21.5 keV: $C_{10B} \cdot CBE_{10P}$ 40; RBE <sub>H</sub> [all] < 3 46.4 keV: $C_{10r}$ 40; PF 40; PF 4
Practically, th sed separate	e parameter valu ly in the followin	g subsecut.	split into two parts, which w
the energies and source neutron en	and values causin	ng non-modal optimal so energy is given, as well	as the deviating
deviet		arou mantionad in -	ine not met, men the
رو رو پر ptimal s ieters. The nium. In the ca the influencing limit is reached cached in skin	and similar data ource neutron en following results use of group 1, the parameters. In the 1 inside the	a found from the 3 mr ergy in each tolerance are therefore $e C_{10^{P}}$ $r^{ost p}$ $r^{ost p}$	n and 7 Le thickness c Le th
PAPER I 91

A parameter study to determine the optimal source neutron energy in BNCT of brain tumours 4289 and CBE<sub>10B</sub> for brod ming parameters. In recording and alcut dose limits, most per neutron fŀ '0 mm and 40 mm. The 3, a mixture of  $C_{10B}$  and  $CBE_{10B}$  for both s. .id the RBEn - the influencing parameters. Although it is clear w. me important in each group, there seem ystematic behaviour when the rualities are interpreted (see table lieved that this is caused due to van. ' effects and actions invol ting the optimal source neutron energy. ms that the value of the RBE<sub>N</sub> parameter plays ... "st for pll phantoms, there is a notable influence at 20 n. and . 4.2.2. Par The ... in table 1 are chosen in such a way that they co parar J BNCT of brain tumours. As discussed in se ctions influence the presence of each paramet ...ght influen ce of certain optimal source neutron ene ...e intention of this su ...rgy is the 'best' source neuti conclude that the modal optimal t is interesting to see how \*' ...s contribute to the modal optimal sou pergy. To indicate ... schematically, figure 6 shows in how many nfigur an parameter value, there is a contribution to th  $_{oy}$ . The values displayed in figure 6 are for the 5 op. resenting the three groups and at four turn for three Jual optimal source Inc To explain these curves. with man which has a modal optimal source per  $^{PE_{10B}}[skin] =$ 320 con+ ... neutron energy. As another example, ... ייר at 4P 191 optimal source neutron energy of 2.15 kev, - 4 contribute to this modal neutron energy. .15 W. , for RBE<sub>H</sub>[bran his occurs in 11% of the 4527600 configura 6 also allows the reade. rides drawing conclusions on influend ters written in the last subsectio. -' effects. That is in group 1, , it seems that the modal optimal source are caused by all value  $\Sigma_{10B}[skin] > 100$ . Except in the case of the 40 ' increasing v concerning <sup>10</sup>B in the skin, the optimal source . -0 ····v: it results every time, like clipped, in the same tow con in groups 2 and 3 for some energy. 11. ... and 40 mm is Generally speak. contrary and different f positions.

### ant of the optimal source neutron energies

sults presented above show a wide spectrum ineutron energy is the question as to whether this study constrained and optimal or alphas generated by each optime ineutron energy is complex index of alphas generated by a neutron the complex index of alphas generate

4290		VA	A Nievaart et al
that most BNCT for "		around	d this energy
(Moss et			05
comparison	, representing groups	s 1 and 3, are	~ 7
im	provement is shown f	for the case of a phanto.	
. 5 mm cram	from group 2 are so	mewhere in between.	
or example, focusing	"oup 1 (top	p graph), it is clear that f	for a ı
mm, there is a parameter con		an optimal source neutro	on enei
V, which produces 600% mc	ore alpin.	V source neutrons.	As a
'e, for group 3 and a tumou	ir at 60 mm:	<sup>b</sup> eV indicate	es th
vrations in which this so	urce neutron energy	uc. more	r'
reutrons. Note that	the improvement for	paramen	.4
optni, mergy at 10	0 keV is of course 0%		
Although		ul	e maximum
improvement is are-		1-11	th), all other
tumour pr	n 200% improvemen	ts. The max	mum
imp <sup>r</sup> _ group 1, whic	h is due to the fact th	at in this, the -'	
. Decreasing the part	rameter values related	to the	
oution of lower energy	neutrons to skin. Th	utrons	s to pre
nas in the tumour than the	10 keV sour-	. suffer from the	e rather h.
ecoil dose in skin. In grou	ps 2 ~	that is set as the me	ore sensiti
these groups, for tume	_ ueep	er, it matters little wheth	er the alr
> produced by 1	ons, which	reach the dose limit in de	eper br
-Att;	Jower energies, w	which reach a dose limit i	n sr
branutrons	produce, in the end,	a corresponding num <sup>1</sup>	
As we.	wements, for single p	parameter of f	_1age
improvements were .			ntigurations
resulting in the same optime!		at a phantom	with 5 mm
skin and 5 mm	J r		<sup>7</sup> but now
ranging			
- 7 is	representative of all	other improvement out	
ins. The	ment for the 3 mm	phantom and 7 mm pha	nton.
1% and 350%, res <sub>F</sub>	wimum imp	provement for both phan	toms ra
ر0% and 500%, respective،			
all phantom cases, the impr	oveme	<sup>-1</sup> source neutron ene	ergies bel
nly in the case of the 7 mm s	kin and cranup.	re cases in wh	nich a se

re cases in which a so regy of more than 100 keV gives an improve...

# 5. Conclusions

This parameter ..... TRv varying many of the rai .... orving ones directly (e.g. KDLo, ... cometry) r⁄ ' ... atent dose, insight can be gained into the behavir ne Agy in BNCT of brain tumours. In selecting special .p to realistic maxima, there was a certain an\*: Liner such an u J- of values, gathered fro would lead to conclusive results. An " ture, was chosen. By ignoring 7 .s, mostly zero values, and obvi ng physically impossible ... number of possible configurations co , unprejudiced, the resulting optimal source <sup>1</sup> significantly 7

A parameter study to determine the optimal s	ource neutron energy in BNCT of brain tumou	rs 4291
energies of the	nt	Nevertheless, the
moda <sup>1</sup>	ies can be seen as Exam	-+ energies.
escribed abo	we, it is clear that there is a sign.	∿f the
ייש for	all parameter configurations. There	is
our of the .	a certain optimal source neutron en	ergy for
40 mm deep seated h	son with the 60 mm and 80 m	mones L
ects the optimal source neuro	Survival the oo min and oo m	skin for a
antom cases and tolerance dose ra	tion min from the	Skill, for u
Eurthermore the optimal source	e neutron tease strong	v influen/
of the tolerance dose in skin	and brain. In this	s dis
refistics in the results	could be identified accoro.	
stisties in the results,		205.
• U1.	dose in every point.	or equal to
the tolerance	· (1)	re of the brain. The
skin tur		
• ´s a higher tol	erance dose in every point of the s-	<b>`</b> int
س. The brain turns ou	t to be the treatment limiting	
<i></i>		
oup 3, which is the same as g	group 2, exercise dose	e in brain is se
ie whole volume of the brain.	Both Jrain can be the	e treatment lim
ssue.		
ma out that t	the telerance does is reached f	inst datarma'
	unt of deviation in the source neut	ron and
n allo	$\mathbf{P}$ concentration $(C_{-})$ the adjusted	
offective CDE	and the biological effective	li. Liocett
(DDE) In conv	and the biblogical ener	. ontimal course
(KDE <sub>H</sub> ). III general,	- toto	
Lower optic		a recon parameters.
10 P	the C as well as the CRE as a	f the
roton pr	The $C_{10B}$ as well as the CDE <sub>10B</sub> factor and the induced gap	mm.
eem up,	stimal source neutron energy	minac
is clear that many	a varied in this study give ve	ary different
le optimal source peutron e	from the perspective	of designing
on filter, this conclusion is over	sha	vement in gen
particles with the optimal so	at the implo	with 10 keV
The improvements are it	are neutron son	1 keV
rite improvements are in	autron energies occurs as the	U KC V
the reutrons lower	r than 1 keV it is possible to achie	ante
On average	Competing a	. t keV can
improve the alpha prov	- 11	keV case Without
averaging as in		to 700% Although
furth	t is annarent already	"Ce neutron
iul eV and 100 eV	will increase the alpha prody	- C IICUIIOII
configurations at all tume	win increase the alpha prov	ally
ular the treatment of shallows	ituated tumour	availabı.
ver source neutrons	nualed tumo" the	avallaUh.
As an overall conclusion the pe	ale that no mottor w	that the inve
ameter values are or will	source neutron energy of	a few keV
a source neutro	er of tens of eV will cover all confi	u iow Kov
t source neur	set of tens of e v will cover an colling	,

V A Nievaart et al

the best alpha production possible in BNCT. A logical next step and future study will be to ignore the non-influencing parameters and investigate new parameters such as, for example, the field size.

## References

4292

Bisceglie E, Colangelo P, Colonna N, Santorelli P and Variale V 2000 On the optimal energy of epithermal neutron beams for BNCT *Phys. Med. Biol.* **45** 49–58

Bleuel D L, Donahue R J, Ludewigt B A and Vujic J 1998 Development of a neutron-biased in-air figure-of-merit for predicting in-phantom BNCT neutron beam characteristics *The 8th Int. Symp. on NCT for Cancer (Ja Jolla, CA, 13–18 Sept. 1998)* 

Briesmeister J F (ed) 2000 MCNP—A General Monte Carlo N-Particle Transport Code, Version 4C, LA-13709-M Duderstadt J J and Hamilton L J 1976 *Nuclear Reactor Analysis* (New York: Wiley)

Hideghéty K et al 2003 Tissue uptake of BSH in patients with glioblastoma in the EORTC 11961 phase I BNCT trial J. Neurooncol. 62 145–56

ICRP 1975 Reference Man: Anatomical, Physiological and Metabolic Characteristics (Oxford: Pergamon) ICRP Publication 23

ICRU 46 1992 Photon, Electron, Proton and Neutron Interaction Data for Body Tissues (Bethesda, MD: International Commission on Radiation Units and Measurements)

Kuni H 1998 The concept of equivalent dose and dose limits—urgent improvements STOA Workshop (Brussels) Moss R L, Aizawa O, Beynon D, Brugger R, Constantine G, Harling O, Liu H B and Watkins P 1997 The requirements

and development of neutron beams for neutron capture therapy of brain cancer J. Neurooncol. **33** 27–40 Nigg D W 1999 Some recent trends and progress in the physics and biophysics of neutron capture therapy *Prog. Nucl. Energy* **35** 79–127

Pinelli T et al 2002 TAOrMINA: from the first idea to the application to the human liver Research and Development Neutron Capture Therapy ed W Sauerwein, R Moss and A Wittig (Bologna: Monduzzi Editore) pp 1065–72

Sauerwein W et al 1999 EORTC BNCT Study Group, Phase I clinical trial with BNCT for patients with glioblastoma at the European High Flux Reactor Petten (EORTC 11961) (11th Int. Congress of Radiation Research (Dublin)) Radiat. Res. 1 26 (Congress Abstracts)

Tanaka K, Kobayashi T, Sakurai Y, Nakagawa Y, Ishikawa M and Hoshi M 2002 Irradiation characteristics of BNCT using near-threshold <sup>7</sup>Li(p,n)<sup>7</sup>Be direct neutrons: application to intra-operative BNCT for malignant brain tumors *Phys. Med. Biol.* **47** 3011–32

van Vliet-Vroegindeweij C, Wheeler F, Stecher-Rasmussen F and Huiskamp R 2000 Microdosimetry model for boron neutron capture therapy: II. Theoretical estimation of the effectiveness function and surviving fractions *Radiat*. *Res.* 155 498–502

Verbakel W F A R 2002 Design and validation of a γ-ray telescope for *in vivo* dosimetry during BNCT *PhD Thesis* NRG, Petten

Wittig A et al 2001 Early phase II study on BNCT in metastatic malignant melanoma using the boron carrier BPA, EORTC BNCT study group Protocol 11011

Yanch J C, Zhou X L and Brownell G L 1991 A Monte Carlo investigation of the dosimetric properties of monoenergetic neutron beams for neutron capture therapy *Radiat. Res.* **126** 1–20

Ye S J 1999 Boron self-shielding effects on dose delivery of neutron capture therapy using epithermal beam and boronophenylalanine Med. Phys. 26 2488–93

# Application of adjoint Monte Carlo to accelerate simulations of mono-directional beams in treatment planning for Boron Neutron Capture Therapy

### V. A. Nievaart<sup>a)</sup>

Department of Applied Sciences, Delft University of Technology, Lorentzweg 1, 2628CJ Delft, The Netherlands and Joint Research Centre of the European Commission, P.O. Box 2, 1755ZG Petten, The Netherlands

# D. Légràdy

Department of Applied Sciences, Delft University of Technology, Lorentzweg 1, 2628CJ Delft, The Netherlands

### R. L. Moss

Joint Research Centre of the European Commission, P.O. Box 2, 1755ZG Petten, The Netherlands

J. L. Kloosterman Department of Applied Sciences, Delft University of Technology, Lorentzweg 1, 2628CJ Delft, The Netherlands

### T. H. J. J. van der Hagen

Department of Applied Sciences, Delft University of Technology, Lorentzweg 1, 2628CJ Delft, The Netherlands and Reactor Institute Delft, Delft University Technology, Mekelweg 15, 2629JB Delft, The Netherlands

### H. van Dam

Department of Applied Sciences, Delft University of Technology, Lorentzweg 1, 2628CJ Delft, The Netherlands

(Received 12 May 2006; revised 8 February 2007; accepted for publication 8 February 2007; published 16 March 2007)

This paper deals with the application of the adjoint transport theory in order to optimize Monte Carlo based radiotherapy treatment planning. The technique is applied to Boron Neutron Capture Therapy where most often mixed beams of neutrons and gammas are involved. In normal forward Monte Carlo simulations the particles start at a source and lose energy as they travel towards the region of interest, i.e., the designated point of detection. Conversely, with adjoint Monte Carlo simulations, the so-called adjoint particles start at the region of interest and gain energy as they travel towards the source where they are detected. In this respect, the particles travel backwards and the real source and real detector become the adjoint detector and adjoint source, respectively. At the adjoint detector, an adjoint function is obtained with which numerically the same result, e.g., dose or flux in the tumor, can be derived as with forward Monte Carlo. In many cases, the adjoint method is more efficient and by that is much quicker when, for example, the response in the tumor or organ at risk for many locations and orientations of the treatment beam around the patient is required. However, a problem occurs when the treatment beam is mono-directional as the probability of detecting adjoint Monte Carlo particles traversing the beam exit (detector plane in adjoint mode) in the negative direction of the incident beam is zero. This problem is addressed here and solved first with the use of next event estimators and second with the application of a Legendre expansion technique of the angular adjoint function. In the first approach, adjoint particles are tracked deterministically through a tube to a (adjoint) point detector far away from the geometric model. The adjoint particles will traverse the disk shaped entrance of this tube (the beam exit in the actual geometry) perpendicularly. This method is slow whenever many events are involved that are not contributing to the point detector, e.g., neutrons in a scattering medium. In the second approach, adjoint particles that traverse an adjoint shaped detector plane are used to estimate the Legendre coefficients for expansion of the angular adjoint function. This provides an estimate of the adjoint function for the direction normal to the detector plane. In a realistic head model, as described in this paper, which is surrounded by 1020 mono-directional neutron/gamma beams and from which the best ones are to be selected, the example calculates the neutron and gamma fluxes in ten tumors and ten organs at risk. For small diameter beams (5 cm), and with comparable relative errors, forward Monte Carlo is seen to be 1.5 times faster than the adjoint Monte Carlo techniques. For larger diameter neutron beams (10 and 15 cm), the Legendre technique is found to be 6 and 20 times

1321 Med. Phys. 34 (4), April 2007

 $0094\hbox{-}2405/2007/34(4)/1321/15/\$23.00$ 

© 2007 Am. Assoc. Phys. Med. 1321

### 1322 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

1322

faster, respectively. In the case of gammas alone, for the 10 and 15 cm diam beams, both adjoint Monte Carlo Legendre and point detector techniques are respectively 2 and 3 times faster than forward Monte Carlo. © 2007 American Association of Physicists in Medicine. [DOI: 10.1118/1.2712573]

Key words: adjoint Monte Carlo, mono-directional beams, treatment planning, BNCT

## I. INTRODUCTION

In c unent planning, the treatment rarameters such as the ∽ used, their acteristics. Inc. 1 and orientation, together w. -nd gans at risk (OAR). The next step 15 to lan by adjusting the weights and/or shapes of ... beams such that the target volume receives the re-'se without exceeding the tolerance doses of "ha first step, it would already be an the improveme... tion direction energy, and shape) were serveciple, possible description of the adjoint theory . ided in this paper and the reader is re-ir literature.<sup>1-3</sup> Furthermore, this paper deals nerapy treatment planning in the field of Boror 'apture Therapy (BNCT). A brief dev ven in Sec. II A. At this str T patients are .... radia-LL. at treatment plans arlo (MC) based programs. can only In normal MC tran., source, lose energy while travenue. tered when passing a d as forward ..... u particles travel from th/ 'octor. This way of trav-"-int MC. Soversed and .nt particles start at the 10. ··---1 s, gaining energy before they are  $1_{c_{c}}$ ve forward source. Actually, in adjoint MC tnc . or becomes the adjoint source and the forward the adjoint detector. (In the remaining text, "ource," "detector," or "particle" is whe. in the real world is written, the meant. With the additive tor or particle is suggested but of While there : . ... order to improve the readability. tim Joint" is always written whenever simulated aticles travel from the adjoint source to the adjoint r.) With the adjoint method the user obtains at t detector the adjoint function which r nition given by Bell and Glas re" of a forward .....ibuting to . the adjoint function .ec. II B. For treatment planning, Whi the adjoint 1u... 41- -tion about the position, angre, ...

Medical Physics, Vol. 34, No. 4, April 2007

and OAR. Instead of performing a 510m .... find the optimal irradiation conditions, the a., `¬ a few calculations. The adjoint approach 18 m the number of beam exits (adjoint detector. than the number of tumors and OAR (adjoin <sup>-1</sup> advantage depends on the number and <sup>D</sup> and on the number and si ели. As stateu .. ... and Difilippo<sup>6-8</sup> can be rev . ine applica-.....g. Although they and ' it is applicable, the adjoint MC metnou 18 sun "inical community. Most of the .....8 h. Carl Jae system MCNP avoratory.<sup>11</sup> Furthermore, in the unes, the irradiated bodies are all surrou spherical or cylindrical surface, which is segmeadjoint detector areas. Such a set-up implies functions are determined at fixed por NP calculation. The adjoint nave the a lattice of smaller -senting an array of pencil beams. This "the lattice is fine strucwould a tured in size, energy, and superposition of the pencil beams,  $\kappa$ am opening, make sense. Even with present a... computer power, this fine structure approach . the due to the time required to achieve dece "hor large lattices need to be used marse grid of optime Div r Jers: irradiate trus. shown that this approac sep, is already a great im ... plan. adjoint detector segments of the contribution for a certain beam. ' heam shaped adjoint 11110 ...... me its assodetector segments and ciated problto be ....joint particles a evue phase space coordinates a ....nent beam has for instance an angular 2 degrees, only adjoint particles hitting the plane within 2° to the normal are valuable æ adjoint particles travel in the opport . particles in the treatment be-. take a long sotained for a specific space coordinates. Such an approach is applied by Wang et al.5 wnof the treatment any, the calculation will take the bear ir " that the possibility of ...e hitun, rlane (=beam perpendicularly is v. +hese mono-directional treatment beau ods are developed and applied to enable .... ent planning. method is based on next event estimators and the 's based on the expansion of the adjoint funcmomials in the MC environment. The tιc second . - the functional expansion technique (FE1) ... Legendre expansion in MC . particles can contrib-... ..... contribution beam to a tumor or OAR , unt function of a certain angle can þ nore quickly everywhere around the ge-.ent

### LS AND METHODS

v starts by giving more T. the adjoint ' ... with the ...ent estimator us II D-II G, the FET approach Sactions III and with Legendre pc. IV, the two developed memory amples. The first example shows the varjoint MC methods and " of the Legen-.... example, a human head r' no-directional neutron The adjoint ar ...... preferred oven ... \*•rms speed. Finally, the optimum a neutron beam are presented which m. in introductory step for further investigations . on of treatment planning.

### A. , ~f Boron Neutron Capture Therap,

Boron Neutron ..., thermal neutrons are delivered at since taining much more boror healthy cells ^ 10B, h - produced that, in principle, can 1. .. Consequently, healthy cells containing . lower doses and will be spared. This nonis a result of the thermal neutrons creating secarticles due to hydrogen and nitrogen reaction . This could give an unwanted dose in its the irradiation time. In BNT 'ed with epitherm? ...e neualize ...n intermediate ine aim is to have a lot of tiss. -f the cancer. In reactor thermal 115 based BNCT, as m.

Medical Physics, Vol. 34, No. 4, April 2007

and a set of this type of the transformed and transformed

# m forward flux and adjo

• in prefer In gener... ward detector is sn... \_\_oint particles that start at a ser-.ve a larger , source, from which "the adjoint function) can be particles start at vumea. ... the tumor or at the OAR and m where outside the body in orde and orientation . arge a. ... be modeled. The tu but me called the forward neutron or  $\_, \mathcal{L}$ ) and the adjoint function  $\phi^+(\underline{r}, \underline{\Omega}, E)$ , Jy. Both quantities are a function of the position direction  $\Omega$  and the energy E of the (adjoint) r interest. From Williams and Engle,<sup>15</sup>  $\phi$  may be quantity related to the "particle distrib" be interpreted as the "effect d' Letector .or neutrons and poltzmann equation. In the perator L. This Call DE WILLI operator includes the characteria -1 scattering terms operating on the flux  $\phi$ .  $C_{\sim}$ transport equation can be defined with a mau. joint transport operator  $L^+$  operating on  $\phi^+$ . Betu the important relation between the forward fly tion  $\phi^+$ , the forward and adjoint transp ...ith accompanying boundary c Using . "fter Bell and Gl? relevant forwas resed defined by  $L\phi(r,\Omega,E) = 0$ (1)

 $r \in S, \Omega \cdot \hat{n} < 0, \tag{2}$ 

where  $\varphi$  is a source and S, of which  $\hat{n}$  is the unit outer source and S, of boundary surface of volume In this paper is a particles crossing Srelevant adjoint problem for  $\phi^+$  s  $\iota \phi^+(\underline{r}, \Omega, E) = -\sum_d(\underline{r}, E), \quad \underline{r} \in V,$ 

$$\phi^{+}(\underline{r},\underline{\Omega},E)=0,\quad \underline{r}\in S,\underline{\Omega}\cdot\hat{n}>0,$$

where  $L^+$  is the adjoint transport  $\dots$  detector response function  $\downarrow$  given by a  $\square$  in the adjoint calcula-

### 1324 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

tion the source of the adjoint illustrate that the

illustrate that the "boundar boundar at solution" involves a specific cit  $\phi$  and  $\phi^+$  is given by

 $\angle_d(\underline{r}, E)\phi(\underline{r}, \underline{\Omega}, E)dEa$ 

wed following the literature on the adjoint - Introduction. Interpreted for treattec11. integral of the adjoint ment plan. function multipliea ... source function over the surf beam exit is locat .....g mward volume V, ar ... me left-hand side of Eq. (5), the ' ...se. This total detector response is the . treatment beam to the flux or dose in, for .or. This is when the adjoint source is located with a source spectrum  $\Sigma_d$  according to the flux e in this paper the focus is on inciri tely or nearly mono-direct

 $\phi$  (6) where  $\delta$  is the \_\_\_\_\_\_\_ the beam. MCNP tracks adjoint neutrons, as well but the adjoint MC implementation of the multigroup created and the method of the me

alizing is acon and the scription of the found in Lewis and the scription of the found in Lewis and the scription of the scri

possible to calculate adjoint functions with sibility is applied in Example I (Sec. foı III). The rov dependent adjoint function at the forward dein forward mus. tector all those scores on ... from the same energy group at the sour sums for each source joint functi-.ms method is similar to th ...ribution function which is defined , contribution of source particles to the course, it is still a forward MC calculation and, .ne statistical uncertainties of an estimate can only when a forward particle reaches the detector

# rent estimator approcession and solution approcession and solution approcession and solution approcession and solution approximately approxima

Medical Physics, Vol. 34, No. 4, April 2007



FIG. I. Adjoint Point Detector a away from the nbaster automotions of adjoint anoutions of adjoint are entrance. This entrance is

detecting all adjoint part ...ulation tions. However good -J adjoint particles wi joint detector. A solution for th Je to force the flight direction along the p just before it crosses the adjoint detector plane has a next-event estimator that closely meets th' ment. The technique is called DXTRAN, w<sup>1</sup> terminology stands for Deterministic TP it involves deterministically + . events ....gnborhood of a contributions to the detector ....er-defined sphere is taken Stain many par-IO DE IIIIII. ticles in a point of interest una. impossible to sample. MCNP tally type 5 is detector. DXTRAN can be used with both 10 joint particles. In this work, the contributions u The second secon - very far away from the phan are parallel. Usir that an . '··stori/ outside of which fectively creates an au, ...e beam exit. This technique ., adjoint ....strated in Fig. 1. ..... in this work, a distance of " the phantoms and the adjoint '> effect disappear point uciecce. and simulate only paraller nce is re-' the lated to the size ~ diameter .ne adjoim point detector in the J. of incidence. This means that an point detector. Due to its forced nature, the re tainty in the APDT decreases rapidly as a number of adjoint particle histories/ever tribution can be made. For the  $\neg$ ..e examples described in +1action in cal-.. point detectors in one

PAPER II 99

single MC run when com menal Q. / adjoint point -' firea. to br ing a so-called heating time, but mation time. for angular interpolation. nsion Technique basic idea behind the functional expansion technic to describe a detector response by a series expanr samples for estimating the coefficients. Since <sup>12</sup> the FET technique has been applied to its convergence propervarious ties analyzed. pressed by  $f(\cdot)$ (7)FIG. 2. Adjoint detector of the crossing set of coordinates in the phase space,  $\psi_n$  are d basis functions, and  $d_n$  are coefficients given by ..., by applying Eq. (10). As is explained product of Liail in the next section,  $\underline{\Omega}_0$  is decomposed intr ). pendent angular variables  $(\alpha \text{ and } \beta)$  and, there dimensional base function set  $\psi_{l,m}(x,y)$  h<sup>2</sup> ;mulat; ... sampling an For the base functions the product of unction (pdf) Ø. A im<sub>k</sub>. mial  $(P_l)$  expansions is used.  $P_m(\beta),$ 

Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

1325

who . As the Legendre  $L^{-}(-1,1)$ , the angular coordi-1- it is more approchure per, priate to introduce the term. Technique (LEXT) instead of using the general

### E. Parameterization of the LEXT for mono hams

ticles traveling parallel to the disk mono-directional beam adjoint uc. rether with direction vector 22 ...sk. According to the last second normal is defined . mever "normal" is adjoint detector disk is menpointing away from the phantom. mot return after crossing the adjoint detect importance is only non-zero ile. The Legend-..er pon. . the hemisphere of the angular co ... ave the disk normal falling in the middle both  $\cos(\alpha)$  and  $\beta$ . Note that this requirement that can be accomplished in several ways chosen that the X- and Y-axis of the are defined within the disk plane r , the normal.  $\beta$  is chosen as ∠-axis and the ... in the ZY plane. Since , Pass the disk from one side,  $\beta$  is in

The phase space (x) and the sample 15 . statistical weight on . can usually be described as an mag chosen detector function <sup>1</sup> mation of the

mber of  $(x_i, w_i)$  samples. If the of the implicitly sampled pdf at a coorum. unction is selected as  $h = \delta(x - x_0)$ . Normally, no. 'es fall exactly at  $x_0$ , unless forced by next-event the number of non-zero scores (and, conseroe rate) is extremely low. This was qu. already duction. After applying the general formula 10. an estimate for the  $d_n$  expansion. some low number,  $M^{n+1}$ to estimate '

$$\sum_{j=0}^{J} \int \varphi(x_0) \psi_j(x) dx \psi_j(x_0)$$

$$\frac{1}{2} \sum_{j=0}^{N} \int \varphi(x) \psi_j(x) dx \psi_j(x_0)$$
(10)

In the . \* cimulation, the estimator of the importance surface, for a certain solid angio -\_\_\_

Medical Physics, Vol. 34, No. 4, April 2007

# 1325

+01

...0-

Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams



the normal direction  $\alpha$ regative for new  $\alpha$  the normal direction.  $\alpha$  the normal direction.  $\alpha$  the normal direction.  $\alpha$  the normal intervals can be  $\alpha$  at  $\pi/2$  in  $[0, \pi]$ . Both intervals can be  $\alpha$  and  $\beta$  linear<sup>1</sup>.

 $\phi^+$ 

1326

$$\frac{\iota}{N} \sum_{i=1}^{N} w_i 4 \sum_{j=0}^{M} \sum_{k=0}^{K} \frac{2j+1}{2} \frac{2_{K-1}}{2}$$
  
  $\times P_j(\cos \alpha_i) P_k(\beta_i) P_j(\cos \alpha_0) P_k(\beta_0).$ (11)

stion provides a factor that belongs to the direcı. which the initial weight of the adjoint MC ha corrected. After this weight correction, the deviation estimated in the regular .... outcome looks like a normal This interpol ....y ii uie impor-..... of the solid angle, and is tance f ue middle of (-1, 1). A more sophistirequires more coefficients for an accurate deacreasing the polynomial order means higher fluca the correction of the weight. Therefore, while 1 truncation error decreases, the estimate in the statistical sense.25 the number of I ... otherred + with zero values , or the Legendre interval anu,  $(-1,1), \iota_{2}$ chosen dependent on the

```
particle type. This ...
```

Medical Physics, Vol. 34, No. 4, April 2007

# $\ldots$ improvement of the LEXT by reducing the ${\bf r}$ of coefficients

The number of Legendre coefficient expanding the angular adjoint ... for  $\alpha$ ....ng the angular and to some extent rapidly conction discontinuities may ' > angle, adjoint anse at me cos particles from the tissue canno. · disk. Likewise the exclusion of rapid gradien. hance the convergence at the cost of shrink. detector (disk) size, meaning less hits to util. T. After rescaling to the Legendre interval, th shosen to be at 0 where good pointwise the Legendre expansion is exalreauy U, The angula. '···mar different for gammas and ..cS that are involved in this we' us scatter ...w shaped angular whilst broad and flat for neu-of  $\alpha$  and  $\beta$  intervals are per-Two examples IOIIIIcu ume... of typical angular adjoint · 1 gammas and the way that ? ייחcated are . ine proces  $\alpha$ . For  $\beta$ , it is the san. ... the intervals are defined by the 2 ...gular adjoint function closest to the norm For example, in Fig. 3(a), the left zero value is  $=\pi/2$  (normal direction of the adjoint detect the symmetrical interval indicated by th interval is rescaled to (-1,1) in illustrated in Fig. 3(b). .an as ... are defined ....joint function closest



Medical Physics, Vol. 34, No. -, ...,

successive angular bin dif-

ing the intervals for  $\alpha$  and  $\beta$ 

ormation

Ch

gular au,

### 1328 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

the techniques and assumptions presented here
 LEXT work fine for the examples discussed
 is worthwhile studying this aspect in more us

G. Adjoint particle the LEXT

...e LEXT to determine the contribution cional beam to a specified region in a phanform an adjoint MC calculation without defining a detector. The adjoint source particles start in a cell which can be a tumor or OAP by traveling through the e the phantor ule sur-....ss a sphere surround. .y, the adjoint particles are recorded whe. a radius of 50 cm. The computer output Inc weight, and energy of the adjoint ~ called a PTRAC file Particle TP user to gather each ind; individual particle events the estimate and ..... uncertainty of a - In or-, the size of the PTRAC file . e filters. In this work only adjoint parus oned sphere around the phantom pass the filter .

1. This sphere will be further called the PTRAC MC calculation is finished, custom made Fortran preprocessing. To obtain the adjoint narresults, the Puss ticles, has to be read twice. . obtain for all adjoint detect of the adjoint .... uncenon which ... or zero values as explained in Lond reading of the PTRAC file is per-.pand the angular adjoint function with Legendre als and to provide an estimate for the normal direcof all adjoint detectors for each energy group overview of all steps involved in the pr XT is given in Fig. 4. Apart fr sing needs a list " ....on of all nrs th , uone with MCNP version 4C2 iao ressing codes written in Fortran, are pertoning

a 3 GHz processor and 512 ivic

Medical Physics, Vol. 34, No. 4, April 2007



which leave a 10 cm diam c cells in a light water gath, with APD<sup>T</sup> c identical result.

...nand shell. This information is given i interpret the absolute times given in the results the calculation times given include the loadir geometries (heating time).

### ATED الاست. III. EXAMPLE 1: سرية IV OF THE APDT تالوند. F OF THE LEXT

APDT and LEXT generations of the show that the APDT and LEXT generations of the adjoint functions calculated with the APD. The way to calculate adjoint functions with the APD. The way to calculate adjoint functions with the adjoint functions with the last paragraph of Sec. II B. Further second with the last paragraph of Sec. II B. Further second with the adjoint functions with the physical signific for the control of the transformer second with the adjoint functions with the adjoint functions with the last paragraph of Sec. II B. Further second with the adjoint functions with the a

· contional gammas coming from a 10 cm diam oss section of the geometry set 'a water sphere 1 <sup>i</sup>cal ... and cu J1 2.0, 1.0, and 0.5 cl. are positioned at different dis the inset in Fig. 5. The goal is to obtain the er adjoint functions in these three cells. Ir set-up the three cells are forward det-, ard source is at the right side. T' forward antom as indisource particles is +-.. arrows in the opposite adjoint particles of the adjoint calcu-

## 1329 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

lation for the APDT. The through the 10 . For both the APDT and LEXT,  $10^{4}$   $\sim$ int sources in which the "--ctions. For the ле tracк. \_ methods, the gamm. cons and group structure from the ME1. The forward and adjoint sources have " vectrum over 12 energy groups between the linn. V up to 20.0 MeV. The detector response functions and adjoint methods are similar and also the 12 energy groups. In this example, the g. A at the central energies of the 12 groups. In . volume tallies of the F4 1. APDT is of type " , ou, as the adjoin the PTRAC file when pa ...ere.

# s: Comparison between the APDT and IC

n in Fig. 5 show six ... are .. Iorward MC herical ... belong to the direcanu tion of u. -+ detector disk. It can be clearly seen that unc To obtain similarly averaged sum 95% confidence interval .... 0.3%, and 2% tt is clear that adjoint MC agains\* r advantage for these , with remain 1-tectors com-.ne size of the forward joint functions in Fig. 5 indica. mmas of a certain energy group leaving ... perpendicularly that contribute to the cells I, 11, example, see the function value indicated with 5. At this point, per source gamma leaving haped source, around 0.0012 source the • 9 and 20 MeV (the 12th gammas energy group) con. every 833 source gammas, there is a an energy between 9 and Concern: - rosmoning of the tbr ....at adjoint function I is the highest med closest to the source. Compared with .11s, many source gammas of all energy groups ribute easily to cell I because they are not affected water thickness (depth). Curve I is higher f a source energies, as these gammas 'lity to scatter in the first few ected towards cell , 1, located .y the higher source 'ntan/ st of the lower energy gammas a. in the water. The smallest cell III is 1u.

Medical Physics, Vol. 34, No. 4, April 2007



numbers of Legendre coefficients "converges towards the APDT, which : "s or more for  $\alpha$  and  $\beta$ 

... ourect gamma beam. Only a very small ... source gammas having the highest energies such that they can contribute to this cell.

### C. Results: Convergence of \*\*

nb	APDT for cell II. For this loulation for the LEXT is
Staticu with	- starting in cell II.
This simulation and writing	7 file takes
131 min of CPU time. In Fig. 6,	, the _ the
adjoint function obtained with th	ie APDT v.
as the reference. To obtain the	e adjoint func
* EXT, the correct number of Leg	gendre coefficie.
In this example, the	M and K for $\alpha \in$
m as 1, 3, 5, 7,	, 9, and 19 coe'
these arents,	the adjoint
shown in Fig.	' orep/'
and one for $\beta$ (written	_cXT
result does not agree with	_ rises for 3
~~~	$30.9 \times 9$ , the adjoint
LEX	T has for this configuration
Se Se	ets of curves for the other
positions	it is observed that
all adjoint functions have	mce with the
corresponding	<sup>4</sup> re co-
efficie-	ner expc he
eter of the	adjoint de.
using $7 \times 7$ coeffic	cients contin
- same adjoint functions as the	e APDT and
When even more coefficients	are used, th
oscillate (see, for example, 19×	< 19 in F <sup>;</sup> ي
relative errors increase. The relat	tive
interval) of the above results	ine more
coefficients are used	error becomes.
	or coefficients has to be





hat the result converges with the minimum -'s approach is further discussed in ' parison with the LEXT 10<sup>5</sup> adjoint uat the ι avein LEXT with 7 ×7 coeths aculation for one adjoint that the detector disk wun LEXT cannot compete regarding LEXT will start to profit when me involved because the the informat: . or now it is concluded that +1 r identical results as f - certain geat the plying a tuned nu.

# LE 2: PHANTOM HEAD IRRADIA

illustrates how the APDT and LEXT perwith forward MC in a real radiotherapy fo. BNC'1 . tion times of the three methods tumors and OAR to obtain the w. are compared when itia. formed with neutrons and gammas three methods is discutions, the cr' and a preliminary resp1 ue optimal neutron beams around After all, the goal of these adjoint methods the number of beam scenarios that can be cala certain amount of time.

### Carlo set-up

ample, the bound	્ર <sub>પ</sub> . (5)],
atm~	вNCT trials at
the	Netherlands. The beam
can be ice	angular divergence
$<2^{\circ}$ ) and consists	

Medical Physics, Vol. 34, No. 4, April 2007

in the source spectrum of the C calculations, the 너 have an detector response function. energy spectrum in accordance w1. "OSSsections of <sup>10</sup>B, H, and N. These cross-sec mon that at thermal energies, the probability is orders of magnitude higher than at high neutron RNCT, determining the total detector r rmal neutrons with energies < of the main (therr ٧١٠. lated) dosc - the response function  $\angle_a$ ..n up to 0.5 eV. The optimizati ... plan from getting a high therand a low thermal neutron пил ш ше с.

For the forward and adjoun ions in this example, a new 17' structure based ne libra. a temsen is taken into account. In th , me adjoint detector response function is the (forward) boundary source function  $\phi^b$  which energy spectrum of the treatment beam in Petter groups representation. Vice versa, the adjoint is similar to the (forward) detector resthis example, for the three me ...gs are ... 1. This means ue to the flux of gammas in with chergy ... 20.0 MeV In this section, "total detector resp sometimes shortened by mentioning only "nu

### **B.** Geometry set-up

<sup>^5</sup> CT images of a patient's head, a v made using the program Scan? The ne. 1 cm, a height of 2' the distance troa. - the 1 measures 22.2 cm. Acco. .el in the CT image, a threshold naterial ... inodeled as soft ..... or skin, muscle, and fat. The ron voxels which contain only me. and cranium. OF IIIIAtures These tissue composition ¹∽n from ICRU46.30 In Fig of the hear ussue ma.

Lat tumor lesions, which is not un , with melanoma metastases in the brain, (dark colored), together with the ten OAR (ligb defined in the BNCT protocol<sup>33</sup> used in Pe<sup>\*</sup> eters of the tumors are all 1 cm. Simi<sup>1</sup>-OAR are modeled as spheres and , inner ears, pituitary gland une thalamic Later terms of the OAR

" actions, at each of the 60 center

# 1331 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams



FIG. 8. Head phantom with all tissues light and the ten tumors dec."

...5 cm of which the last merely indicates a 20 tumors/OAR are the adjoint sources in the schniques and forward detectors in the formulas.

for the LEXT, there erection is a sphe. Inter points are 20 cm aw point and point and point and point and point and go an angles, as shown in the first and point and

added. As shown in Fig. 9(c), eight dis. 4 ing angles of 10° with the initial disk normal, numbered 2-9. Another eight normals make a with the initial normal [numbered 10-17 in F. adjoint detectors are generated and n tine for each of the 60 point <sup>1</sup> orientations for *e* nc.. positions to 2 actors 2' function is determine ı020 ed. In the سی s are the forward میں "Petten" neutron/gamma ' In the APDT these 1020 an Fig. 1. In uisto aci ince . order to observe the influence treatment beam on the three beam exit diam un. In M "ward detectors are h neutro ... F4 type and the point detectu . .ype F5.

# C. Strategy for comparing the calculation tir three methods



... ou center points of the ....yount detector disks around the head. (b) These points are described with ' azimuthal angles. (c) Orien-''aint detector disks 's similarly "anter " anter 7

=1020 beams sn. have diameters of 5,

Medical Physics, Vol. 34, No. 4, April 2007

### 1332 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

# 

valative errors in the total detector responses due to the flux of (<0.5 eV) in all tumors and OAR. The errors for the three  $^{-4}$  to a maximum of 5% and averaged over 255 neutro..  $^{-1}$ culation times of the methods can be adapted and com<sub>P</sub>.

at the relative error in " root of the -Jan be calc. ... runs should ...or. The pre-processing be in orus bar of histories time of the LEA. to be processed. It is chosen relative error in the tumors and OAP to t interval). The comparion the adapt LEAT to obtain the A adapted calculation results tir л ms are multi-

## Calculation times of the three me.

ase of the neutrons, for the LEXT, a 20 Gb vritten for the tumors and OAR. For the gam-'re of 130 Gb. More gamma samples are h and "narrow" angular adjoint neede Since gammas are functions w. tracked faster, it take. 130 Gb file in comparison with the new time to obtain the optimum (in this examr'  $\dots$   $i \land i$  for the gamm ure total time for reasons disves an example of the averaged relative errors leutron beams with a diameter of 15 cm, after the calculation times of the three methods in c same maximum relative error. 'ative errors of the forward -<sup>R</sup> 1, 2, 6, and <sup>7</sup> ...y large glar .act that forward large detectors. Overall, pa. Fig. 10 sn. - of the three methods are

in the same range and

Medical Physics, Vol. 34, No. 4, April 2007

for the three methods for neutrons. All nmand shell on a personal ` h of memory. computer with a Penni Time 15 normalizeu Diameter of adjoint detector/ beam exit (cm) Forward MC AP 1.00(=140 days) 27. 1.00(=113 days) 1 1 00(=88 days) -1-4 . Jameters and gam-- similar to Fig. 10. the three meth-10010 ods, for 1020 beams, to obtain ses of the thermal neutron flur relative error c' ...e inter . Ine calculation times the ter Jers. Regarding Table I, due to neutron events, which are all tracked but sie to the point detector, the APDT takes 27 to longer when compared with the forward calculat calculations would take years. For 1020 1 tumors/OAR, the calculation of the conwith a diameter of 5 cm is bo ...d cal-... 15 cm can be ---1--4 \_\_\_\_ more than 6 and 20 times the advantage of the LEXT Decomes more meter becomes larger. In spite of the results ocmle. it is demonstrated here that the LEXT bec. APDT whenever more beam positions are ca. In Table II, the gamma results for the head p. rented. As there are far less gamma events in a T improves greatly from that compa ramma beams with diameters 15 сп., T are roughly 2 r faster, respective Note will become advantageo. .an 1020 beams are required Tucessing , w get similar stame nas to be obtained which is

TABLE II. The total calculations are	com.	512	°≤. All 1
	Time to calc normalized	culate 10 1 to forw	
Diameter of adjoint detector/ beam exit [cm]	Forward MC		
5	1.00(=53 '	-	.1
10	1.0		0.55
15		<b>U.33</b>	0.30

PAPER II 107



FIG. 11. An example outcome for treatment planning with a containing and the head, the maximum respinal ratio of thermal neutrons in the better is the ratio.

$T^{i}$	atu.	° ⊆ cm diam-
e'	quickly obtained	10

### timum neutron beams for irradiation

the result of what are termed the best "Per s (diameter 15 cm) out of 17 orientaare shown in Fig. 11. These tions at c in total 1020 Inc 'ated with the LEXT, which is 20 times fasu. gray bubble in Fig. 11, the larger 18 und average thermal neutron firm thermal neutron P ..., iaige gray bubbles r ... use tumor and a low dose in the <sup>2</sup> arors of the ratios presented in Fig. 11 5% confidence interval. The numbers writ-.ch gray bubble correspond to the optimum the result is checked by the 255 forward calor the differences are according to the

 ting at some product
 ... reg. 1, the

 h.
 rest '

 average
 ... ound 35 times higher

 in the tumore
 ... reg. 11, the

 with "6" in the faciar ...
 ... reg. 11

Medical Physics, Vol. 34, No. 4, April 2007

the average thermal neutron flux in the tumors . times less than in the OAR. Nevertheless, this is the rm out of 17 orientations from this position. To av which the average flux in the tumors is sma ' zero, which would give anyway antation is chosen such th ratio, ... or more of the ... mal neutror tumors is always provide For the optimization presented est beams it is chosen to ...mal neu-.... ine thermal neutron ....R. It needs to be investigated (optimum) treatment plan when compared to ... <sup>:</sup>bution of every beam to each tume ing an optimization *и* и. subject ' uneless, optimiz. , with the LEXT and APDT at

## SCUSSION

### 1334 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

MC based radiotherapy treater with forward MC be calcer' are treatment plans for monod' creatistic scenarios, the an APD a creatistic scenarios, the as For the LEXT, so... creations which need some further and, the next two sections.

### **efficient training**

and a solution times of the LEXT, the times to fus. mdre coefficients for which the -+ this number of results conveis Legendre coefficients ucrand type of particle it adjoint detect/ ... w be expected that when ir ....toms, such as head phantoms, .ons, this "training for coefficients" up e only once. Nevertheless, although this necked for the geometries presented in Ex-2, in which the adjoint sources are at var rther investigation is necessary. In ul.25 describe a first app that autom .ue optimun , pe tested and implement oroperly for the radiotherapy application.

### **B.** Improvements

A way to im .sugate other function exr harmonics or Fourier or ngular adjoint y definition. ... in the dioint functions between Inthe selection of a function expansio. he particle type, will improve the outcome. Su on, in conjunction with investigating how to mples for the estimates, as described in Sec.  $\mathbf{n}$  . ber research. As well as improvements, work ca. >\* the LEXT such that it can be used for more ... neutron beam in Petten has a 2° angular une differences when simulated by a mobeam. Actually, the difference for the distr and the irradiated bod-... not been applied, as it was cho-... the methods by using Example 2, which at simplified. Note that in this respect, in Ex-.ie distance of the detector disks is chosen as fixed ect to the center of the phantom instead of " he skin.

### NOIS

tion rate of the flux, dose or reacdue to a known source, can be obtained when

Medical Physics, Vol. 34, No. 4, April 2007

1 of the adjoint function and the mace variables provides this total (forwara, 'n general, whether an adjoint MC calculation er a forward MC calculation, in terms of spec number of sources, detectors, and their respec ~xample, the results for many large forward sour <sup>11</sup> forward detectors (=adjoint sources), adjoint MC. Apart from the e-" the number of sor 0115 and their Su. nges. ; ward to know when. \_\_oint MC or vice versa. It will ...rmine this us work can, thereanded as an example for rahead. This exunununup, ... ample shows that 20 adjoint s. m still be regarded as a "few (beams) are sir Un ' .autotherapy treatmen. ...ons are impossible when the ta ...ono-directional. This is in despite of the - circumstances are in favor of the adjoint as d above. The LEXT and APDT are two techniques an estimate for the value of the angular adir the direction of the adjoint detector's pr . geomof the LEXT is that for diffnas to be deter-... converged. The APDT is disevents which are not contributing ... ms in light water. From the BNCT example of a human. 20 tumors/OAR and 1020 different position. directional neutron or gamma beam, it seems ms (exit diameters of 5 cm) are calculated mos forward MC calculations. This result is routrons. For the larger beam and 1. the adjoint MC based on the LL s fast than the forward MC. .1S. both the LEXT and APDT unan the ....eless, the APDT whenever fewer beam positions ms directed to the face in the "v i.e., are unacneau phano... ceptable. To conclude. r inns always <sup>1</sup> ior MC be suffered when mono-di. ...e considered. From the work pre rublem has now been solved and the opt treatment plans and/or studying its mechanis

possible.

# ACKNOWLEDGMENT

This work has been are Joint Re-

#### 1335 Nievaart et al.: Application of adjoint Monte Carlo for mono-directional beams

a)Electronic mail: victor.nievaart@jrc.nl

<sup>1</sup>G. B. Arfken and H. J. Weber, Mathematical Methods For Physicists (Academic, San Diego, 2005).

<sup>2</sup>B. Davison and J. B. Sykes, Neutron Transport Theory (Clarendon, Oxford, 1957)

<sup>3</sup>G. I. Marchuk, Adjoint Equations and Analysis of Complex Systems (Klu-<sup>4</sup>G. I. Bell and S. Glasstone, *Nuclear Reactor Theory* (Van Nostrand Re-

inhold, New York, NY, 1970), pp. 252-313.

<sup>5</sup>B. Wang, M. Goldstein, X. G. Xu, and N. Sahoo, "Adjoint Monte Carlo method for prostate external gamma beam treatment planning: An appli-cation to 3D patient anatomy," Phys. Med. Biol. **50**, 923–935 (2005).

F. C. Diflippo, "Forward and adjoint methods for radiotherapy planning," Med. Phys. 25, 1702–1710 (1998). <sup>7</sup>F. C. Difilippo, M. Goldstein, B. A. Worley, and J. C. Ryman, "Adjoint

Monte Carlo methods for radiotherapy treatment planning," Trans. Am. Nucl. Soc. 74, 14–16 (1996).

<sup>8</sup>M. Goldstein, "The adjoint Monte Carlo-a viable option for efficient radiotherapy treatment planning," *Proceedings of the 19th Conference of the Israeli Nuclear Society 11-2* (1996).
 <sup>9</sup>D. L. Henderson, S. Yoo, M. Kowalok, T. R. Mackie, and B. R. Thomad-umper strength and the strength of the strength

- sen, "The adjoint method for the optimization of brachytherapy and ra-diotherapy patient treatment planning procedures using Monte Carlo calculations," Final Report for DOE-NEER Grant: DE-FG07-99ID13774 (2001).
- <sup>(2001)</sup>.
   <sup>10</sup>R. Jeraj and P. Keall, "Monte Carlo-based inverse treatment planning," Phys. Med. Biol. 44, 1885–1896 (1999).
   <sup>11</sup>"MCNP–A general Monte Carlo N-particle transport code, version 4C," LA-13709–M, edited by J. F. Briesmeister (2000).
- <sup>1</sup> Zh. L. Beers and V. W. Pine, "Functional expansion technique for Monte Carlo electron transport calculations" IEEE Trans. Nucl. Sci. 23, 1850– 1856 (1976)
- <sup>13</sup>R. F. Barth, J. A. Coderre, M. G. H. Vicente, and T. E. Blue, "Boron neutron capture therapy of cancer: Current status and future prospects," Clin. Cancer Res. **11**, 3987–4002 (2005).
- <sup>14</sup>Y. Nakagawa, K. Pooh, T. Kobayashi, Y. Sakurai, T. Kageji, S. Uyama, A. Matsumura, T. Yamamoto, and H. Kumada, "Clinical review of the Japa-
- Matsumura, I. ramamoto, and H. Kumada, Chincai review of the Japanese experience with boron neutron capture therapy and a proposed strategy using epithermal neutron beams," J. Neuro-Oncol. 62, 87–99 (2003).
  <sup>15</sup>M. L. Williams and W. W. Engle, Jr., "The concept of spatial channel theory applied to reactor shielding analysis," Nucl. Sci. Eng. 62, 92–104 (2007).
- (1977).
   <sup>16</sup>J. C. Wagner, E. L. Redmond II, S. P. Palmtag, and J. S. Hendricks, "MCNP: Multigroup/adjoint capabilities," LA-12704, Los Alamos National Laboratory (1994).
- <sup>17</sup>E. E. Lewis and W. F. Miller, Jr., Computational Methods of Neutron Transport (Wiley, New York, 1984).
- <sup>18</sup>J. E. Hoogenboom, "Methodology of continuous-energy adjoint Monte Carlo for neutron, photon, and coupled neutron-photon transport," Nucl.

Sci. Eng. 143, 99-120 (2003).

<sup>19</sup>http://www-rsicc.ornl.gov/rsic-cgi-bin/enote.pl?nb=mcnp5&action=view &page=116; MCNP5 Notebook-page 116; F. Wasastjerna and T. Booth. <sup>20</sup>A. Noel and H. S. Wio, "A new series-expansion approach in Monte Carlo: Application to neutron shielding," Ann. Nucl. Energy **11**, 225–227

- (1984)<sup>21</sup>D. P. Griesheimer and W. R. Martin, "Monte Carlo based angular flux
- distribution with orthogonal function expansion," Trans. Am. Nucl. Soc. 89, 370-372 (2003).
- <sup>22</sup>D. P. Griesheimer and W. R. Martin, "Two dimensional functional expansion tallies for Monte Carlo calculations." PHYSOR 2004-The Physics of Fuel Cycles and Advanced Nuclear Systems: Global Developments (2004)
- <sup>23</sup>D. Legrady and J. E. Hoogenboom, "Midway Monte Carlo forwardadjoint coupling with Legendre polynomials, "New Frontiers of Nuclear Technology: Reactor Physics, Safety and High-Performance Computing (2004).
- <sup>24</sup>D. Legrady and J. E. Hoogenboom, "Visualization of space-dependency of responses of Monte Carlo calculations using Legendre polynomials, Proceedings PHYSOR 2004, Chicago, IL, April 25-29 (Elsevier Science, Chicago, 2004), pp. 1-7.
- <sup>25</sup>D. P. Griesheimer, W. R. Martin, and J. P. Holloway, "Convergence prop-erties of Monte Carlo functional/expansion techniques," J. Comput. Phys. 211, 129-153 (2006).
- <sup>26</sup>J. P. Boyd, Chebyshev and Fourier Spectral Method (Dover, New York, 2000).
- <sup>27</sup>R. C. Little and R. E. Seamon, "New MENDF5 and MENDF5G," Los Alamos Internal Memorandum X-6:RCL-86-412 (1986).
- <sup>28</sup>JEF-2.2 Nuclear Data Library, JEFF Report 17, Paris, France (2000).
  <sup>29</sup>K. A. Van Riper, "A CT and MRI scan to MCNP input conversion pro-
- " Radiat. Prot. Dosim. 115, 513-516 (2005). gram," Radiat. Prot. Dosim. 115, 513–510 (2005). tissues. International Commission on Radiation Units and Measurements,
- Bethesda MD (1992) <sup>31</sup>J. T. West III, SABRINA, "An interactive three-dimensional geometrymodeling program for MCNP," Los Alamos National Laboratory Report,
- LA10688M (1986). <sup>32</sup>K. A. Van Riper, "New features in Sabrina," *Proceedings of the Topical Meeting on Radiation Protection for our National Priorities*, Spokane,
- WA, Sept. 17-21, (2000), pp. 316-323. <sup>33</sup>EORTC protocol 11011 (version 2.0/February 2, 2004): Early phase II study on BNCT in metastatic malignant melanoma using the boron carrier BPA. European Organization for Research and Treatment of Cancer, Brussels (2004).
- 34 R. H. Hardin, N. J. A. Sloane and W. D. Smith, "Tables of spherical codes with icosahedral symmetry," published electronically at http:// www.research.att.com/~njas/icosahedral.codes/. These tables are copy-right R. H. Hardin, N. J. A. Sloane, and W. D. Smith, 1994 (2000).

Medical Physics, Vol. 34, No. 4, April 2007

RADIATION RESEARCH **166**, 81–88 (2006) 0033-7587/06 \$15.00 © 2006 by Radiation Research Society. All rights of reproduction in any form reserved.

# Design of a Rotating Facility for Extracorporal Treatment of an Explanted Liver with Disseminated Metastases by Boron Neutron Capture Therapy with an Epithermal Neutron Beam

### V. A. Nievaart,<sup>a,b,1</sup> R. L. Moss,<sup>b</sup> J. L. Kloosterman,<sup>a</sup> T. H. J. J. van der Hagen<sup>a,c</sup> H. van Dam<sup>a</sup> A. Wittig,<sup>d</sup> M. Malago<sup>a</sup> and W. Sauerwein<sup>d</sup>

<sup>a</sup> Department of Applied Sciences, Delft University of Technology, 2628CJ Delft, The Netherlands; <sup>b</sup> Joint Research Centre of the European Commission, 1755ZG Petten, The Netherlands; <sup>c</sup> Reactor Institute Delft, Delft University of Technology, 2629JB Delft, The Netherlands; <sup>e</sup> Clinic and Policlinic for Radiotherapy, Essen University Hospital, 45122 Essen, Germany; and <sup>e</sup> Department of General and Transplantation Surgery, Essen University Hospital, 45122 Essen, Germany

Nievaart, V. A., Moss, R. L., Kloosterman, J. L., van der Hagen, T. H. J. J., van Dam, H., Wittig, A., Malago, M. and Sauerwein, W. Design of a Rotating Facility for Extracorporal Treatment of an Explanted Liver with Disseminated Metastases by Boron Neutron Capture Therapy with an Epithermal Neutron Beam. *Radiat. Res.* 166, 81–88 (2006).

In 2001, at the TRIGA reactor of the University of Pavia (Italy), a patient suffering from diffuse liver metastases from an adenocarcinoma of the sigmoid was successfully treated by boron neutron capture therapy (BNCT). The procedure involved boron infusion prior to hepatectomy, irradiation of the explanted liver at the thermal column of the reactor, and subsequent reimplantation. A complete response was observed. This encouraging outcome stimulated the Essen/Petten BNCT group to investigate whether such an extracorporal irradiation could be performed at the BNCT irradiation facility at the HFR Petten (The Netherlands), which has very different irradiation characteristics than the Pavia facility. A computational study has been carried out. A rotating PMMA container with a liver, surrounded by PMMA and graphite, is simulated using the Monte Carlo code MCNP. Due to the ro-tation and neutron moderation of the PMMA container, the initial epithermal neutron beam provides a nearly homogeneous thermal neutron field in the liver. The main conditions for treatment as reported from the Pavia experiment, i.e. a thermal neutron fluence of  $4 \times 10^{12} \pm 20\%$  cm<sup>-2</sup>, can be closely met at the HFR in an acceptable time, which, depending on the defined conditions, is between 140 and 180 min. © 2006 by Radiation Research Society

### INTRODUCTION

The liver is the first major organ reached by the blood draining the gastrointestinal tract. Consequently, it is a common site of hematogenous metastasis from gastrointes-

<sup>1</sup> Address for correspondence: Joint Research Centre of the European Commission, P.O. Box 2, 1755ZG Petten, The Netherlands; e-mail: victor. nievaart@jrc.nl. tipe  $\gamma$  res. Colorectal cancer is the standard radius  $\gamma$  and  $\gamma$  death in developed count is the standard radius  $\gamma$  and  $\gamma$  death in developed count is  $\gamma$  and  $\gamma$ 

**Q**1. gery offers ... baraneutic or ... an im-- and a 20% proved medi-10-yer . (J). However, in cases Of mor cricted to the liver that capr le' nepatectomy, chemother2only trea in be offered. Even " , newest chemot ue of around 40%, the drugs have a r en shown to be marginal ival ber . to loco-regional techniques " ativ ' ""roical procedures " ....tially er with curative. One sus ......urporal treatment of the 1' TV (BNCT).

uon presented here is mou. Th suctreatment using BNCT on 0. С ~ 48-year-old patient dev ıa (Italy) ronous diffuse h. tases from an adeno of the sigmoid colon. ver metastases after conventional chemo. fter jr weriments (6, 7), the patients pru n1، of a 0.1Jution (300 mg BPA  $_{\mu\nu}.$ over a pr and tumor sar . taken to measure the borop ration Jr to hepatectomy. The expl-٦d C and transported to the .eactor of 1. ity of Pavia where \* uated at the therm

<sup>&</sup>lt;sup>2</sup> Private communication..., r r nelli and S. Altieri.

NIEVAART ET AL.



liver ...

tumor respon . an observation period of 12 · ... the liver showed a normal liver v . tesions in the former tumor areas. Aftumor recurrence was observed in the form atside the liver but in the area of the former It was interpreted as iatrogenic due \* mor tissue prior to the liver er moved. An adim au. BNC'ı, . uns after his in ..... shortly followed by intraperiton. 2005, 44 months after from diffuse liver metastases of a con-

treated following the same ter the auto-trr .... problems that were int m prior chemotherapy. Τ<sup>1</sup> any viable Jr (communicaus rvations stimulated not on but also similar studies in Finland 10, 'J.S. (10). the reaction that occurs between the boronarmal neutrons. In this reaction, the thermaı <sup>1</sup> by the nucleus, and the resulting boron-lı the emission of high-LET and high-RBE  $\alpha$  a... ticles are only 9 and 5  $\mu$ m,  $\mu$ , particle traverses the cell **nucl** death. With tare lective can be obtained (11). This

Possibility to treat multiple metastases ach more efficiently than with existing conuerapies. Moreover, BNCT has the potential to very small metastases or satellite nodules that ally detectable and are often responsible sease. A prerequisite, how of 10B inside the

with an array of the state of t

<sup>3</sup> Europhysic. September 5–7, .

actively transported by the L-all form into cells (12) and which (13-18). Another are tumor cells. 1. aus of this paper. The proposed ake place at the BNCT facility (19) of arriux Reactor (HFR) in Petten (The Netherlands) ing the available, well-collimated epithermal neutror The aim of the study is therefore to demonst homogeneous thermal neutron flux distribution livered to the liver using for a thermal neutron constitution to the study is therefore to demonst homogeneous thermal neutron flux distribution constitution flux distribution flux distribution and thermal neutron flux distribution and

### MATERIALS AND METHO.

The Liver

<sup>14</sup> verify is the largest internal organ in the human body and
<sup>15</sup> songy, wedge-shaped organ. According to the
<sup>15</sup> verify and liver is around 1.5
that or a solution of a healthy male liver is around 1.5
that or a solution of the material set of a weather of vertically. \*
or right surface, up to 17...
12.5 cm. A liver with metastases.
However, the evolving calculation
Low with a spheroidLow with a spheroidsolution of a plastic bag, as is usually
solution of 1.6 liver
mold with a volume of 1.6 liver

The pictures were 'r' General and 'r'' and University (Gr'' and prior to the discara, arefore, under German law, exempt fi arot the welfare of human subjects.

### Medical Requirements

The University Hospital Essen is one of the main ' both diseased and living donor liver transm'

82

r

at i.

### FACILITY FOR EXTRACORPORAL LIVER IRRADIATIONS

rating and participating in the second secon

### Radiation C.

In BNCT, the absorbed dose ... eral dose components.  $\sim_{\mathcal{F}}$  nuclear reactions. The valuated are the boron dose  $(D_{\rm B})$ , beam, others g four dosu doses  $(D_p \text{ and } D_n)$ , and the  $\gamma$ -ray dose  $(D_{\gamma})$ , different dose components and to take into considthe ent biological effectiveness, the term total biologically  $D_{w}$ , has been proposed.<sup>5</sup> Pinelli *et al.* estimated the limthe liver to be  $D_w = 15$  Gy (23) when applying the westing from *in vitro* and *in vivo* experiments and skin (24–26). Apart from <sup>+1</sup> 'anted liver irradia\*-....e of ....ss follows the 10.00 propu The limiting antablished through appropriate animal experiments. It has to be stressed that only use to the tumor. All other dose components are no duce an absorbed dose in the L the limiting dose ors: the minimization of the non-Succer SP<sup>4</sup> stases. 1. the uptake and probably vary between I for .ase; this is currently being studied

### 'oals and Limits in this Computational Study

' of this study is to evaluate and compare the fluence in the reported in Pavia. Nevertheless, the total weighted do... do... ' to that reported in Pavia using the estimates describe... binelli *et al.* (4) report a  $D_w$  of  $8.0^{\circ} \pm 0.5$  Gy in the .... Since the metastases are spread throughour .... flux distribution should be ac.'

Roter

, an a thermal column, as in the Pavia setup, the opti-, for the liver is in its natural form when placed on a flat dhermal column allows for the organ to be placed in effectively ' of thermal neutrons. The thermal neutron flux will be a the surface and will decay rather strongly into  $t^{\mu}$ ' n the surface, the thermal neutron  $t^{\mu}$ .

# int status contraction and an AA-TECDOC-

Market in different tumours using . Survey Study coordinator: W. Sauerwein.



FIG. 2. Isometric view of the torus-shaped tallies to a whole liver of

....e) value. An initial design considered fitting the anaginary cylindrical container with a diameter of 25 cr neight of 8 cm. In this way, the thermal neutrons, coming f sides, will deliver a homogeneously distributed dose in the ' When dealing with an epithermal neutron beam. " completely re-evaluated. With respect to the P energy spectrum consists mainly of around 30 keV , nas a very small divergence collimators of diameters 8, 12 and 15 '---ation of the epithermal '-alf this maxneutrons has a maximu imum value at 0.5 cm and 6.5 cm. And the cylindrical approach as mentioned above, it use ding most effective part of the neutron fluence, i.e. between cm, can be used if the liver is inserted into a spheroid-shah and then rotated around the polar axis. In this way, when the  $\gamma$  is moved slightly nearer to the surface by the scatter. iner material, the thermal fluence will be spr tor of the spheroid, the thermal findistributed over a sm its lowe. when rotating. When using one of the cu. designed for BNCT, such as BNC, (29), it is not possible r r'lan calculate the as therefore necessary to "var. torus-shaped volumes were "'as in MCNP), thereргодта by simulating the rotation (see . As illustrated in Fig. ? is positioned sw ...at of the ... ine thermal flux ... more accurately, tori with su .... at the vertical axis are spheres ...  $_{\rm LG}$  tori, only one calculation is needed in which the better from one direction. MCNP adds all the track J particles inside the torus volume and gives, when dividir of the torus, the flux. This averaging of the calculate

<sup>7</sup> J. T. West III, SABRINA: An <sup>Jac</sup> ... Geometry-Modeling Program

NIEVAART ET AL. 4 " the final setup: Right (surrounded by dashed line)." 5. Wall possio. Graphite containing boron reflect ent simulation of the flux distribution for the rotating can b , the thermal neutrons in a ra. sc scattering cross section of the same hya a consequence, the flux becomes a time-averaged flux ng liver. flux is uniform in the whole torus after e utermal neutrons in the BNCT treatment of brain tumors. In 'es to make the flux unifordesigns for the 2.4-liter container, the wall of the PMMA contr the thickest (5 mm) at the equator and decreased to 2 mm to and bottom of the spheroid. Although the neutron flux be tuned even better in this way, for engineer Liver Irrau constant PMMA thickness. . ugether with a The design concept 10. does not require any changes to the .....o the container, the container "I liquid (water or more conservation ment. The neutron source spectrum and source to be adapted. The liver source · · · · · · gaps that would to install and rem mponents. To keep the liver influence the irradiation new. In this computational study, the liver  $com_{\lambda}$  modeled human are taken from ICRU Report 40 , regard to on of the organ will cor es a precise volume inside the container is modeled as liver tissue, we in such a way that it doc. as evolved from surrounding the hyperbolic sectors and the surrounding the sector sectors and the sector sectors and the sector sectors are set of the sectors and the sectors are set of the sectors and the sectors are set of rable with water and the conservation solution. During the tra \*ions, a conservatively chosen 15 ppm of <sup>10</sup>B is distributed \*he liver. From the brain studies, it is known that the r, moderator and coolant) toward a surrounding he same characteristics. It appears that the solid surround depth in the brain will be 4% higher who homogeneously (33). Therefore negligible difference in yield. However, it is easier to roarison with something that must be watertight. After 18 、 servatively ontration in health i.. ⇒ he 'culations with different materials and configurais lower than 15 ppm, tions in this study. ating a voi... in Fig. 3. The deriv... 1 6 liters, respectively, as shown e given in the next section. Figure 3 is a collection of three geometrice: left, a cross section of the lot 'mo optimal setups have ha upper lim-- .... inty for the liver been considered un is give-its. The lower limit is an :liters; it is not ~ .ped container holding the liver is placed in a PMMA .th sides of 26 cm. The PMMA cube acts as a neutron with the . conservation . pecially for the top and bottom of the spheroid. For back the cube is surrounded by pure graphite blocks (1.6 at are smaller than this vo an irradiation volume capacity of 2.4 ntrance a graphite cone. 'heroid-shaped container for holdir ...strained by the existing beam characteristics. onsists of two equal part Thermal Neutron Flux Distribution maped easily, is vantages In BNCT, three of the four ma and 15 based on thermal neut ing device. A duction of y rays due to the  $(n,\gamma)$  reaction from  $m_{J-}$ ....g parameter.

~f16

are

.... neutron



FACILITY FOR EXTRACORPORAL LIVER IRRADIATIONS

... Panel a: Thermal neutron flux inside the 2.4-liter container when irradic anal neutron flux with the container rotating.

shows thermal neutron flux distribution f-4-liter volume: an equatorial ro 'ius of 8.0 cm. There energy --uportionally 5. ... is given as background to u. -fictical uncertainties in the MCNP results pro-Figure 4a gives the thermal neuron cross section along the botionary. The per ... une rert; nence it is pr whest flux values at the th "the liver. In Fig. this ro-.ver containe1 . s flux distribution product

. Note that the gradient of the therman n the left panel gives a disturbed view winth the right panel due to a different step size



FIG. 5. Thermal neuro

... ines. Clearly, from Fig. 4b, the thermal ... is around  $4.4 \times 10^8$  cm<sup>-2</sup> s<sup>-1</sup> at 3 cm from th face. Similar results can be seen in Fig. 5, where 'mal neutron flux contours are plotted for the 'container. In this design, the loss of vc' for an more homogeneous "

uermal neutron at the center of the liver. Increasares can be accomplished by more PIVIIVIA, י~w down more epithermal neutrons. To compthermal neutrons at the center, the equator. decreased. The optimal dimensions turn out to be radius of 7.0 cm and polar radius of 8.0 c. mothy related to the size of the beam. In fa ' 5-liter spheroid is rotated whe ainer. Of course, th pareu and live absolutely no u... 4 and 5 are 2D represent. essary to give information -.on flux .. containers. Note ., nom the center describes a bigger ' or to the center. Figure 6 shows the tnerma. rtograms for both containers. In 7% of the mal -.... 1.6-liter container, apart from . at, all thermal neutron flux intervals are sess equal in filling the volume. The smaller wi spectrum for the small container shows the gr geneity. Note that the lowest thermal nev+ in the 2.4-liter container looks quite ∠wn as an iso-contour. as in T ...nter), but .....e (1.1%). For what



To define  $D_{w}$  or  $z_{w}$ 

<sup>c</sup> <sup>1</sup> × 10<sup>12</sup> cm<sup>-2</sup>

### FACILITY FOR EXTRACORPORAL LIVER IRRADIATIONS

	Total biologically				_	
		althy	Physical dose rates			
	(47 ppm <sup>10</sup> B)		<i>D<sub>B</sub></i> (Gy/h) <sup>10</sup> B)	$\dot{D}_n$ (Gy/h)	$\dot{D}_p$ (Gy/h)	$\dot{D}_{\gamma}$ (Gy/h)
1.6 liters		_				
Minimum	23.7	5.2			0.3	2.2
Maximum	29.0	6.5	1.1			3.3
Average	26.5	6.0	1.0	U		2.8
		4.8	0.7	0.1		
Мали						2.9
Average						2 A

. as to what the influence of the irradiation , which is one of the objectives of the analed in this paper. , ver holders, the desired thermal neutro

 $m^{-2} \pm 20\%$  can be reac<sup>1</sup>

respect to the . w the liv. ...ued. This is a very cru. -gestion made by Pinelli et al. in 2001 (22) movimum in a small volume. In 2002, 1 .... treated in Pavia received a total .....  $\pm$  0.5 Gy (4) but dial received +1-. resented here probably lea/1 ion of all dose comadvantage .e liver, . to the tolerance dost ider the conditions used to irrau. nimal experiments will be necessary to esu e dose to a healthy explanted liver after BPA nost suitable model would appear to be the organs are of similar size to those of huma. Possible me. <sup>1</sup> Amence include a further reduction of the highest thermal flux value in the liver

γ rays would reading the provided for the spectrum and flux. Decreasing the maximal neutron flux at the thermal peak is impossible is thermal peak is inherent to the use of an epitheam. Although it is feasible to shift the slightly, the procedure is limiting on the fluxes at the center

has lea to have encouraging the next step in the project, ...

a program of . Luts to valida

, of boron-10 throughout the irrad also be evaluated in detail to better deter me dose delivered to the healthy tissue. A first st been undertaken by the clinical trial EORTC 11 which investigates the uptake of BPA in ' liver metastases. In conclusion 4o an extra-.... epithermal neutron he closely met. With the therand 10B concentrations allow a safe and . "acorporal treatment of liver cancer by Bins BNCT facility in Petten is feasible from the tec of view.

## **~KNOWLEDGMENTS**

This work ... "w by the Euror sion, contract no QLK.

Received: July 8, 2005: accord

### TYCES

4. T. Pinelli, A. Zonta, S. Altieri, S. Barni, A. Branni, A. Branni, P. Chiari, C. Ferrari and C. Zonta, and Development Neuronal Action to the solution provide structure, Bologna, 2002.

### NIEVAART ET AL.

- E. Wilkinson, An out-of-body experience. Lancet Oncol. 4, 64 (2003).
- L. Roveda, U. Prati, J. Bakeine, F. Trotta, P. Marotta, P. Valsecchi, A. Zonta, R. Nano, A. Faccetti and P. Pedroni, How to study boron biodistribution in liver metastases from colorectal cancer. J. Chemother. 16, 15–18 (2004).
- 7. R. Nano, S. Barni, P. Chiari, T. Pinelli, F. Fossati, S. Altieri, C. Zonta, U. Prati, L. Roveda and A. Zonta, Efficacy of boron neutron capture therapy on liver metastases of colon adenocarcinoma: optical and ultrastructural study in the rat. *Oncol. Rep.* **11**, 149–153 (2004).
- P. Kotiluoto and I. Auterinen, MCNP study for epithermal neutron irradiation of an isolated liver at the Finnish BNCT facility. *Appl. Radiat. Isot.* 61, 781–785 (2004).
- M. Suzuki, Y. Sakurai, S. Masunaga, Y. Kinashi, K. Nagata and K. Ono, Dosimetric study of boron neutron capture therapy with borocaptate sodium (BSH)/lipiodol emulsion (BSH/lipiodol-BNCT) for treatment of multiple liver tumors. Int. J. Radiat. Oncol. Biol. Phys. 58, 892–896 (2004).
- H. Koivunoro, D. L. Bleuel, U. Nastasi, T. P. Lou, J. Reijonen and K. N. Leung, BNCT dose distribution in liver with epithermal D-D and D-T fusion-based neutron beams. *Appl. Radiat. Isot.* 61, 853– 859 (2004).
- W. Sauerwein, Principles and history of neutron capture therapy. Strahlenther. Onkol. 169, 1–6 (1993).
- A. Wittig, W. Sauerwein and J. Coderre, Mechanisms of transport of p-borono-phenylalanine through the cell membranein vitro. *Radiat. Res.* 153, 173–180 (2000).
- W. Sauerwein and A. Zurlo, The EORTC Boron Neutron Capture Therapy (BNCT) Group: Achievements and future projects. *Eur. J. Cancer* 38, S31–S34 (2002).
- 14. S. Obayashi, I. Kato, K. Ono, S. Masunaga, M. Suzuki, K. Nagata, Y. Sakurai and Y. Yura, Delivery of (10)boron to oral squamous cell carcinoma using boronophenylalanine and borocaptate sodium for boron neutron capture therapy. *Oral Oncol.* 40, 474–482 (2004).
- I. Kato, K. Ono, Y. Sakurai, M. Ohmae, A. Maruhashi, Y. Imahori, M. Kirihata, M. Nakazawa and Y. Yura, Effectiveness of BNCT for recurrent head and neck malignancies. *Appl. Radiat. Isot.* 61, 1069– 1073 (2004).
- A. Z. Diaz, Assessment of the results from the phase I/II boron neutron capture therapy trials at the Brookhaven National Laboratory from a clinician's point of view. J. Neurooncol. 62, 101–109 (2003).
- 17. H. Joensuu, L. Kankaanranta, T. Seppala, I. Auterinen, M. Kallio, M. Kulvik, J. Laakso, J. Vahatalo, M. Kortesniemi and S. Savolainen, Boron neutron capture therapy of brain tumors: Clinical trials at the Finnish facility using boronophenylalanine. *J. Neurooncol.* **62**, 123–134 (2003).
- P. M. Busse, O. K. Harling, M. R. Palmer, W. S. Kiger, 3rd, J. Kaplan, I. Kaplan, C. F. Chuang, J. T. Goorley, K. J. Riley and R. G. Zamenhof, A critical examination of the results from the Harvard-MIT NCT program phase I clinical trial of neutron capture therapy for intracranial disease. J. Neurooncol. 62, 111–121 (2003).
- R. L. Moss, O. Aizawa, D. Beynon, R. Brugger, G. Constantine, O. Harling, H. B. Liu and P. Watkins, The requirements and development of neutron beams for neutron capture therapy of brain cancer. *J. Neurooncol.* 33, 27–40 (1997).
- H. Gray, Anatomy of the Human Body. Lea & Febiger, Philadelphia, 1918. [Thoroughly revised and re-edited by W. H. Lewis, New York, 2000; available online at http://www.bartleby.com/107]
- M. Malago, G. Testa, A. Frilling, S. Nadalin, C. Valentin-Gamazo, A. Paul, H. Lang, U. Treichel and C. E. Broelsch, Cicinnati right living donor liver transplantation: An option for adult patients: Single

institution experience with 74 patients. Ann. Surg. 238, 853-862 (2003).

- J. Rassow, W. Sauerwein, A. Wittig, E. Bourhis-Martin, K. Hideghéty and R. Moss, Advantage and limitations of weighting factors and weighted dose quantities and their units in boron neutron capture therapy. *Med. Phys.* 31, 1128–1134 (2004).
- uerapy. *mea. Phys.* **31**, 1128–1134 (2004).
  23. T. Pinelli, S. Altieri, F. Fossati, A. Zonta, C. Ferrari, U. Prati, L. Roveda, S. Ngnitejeu Tata and D. M. Ferguson, Operative modalities and effects of BNCT on liver metastases of colon adenocarcinoma. In *Frontiers in Neutron Capture Therapy* (M. F. Hawthorne, K. Shelly and R. J. Wiersema, Eds.), pp. 1427–1440. Kluwer Academic, Plenum Publishers, New York, 2001.
- 24. J. A. Coderre, M. S. Makar, P. L. Micca, M. M. Nawrocky, H. B. Liu, D. D. Joel, D. N. Slatkin and H. I. Amols, Derivations of relative biological effectiveness for the high-LET radiations produced during boron neutron capture irradiation of the 9L rat gliosarcoma in vitro and in vivo. Int. J. Radiat. Oncol. Biol. Phys. 27, 1121–1129 (1993).
- G. M. Morris, J. A. Coderre, J. W. Hopewell, P. L. Micca, M. M. Nawrocky, H. B. Liu and A. Bywaters, Assessment of the response of the central nervous system to boron neutron capture therapy using a rat spinal cord model. *Radiother. Oncol.* **30**, 249–255 (1994).
- 26. H. Fukuda, J. Hiratsuka, C. Honda, T. Kobayashi, K. Yoshino, H. Karashima, J. Takahashi, Y. Abe, K. Kanda and M. Ichihashi, Boron neutron capture therapy of malignant melanoma using "0B-paraboronophenylalanine with special reference to evaluation of radiation dose and damage to the normal skin. *Radiat. Res.* **138**, 435–442 (1994).
- D. E. Wessol, R. S. Babcock, F. J. Wheeler, G. J. Harkin, L. L. Voss and M. W. Frandsen, BNCT-Rtpe: BNCT Radiation Treatment Planning Environment User's Manual, Version 2.2, 1997. [Available online at http://d.inel.gov/rtpe-manual/sec00.html]
- B. D. E. Wessol, C. A. Wemple, F. J. Wheeler, M. T. Cohen, M. B. Rossmeier and J. J. Cogliati, SERA: Simulation Environment for Radiotherapy Applications User's Manual, version 1C0, 1999. [Available online at http://www.cs.montana.edu/~bnct/manual/sec00.html]
- R. Zamenhof, E. Redmond, G. Solares, D. Katz, K. Riley, S. Kiger and O. Harling, Monte Carlo-based treatment planning for boron neutron capture therapy using custom-designed models automatically generated from CT data. *Int. J. Radiat. Oncol. Biol. Phys.* 35, 383– 397 (1996).
- J. F. Briesmeister, Ed., A General Monte Carlo N-Particle Code, version 4C. Report LA-13079-M, Los Alamos National Laboratory, Los Alamos, NM, 2000.
- 31. K. A. Van Riper, New features in Sabrina. In *Radiation Protection for our National Priorities*, pp. 316–323. American Nuclear Society, La Grange Park, IL, 2000.
- ICRU, Photon, Electron, Proton and Neutron Interaction Data for Body Tissues. Report 46, International Commission on Radiation Units and Measurements. Bethesda. MD, 1992.
- S. J. Ye, Boron self-shielding effects on dose delivery of neutron capture therapy using epithermal beam and boronophenylalanine. *Med. Phys.* 26, 2488–2493 (1999).
- 34. A. Zonta, U. Prati, L. Roveda, S. Zonta, C. Ferrari and C. Zonta, Clinical lessons from the first applications of BNCT on unresectable liver metastases. In *Europhysics Conference Abstracts* (A. Fontana, G Viesti and A. Zenoni, Eds.), pp. 97–98. European Physical Society, Pavia, 2005.
- 35. T. Pinelli, S. Altieri, F. Fossati, A. Zonta, D. Cossard, U. Prati, L. Roveda, G. Ricevuti and R. Nano, Development of a method to use boron neutron capture therapy for diffused tumours of liver (TA-ORMINA project). In *Cancer Neutron Capture Therapy* (Y. Mishima, Ed.), pp. 783–794. Plenum Press, New York, 1996.

PAPER III 119