

## Towards an ultrasonically powered efficient multichannel neurostimulator implant

Kolovou Kouri, Konstantina; Serdijn, W.A.; Giagka, Vasiliki

**Publication date**

2022

**Document Version**

Accepted author manuscript

**Citation (APA)**

Kolovou Kouri, K., Serdijn, W. A., & Giagka, V. (2022). *Towards an ultrasonically powered efficient multichannel neurostimulator implant*. Abstract from International Winterschool on Bioelectronics (BioEI) 2022, Tirol, Austria, Tirol, Austria.

**Important note**

To cite this publication, please use the final published version (if applicable).  
Please check the document version above.

**Copyright**

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

**Takedown policy**

Please contact us and provide details if you believe this document breaches copyrights.  
We will remove access to the work immediately and investigate your claim.

# Towards an Ultrasonically Powered Efficient Multichannel Neurostimulator Implant

Konstantina Kolovou-Kouri<sup>a,b</sup>, Wouter A. Serdijn<sup>b</sup>, Vasiliki Giagka<sup>a,b</sup>

<sup>a</sup> *Bioelectronics Section, Department of Microelectronics, Faculty of Electrical Engineering, Mathematics and Computer Science, Delft University of Technology  
Mekelweg 4, 2628 CD, Delft, The Netherlands*

<sup>b</sup> *Technologies for Bioelectronics Group, Department of System Integration and Interconnection Technologies, Fraunhofer Institute for Reliability and Microintegration IZM  
Gustav-Meyer-Allee 25, 13355, Berlin, Germany*

e-mail: K.KolovouKouri@tudelft.nl

During the last few decades, electrical neural stimulators have successfully been employed as a means of treatment for a wide range of neurological disorders. By targeting the peripheral and central nervous systems, electrical neurostimulators activate/inhibit neural activity by manipulating the stimulus-induced electric field arising at the targeted area through diverse electrode configurations. Aiming at reducing the overall size of implanted stimulator systems, these are being designed to be wirelessly powered and batteryless. Technologies for wireless power transfer to implants are mainly based on inductive coupling and, more recently, ultrasonic waves.

To increase the power efficiency and thereby minimize the power consumption of the implant, we have previously proposed a stimulation technique that alters the electric field at the tissue by delivering charge in small packets in a very rapid manner (e.g. 1 Mpps). This charge is consequently accumulated by the tissue's integrating nature.<sup>1</sup> This approach removes the need to ensure continuous accurate control of the stimulus current amplitude, resulting in power savings. To the same end, in an ultrasonically powered system, unnecessary power-conversion blocks can be eliminated, as the incoming ultrasonic wave is harvested, converted into an electrical signal, rectified, and then directly used for stimulation.<sup>2</sup>

Inspired by the above, this work focuses on providing a discrete-component multi-channel neural stimulator, powered wirelessly through ultrasound (US), to activate and inhibit neural activity. The use of mostly commercially available discrete components will ensure reproducibility by other research labs and help provide a platform technology that can be used as an experimental tool in a variety of applications.<sup>3</sup> The US pressure wave obtained at the receiving US transducer will be converted into electrical energy and used both for powering the system as well as to shape the charge packets of the eventual stimulus pulse. In this manner, the need for DC-DC up-conversion, typically needed for neuromodulation, will be eliminated.<sup>4</sup> A charge-balancing technique, matching the fast pulse repetition rates (PRR) required for inhibition of neural activity (typically  $\geq 5$  kHz), will ensure the safety of the implant, while the injected charge will be constantly monitored and adapted for safe and efficacious activation/inhibition.

1. M. N. van Dongen, F. E. Hoebeek, S. K. E. Koekkoek, C. I. De Zeeuw and W. A. Serdijn, "High frequency switched-mode stimulation can evoke postsynaptic responses in cerebellar principal neurons," *Frontiers Neuroeng.*, vol. 8, p. 2, 2015, doi: 10.3389/fneng.2015.00002.
2. L. Tacchetti, W. A. Serdijn, and V. Giagka, "An ultrasonically powered and controlled ultra-high-frequency biphasic electrical neurostimulator," in *Proc. IEEE BioCAS 2018*, Cleveland, OH, 2018, pp. 1-4, doi: 10.1109/BIOCAS.2018.8584718.
3. K. Kolovou-Kouri, S. Soloukey, B. S. Harhangi, W. A. Serdijn, and V. Giagka, "Dorsal Root Ganglion (DRG) Multichannel Stimulator Prototype Developed for Use in Early Clinical Trials," in *Proc. IEEE NER 2021*, Virtual, May 2021. doi: 10.1109/NER49283.2021.9441101.
4. G. E. Ólafsdóttir, W. A. Serdijn, and V. Giagka, "An energy-efficient, inexpensive, spinal cord stimulator with adaptive voltage compliance for freely moving rats," in *Proc. 40<sup>th</sup> IEEE EMBC 2018*, Honolulu, Hawaii, USA, Jul. 2018, pp. 2937 – 40, doi: 10.1109/EMBC.2018.8512904.