

High-resolution imaging and inversion of 3D wavefield data for layered media

Slob, Evert; Yang, Bingkun; Zhang, Lele

Publication date

2019

Document Version

Final published version

Citation (APA)

Slob, E., Yang, B., & Zhang, L. (2019). *High-resolution imaging and inversion of 3D wavefield data for layered media*. Poster session presented at EGU General Assembly 2019 , Vienna, Austria.

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.



Distributed Temperature Sensing (DTS) calibration with confidence intervals

Bas des Tombe (1), Bart Schilperoort (1), Zoltán Perkó (2), Mark Bakker (1), Nick van de Giesen (1), and John Selker (3)

(1) Delft University of Technology, Water Resources Section, Delft, Netherlands (b.f.destombe@tudelft.nl), (2) Delft University of Technology, Reactor Physics and Nuclear Materials Section, Delft, Netherlands, (3) Oregon State University, Department of Biological and Ecological Engineering, Corvallis (OR), United States of America

Temperature measurements along fiber-optic cables have shown their value in a range of geophysical applications. DTS devices are becoming so user-friendly that ‘everyone’ is able to perform temperature measurements along kilometers of fiber, but it remains difficult to obtain *accurate* temperature measurements. Quantification of the accuracy is important as it varies along the cable and over time and is different for each setup. The accuracy decreases as the measured signal decays farther down the fiber, the accuracy decreases with higher temperatures, and the device’s sensitivity fluctuates over time. Furthermore, the uncertainty in the estimated temperature is affected by the estimation of the parameters that relate the measured Stokes and anti-Stokes scattering to temperature.

We present a calibration framework that provides confidence intervals to the calibrated temperature. First, the noise variance in the (anti-) Stokes scattering is estimated directly from the measurements. Second, the temperature along the fiber is calibrated to reference sections with a known temperature. The calibration is weighted by the noise variance in the signal, so that reference sections close to the DTS device have a larger influence on the parameter estimation than sections with a weaker signal. Third, the uncertainty from the (anti-) Stokes scattering and the parameter uncertainty are projected to confidence intervals.

The procedures for single-ended and double-ended setups are wrapped into a Python package, available on github.com/dts calibration together with documentation and example notebooks.