



Delft University of Technology

Document Version

Final published version

Citation (APA)

Moghaddasi Kelishomi, H., Hunyadi, B., van der Veen, A. J., de Groot, N. M. S., & Hendriks, R. C. (2022). Surface Electrocardiogram Reconstruction Using Intra-operative Electrograms. In J. Louveaux, & F. Quitin (Eds.), *42nd WIC Symposium on Information Theory and Signal Processing in the Benelux (SITB 2022)* (pp. 136)

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

In case the licence states "Dutch Copyright Act (Article 25fa)", this publication was made available Green Open Access via the TU Delft Institutional Repository pursuant to Dutch Copyright Act (Article 25fa, the Taverne amendment). This provision does not affect copyright ownership.
Unless copyright is transferred by contract or statute, it remains with the copyright holder.

Sharing and reuse

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

This work is downloaded from Delft University of Technology.

Surface Electrocardiogram Reconstruction Using Intra-operative Electrograms

Hanie Moghaddasi

Circuits and Systems

Delft University of Technology

Delft, The Netherlands

H.moghaddasi@tudelft.nl

Borbála Hunyadi

Circuits and Systems

Delft University of Technology

Delft, The Netherlands

B.Hunyadi@tudelft.nl

Alle-Jan van der Veen

Circuits and Systems

Delft University of Technology

Delft, The Netherlands

A.J.vanderVeen@tudelft.nl

Natasja M.S. de Groot

Department of Cardiology

Erasmus Medical Center

Rotterdam, The Netherlands

N.m.s.degroot@erasmusmc.nl

Richard C. Hendriks

Circuits and Systems

Delft University of Technology

Delft, The Netherlands

R.C.Hendriks@tudelft.nl

ABSTRACT

Atrial Fibrillation (AF) is the most sustained arrhythmia in the heart. On the surface electrocardiogram (ECG), AF is characterised by the irregular RR intervals and by fibrillatory waves or the absence of a P wave. Since AF is a progressive disease, timely and correct detection is crucial for AF treatment. Detailed insight into the areas of arrhythmia-related electropathology can be obtained by analyzing high-resolution (inter-electrode distance 2mm) electrograms (EGMs). However, these measurements are rather invasive. By integration of high-resolution epicardial mapping data and surface ECG data, we hope to learn how different stages of AF represent themselves on the ECG. Eventually this can help to guide to identify areas of electropathology as target sites of ablation therapy on the less invasive ECG. A first step in this direction is to learn how to reconstruct the ECG based on EGM measurements. In practice, however, EGMs are only measured at few atrial locations, not covering the complete atria. An important question therefore is: How can we reconstruct ECG based on the observations from a limited part of the heart? To answer this question, we propose two methods. In the first method, we increase the number of observations from a part of the right atrium (RA) to the whole RA by synchronizing EGMs that are measured at different moments in time based on the local activation time (LAT). In the second method, under the assumption that atrial EGMs measured at different spatial areas are linearly related, the conductivity matrix is estimated for the whole atrium which enables us to reconstruct the ECGs from the limited observations. The second method brings twofold benefits. First, the conductivity matrix can be used as a novel diagnostic tool to detect AF as well as areas of electropathology. Second, it provides a practical solution to reconstruct epicardial potentials from ECGs, non-

invasively. The results show that method one increases the reconstruction accuracy. Furthermore, the conductivity matrix reveals the structural differences between sinus rhythm (SR) and AF episodes which could be the first step to interpret the underlying electropathology of AF.

This research was funded in part by the Medical Delta Cardiac Arrhythmia Lab (CAL), the Netherlands.