

B1-based SAR reconstruction using contrast source inversion–electric properties tomography (CSI-EPT)

Balidemaj, Edmond; van den Berg, Cornelis A.T.; van Lier, Astrid L.H.M.W.; Nederveen, Aart J.; Stalpers, Lukas J.A.; Crezee, Hans; Remis, Rob F.

DOI

[10.1007/s11517-016-1497-6](https://doi.org/10.1007/s11517-016-1497-6)

Publication date

2016

Document Version

Final published version

Published in

Medical & Biological Engineering & Computing

Citation (APA)

Balidemaj, E., van den Berg, C. A. T., van Lier, A. L. H. M. W., Nederveen, A. J., Stalpers, L. J. A., Crezee, H., & Remis, R. F. (2016). B1-based SAR reconstruction using contrast source inversion–electric properties tomography (CSI-EPT). *Medical & Biological Engineering & Computing*, *55*(2), 225-233. <https://doi.org/10.1007/s11517-016-1497-6>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

B1-based SAR reconstruction using contrast source inversion–electric properties tomography (CSI-EPT)

Edmond Balidemaj¹ · Cornelis A. T. van den Berg² · Astrid L. H. M. W. van Lier² · Aart J. Nederveen³ · Lukas J. A. Stalpers¹ · Hans Crezee¹ · Rob F. Remis⁴

Received: 8 October 2015 / Accepted: 24 March 2016 / Published online: 23 April 2016
© The Author(s) 2016. This article is published with open access at Springerlink.com

Abstract Specific absorption rate (SAR) assessment is essential for safety purposes during MR acquisition. Online SAR assessment is not trivial and requires, in addition, knowledge of the electric tissue properties and the electric fields in the human anatomy. In this study, the potential of the recently developed CSI-EPT method to reconstruct SAR distributions is investigated. This method is based on integral representations for the electromagnetic field and attempts to reconstruct the tissue parameters and the electric field strength based on B_1^+ field data only. Full three-dimensional FDTD simulations using a female pelvis model are used to validate two-dimensional CSI reconstruction results in the central transverse plane of a 3T body coil. Numerical experiments demonstrate that the reconstructed SAR distributions are in good agreement with the SAR distributions as determined via 3D FDTD simulations and show that these distributions can be computed very efficiently in the central transverse plane of a body coil with the two-dimensional approach of CSI-EPT.

Keywords B1 maps · Specific absorption rate (SAR) · Contrast source inversion (CSI) · Electric properties tomography (EPT) · Conductivity

1 Introduction

Assessment of the specific absorption rate (SAR) due to electromagnetic (EM) fields in human tissue is relevant in many applications such as hyperthermia [9, 10, 16], telecommunications [23], and high field MRI [5, 8, 21, 30, 35]. However, for reliable SAR assessment, knowledge of the electric properties (EPs) of biological tissues is required (in particular, the conductivity σ and permittivity ϵ) and the electric field strength must be known as well. This information is usually not directly available and therefore has to be determined by other means. In MRI, various implementations of electric properties tomography (EPT) methods have been developed to extract this information from the B_1^+ field [1, 12, 13, 19, 24, 33, 34, 36]. This field is accessible to measurement and present-day EPT methods attempt to reconstruct the electric tissue parameters from measured B_1^+ field maps, while the corresponding electric field strength is determined by forward modeling in which the reconstructed conductivity and permittivity profiles serve as a model for the patient's anatomy.

One of the drawbacks of the EPT methods mentioned above is that these methods typically suffer from reconstruction artifacts especially near tissue boundaries. These artifacts occur mainly because currently used EPT methods are based on local field equations (either Maxwell's equations or Helmholtz's equation) and do not take the electromagnetic boundary conditions into account. Furthermore, these methods are very sensitive to noise or other perturbations in the data, since differential operators act on measured B_1^+ field data. Different studies have focused on minimizing the

✉ Edmond Balidemaj
e.balidemaj@amc.nl

¹ Radiotherapy Department, Academic Medical Center, University of Amsterdam, Amsterdam, The Netherlands

² Radiotherapy Department, University Medical Center, Utrecht University, Utrecht, The Netherlands

³ Radiology Department, Academic Medical Center, University of Amsterdam, Amsterdam, The Netherlands

⁴ Circuits and Systems Group, Faculty of Electrical Engineering, Mathematics and Computer Science, Delft University of Technology, Delft, The Netherlands

reconstruction artifacts by using either the gradient of EP profiles in conjunction with a multi-channel transmit/receive array RF coil [20] or arbitrary-shaped kernels based on voxel position [15]. However, these ad hoc solutions are still based on a local differential operator approach, which may yield less accurate SAR predictions due to potential reconstruction errors in the EP profiles that immediately affect the computed electric field strength in the forward modeling step.

As an alternative to local EPT methods, we have recently proposed an iterative contrast source inversion EPT method (CSI-EPT) [2], which is based on global integral representations for the electromagnetic field [2, 3]. The electromagnetic boundary conditions are then automatically satisfied and reconstruction results near tissue interfaces are significantly improved [2]. Furthermore, CSI-EPT is less sensitive to noise since in CSI-EPT integral operators act on measured field data (instead of differential operators as in local EPT methods) and CSI-EPT reconstructs the electric field strength inside the region of interest as well. This latter property makes CSI-EPT an ideal candidate for SAR reconstructions based on B_1^+ field data, since it attempts to simultaneously reconstruct the EP profiles and the electric field strength within the human anatomy.

The electromagnetic wave field inside the human body is obviously a fully vectorial three-dimensional wave field. However, as earlier described by van de Bergen [31], the electromagnetic field in the central transverse plane of a 3T or 7T body coil can be treated as a two-dimensional wave field where only H_x , H_y , and E_z are present. The case where only H_x , H_y , and E_z are considered is also referred to as the TM-polarized case. Reconstructing the SAR distribution based on two-dimensional instead of three-dimensional fields obviously leads to significant speedups in computation time and may even allow for online SAR reconstructions. Our approach is therefore to reconstruct the SAR distribution in the neighborhood of the central transverse plane of a body coil using a two-dimensional CSI-EPT reconstruction method. To validate our approach, we compare the reconstructed profiles, electric fields, and SAR distributions with 3D models and fully vectorial 3D FDTD simulations. We use a static field of 3T in all numerical experiments. The approach is equally applicable for 7T or other static background field strengths, as long as the two-dimensional field approximation in the central slice remains valid.

2 Methods

2.1 The CSI-EPT method

In this section, we briefly discuss the main features of the CSI-EPT method. The method is fully described in [2] and further mathematical details can be found in [27] and [28].

As a starting point, we first write the RF field $\{E, B_1^+\}$ that is present in the MR system as a superposition of the electromagnetic background field and the scattered field. The background field $\{E^b, B_1^{+;b}\}$ is the field that is present within the MR system in the absence of a dielectric object or body, whereas the scattered field $\{E^{sc}, B_1^{+;sc}\}$ is the field induced by the object or body. The object occupies a bounded domain D and is characterized by a conductivity $\sigma(r)$, a permittivity $\varepsilon(r)$, and a permeability $\mu(r)$, with $r = (x, y, z)$ the position vector. In this work, we have ignored relative permeability variations as they are considered negligible for biological tissue [7]. In practice, the background field $\{E^b, B_1^{+;b}\}$ can be acquired by forward modeling.

Using the linearity of Maxwell's equations, the scattered electric field at a point with position vector r can be written as [27]

$$E^{sc}(r) = \int_{r' \in D} G^{EJ}(r, r') w(r') dV, \quad (1)$$

while the scattered $B_1^{+;sc}$ can be written as

$$B_1^{+;sc}(r) = \int_{r' \in D} G^{+;HJ}(r, r') w(r') dV. \quad (2)$$

In these equations, G^{EJ} denotes the Green's tensor relating the electric current to electric field and the tensor $G^{+;HJ}$ relates the electric current to the B_1^+ field. Furthermore, w is the contrast source given by

$$w = \chi E, \quad (3)$$

where $\chi = \eta/\eta_b - 1$ is the contrast function, with $\eta(r) = \sigma(r) - i\omega\varepsilon(r)$, and $\eta_b = -i\omega\varepsilon_0$.

Obviously, the goal is to reconstruct the contrast function χ and the electric field E based on B_1^+ data. A solution to this inverse problem is formulated by iteratively minimizing the cost function given by

$$F = F_{\text{data}} + F_{\text{obj}} \quad (4)$$

where

$$F_{\text{data}}^{[n]} = \frac{B_1^{+;sc} - \hat{G}^{+;HJ}\{w^{[n]}\}}{B_1^{+;sc}} \quad (5)$$

and

$$F_{\text{obj}}^{[n]} = \frac{\chi^{[n]} E^{[n]} - w^{[n]}}{\chi^{[n-1]} E^b} \quad (6)$$

where we have introduced the operator $\hat{G}^{+;HJ}\{w\}$ as

$$\hat{G}^{+;HJ}\{w\}(r) = \int_{r' \in D} G^{+;HJ}(r, r') w(r') dV. \quad (7)$$

The subscript $[n]$ in (5) and (6) represents the iteration number. The CSI method updates both the contrast source ($w^{[n]}$) and the contrast function ($\chi^{[n]}$) using a two-step updating procedure. In the first step, the contrast function is fixed ($\chi = \chi^{[n-1]}$) while the contrast source ($w^{[n]}$) is updated by minimizing Eq. (5). In the second step, a new contrast function ($\chi^{[n]}$) is obtained by using the updated contrast source $w^{[n]}$ from the first step. Moreover, the electric field corresponding to the updated $w^{[n]}$ can be computed by

$$E^{[n]} = E^b(r) + \hat{G}^{EJ} \{w^{[n]}\} \tag{8}$$

with

$$\hat{G}^{EJ} \{w\}(r) = \int_{r' \in D} G^{EJ}(r, r') w(r') dV. \tag{9}$$

Finally, the contrast function is then obtained by minimizing Eq. (6) with respect to χ ; hence, the new contrast function is computed as

$$\chi^{[n]} = \frac{w^{[n]} \bar{E}^{[n]}}{E^{[n]} \bar{E}^{[n]}}. \tag{10}$$

The overbar in Eq. (10) denotes the complex conjugate. The iterative process is terminated once the cost function, Eq. (4), reaches a user-specified tolerance level. Elsewhere we reported a more detailed description of the CSI-EPT algorithm [2] which includes the multiplicative total variation factor for noise suppression and the ability to include more than one B1 data set in the iterative process.

2.2 3D and 2D electromagnetic modeling

We have performed 3D field simulations using in-house developed finite-difference time domain (FDTD) tools [25] and the 3T body coil model as described in [29]. The coil was tuned at 128 MHz (i.e., the Larmor frequency at 3T) and was driven in quadrature mode. The female body model (Ella) from the virtual family provided by IT'IS [6] has been used, and the assigned conductivity and permittivity values are based on [11] at 128 MHz. The tissue density values reported in [14] were used for SAR computations. The computed SAR by 3D field simulations serves as a benchmark to which the 2D simulations will be compared.

The 2D simulations (for a TM-polarized configuration) were conducted using the integral equation method. In the TM-polarized configuration, the electric field vector is parallel to the invariance direction. The fields were generated by 16 RF line sources driven at 128 MHz, which corresponds to an operating frequency of the RF body coil in a 3T MR system. The line sources were located on a circle ($R = 0.34$ m) symmetrically positioned around the female pelvis model with an isotropic voxel size of 2.5 mm. A

homogeneous medium (free space) is taken as a background model. In the current implementation, we have assumed exact knowledge of the B_1^+ phase.

The CSI-EPT algorithm is implemented as we previously described in [2]. The CSI-EPT software code was implemented in MATLAB (MathWorks, Natick, Massachusetts, USA). The computational time for 5000 iterations of the presented method, with a grid size of 2.5 mm, is around 110 s on an Intel Core i7 operating at 1.9 GHz. Furthermore, SAR_{1g} and SAR_{10g} , representing the average SAR over a mass of 1 and 10 g, respectively, are computed based on [4] and take approximately 20 and 10 s, respectively.

We have compared the results for the mid-plane slice ($z = 0$ cm) as the 2D modeling is likely to be a valid approximation in this region. However, we have also explored the reconstruction at two off-central slices (i.e., $z = +7.5$ cm and $z = -2.5$ cm).

3 Results

To test the SAR reconstruction results of our algorithm, we first compute the fully three-dimensional electromagnetic field inside the 3D Ella body model using FDTD and focus on the field and SAR distributions in three slices located at $z = 0$ cm (midplane), $z = +7.5$ cm, and $z = -2.5$ cm. The conductivity and permittivity profiles within these three slices are shown in Fig. 1, while the magnitude of the Cartesian components of the corresponding 3D electric field strength is shown in Fig. 2. In these figures, the amplitudes of the field components E_x and E_y are normalized with respect to the maximum amplitude of the E_z field of the corresponding slice. We observe that E_z is the dominant field component in all three slices indicating that it is reasonable to assume a two-dimensional E-polarized field structure in and around the midplane of the body coil. The field is not exactly two-dimensional, of course, which is particularly noticeable for the x -component of the electric field strength (first column of Fig. 2). This component vanishes for a two-dimensional E-polarized field, but it clearly does not in the fully three-dimensional case especially around the center of the slices and within the slice located closed to the legs (slice at $z = -2.5$ cm). These deviations from 2D are due to anatomical variations in the longitudinal z -direction, which are especially large around the slice located at $z = -2.5$ cm, since here we transitioned from the torso to the upper legs. Finally, the 3D- and 2D-normalized $|B_1^+|$ maps of the mid-plane slice are shown in Fig. 3a, b, respectively. We observe that both maps have a similar field pattern, apart from some local differences mainly at the central region. This observation again confirms that it is reasonable to assume that the electromagnetic field essentially has a two-dimensional E-polarized field structure in the midplane of the body coil.

Fig. 1 Target electric conductivity (*left*) and permittivity maps (*right*) of the midplane slice (*top row*), the slice at $z = +7.5$ cm (*middle row*) and the slice at $z = -2.5$ cm (*bottom row*)




Fig. 2 $|E_x|$, $|E_y|$, and $|E_z|$ (*left to right*) distributions in the mid-plane slice (*top row*), the slice at $z = +7.5$ cm (*middle row*), and the slice at $z = -2.5$ cm (*bottom row*). The $|E_x|$, $|E_y|$ field distributions are normalized with respect to the maximum amplitude of the corresponding $|E_z|$ field distribution





Fig. 3 Normalized $|B_1^+|$ field distribution in the midplane slice based on 3D FDTD (**a**) and the 2D integral equation method (**b**)



In practice, the measured B_1^+ field is not known exactly, of course, and we therefore contaminate the 2D-simulated B_1^+ field with additive Gaussian noise (SNR 20). This field now serves as an input for our CSI-EPT algorithm. The

reconstructed conductivity and permittivity maps obtained after 5000 iterations of the CSI-EPT algorithm are shown in Fig. 4a, b, respectively. We note that these results were obtained by incorporating multiplicative total variation regularization into




Fig. 4 The reconstructed conductivity (a) and permittivity (b) maps after 5000 iterations of the CSI-EPT algorithm. c The normalized $|E_z|$

our CSI-EPT algorithm (for details, see [2]) to suppress the effects of noise in the data. From Fig. 4a, b, we observe that the conductivity and permittivity reconstructions are in good agreement with the target maps of Fig. 1a, b. Furthermore, in Fig. 4c, the reconstructed $|E_z|$ field is shown which is used together with the reconstructed conductivity map of Fig. 4a to determine the SAR distribution as reconstructed by CSI-EPT.

Figure 5a shows the voxel-wise SAR distribution based on 3D field simulations, while the SAR_{10g} and SAR_{1g} are depicted in Fig. 5d, g, respectively. In Fig. 5b, e, h (second column of Fig. 5), the computed SAR distributions based on the 2D field simulations are shown, which are in good agreement with the distributions based on the 3D simulations (1st column of Fig. 5). Only slight deviations are observed on the right bottom part of the anatomy. Finally, the CSI-SAR reconstructions using only B_1^+ field information are shown in Fig. 5c, f, i (third column of Fig. 5). As mentioned above, this B_1^+ field is contaminated with additive Gaussian noise (SNR 20). Comparing the different reconstructed SAR distributions with the 3D (first column of Fig. 5) and 2D (second column of Fig. 5) SAR distributions, we observe that the CSI-SAR reconstructions are in good agreement with the 3D- and 2D-modeled SAR distributions. The relative error between the reconstructed SAR distributions based on CSI-EPT and 3D FDTD is shown in the fourth column of Fig. 5.

The SAR distributions within the non-central slices are depicted in Fig. 6 (slice at $z = +7.5$ cm) and Fig. 7 (slice at $z = -2.5$ cm). The reconstructed SAR distribution of the transversal slice at $z = +7.5$ cm, where $|E_z|$ is the dominant




Fig. 5 The normalized voxel-based SAR distribution (top row), the normalized SAR_{10g} distribution (middle row), and the SAR_{1g} distribution (bottom row) of the midplane slice ($z = 0$ cm). The distributions based on 3D FDTD field simulations are shown in (a, e, i), while the distributions based on a 2D integral equation approach are shown in

(b, f, j). The reconstructed SAR distributions based on CSI-EPT are presented in (c, g, k). The relative error between the reconstructed SAR distributions based on CSI-EPT and 3D FDTD is shown in (d, h, l)




Fig. 6 The normalized voxel-based SAR distribution (*top row*), the normalized SAR_{10g} distribution (*middle row*), and the SAR_{1g} distribution (*bottom row*) of the slice at $z = +7.5$ cm. The distributions based on 3D FDTD field simulations are shown in (a, e, i), while the distributions based on a 2D integral equation approach are shown in

(b, f, j). The reconstructed SAR distributions based on CSI-EPT are presented in (c, g, k). The relative error between the reconstructed SAR distributions based on CSI-EPT and 3D FDTD is shown in (d, h, l)




Fig. 7 The normalized voxel-based SAR distribution (*top row*), the normalized SAR_{10g} distribution (*middle row*), and the SAR_{1g} distribution (*bottom row*) of the slice at $z = -2.5$ cm. The distributions based on 3D FDTD field simulations are shown in (a, e, i), while the distributions based on a 2D integral equation approach are shown in

(b, f, j). The reconstructed SAR distributions based on CSI-EPT are presented in (c, g, k). The relative error between the reconstructed SAR distributions based on CSI-EPT and 3D FDTD is shown in (d, h, l)

field, is in good agreement with the SAR distributions calculated by 3D and 2D forward modeling as shown in the 1st and 2nd column of Fig. 6, respectively. In the fourth column of Fig. 6, the relative error between the reconstructed SAR distributions based on CSI-EPT and 3D FDTD is shown. However, the SAR reconstruction within the slice located at $z = -2.5$ cm, where the transverse electric field components were not negligible, shows a discrepancy in the central region

in a comparison between the 1st and 3rd column of Fig. 7. The discrepancy is due to the fact that transverse electric fields are not considered in a 2D approach, and discrepancies in reconstructed SAR may therefore appear in regions where these transverse fields are not negligible. However, comparison of the 1st and 3rd column of Fig. 7 still shows a good agreement outside the central region as confirmed by the relative error shown in the fourth column of Fig. 7.

4 Discussion

Hot spots are a potential risk of high field clinical MRI; prediction of SAR distribution may help to reduce this hazard and is thus essential for MRI quality assurance and patient safety. In this paper, we have exploited the CSI-EPT method to reconstruct electric field and tissue properties and investigated the performance to reconstruct SAR distributions based on B_1^+ information only. This method takes the integral representations for the electromagnetic field as a starting point, and the electric field and tissue parameters are obtained by iteratively minimizing an objective function which measures the discrepancy between measured and modeled data and the discrepancy in satisfying a consistency equation known as the object equation.

Numerical results illustrate that SAR distributions can be reconstructed based on B_1^+ information using a 2D implementation of CSI-EPT. In general, a good performance was observed for slices where the transverse components of the electric field were negligible. These results clearly illustrate the ability of CSI-EPT to reconstruct SAR distributions within slices where E_z is the dominant field component, which is in general the case for the midplane slice of an RF body coil model [31]. Our studies indicate, however, that a two-dimensional field approximation may also be applied for off-central transverse slices (see Fig. 6). In such cases, a 2D implementation of CSI-EPT would yield reliable SAR reconstruction as well. Unfortunately, it is not a priori known on which off-central slices, the transverse components of the E-field are negligible and we therefore restrict ourselves to the midplane slice when we use a 2D implementation of CSI-EPT. Despite this restriction, the current 2D implementation of CSI-EPT seems to be a promising tool to improve current SAR assessment, since a good agreement was observed between reconstructed SAR distributions and 3D FDTD-based SAR distributions. As can be seen from Figs. 5, 6, and 7, the reconstructed voxel-based, 10 and 1 g SAR distributions show a good overall agreement. To quantify the error in all three cases, we have computed the relative error between the two-dimensional reconstructed SAR based on CSI-EPT (third column in Figs. 5, 6, and 7) and the true SAR distribution as determined by the full 3D FDTD model (first column in Figs. 5, 6, and 7). We observe that the error is small throughout the slice except in some highly isolated regions. These error regions occur mainly because the size of the hot spots is not precisely predicted by our 2D model. Our model does indicate, however, where hot spots can be expected and gives a good overall qualitative indication of the SAR distribution within the slices of interest. Moreover, CSI-EPT is applicable at all fields strength and is not limited to the demonstrated performance at 3T.

In its present form, the CSI-EPT algorithm takes perturbed B_1^+ field as input and effects due to noise are suppressed by incorporating multiplicative total variation regularization into the CSI-EPT algorithm (see [2]). Additional uncertainties in the B_1^+ phase may also be taken into account [2]. In practice, measurements of the B_1^+ phase are based on assumptions regarding the object and coil geometry [17, 32] and this transceive phase assumption can be considered as an uncertainty in the B_1^+ phase as well. These uncertainties can be taken into account in CSI-EPT by modifying the objective function in a similar manner as in [26]. However, in a number of recent studies [18, 19, 24, 37], the transceive phase assumption is avoided by using multiple independent transmit/receive channels. This opens up possibilities for EPT reconstruction and local SAR estimation [18, 37, 38] free of assumptions regarding the B_1^+ phase. Although we have presented reconstruction results for a quadrature coil configuration only, CSI-EPT is actually suitable for various antenna settings and can therefore benefit from multiple independent transmit/receive systems as exploited in [18, 37, 38] for assumption-free phase data.

The applicability of the EPT method to electric properties mapping has recently been confirmed in a series of phantom and in vivo experiments with MRI systems [17, 20, 22, 34]. Present work is therefore focused on extending the current implementation of CSI-EPT toward a practical MRI setting using both 2D and 3D field models. Three-dimensional models obviously do not suffer from a restriction to the midplane of the body coil and will provide more accurate reconstruction results in regions where two-dimensional field approximations fail. On the other hand, computation times in 3D will be significantly larger than in 2D due to an increase in the number of unknowns and the application of 3D FFTs. If possible, it is therefore beneficial to use 2D CSI-EPT, which may even provide online SAR reconstructions in the midplane of a body coil.

5 Conclusion

Whether a two- or three-dimensional CSI-EPT method is applied, the CSI-EPT method reconstructs, besides the electric properties, also the electric field at no additional computational costs. Given the promising results presented in this paper, we believe that CSI-EPT may prove an important tool toward MR-based SAR reconstruction. In future work, we will therefore focus on developing an efficient implementation of 3D CSI-EPT that allows for complete local SAR assessment inside and outside the midplane of the RF transmit coil.

Acknowledgments The research presented in this paper was supported by the Dutch Cancer Society (Grant Number: UVA 2010-4660).

Open Access This article is distributed under the terms of the Creative Commons Attribution 4.0 International License (<http://creativecommons.org/licenses/by/4.0/>), which permits unrestricted use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made.

References

1. Balidemaj E, Van Lier ALHMW, Crezee H, Nederveen AJ, Stalpers LJA, Van Den Berg CAT (2014) Feasibility of electric property tomography of pelvic tumors at 3T. *Magn Reson Med* 73:1505–1513
2. Balidemaj E, van den Berg CAT, Trinks J, van Lier A, Nederveen AN, Stalpers LJA, Crezee H, Remis RF (2015) CSI-EPT: a contrast source inversion approach for improved MRI-based electric properties tomography. *IEEE Trans Med Imaging* 34(9):1788–1796
3. Caorsi S, Frattoni A, Gragnani GL, Nortino E, Pastorino M (1991) Numerical algorithm for dielectric-permittivity microwave imaging of inhomogeneous biological bodies. *Med Biol Eng Comput* 29(6):NS37–NS44
4. Carluccio G, Erricolo D, Oh S, Collins CM (2013) An approach to rapid calculation of temperature change in tissue using spatial filters to approximate effects of thermal conduction. *IEEE Trans Biomed Eng* 60(6):1735–1741
5. Chen J, Feng Z, Jin JM (1998) Numerical simulation of SAR and B1-field inhomogeneity of shielded RF coils loaded with the human head. *IEEE Trans Biomed Eng* 45(5):650–659
6. Christ A, Kainz W, Hahn EG, Honegger K, Zefferer M, Neufeld E, Rascher W, Janka R, Bautz W, Chen J, Kiefer B, Schmitt P, Hollenbach H-P, Shen J, Oberle M, Szczerba D, Kam A, Guag JW, Kuster N (2010) The virtual family-development of surface-based anatomical models of two adults and two children for dosimetric simulations. *Phys Med Biol* 55(2):N23–N38
7. Collins CM, Yang B, Yang QX, Smith MB (2002) Numerical calculations of the static magnetic field in three-dimensional multi-tissue models of the human head. *Magn Reson Imaging* 20(5):413–424
8. Collins CM, Liu W, Wang J, Gruetter R, Vaughan JT, Ugurbil K, Smith MB (2004) Temperature and SAR calculations for a human head within volume and surface coils at 64 and 300 MHz. *J Magn Reson Imaging* 19(5):650–656
9. D'Ambrosio V, Dughiero F (2007) Numerical model for RF capacitive regional deep hyperthermia in pelvic tumors. *Med Biol Eng Comput* 45(5):459–466
10. de Greef M, Kok HP, Correia D, Borsboom P-P, Bel A, Crezee J (2011) Uncertainty in hyperthermia treatment planning: the need for robust system design. *Phys Med Biol* 56(11):3233–3250
11. Gabriel S, Lau RW, Gabriel C (1996) The dielectric properties of biological tissues: II. Measurements in the frequency range 10 Hz to 20 GHz. *Phys Med Biol* 41:2251–2269
12. Gamba HR, Delpy DT (1998) Measurement of electrical current density distribution within the tissues of the head by magnetic resonance imaging. *Med Biol Eng Comput* 36(2):165–170
13. Haacke EM, Petropoulos LS, Nilges EW, Wu DH (1991) Extraction of conductivity and permittivity using magnetic resonance imaging. *Phys Med Biol* 36(6):713–734
14. Hasgall P, Di Gennaro F, Baumgartner C, Neufeld E, Gosselin M, Payne D, Klingeböck A, Kuster N (2015) IT'IS database for thermal and electromagnetic parameters of biological tissues. Version 2.6, January 13th, 2015. www.itis.ethz.ch/database
15. Huang L, Schweser F, Herrmann KH, Krämer M, Deistung A, Reichenbach JR (2014) A Monte Carlo method for overcoming the edge artifacts in MRI-based electrical conductivity mapping. In: Proceedings of 22th Annual Meeting ISMRM, p 3190
16. Kato H, Ishida T (1993) Present and future status of noninvasive selective deep heating using RF in hyperthermia. *Med Biol Eng Comput* 31(1):S2–S11
17. Katscher U, Voigt T, Findeklee C, Vernickel P, Nehrke K, Dössel O (2009) Determination of electric conductivity and local SAR via B1 mapping. *IEEE Trans Med Imaging* 28(9):1365–1374
18. Katscher U, Findeklee C, Voigt T (2012) B1-based specific energy absorption rate determination for nonquadrature radiofrequency excitation. *Magn Reson Med* 68(6):1911–1918
19. Liu J, Zhang X, Van de Moortele P-F, Schmitter S, He B (2013) Determining electrical properties based on B1 fields measured in an MR scanner using a multi-channel transmit/receive coil: a general approach. *Phys Med Biol* 58(13):4395–4408
20. Liu J, Zhang X, Schmitter S, Van de Moortele P-F, He B (2015) Gradient-based electrical properties tomography (gEPT): a robust method for mapping electrical properties of biological tissues in vivo using magnetic resonance imaging. *Magn Reson Med* 74(3):634–646
21. Makris N, Angelone L, Tulloch S, Sorg S, Kaiser J, Kennedy D, Bonmassar G (2008) MRI-based anatomical model of the human head for specific absorption rate mapping. *Med Biol Eng Comput* 46(12):1239–1251
22. Marques JP, Sodickson DK, Ipek O, Collins CM, Gruetter R (2015) Single acquisition electrical property mapping based on relative coil sensitivities: a proof-of-concept demonstration. *Magn Reson Med* 74(1):185–195
23. Peyman A, Gabriel C, Grant EH, Vermeeren G, Martens L (2009) Variation of the dielectric properties of tissues with age: the effect on the values of SAR in children when exposed to walkie-talkie devices. *Phys Med Biol* 54(2):227–241
24. Sodickson DK, Alon L, Deniz CM, Ben-Eliezer N, Cloos M, Sodickson LA, Collins CM, Wiggins GC, Novikov DS (2013) Generalized local maxwell tomography for mapping of electrical property gradients and tensors. In: Proceedings of 21th Annual Meeting ISMRM, Salt Lake City, USA, p 4175
25. Van De Kamer JB, Van Wieringen N, De Leeuw AAC, Lagendijk JJW (2001) The significance of accurate dielectric tissue data for hyperthermia. *Int J Hypertherm* 17(2):123–142
26. van den Berg PM (1999) Reconstruction of media posed as an optimization problem. In: Wirgin A (ed) *Wavefield inversion*. Springer, New York, pp 191–240
27. van den Berg PM, Abubakar A (2001) Contrast source inversion method: state of art. *Prog Electromagn Res* 34:189–218
28. van den Berg PM, Kleinman RE (1997) A contrast source inversion method. *Inverse Probl* 13:1607–1620
29. van den Berg CAT, Bartels LW, van den Bergen B, Kroeze H, de Leeuw AAC, van de Kamer JB, Lagendijk JJW (2006) The use of MR B1 imaging for validation of FDTD electromagnetic simulations of human anatomies. *Phys Med Biol* 51:4735–4746
30. van den Bergen B, Van den Berg CAT, Bartels LW, Lagendijk JJW (2007) 7T body MRI: B1 shimming with simultaneous SAR reduction. *Phys Med Biol* 52(17):5429–5441
31. van den Bergen B, Stolk CC, Van Den Berg JB, Lagendijk JJW, Van den Berg CAT (2009) Ultra fast electromagnetic field computations for RF multi-transmit techniques in high field MRI. *Phys Med Biol* 54(5):1253–1264
32. van Lier ALHMW, Brunner DO, Pruessmann KP, Klomp DWJ, Luijten PR, Lagendijk JJW, van den Berg CAT (2012) B1(+) phase mapping at 7 T and its application for in vivo electrical conductivity mapping. *Magn Reson Med* 67(2):552–561
33. Van Lier ALHMW, Raaijmakers A, Voigt T, Lagendijk JJW, Luijten PR, Katscher U, Van Den Berg CAT (2014) Electrical properties tomography in the human brain at 1.5, 3, and 7T: A comparison study. *Magn Reson Med* 71(1):354–363
34. Voigt T, Katscher U, Doessel O (2011) Quantitative conductivity and permittivity imaging of the human brain using electric properties tomography. *Magn Reson Med* 66(2):456–466

35. Voigt T, Homann H, Katscher U, Doessel O (2012) Patient-individual local SAR determination: in vivo measurements and numerical validation. *Magn Reson Med* 68(4):1117–1126
36. Wen H (2003) Invasive quantitative mapping of conductivity and dielectric distributions using the RF wave propagation effects in high field MRI. *Med Imaging 2003* 5030:471–477
37. Zhang X, Schmitter S, Van de Moortele P-F, Liu J, He B (2013) From complex B(1) mapping to local SAR estimation for human brain MR imaging using multi-channel transceiver coil at 7T. *IEEE Trans Med Imaging* 32(6):1058–1067
38. Zhang X, Van de Moortele P-F, Schmitter S, He B (2013) Complex B1 mapping and electrical properties imaging of the human brain using a 16-channel transceiver coil at 7T. *Magn Reson Med* 69(5):1285–1296



Edmond Balidemaj finished his master studies at the Delft University of Technology at the department of Electrical Engineering. He is currently a PhD candidate at the Department of Radiotherapy of the AMC.



Cornelis A. T. van den Berg is an Associate Professor in the Imaging Division of the University Medical Centre Utrecht. His research themes are MR methods in ultra high field MRI and MRI guided radiotherapy.



Astrid L. H. M. W. van Lier Ph.D. graduated from Utrecht University. Currently, she is combining a residency in Medical Physics with research. Her research interests are quantitative MRI and image-based response assessment of cancer.



Aart J. Nederveen is a Principal Investigator at the Department of Radiology of the AMC. His research interests are designing and validating new MRI methods, mainly in the field of neurovascular, muscular, and abdominal imaging.



Lukas J. A. Stalpers is Principal Investigator at the Department of Radiotherapy of the AMC and his clinical focus is gynaecological oncology and brain tumors.



Hans Crezee has a MSc and a PhD in experimental physics. He is Principal Investigator at the Department of Radiotherapy of the AMC and project leader for 8 Dutch Cancer Society research projects.



Rob F. Remis is a member of the Circuits and Systems Group at Delft University of Technology. His research interests are in the area of computational electromagnetics with a special focus on imaging and model-order reduction techniques.