

Erratum: Lamb Waves and Adaptive Beamforming for Aberration Correction in Medical Ultrasound Imaging

Lamb Waves and Adaptive Beamforming for Aberration Correction in Medical Ultrasound Imaging (IEEE Trans.Ultrason., Ferroelectr., Freq. Control, early access (2020) DOI: 10.1109/TUFFC.2020.3007345)

Mozaffarzadeh, M.; Minonzio, C.G.; de Jong, N.; Verweij, M.D.; Hemm, Simone; Renaud, G.G.J.; Daeichin, V.

DOI

[10.1109/TUFFC.2020.3045936](https://doi.org/10.1109/TUFFC.2020.3045936)

Publication date

2020

Document Version

Final published version

Published in

IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control

Citation (APA)

Mozaffarzadeh, M., Minonzio, C. G., de Jong, N., Verweij, M. D., Hemm, S., Renaud, G. G. J., & Daeichin, V. (2020). Erratum: Lamb Waves and Adaptive Beamforming for Aberration Correction in Medical Ultrasound Imaging: Lamb Waves and Adaptive Beamforming for Aberration Correction in Medical Ultrasound Imaging (IEEE Trans.Ultrason., Ferroelectr., Freq. Control, early access (2020) DOI: 10.1109/TUFFC.2020.3007345). *IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control*, 68(2), 352. Article 9298803. <https://doi.org/10.1109/TUFFC.2020.3045936>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

Green Open Access added to TU Delft Institutional Repository

'You share, we take care!' – Taverne project

<https://www.openaccess.nl/en/you-share-we-take-care>

Otherwise as indicated in the copyright section: the publisher is the copyright holder of this work and the author uses the Dutch legislation to make this work public.

Erratum to “Lamb Waves and Adaptive Beamforming for Aberration Correction in Medical Ultrasound Imaging”

Moein Mozaffarzadeh^{1b}, Claudio Minonzio, Nico de Jong^{1b}, Martin D. Verweij^{1b},
Simone Hemm, Guillaume Renaud, and Verya Daeichin^{1b}

In the above article [1], we mentioned that the superposition of the different symmetric (S) modes in the frequency-wavenumber (f-k) domain results in a high-intensity region where its slope corresponds to the longitudinal wave speed in the slab. However, we have recently understood that this high-intensity region belongs to the propagation of a wave called lateral wave or head wave [2]–[5]. It is generated if the longitudinal sound speed of the aberrator (i.e., the PVC slab) is larger than that of water and if the incident wavefront is curved. When the incidence angle at the interface between water and PVC is near the critical angle, the refracted wave in PVC reradiates a small part of its energy into the fluid (i.e., the head wave). As discussed in [4], if the thickness of the waveguide is larger than the wavelength, the first arriving signal is the head wave. This is also the case in our study [1] where the ultrasound wavelength of a compressional wave in PVC was close to 1 mm, and a PVC slab with a thickness of 8 mm was used.

In this Erratum, numerical simulations (with SimSonic solver [5]) and experimental measurements (with the same PVC slab used in [1]) are conducted to investigate the propagation of the Lamb waves and head wave in detail, for the specific configuration studied in [1]. The pitch and element width of the P4-1 probe were used to assemble the numerical signals [see Fig. 1(a)]. If all the data simulated for the P4-1 probe is used, there indeed is a region with a slope [see Fig. 1(b)], but this has a low intensity, meaning

that the head wave has a relatively low amplitude compared to the specular reflections. Once the head wave is isolated, the sound speed can be estimated with a 0.3% error from the f-k domain plot [see Fig. 1(c)]. No significant difference is observed between Fig. 1(b) and (d), in which the head wave is muted.

Our experimental results show that if only the head wave (the first arriving signal) is used [see Fig. 2(b)], the slope of the linear fitting in the f-k domain also yields the longitudinal sound speed of the PVC with a 0.3% error.

Of note, the signal processing (i.e., linear fitting in the f-k domain) used in our study [1] still works for the head wave and is correct provided that the aberrator is parallel to the probe [6].

Also, in [1, p. 6], it is mentioned that “the curved structure of the skull might lead to other types of modes, such as the torsional modes.” Here, we acknowledge that this sentence is not correct, as torsional modes only exist in cylindrical waveguides or rectangular bars.

We would like to mention that Guillaume Renaud is added as a coauthor to acknowledge his contribution to the findings reported in this Erratum.

REFERENCES

- [1] M. Mozaffarzadeh, C. Minonzio, N. D. Jong, M. D. Verweij, S. Hemm, and V. Daeichin, “Lamb waves and adaptive beamforming for aberration correction in medical ultrasound imaging,” *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, early access, Jul. 6, 2020, doi: 10.1109/TUFFC.2020.3007345.
- [2] E. Camus, M. Talmant, G. Berger, and P. Laugier, “Analysis of the axial transmission technique for the assessment of skeletal status,” *J. Acoust. Soc. Amer.*, vol. 108, no. 6, pp. 3058–3065, Dec. 2000.
- [3] E. Bossy, M. Talmant, M. Defontaine, F. Patat, and P. Laugier, “Bidirectional axial transmission can improve accuracy and precision of ultrasonic velocity measurement in cortical bone: A validation on test materials,” *IEEE Trans. Ultrason., Ferroelectr., Freq. Control*, vol. 51, no. 1, pp. 71–79, Jan. 2004.
- [4] E. Bossy, M. Talmant, and P. Laugier, “Effect of bone cortical thickness on velocity measurements using ultrasonic axial transmission: A 2D simulation study,” *J. Acoust. Soc. Amer.*, vol. 112, no. 1, pp. 297–307, Jul. 2002.
- [5] E. Bossy, M. Talmant, and P. Laugier, “Three-dimensional simulations of ultrasonic axial transmission velocity measurement on cortical bone models,” *J. Acoust. Soc. Amer.*, vol. 115, no. 5, pp. 2314–2324, May 2004.
- [6] G. Renaud, P. Kruizinga, D. Cassereau, and P. Laugier, “In vivo ultrasound imaging of the bone cortex,” *Phys. Med. Biol.*, vol. 63, no. 12, Jun. 2018, Art. no. 125010.

Manuscript received December 2, 2020; accepted December 16, 2020. Date of publication December 18, 2020; date of current version January 26, 2021. This work was supported in part by a joint grant from the Netherlands Organisation for Scientific Research (NWO)/the Netherlands Organisation for Health Research and Development (ZonMw) and in part by the Department of Biotechnology, Government of India, through the Program Medical Devices for Affordable Health (MDAH) as Project Imaging Needles under Grant 116310008. (Corresponding author: Moein Mozaffarzadeh.)

Moein Mozaffarzadeh, Claudio Minonzio, Guillaume Renaud, and Verya Daeichin are with the Laboratory of Medical Imaging, Department of Imaging Physics, Delft University of Technology, 2628 CJ Delft, The Netherlands (e-mail: m.mozaffarzadeh@tudelft.nl).

Nico de Jong and Martin D. Verweij are with the Laboratory of Medical Imaging, Department of Imaging Physics, Delft University of Technology, 2628 CJ Delft, The Netherlands, and also with the Department Biomedical Engineering, Thoraxcenter, Erasmus Medical Center, 3015 GD Rotterdam, The Netherlands.

Simone Hemm is with the School of Life Sciences, Institute for Medical Engineering and Medical Informatics, University of Applied Sciences and Arts Northwestern Switzerland, 4132 Muttens, Switzerland.

Digital Object Identifier 10.1109/TUFFC.2020.3045936




Fig. 1. (a) Numerical model used in SimSonic. F_0 , BW , V_L , and V_s are the central frequency, bandwidth, longitudinal wave velocity, and transverse wave velocity, respectively. (b)–(d) Numerical space–time domain data and the corresponding f-k representation. (b) All the data. (c) Only the head wave. (d) Head wave is removed.




Fig. 2. Experimental space–time domain data and the corresponding f-k representation. A P4-1 ultrasound probe and a PVC slab with a thickness of 8 mm were used in the experiment. (a) All the data. (b) Only the head wave. (c) Head wave is removed.