

Challenges and solutions for application and wider adoption of wearable robots

Babič, Jan; Laffranchi, Matteo; Tessari, Federico; Verstraten, Tom; Novak, Domen; Šarabon, Nejc; Ugurlu, Barkan; Peternel, L.; Torricelli, Diego; Veneman, Jan F.

DOI

[10.1017/wtc.2021.13](https://doi.org/10.1017/wtc.2021.13)

Publication date

2021

Document Version

Final published version

Published in

Wearable Technologies

Citation (APA)

Babič, J., Laffranchi, M., Tessari, F., Verstraten, T., Novak, D., Šarabon, N., Ugurlu, B., Peternel, L., Torricelli, D., & Veneman, J. F. (2021). Challenges and solutions for application and wider adoption of wearable robots. *Wearable Technologies*, 2, Article e14. <https://doi.org/10.1017/wtc.2021.13>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

REVIEW ARTICLE

Challenges and solutions for application and wider adoption of wearable robots

Jan Babič^{1*} , Matteo Laffranchi² , Federico Tessari², Tom Verstraten³ , Domen Novak⁴ ,
Nejc Šarabon⁵, Barkan Ugurlu⁶ , Luka Peternel⁷ , Diego Torricelli⁸ and Jan F. Veneman⁹

¹Laboratory for Neuromechanics and Biorobotics, Department of Automation, Biocybernetics and Robotics, Jožef Stefan Institute, Ljubljana, Slovenia

²Rehab Technologies Lab, Istituto Italiano di Tecnologia, Genoa, Italy

³Robotics & Multibody Mechanics Research Group, Vrije Universiteit Brussel and Flanders Make, Brussels, Belgium

⁴University of Wyoming, Laramie, Wyoming, USA

⁵Faculty of Health Sciences, University of Primorska, Izola, Slovenia

⁶Biomechatronics Laboratory, Faculty of Engineering, Ozyegin University, Istanbul, Turkey

⁷Delft Haptics Lab, Department of Cognitive Robotics, Delft University of Technology, Delft, The Netherlands

⁸Cajal Institute, Spanish National Research Council, Madrid, Spain

⁹Hocoma AG, Volketswil, Switzerland

*Corresponding author. Email: jan.babic@ijs.si

Received: 02 March 2021; **Revised:** 25 August 2021; **Accepted:** 18 September 2021

Key words: exoskeleton; user acceptance; wearable; wearable robotics

Abstract

The science and technology of wearable robots are steadily advancing, and the use of such robots in our everyday life appears to be within reach. Nevertheless, widespread adoption of wearable robots should not be taken for granted, especially since many recent attempts to bring them to real-life applications resulted in mixed outcomes. The aim of this article is to address the current challenges that are limiting the application and wider adoption of wearable robots that are typically worn over the human body. We categorized the challenges into mechanical layout, actuation, sensing, body interface, control, human–robot interfacing and coadaptation, and benchmarking. For each category, we discuss specific challenges and the rationale for why solving them is important, followed by an overview of relevant recent works. We conclude with an opinion that summarizes possible solutions that could contribute to the wider adoption of wearable robots.

Introduction

Wearable robots are slowly but persistently entering our everyday lives—either at work, in sports, during rehabilitation, and in other situations. Even if we do not yet use such a robot, we can certainly imagine how nice it would be to wear a device that could reduce the burden of manual work on our body, increase our abilities during sports, or allow us to perform effective rehabilitation at our homes. With continuing advancements in mechanical design, modern materials, improved power sources, miniaturization of electronics, increased rate of data processing, and better understanding of biomechanics, widespread adoption of wearable robots seems to be within reach. Nevertheless, despite all these scientific and

technological advancements, many recent pioneering efforts to bring wearable robots to real-life applications were accompanied by mixed feelings of those who were supposed to benefit from them (Hensel and Keil, 2019; Davis et al., 2020; Mortenson et al., 2020).

The aim of this review article is to address the current challenges that are limiting application and wider adoption of wearable robots in various domains (e.g., industry, rehabilitation, sports, healthcare, for physically challenged individuals, harsh environments, and aeronautics) and offer an outlook on relevant solutions and approaches that could overcome these challenges. We focus on wearable robots that are typically worn over the human body—for example, stiff exoskeletons, soft actuated suits, and hybrid soft-stiff exoskeletons. However, we also mention robotic prosthetic devices wherever relevant.

In the following seven sections, we review various aspects of wearable robots and their interaction with the human body (Figure 1). The sections start with the challenges and the rationale for why solving them is important, followed by an overview of relevant recent works related to the specific challenges. Each section then concludes with an opinion that summarizes possible solutions to advance wearable robots beyond the state of art and contribute to their wider adoption.

Mechanical Layout

Today's wearable robotic devices feature a diverse variety of mechanical designs depending on the specific task they are intended to perform. Although examples of hybrid designs exist, most design approaches for wearable robots can be classified into two main categories: stiff devices, which consist of rigid, mostly metallic, structural elements, and so-called “exosuits,” which consist of flexible or even textile elements (Sanchez-Villamañan et al., 2019). The main challenge that engineers face in the development of wearable robot mechanics is to realize compact and light structures, combined with suitable kinematics so that the robot does not hinder the human's musculoskeletal kinematics. In other words, good exoskeleton mechanical design must be able to effectively replicate or support the complex movements of the human limbs while being as unobtrusive as possible in their utilization context and in their management.




Figure 1. Seven aspects of wearable robots that are addressed in the following sections.

Stiff Exoskeleton Designs

Exoskeletons with stiff structure and active joint actuation are the most widely known wearable devices and will be addressed in this section by analyzing applications for both the upper and lower limb. Depending on the amount of net active torque they can generate, they can be classified as “high-torque” or “low-torque” solutions. Specifically, “high-torque” stiff exoskeletons can generate sufficient torque to realize absolute control over the limb to which the exoskeleton is attached. This subcategory is ideal either when the wearer needs substantial support by the machine or when high levels of controllability are needed on the limb. High-torque use cases include lower limb assistance for severely impaired people, (Esquenazi et al., 2012; Wang et al., 2015; Milia et al., 2016; Vouga et al., 2017; van der Kooij et al., 2019; Jyräkoski et al., 2020; Figure 2c; www.indego.com) or power augmentation applications in which also upper limbs might be involved (Kazerooni, 2005; Marcheschi et al., 2011; www.sarcos.com; www.raytheon.com; Figure 2a,b). While kinematics that enables high levels of controllability of the limb is frequently used in rehabilitation applications since it allows the movement pattern to be precisely tuned to the patient’s needs (Bai and Christensen, 2017; Molteni et al., 2018), power augmentation applications typically do not demand total control of the biological joints. Indeed, within the “high-torque” class of wearable robots, some devices do not constrain the full limb of the wearer and deliberately leave freedom




Figure 2. Stiff exoskeletons: (a) the power augmentation exoskeleton Guardian XO from Sarcos (www.sarcos.com), (b) the Raytheon XOS 2 (www.raytheon.com), (c) the lower limb rehabilitation exoskeleton Indego (www.indego.com; Farris et al., 2011; Jyräkoski et al., 2020), (d) assistive lower limb exoskeleton GEMS (Lim et al., 2019), (e) the rehabilitation upper limb Harmony Exoskeleton by ReNeu Lab (www.reneu.robotics.utexas.edu/harmony-exoskeleton; Kim and Deshpande, 2017), (f) rehabilitation upper limb exoskeleton ANYexo from ETH (www.sms.hetz.ethz.ch/research/current-research-projects/armin-robot/ANYexo.html; Zimmermann et al., 2019), (g) the occupation active trunk exoskeleton Robo.Mate (Huysamen et al., 2018), and (h) the occupational passive upper limb exoskeleton EksoVest (www.eksobionics.com/ekso-evo/; Kim et al., 2018a,b).

of movement to some biological joints. A similar approach can be found in underconstrained exoskeletons, where dedicated kinematic solutions and redundancy are exploited to maximize the wearability and usability of the device (Lo and Xie, 2012; Tiseni et al., 2019). In fact, nonfully constrained architectures may also be employed in other applications where absolute control of limb joints is not needed, such as assistive/industrial scenarios (Guizzo and Goldstein, 2005; Bogue, 2018). In these applications, biological joints might not exactly match those of the robot, resolving possible joint alignment issues.

In all these high-torque lower/full-body exoskeletons, the mechanical structure needs to be in direct physical contact with the ground (i.e., through the feet) to support the load of the machine and the user.¹ Although significant engineering efforts have been made to lighten the structures of these devices (Farris et al., 2011; Vouga et al., 2017), their mechanical structure still tends to be cumbersome due to the considerable load they need to bear as well as the high torques that have to be generated and transferred through the structure to the ground, the load, and possibly to human limbs.

On the other side of the spectrum, “low-torque” stiff exoskeletons are not capable of generating sufficient torque to realize total control over either the upper or lower limb, and thus provide only partial assistance to the user. Application domains of this class include gait assistance devices (Lim et al., 2019; Zhang et al., 2019; Figure 2d) and occupational exoskeletons, either active (Figure 2g) or passive (Figure 2h; Bogue, 2018; Huysamen et al., 2018; Theurel and Desbrosses, 2019; www.eksobionics.com/ekso-evo/). The main function of this “low-torque” category is to assist the wearer in performing a specific motor task. As a result of the smaller force/torque exchange required by these applications, their structure is typically slimmer. These “low-torque” exoskeletons are considerably lighter than those in the “high torque” group, and to a certain extent their weight can thus be borne by the user.

Concerning the upper limb, most actuated exoskeletons for this limb are used for robot-assisted rehabilitation (Nef et al., 2007; Gopura et al., 2016; Kim and Deshpande, 2017; Rehmat et al., 2018). In upper-body exoskeletons, the device can be fixed on an external platform (Caldwell et al., 2007; Noda et al., 2014), which severely limits its mobility; conversely, it can be carried entirely by the human (Balasubramanian et al., 2008), which requires sufficient fitness to support the weight.

Upper limb rehabilitation devices need to deliver torque to the wearer during physical exercise (Lo and Xie, 2012; Mehrholz et al., 2012; Wendong et al., 2020), and typically involve designs that allow high levels of controllability of the limb (Nef et al., 2007; Gopura et al., 2016; Bai and Christensen, 2017). The robot is usually used in one of three modes: (a) passive mobilization, where the user relaxes their arm and the robot moves it to improve its range of motion (Brahmi et al., 2019); (b) assist as needed, where the robot assists voluntary movements of the patient (Reinkensmeyer et al., 2012); or (c) resistive or error augmentation mode, where the robot acts in opposition to the wearer’s movements (Marchal-Crespo and Reinkensmeyer, 2009). Please refer to Section “Control” for further details. In all three modes, the torques applied to the user are relatively low compared to lower limb or power augmentation scenarios. This is also true in the case of upper limb exoskeletons for teleoperative haptic applications, which need to generate feedback torques toward the user (Mallwitz et al., 2015; Buongiorno et al., 2019). It is worth mentioning that, although the assistive torques that need to be generated are much smaller compared to those in lower limb or power augmentation applications, the structure of these exoskeletons may need to support significant load due to the “cantilever” effect generated when the arm is extended—for example, during reaching. Therefore, despite the low required net torque, the structures of these exoskeletons may become bulky due to the need to withstand considerable gravitational load. In addition to this, upper limb exoskeleton kinematics are generally far from trivial due to the high number of degrees of freedom of the human and the complexity of the shoulder joint (Figure 2e,f; www.reneu.robotics.utexas.edu/harmony-exoskeleton; www.sms.hest.ethz.ch/research/current-research-projects/armin-robot/ANYexo.html; Kim and Deshpande, 2017; Zimmermann et al., 2019). Therefore, this exoskeleton type may need dedicated

¹ Please note that ensuring mechanical contact of the machine structure with the ground is also needed in several types of upper limb exoskeletons, as will be explained later in this paragraph.

designs to allow correct replication of biological kinematics and consequently facilitate their control (Perry et al., 2009). In contrast to the fully actuated and typically heavy upper limb machines addressed above, there is a growing interest in exoskeletons for arm support or gravity compensation in industrial applications (Spada et al., 2017; Kim et al., 2018a,b; Van Engelhoven et al., 2018; Pacifico et al., 2020; Goršič et al. 2020). These devices are typically passive and use elastic elements to provide the needed compensation torque.

Regardless of the application domain or body area, stiff exoskeleton designs may present wearability issues since mechanical stress is generated on the user's skin due to the generation of shear forces at the interface point. These shear forces may occur when there is misalignment between the exoskeleton rotation axes and the corresponding human joints. This represents a nonnegligible aspect of stiff exoskeletons, where the joints' centers of instantaneous rotation are often fixed while the human joints' (e.g., knee or shoulder) centers of instantaneous rotation change their relative position during the motion (Veeger, 2000; Hollman et al., 2002). As a consequence, aligning a stiff exoskeleton to a human limb might lead to undesired stresses. The amount of these stresses depends on the quality of the alignment and the capability to maintain it during motion. To deal with this, researchers are increasingly adopting original solutions that allow the exoskeleton structure to self-align to the anatomical kinematics of the wearer using additional passive degrees of freedom (Ball et al., 2007; Toxiri et al., 2015; Näf et al., 2018), compliant actuation systems (Vitiello et al., 2013), or specific mechanisms (Stienen et al., 2009; Ergin and Patoglu, 2012). However, these solutions require additional space and volume and are thus not always feasible in portable devices.

Soft Exoskeleton Designs

In contrast to stiff devices, soft exoskeletons (also called exosuits) deliver mechanical power to the user using anchors that are typically made of clothing textiles and powered by cable-driven actuation (Awad et al., 2017; Schmidt et al., 2017; Li et al., 2018; Di Natali et al., 2019; Kim et al., 2019a,b; Xiloyannis et al., 2019; Yang et al., 2019; Goršič et al., 2021; Figure 3). Furthermore, soft exoskeletons are typically ergonomic, lightweight, and often designed following bioinspired techniques to better reproduce human physiological movements (Lessard et al., 2018; Figure 3b). They can be employed in several domains, from medical/rehabilitation scenarios to improving the physical strength of healthy individuals (Kim et al., 2019a,b; Figure 3c). The same work also demonstrated how a lower limb soft exosuit can reduce the metabolic cost of walking and running. Another important advantage is that, compared to stiff designs, exosuits largely maintain the natural freedom of movement of the wearer (Wehner et al., 2013; Figure 3d). However, the preservation of the freedom of movement offered by soft exosuits, as well as their comfort, depends on ergonomically appropriate design. Indeed, several recent works have focused on the design of ergonomic soft exoskeletons to minimize user discomfort (Andrikopoulos et al., 2015; Varghese et al., 2018; Choi et al., 2019; Chiaradia et al., 2021). These studies showed how the combination of textile fibers, 3D printed plastic materials and sometimes even metal parts, should be carefully designed to offer appropriate ergonomics and avoid possible limitations of the device's functionality.

The main disadvantage of soft exoskeletons is that the mechanical power they can deliver is still limited as compared to stiff designs (Schmidt et al., 2017). This is mostly because the forces produced by soft exosuits inherently increase the shear stress on the human body. Indeed, although these devices may potentially offer higher comfort than stiff systems due to their compliant nature, soft exoskeletons often present problems of slipping and chafing at the anchor points due to the generation of shear forces at the interface point with the wearer (Asbeck et al., 2013; Yandell et al., 2017). On the other hand, the effects of the mentioned shear forces on joint loading are usually negligible: the radius over which they apply forces is significantly greater than that exploited by the biological tendon, and the joint compression force decreases as a consequence of this (Lamers et al., 2018). The shear force problem is instead minimized in stiff designs, where forces are applied orthogonally to the human segment (Yandell et al., 2017). Another common issue of soft design approaches is the friction introduced by the employed cable-driven actuation systems (e.g., Bowden-cable transmission), which typically results in low efficiency and bad control



Figure 3. Examples of soft exoskeletons: (a) the research upper limb soft exosuit from Aries Lab (Xiloyannis et al., 2019), (b) the CRUX soft upper limb exosuit (Lessard et al., 2018), (c) the research lower limb soft exosuit from the Wyss Institute (Kim et al., 2019a,b), and (d) a lower limb pneumatic soft exoskeleton (Wehner et al., 2013).

performance of the overall system if not compensated (Jeong et al., 2020). Furthermore, given their soft nature, the interface points where the human limb is anchored introduce a low-stiffness series compliance that would need high actuation power and appropriate bandwidth to be overcome (Chiaradia et al., 2019). Figure 3 shows examples of upper and lower limb soft exoskeletons.

How Can Design Choices Facilitate the Adoption of Exoskeletons?

The mentioned size/weight issues typical of stiff designs have motivated a growing number of research groups to adopt methods from the field of industrial design to truly understand user needs and design exoskeletal solutions tailored to the specific application (Schmidt et al., 2017; Laffranchi et al., 2018). In most of these works, the main user needs that still need to be overcome are diverse: reduced donning time, lighter structures, higher autonomy of use, higher freedom of movement, increased comfort, and so forth (Gorgey, 2018; Wang et al., 2018; Fritz et al., 2019; Davis et al., 2020). Researchers and companies are addressing the needs with different solutions. One approach is the utilization of high-performance materials such as carbon fiber, aluminum or aluminum–titanium alloys, which may enable the development of lighter and more flexible structures for both upper and lower limb exoskeletons, thus reducing the overall weight of the device (Frisoli et al., 2005; Meijneke et al., 2014; Bianchi et al., 2016). A few devices propose modular architectures that allow the exoskeleton structure to be disassembled into multiple small and light modules to

facilitate donning, transportation and autonomous usage by impaired subjects (Laffranchi et al., 2018; Vassallo et al., 2020; <http://www.indego.com/indego/us/en/home>). Finally, recent works show a growing interest in passive exoskeletons (Spada et al., 2017; Kim et al., 2018a,b; Van Engelhoven et al., 2018; Pacifico et al., 2020). As these devices are typically lighter than their active counterparts, they may be able to solve the weight issue. However, as their purpose is to counteract gravity in semi-static scenarios by means of passive mechanical components, they lack the versatility of active devices.

Regarding soft exoskeletons, their application is still limited to cases where the amount of needed assistance is limited—as in “low-torque” stiff exoskeletons—due to the problems mentioned in the previous section. A limited torque generation does not necessarily represent a constraint in the application of such devices, as seen in gait assistance exoskeletons where full torque compensation is not required. Nonetheless, it can be argued that considerable efforts are still needed to design appropriate ergonomics and realize systems that can effectively increase their load capacity without generating high stress on the wearer’s biological tissues (Lotti et al., 2020).

The search for novel solutions, both in stiff and soft exoskeletons, may be facilitated by careful analysis of users’ feedback and the inclusion of a diverse range of design disciplines besides engineering. Indeed, many innovative works are driven by approaches that dedicate considerable effort to understanding user needs and proposing solutions that are tailored to specific end-user needs. For example, user-centered design techniques have been applied to increase adoption of exoskeleton technologies in applications such as medical (Almenara et al., 2017; Laffranchi et al., 2018; Meyer et al., 2019; Vassallo et al., 2020), industrial (Babič et al., 2017; Davis et al., 2020), and assistive (Wang et al., 2018). Such user-centered design approaches, which may additionally incorporate industrial and ergonomics design in the development process, could serve as useful and effective tools to foster the adoption of exoskeleton technology (Hill et al., 2017).

A final consideration must be given to hybrid exoskeletons. The aim of these devices is to combine the light weight and comfort of soft exoskeletons with the generally higher torques generated by stiff solutions. Hybrid exoskeletons have been developed for both the upper and lower limb. For the upper limb, Rose and O’Malley (2019) proposed a hybrid hand exoskeleton that favors ergonomics, self-donning and doffing while satisfying the required range of motion and grasping force. Instead, for the lower limb, the reader might refer to the MyoSuit device, which demonstrated effective hip assistance without altering gait kinematics (Schmidt et al., 2017; Haufe et al., 2019). Moreover, an interesting hybrid knee exoskeleton with considerable high-torque capabilities has been proposed by Su et al., and showed contained mass properties and high backdrivability (Yu et al., 2019). As of today, there are no general and quantitative guidelines on how to combine stiff and soft elements in hybrid exoskeleton designs. In fact, each stiff or soft subcomponent embedded in a given hybrid design inherits the same pros and cons of stiff or soft designs seen in previous subsections. Therefore, the successful combination (and selection) of soft and stiff components for hybrid exoskeletons depends on the requirements of the specific application.

Actuation

Actuators are a major bottleneck for the development of active wearable robots. The demanding requirements imposed upon them include high torque density (torque per unit mass), backdrivability, compactness and, for untethered devices, high energy efficiency. The actuator’s torque density is especially important for actuators at distal joints due to the large metabolic cost associated with carrying mass in these locations (Royer and Martin, 2005; Browning et al., 2007). Highly backdrivable actuators allow the wearer to move unimpeded when the robot is not providing assistance, for example, during the swing phase of walking or during donning and doffing. Finally, wearable robots should be as unobtrusive as possible. Actuators, which tend to protrude from the device, therefore need to be designed in small and compact form factors.

The majority of actuators for wearable robotics incorporate electric motors as prime movers (Windrich et al. 2016; Gopura et al., 2016; Veale and Xie, 2016). The main competing technologies, hydraulic and pneumatic actuators, are typically less efficient and reliable, tend to be noisy, and require pressurized fluid

supplies that are difficult to make portable in an acceptable form factor (Madden, 2007; Veale and Xie, 2016). Hydraulic power units driven by combustion engines do have very high power densities, which makes them interesting when high power output is required, for example, for exoskeletons intended for load carrying and human performance augmentation (National Research Council, 2004). However, the high power density of hydraulic actuators comes at the cost of high impedance (Hollerbach et al., 1992; Zinn et al., 2004), which is undesirable for wearable robots. Overall, for most wearable robots that are not used for human performance augmentation, electric motors are considered to present the best trade-off between the different requirements.

Interestingly, in terms of power-to-mass ratio and energy efficiency, electric motors perform better than human muscles (Madden, 2007). However, in wearable robots, motors need to operate dynamically at varying torques and speeds. In these conditions, their efficiency is far below the nominal value (Verstraten et al., 2018). Moreover, the torque output of electric motors is low relative to the demands of wearable robots, which is why they are typically paired with high-reduction gearboxes. The dominant gearbox technologies in wearable robotics, harmonic gears and planetary gearboxes, offer high reduction ratios in compact sizes. However, their low efficiency and poor backdrivability are detrimental to the overall performance of the actuators and the wearable robot as a whole (Kazerooni, 1995; Nef and Lum, 2009; Lv et al., 2018; García et al., 2020).

Two paths are currently followed by the research community to tackle this problem. On one hand, novel gearbox technologies with improved characteristics are being explored, as summarized by García et al. (2020). Recent examples include the Gear Bearing Drive (Brassitos and Jalili, 2019), and the Bilateral Drive Gear (Matsuki et al. 2019). These gearboxes achieve high gear reductions at high efficiencies and in compact sizes through the use of the Wolfrom gearbox topology, combined with performance-enhancing modifications such as a carrierless design (Gear Bearing Drive) and profile shift modification (Bilateral Drive). On the other hand, quasi-direct drive has emerged as a new actuation paradigm. It combines high torque-motors with low-reduction gearboxes, resulting in energy-efficient actuators with high backdrivability (Sensinger et al., 2011). Ideally, these gearboxes are single-stage (gear ratio below 10) since efficiency is mostly related to the number of stages. Following their success in legged robots (Seok et al., 2015; Ding and Park, 2017), quasi-direct drive actuators are currently being implemented in wearable robots (Wang et al., 2018; Elery et al., 2020; Yu et al., 2020; Zhu et al., 2021).

To further compensate for the inherent limitations of commercial motors and gearboxes, researchers have resorted to specialized actuator designs. Our musculoskeletal system is built in such a way that elastic tissues in our body (tendons, aponeuroses, elastic tissues between muscle fibers, etc.) can perform a large amount of work by passive stretch and recoil (Alexander, 1990). This is possible because the joint torques applied during human movements typically exhibit a linear relationship with respect to the joint angle during consistent parts of the movement (Feldman, 1980; Hasan, 1986; Frigo et al., 1996; Toxiri et al., 2015). Engineered actuators for wearable robots, which are designed to provide assistive torques with a similar profile, can therefore also benefit from incorporating elastic elements with well-chosen stiffness. These specialized actuators are called elastic actuators.

Springs with well-designed stiffness allow actuators to deliver peak powers well beyond the capabilities of the motor itself (Paluska and Herr, 2006). Elastic actuators can therefore incorporate a smaller electric motor and gearbox, resulting in an overall reduction of weight and size compared to a traditional, stiff actuator. Moreover, springs offer the possibility to mechanically store energy when the actuator performs negative work and release it when positive work is required. As such, the average power flow through the actuator is reduced, resulting in an improvement in energy efficiency of the actuator (Hollander et al., 2006; Verstraten et al., 2017; Bolivar Nieto et al., 2019).

Elastic actuators can be subdivided into two main categories based on the placement of the spring element. By adding a (unidirectional) parallel spring to a stiff motor, one obtains a parallel elastic actuator (PEA). In this case, the parallel spring takes over a part of the load, leading to a reduction of the motor torque. The required motor speed remains unchanged since the motor is directly coupled to the output. Another option is to place the spring in between the motor and load, that is, in a series configuration. This type of actuator is called a series elastic actuator (SEA). In contrast to the PEA, a well-tuned spring lowers the speed

required from the motor while the motor torque is unaffected. Although SEA and PEA provide similar levels of peak power amplification and increased energy efficiency during locomotion tasks (Verstraten et al., 2017), the PEA has the specific ability to compensate for static torques by tuning the equilibrium angle of the spring (Beckerle et al., 2016). This makes the PEA particularly suitable for upper-body exoskeletons that assist with overhead work or back-support exoskeletons for lifting tasks. Here, the parallel spring can provide the torque required to support the weight of the arms or torso as well as the load.

The performance of elastic actuators can be further enhanced by adding locking mechanisms and clutches (Plooiij et al., 2017). Passive occupational exoskeletons often allow the user to disengage the spring providing the assistance so that it does not impede the user by generating a torque when it is not needed (Zhang and Huang, 2018). In a similar fashion, since little torque is required during the swing phase of gait, lower limb exoskeletons and prostheses can benefit from clutched PEAs (Haeufle et al., 2012; Geeroms et al., 2017) and SEAs (Rouse et al., 2014; Convens et al., 2019) that only engage the spring during stance. In robotic ankles, a catapult-like mechanism consisting of a spring and locking mechanism could be used to deliver the explosive burst of power required during push-off in an energy-efficient way (Collins and Kuo, 2010; Chernelle et al., 2014). Clutches have also been implemented in new actuator designs with multiple springs, which can be recruited progressively to contribute to the output torque (Mathijssen et al., 2015; Krimsky and Collins, 2020).

Most springs are manufactured to exhibit a linear relationship between torque and deflection. However, in wearable robotics, the required torques are rarely perfectly linear with respect to the output angle over the entire operating range of motion. The joint stiffness profiles required for walking, for example, exhibit low stiffness at low joint flexion and high stiffness at high joint flexion (Seyfarth et al., 2006). A closer match between the spring characteristic and the required torque can be achieved by means of nonlinear transmissions. The simplest way to generate a nonlinear transmission is by providing torque to the joint through a lever (Sugar and Holgate, 2013). The linear actuators driving the lever often consist of brushless DC motors combined with ball screw drives, a solution which offers relatively high reductions at high efficiencies. Other common nonlinear transmissions are cams (Realmuto et al., 2015; Gao et al., 2018) and four-bar mechanisms (Bergelin and Voglewede, 2012; Hyun et al., 2017).

Specialized actuator designs achieve impressive performance by relying on springs and nonlinear transmissions optimized for a specific task. The downside of this approach is significantly worse performance in off-design conditions. This issue can be mitigated by variable stiffness springs (Vanderborght et al., 2013; Shepherd and Rouse, 2017), variable transmissions (Mooney and Herr, 2013; Tran et al., 2019), multispeed gearboxes (Sugar and Holgate, 2013) or multimotor solutions (Verstraten et al., 2019), which allow the relationship between the required output torque and the spring characteristic to be shaped depending on the task. However, these solutions require additional motors or mechanisms, which increase the weight of the actuator. The metabolic penalty resulting from the additional weight can be avoided by remote actuation through, for example, cable- or wire-based transmissions, as implemented in several rehabilitation exoskeletons (Veneman et al., 2006; Mao and Agrawal, 2012). In portable devices, these transmissions can decrease metabolic cost by enabling a more proximal placement of the actuators, near the human's center of mass (Asbeck et al., 2015).

Motor and gearbox technologies continue to improve and these improvements will promote further adoption of wearable robots in the near future. However, commercial devices for specific and simple tasks (e.g., occupational exoskeletons) may very well continue to rely on passive solutions until actuators with superior torque densities and energy efficiencies become available. The major breakthrough may arrive through the development of radically new actuator technologies such as dielectric elastomer actuators (Veale and Xie, 2016). Until these technologies reach a mature state, we expect that advances will most likely come from specialized actuator designs as those described in this section.

Sensing

Wearable robots require advanced sensors so that they can determine the user's actual and intended motions, determine the state of the robot itself, and use this information as a basis for robot control. Any

limitations in the sensing lead to suboptimal control, which limits the overall performance of the wearable robot. These limitations can occur at one of two levels: at the lower level of individual sensors and at the higher level of multisensor fusion.

Sensor Hardware and Low-Level Processing

While wearable robot sensors suffered from significant noise-related issues in the past, extensive work has been done to address them. Challenges on the level of individual sensors now primarily involve addressing remaining gaps in knowledge related to specific sensor modalities, particularly those that are just becoming widespread in the field.

Wearable robot systems commonly include standard robot sensors integrated into the robot, such as joint encoders and strain gauges (Huang et al., 2019; Singer et al., 2020). Assuming that the robot is tightly attached to the wearer and its morphology matches the wearer, we can assume that human body kinematics can be obtained from robot body kinematics via a constant transformation (Sposito et al., 2020). Commonly, pressure sensors are also built into points where the robot touches the wearer, allowing interaction forces between the wearer and robot to be measured. While these sensors have traditionally been rigid, soft sensors are becoming increasingly common in wearable robotics, potentially providing better robustness and comfort. For example, Chiaradia et al. (2018) used soft strain sensors to monitor robot joint angles instead of traditional rigid encoders, and Kim et al. (2019a,b) combined deep learning with a full suit of soft strain sensors to extract full-body kinematics. Most soft strain sensors are resistive (i.e., change resistance in response to strain), though capacitive and optical sensors have also been developed for wearable robotics (Atalay et al., 2017; Jin et al., 2020; Kim et al., 2020). Such soft sensors will probably never entirely replace rigid sensors (since soft technologies are not needed in all applications), but are likely to be a major aspect of the next generation of wearable robotics. One limitation they currently face is that the soft sensors generally still need to be connected to “hard” electronics, though such electronic components will likely become more integrated into soft materials over the next few years as well (Sanchez et al., 2021). The weakness of sensors built into the robot is that they only obtain knowledge about wearer body parts that are attached to the wearable robot and that they require the wearer to either already move the robot (in the case of position sensors) or apply forces to the robot (in the case of force sensors) before the robot can react, reducing its assistive benefits. The latter weakness can be reduced by reducing the amount of wearer force/movement required to trigger robotic assistance, but this may cause false positives (Novak and Riener, 2015; Tucker et al., 2015). To obtain a more complete picture of the wearer, wearable robots are thus commonly augmented with two types of sensors: inertial measurement units (IMUs) and electromyography (EMG) sensors.

IMUs consist of accelerometers, gyroscopes, and (optionally) magnetometers, whose outputs can be combined to obtain IMU orientation. IMU technology is cheap and can be easily attached to wearers; thus, it is commonly placed on parts of the body not connected to the wearable robot—for example, in a leg exoskeleton, an IMU can be placed on the arm or trunk (Lazzaroni et al., 2020). Alternatively, the IMUs can be embedded in the robot itself, providing an alternative to classic robot sensors (Naseri et al., 2020). As the individual components of the IMU tend to have noisy outputs, obtaining IMU orientation requires the use of techniques such as Kalman filtering (Nazarahari and Rouhani, 2021). IMUs that incorporate magnetometers are generally accurate in magnetically homogenous environments, as the magnetometer provides a reference estimate of magnetic north. In real-world environments with magnetic disturbances, however, IMU noise increases significantly and requires the use of advanced compensation algorithms (Slajpah et al., 2017; Wittmann et al., 2019). These algorithms are present in high-grade IMU systems and can be transferred to lower-cost IMUs, but are not present in such off-the-shelf devices by default.

EMG sensors measure the electrical activity of human muscles using electrodes placed on the skin. EMG levels are proportional to muscle contraction levels, but appear prior to mechanical contraction, allowing the wearer’s intentions to be anticipated (Novak and Riener, 2015). EMG can also be combined with measurements such as joint angles to obtain estimates of muscle dynamics and torque (Lotti et al., 2020). Furthermore, if the wearer is too weak to move their limb themselves, EMG may still allow motion

intention to be detected and the wearable robot to be moved accordingly (Yun et al., 2020). Additionally, EMG can be used as an effectiveness measure of the wearable robot: since wearable robots are meant to support the wearer and reduce human workload, reductions in EMG level can be considered proportional to reductions in human muscle demand (Goršič et al., 2021; Kermavnar et al., 2021). However, EMG measurements also suffer from significant variability, both intersubject (due to, e.g., different fitness levels) and intrasubject (due to, e.g., fatigue and sweat); they are also prone to noise due to mechanical movement and electromagnetic interference. Methods to reduce noise have been demonstrated in the literature (Gokcesu et al., 2019), but not adequately transferred to wearable robotics. Similarly, alternatives to EMG such as mechanomyography and force myography (Xiao and Menon, 2019) have promise as a less noisy measurement approach, but have seen limited use in wearable robotics.

Aside from these common sensor modalities, many others have been used in wearable robotics to a lesser degree but suffer from low accuracy, difficulty of incorporating into real-time control, or simply lack of evidence. For example, brain measurements (e.g., electroencephalography [EEG]) can be used to infer the wearer's intended motions without any muscle activation, which could be used to control wearable robots for people with disabilities who cannot move on their own. Unfortunately, such brain measurements provide only limited information (e.g., left vs. right arm movement), and their utility as a stand-alone method is thus questionable except for severely paralyzed individuals (AL-Quraishi et al., 2018); however, they could potentially be combined with more popular modalities such as EMG (Li et al., 2019). As another example, heart rate, respiration and other physiological measures are commonly used as a metric of wearer workload (physical or mental) and thus robot performance (Torricelli et al., 2020), and could be used to adapt the amount of robot assistance; however, while studies have suggested complex approaches to doing this (Schürmann et al., 2019), only limited work in this area has been done, mostly with stationary wearable robots (e.g., the Lokomat) (Koenig et al., 2011a,b).

Finally, a particularly promising case of underexplored sensors should be acknowledged: all the above sensors are meant to obtain information about the wearer. However, sensors for wearable robots should also obtain information about the environment—for example, possible obstacles in the path of a person wearing a leg exoskeleton, or objects that a person wearing an arm exoskeleton may wish to manipulate. Work on such “exteroceptive” sensors has been done in multiple areas of wearable robotics. For example, in lower limb prosthetics, researchers have used depth sensors to identify obstacles in the user's path (Liu et al., 2016; Zhang et al., 2019) and segment stairs in the user's path (Krausz et al., 2015); similar approaches have been used in lower limb exoskeletons (Laschowski et al., 2020) and upper limb prosthetics (Shi et al., 2020; Mouchoux et al., 2021). Nonetheless, further work is critically needed to incorporate environment sensing into wearable robot sensing and control.

Sensor Fusion

Sensor fusion refers to the process of combining data from multiple sensors into an estimate of the user's motor intention or physical state; the main challenge with such fusion is that existing methods are not sufficiently robust for real-world use. Existing algorithms range from simple thresholding (e.g., activate assistance if user bends forward) to more complex approaches based on classification and regression using machine learning (Novak and Riener, 2015; Tucker et al., 2015). For example, classifiers can learn to identify desired gestures from EMG-based on a training dataset of previously recorded and manually labeled EMG data (Yun et al., 2020); similarly, regression algorithms can learn to estimate desired robot torque from ground reaction force (Naseri et al., 2020) or EMG (Gui et al., 2019).

Such classification and regression have achieved high accuracies in laboratory experiments; however, the high accuracies in laboratory experiments (where conditions are largely stable and only a small set of wearer behaviors is expected) tend to not generalize to more chaotic real-world environments. For example, a wearable robot designed to support lifting tasks can, in a laboratory environment, simply classify wearer behavior as “lifting” or “not lifting,” and provide assistance only when the “lifting” class is detected (Hlucny and Novak, 2020). In the real world, however, nonlifting behaviors can include a variety of actions, many of which may appear similar to lifting (e.g., squatting or stretching the arms). One

possible solution is to first classify a common denominator movement, such as a pre-lift and in later stages distinguish lifting versus squatting (Jamšek et al., 2020). Alternatively, a multilevel sensor fusion algorithm is needed to first categorize the wearer's behavior into broad classes, then more specific actions—for example, the three-level sensor fusion and control system proposed by Lazzaroni et al. (2020) for back exoskeletons or the multilevel stumble detection system proposed by Zhang et al. (2011) for artificial legs. In such multilevel sensor fusion systems, the different levels may be based on different sensor types—for example, using information about walking terrain to switch between different EMG-based classifiers in a prosthesis (Liu et al., 2016).

Additionally, advanced sensor fusion for wearable robots will likely need to incorporate aspects such as active and semi-supervised learning—the wearer may perform an action that the robot does not understand, and the robot must then decide whether to actively query the wearer about what they are trying to do (which may interrupt the intended action) or try to identify the action on its own (which may be inaccurate). In both cases, the robot then needs to adapt its sensor fusion rules so that it can deal with this new motion in the future. This may involve a significant but reasonable adaptation—for example, if the baseline of one of the sensors has drifted and the sensor data are thus a shifted version of previously seen data. In such cases, several authors have presented online learning strategies that can adapt robot sensor fusion rules (Spanias et al., 2018; Gui et al., 2019; Hu et al., 2019; Huang et al., 2019; Zheng et al., 2019). It may, however, involve a major change to the sensor fusion rules—for example, the observed action may be an entirely new type of motion, and the robot may then need to add it to its “dictionary” of possible motions. This challenge of dealing with numerous possible wearer behaviors, some of which may be unknown to the robot, represents a major challenge in wearable robotics, and improvements in this area have an immense potential.

Finally, advanced sensor fusion for wearable robots will need to consider advanced planning that does not rely on physical contact between the robot and environment. Such advanced planning can be done using, for example, the depth sensors discussed in the previous section, but remains largely underexplored (Krausz et al., 2019).

Wearable robots have made enormous strides in sensing hardware and software. Sensors can be unobtrusively integrated into the robot and/or placed on the wearer, and information from diverse sensors can be integrated into a reasonably accurate estimate of the wearer's current and intended activity as well as physical state. While improvements can be made on the level of individual sensors (e.g., reducing IMU noise in magnetically nonhomogenous environments), we believe that the greatest gains can be made through broader adoption of less popular sensor modalities (e.g., environment sensors) and on the level of sensor fusion (e.g., multilevel sensor fusion and semisupervised learning). Such improvements will allow wearable robots to be effectively used in dynamic, unpredictable environments where they can have a major effect on human wellbeing and productivity.

Body Interface

Wearable robots cannot provide any benefits if humans do not use them. State-of-the-art wearable robots can effectively reduce workload; however, limitations related to user comfort and acceptance are a major remaining barrier to widespread use (Kermavnar et al., 2021), together with high cost. It was recognized in early exoskeleton research that a certain amount of discomfort was experienced by the users (Abdoli-E and Stevenson, 2008; Wehner et al., 2009), while newer models have been reported to minimize this discomfort to a tolerable level (Sirawattanakul and Sanngoen, 2020; Kermavnar et al., 2021). Some wearable robots are heavy and bulky (Stadler et al., 2016; Näf et al., 2018; García et al., 2020), which can interfere with the user's freedom of movement. Even with lighter models, a certain level of discomfort and movement hindrance has been reported (Baltrusch et al., 2018; Goršič et al. 2020). Inevitably, the unloading of one part of the body results in an increased loading (i.e., force) on another. In spinal exoskeletons, the reduction of the load to the back is typically achieved via increased loading to the lower limbs, which can be sometimes reflected in increased leg muscle activity (Luger et al., 2021) and discomfort to the chest area (Hensel and Keil, 2019), although back exoskeletons may also reduce leg

activity (Bosch et al., 2016; Jeong et al., 2020). In comparison, upper limb exoskeleton interfaces can increase the loading to the spine (Weston et al., 2018). This trade-off, as well as the effects of wearable robots on task performance, discomfort and perceived freedom of movement, must be taken into account when designing wearable robots and studying their effects. While these functional and subjective aspects of wearable robots are increasingly being addressed, research is predominantly done in laboratory environments with healthy young male participants (Kermavnar et al., 2021). Therefore, on-field research with different groups is needed to support the development of body interfaces, suitable for broader population. Interface design has also begun to address the issue of load transfer to other body parts. One possible way to address this is by extending the legs of the exoskeleton all the way to the ground. For instance, a recent prototype of a semi-active exoskeleton, designed in this fashion, has been reported to decrease the activity of both back and leg muscles (Wang et al., 2021). Novel exoskeletons such as the Guardian XO (CRI Writers, 2018), are also being developed to support immense loads without stressing any human body parts. On the other hand, back exoskeletons with good design seem to decrease the load to the legs in addition to unloading the back (Bosch et al., 2016; Jeong et al., 2020).

Exoskeleton interfaces should enable efficient transmission of mechanical power without compromising user comfort. Langlois et al. (2018) have proposed three crucial design requirements that need to be fulfilled to achieve this. First, the attachment points of the device on the human body segments should be as distant to the joint as possible to reduce interaction forces (e.g., attachment points of the hip exoskeleton on the thighs should be as close to the knees as possible). At the same time, contact areas should generally be maximized. However, caution is needed as some parts of the body may be more sensitive to pressure and increasing the surface area may not be desirable in such cases. Next, the exoskeleton should be designed in a way that prevents interface migration (e.g., sliding along the user's body), as misalignments caused by sliding commonly reduce power transmission efficiency (Langlois et al., 2018). Power transmission is also increased with larger contact point stiffness (Asbeck et al., 2015; Langlois et al., 2020). This can be achieved by placing the contact points over the bony prominences, such as tibial surface in case of lower limb exoskeletons (Mooney and Herr, 2016) or pelvis bones in case of trunk exoskeletons (Näf et al., 2018). Alternatively, the pads can be tightly attached to the user, but this can cause discomfort. Keeping in mind the importance of contact point placement, prevention of interface migration and contact stiffness, the compactness of the design should not be overlooked. For instance, the additional inertia due to the device's weight can increase actuation demands (Langlois et al., 2018).

Wearable robots for industrial environments are usually designed primarily with a focus on load reduction during strenuous physical tasks such as load lifting, overhead work and trunk bending (de Looze et al., 2016; Bao et al., 2019). However, to allow efficient work and freedom of movement, the interface should not interfere heavily with other movements such as walking, stair climbing and sitting down or standing up. Indeed, wearable robots often decrease performance in tasks that require quickness and range of motion (Kermavnar et al., 2020). While these aspects were partially addressed in earlier research (Wehner et al., 2009; Bosch et al., 2016), assessments involving subjective reports have only recently been introduced (Kozinc et al., 2020a). Such subjective reports allow potential drawbacks of specific wearable robots, especially those related to the interface and discomfort, to be identified. For instance, specific contact points that exert pressure on the user or limit their range of motion can be identified and considered in subsequent design (Abdoli-Eramaki et al., 2007; Baltrusch et al., 2018; Baltrusch et al., 2021; Kozinc et al., 2020b). Identifying the drawbacks of state-of-the-art devices should help developers optimize subsequent models. One recent example of approaches to improve the exoskeleton interface is found in the SPEXOR spinal exoskeleton. Specifically, a self-aligning mechanism was integrated into the hip joint, which prevents the sliding of the interface (Näf et al., 2018; Babič et al., 2019). With such improvements, exoskeleton interfaces could become more user-friendly. Another possible consideration for interface improvement is the option to adjust and switch off the exoskeleton's support (Babič et al., 2019; Wang et al., 2021), which allows the user to freely and efficiently perform the tasks that do not require support (e.g., walking).

As mentioned, the unloading of one part of the body inevitably results in an increased loading somewhere else. For instance, the mechanism by which spinal exoskeletons contribute to reduced

workload is generation of torque around the spine and/or the hips (Bosch et al., 2016; Stadler et al., 2016; de Looze et al., 2017; Näf et al., 2018; Babič et al., 2019; Kermavnar et al., 2020). Even though spinal exoskeletons are designed to compensate for relatively small amounts of required spinal torque (Näf et al., 2018), this unloading of the spine results in increased levels of force (and consequently pressure) on the exoskeleton attachment sites on the body (i.e., distal thighs and shoulder girdle). Similar observations have been reported for upper limb (Huysamen et al., 2018) and lower limb (Luger et al., 2019) exoskeletons. Therefore, it is crucial to consider discomfort and pain tolerance thresholds at typical attachment sites for wearable robotic devices, and to design the devices in a way that minimize the chances of these thresholds being reached. While many studies have investigated pain thresholds with algometry (Sterling, 2011), their relevance to wearable robotics is limited, as the surface of force application in case of standard algometry is typically set at 1 cm² (Graven-Nielsen and Arendt-Nielsen, 2010). Potentially more useful evidence is found in studies using inflatable cuffs to induce pressure at different sites, typical for attachments of wearable robots to the human body (Kermavnar et al., 2018). There are only two studies to date that explore this issue directly. Yandell et al. (2020) applied the pressure to the human body with fabricated exosuit interfaces for the shoulders, thigh, and shank, and thereby provided normative data for discomfort thresholds. Similarly, Kozinc et al. (2021) applied a pressure using exoskeleton pads and straps to the thigh, pelvic and chest/shoulder areas. Exoskeleton developers are encouraged to consider these values during the design of wearable robots. One of the most important findings of both studies is the large (up to 10-fold) interindividual variability in discomfort thresholds. Thus, further studies should critically include sufficient sample sizes to encompass this variability. Additionally, an issue that remains largely unexplored is discomfort during longer-term use of exoskeletons (i.e., over a workday), where higher discomfort or pain levels might be seen due to temporal summation (Nie et al., 2005).

It is unquestionable that wearable robots have the potential to increase work performance and reduce physical loads to the body. Some exoskeletons with inferior design and poor kinematic structures and human interfaces have been shown to reduce freedom of movement and increase discomfort, which hinders widespread use. On the other hand, several models have achieved high levels of support without compromising overall user experience. Recent advances in methods to assess functional performance and discomfort, as well as the availability of normative data on discomfort tolerance, should help the developers to design better interfaces in the future.

Control

When synthesizing controllers for advanced robotic systems, it is common to create a hierarchical controller scheme where high-, mid-, and low-level control actions are locally isolated from each other, thus providing a broader understanding of the control problem in question. Wearable robots and exoskeleton systems are no exceptions to that; it is important to separately plan these controllers on different levels so that interrelations between high-level decision-making, mid-level planning and low-level actuator control can be clearly identified (Jimenez-Fabian and Verlinden, 2012; Tucker et al., 2015). In this section, we succinctly cover low- and mid-level controllers, while high-level control is discussed in the next section.

Low-Level Control

The central problem in low-level control is to achieve high-fidelity servo tracking for actuators. This problem can be addressed via two different strategies: (a) precise position control through the use of nonbackdrivable mechanisms and high-gain controllers, and (b) force and torque control. The latter has certain advantages in terms of safe and dependable physical human–robot interaction (Vallery et al., 2008) while the former can guarantee performance metrics if trajectory tracking is prioritized (Ohnishi et al., 1996).

On one hand, force-controlled systems may require hardware-level modifications in actuators, for example, deployment of custom-built springs or strain gauges (Kashiri et al., 2017) and thus lead to

hardware complexity. On the other hand, position-controlled systems generally possess high mechanical impedance at the output and are therefore deemed inappropriate for wearable robots and exoskeleton systems other than except rehabilitation procedures such as passive trajectory tracking tasks (Brahmi et al., 2019). Although the list of pros and cons can be further extended, force controllability is essential to establish dependable human–robot interaction control (Haddadin et al., 2009).

To implement force control at the actuator level, most systems employ SEAs in which a spring is attached between the mechanical output and motor (Pratt and Williamson, 1995); see the Section “Actuation.” The direct use of spring deflection feedback leads to instability issues (Pratt et al., 2004). To obtain fine torque tracking with guaranteed stability, one could employ cascade control strategies (Vallery et al., 2007), sliding mode with disturbance observer (Sariyildiz et al., 2020), or other observer-based robust control methods that are readily available for online simulation (Ozcinar and Ugurlu, 2021; Ugurlu et al. 2021). To this end, observer-based techniques were proven to be effective in handling perturbations due to uncertainties and unknown environmental effects (Haninger et al., 2020).

The performance of low-level force control depends on several factors: for example, actuator bandwidth, whole-body system dynamics, sensor signal quality, inner loop (current, pressure, etc.) control bandwidth, unmodeled dynamics, and sampling rate. Yet, given a completed hardware setup, one of the major contributions of a control engineer is to obtain a sufficient description of actuator dynamics. Once a sufficiently descriptive mathematical model is obtained, one can achieve fine force tracking for robot joints with pneumatic muscles (Ugurlu et al., 2019), SEAs (Vallery et al., 2007, Yildirim et al., 2019), hydraulic systems (Otten et al., 2015) and variable impedance actuators (Kaneishi et al., 2018). Through a detailed mathematical model, sensorless torque control is possible even for robot joints with nonbackdrivable mechanisms (Ugurlu et al., 2015). With this in mind, we argue that the implementation of force control is possible for most of systems to an extent, although it may not be a trivial task.

In summary, low-level force control at the actuator level became a standard in the wearable robotics community and there is a significant amount of literature on the topic (van der Kooij et al., 2006; Veale and Xie, 2016). The use of SEAs is arguably the most common approach owing to its relatively simple mathematical model and practical feasibility. Regardless of the actuator type, it can be argued that there is no open research question in relation to low-level control; however, engineering problems do persist. For instance, the bandwidth of mid-level control strongly depends on the low-level controller (Buerger and Hogan, 2007); thus, physical low-level controllers with deterministic real-time capability and higher sampling frequencies (~5 kHz) are required. Moreover, the integration of force sensing elements always introduces certain technical challenges to be overcome (Ham et al., 2009; Kim et al., 2018a,b; Yildirim et al., 2021).

Mid-Level Control

Mid-level controllers are mostly task-specific and built on the basis of realizable actuator commands. To this end, the output of a mid-level controller is usually a link torque or position command that must be attained by the low-level controller. As nearly all wearable robotics applications involve physical interaction, impedance/admittance control is widely used (Hogan, 1985). In basic terms, an impedance controller outputs a certain amount of force command to render a given mechanical impedance, thus requiring force control. An admittance controller performs the same duty; instead, its transfer function is the inverse of impedance controller and it outputs a certain amount of displacement that corresponds to the aforementioned force command to render mechanical impedance. In practical applications, admittance control is preferred mostly when force-controllable actuators are not available or the torque command is not accessible; however, its tuning is more challenging and usually entails complex algorithms (Zheng et al., 2019).

Regardless of the impedance/admittance formulation, both controllers enable us to manage the trade-off between position and force tracking through the rendered output impedance. Hence, wearable robots for walking assistance need to possess active impedance control to cope with ground reaction during touch-down (Veneman et al., 2007). When employed via impedance approach with highly backdrivable

robots (Lv et al., 2018), or via admittance approach with nonbackdrivable robots (Ugurlu et al., 2020), the locomotion control performance sufficiently increases. However, in these locomotion scenarios, human subjects are considered to be passive and the robot completely handles the walking behavior, regardless of the user-intended forces.

When it comes to robot-aided rehabilitation in general, training paradigms are constructed either in assistive or resistive configurations (Marchal-Crespo and Reinkensmeyer, 2009). Assistive controllers usually exert additional torques in a way so as to help the user to complete the task. For instance, force field controllers can be used to attain assist-as-needed strategies (Banala et al., 2009). Intuitively speaking, an assistive controller is always necessary to help the user to carry out the desired tasks. On the contrary, some researchers argue that resistive controllers can provoke neuro-motor plasticity and are thus more favorable in robot-aided rehabilitation for patients that possess some degree of strength and control (Washabaugh et al., 2019). From a control engineering perspective, both strategies can be effectively achieved using various forms of impedance control and/or rendered force fields.

The mid-level control strategies vary significantly across different disciplines since impedance and force control techniques are generally tailored to meet task-specific requirements. These requirements vary from feasible and dynamically equilibrated locomotion controllers (Ugurlu et al., 2016; Ugurlu et al., 2020) to minimization of a user's deficit in human-machine interaction (Fitzsimons et al., 2019). Therefore, synthesis of a mid-level controller requires the determination of task-specific needs in accordance with the target use of the wearable robot. The reader can refer to review articles for further information in this matter (Marchal-Crespo and Reinkensmeyer, 2009; Yan et al., 2015; Proietti et al., 2016; Tucker et al., 2016; Chaparro-Cárdenas et al., 2018).

Human-Robot Interfacing and Coadaptation

Low-level control of wearable hardware is an essential basis for practically applicable exoskeletons. Moreover, exoskeletons can only successfully perform various tasks in synergy with their human users if they are able to adapt to the user while considering the fact that the user simultaneously adapts to the exoskeleton. To do this, a high-level control system is required to infer human intention and create and adapt reference actions to make sure that the desired task is executed as intended by the user and assisted as intended by the exoskeleton. On the other hand, for the human to understand the exoskeleton's behavior, it is also important that it behaves consistently and predictably; thus, the high-level control system needs to be designed with human-in-the-loop aspects in mind.

If a wearable device, such as an exoskeleton, is to perform actions to accommodate and assist humans in performing their tasks, the high-level control system should be able to extract the human user's intention. Several signals can be employed to infer the intention, and the high-level control system thus tightly connects to the sensing hardware and methods. While the Sensing section focused on how to obtain the signals, this section focuses on how to use the signals in the high-level controller in order to achieve effective interfacing and coadaptation between the user and the exoskeleton.

Since the human and wearable device are physically coupled at least to some degree, haptic sensing is a common choice to detect the intended actions of the user. Such signals can be obtained by force/torque sensors or pressure sensors that are mounted on or inside the exoskeleton and can measure the interaction forces between its structure and the user's body (Kazerooni and Guo, 1993; Nagai et al., 1998; Kong and Jeon, 2006; Lee et al., 2014). When the exoskeleton contains elastic elements like springs, the interaction force and thus control intention can also be estimated indirectly through the position of the spring (Pratt et al., 2004). An alternative is to simply use the measured position or motion induced by the user (Kazerooni et al., 2006). Position and motion states can also be used to detect the human intention and then play out a specific predefined feedforward behavior of an exoskeleton (Kawamoto et al., 2003a; Varol et al., 2010; Li and Hsiao-Wecksler, 2013). Information about the Center or Pressure or Center of Gravity can be used to a similar end for walking tasks (Farris et al., 2014).

Another common type of signals used for high-level control involves measuring the human body's electrical activity: for example, muscle activity by EMG (Rosen et al., 2001; Kawamoto et al., 2003b;

Fleischer and Hommel, 2008; Huang et al., 2009; Kiguchi and Hayashi, 2012), brain activity by EEG (Soekadar et al., 2015; AL-Quraishi et al., 2018) and eye activity by electrooculography (EOG) (Soekadar et al., 2015). Such electrical signals are advantageous since they can provide the information slightly prior to the actual human movement and potentially even if the human cannot make a physical movement (Fleischer and Hommel, 2008). On the other hand, their main disadvantage compared to haptic or position signals is that they require probes to be attached to the user, which usually comes with complex equipment and time-consuming calibration procedures.

Based on the detected human intention, the exoskeleton can devise appropriate physical actions to adapt and assist the users in their tasks. The simplest method is to amplify a signal that represents the user's effort. For example, the measured human interaction force in a given direction can be amplified by the exoskeleton to assist the motion in that direction (Kazerooni et al., 2006; Kong and Jeon, 2006). If the user induces motion in the desired direction, then the exoskeleton can derive the assistive movement through a dynamical model (Kazerooni et al., 2006). In case of EMG, the intention signal (i.e., muscle activation) first has to be transformed into force by biomechanical models of muscles before it can be amplified (Rosen et al., 2001; Fleischer and Hommel, 2008). However, biomechanical models are user-dependent and can be hard to obtain. In some cases, a proportional mapping between EMG-derived muscle activation and the resulting exoskeleton assistive force can be sufficient for practical applications (Kawamoto et al., 2003b; Ferris et al., 2006; Lenzi et al., 2012). In complex tasks like locomotion, effort amplification alone might not be sufficient and dynamical models might be required; furthermore, different phases of the task may require distinct control modes (e.g., separate mode for a stance phase and a swing phase) (Kazerooni et al., 2006).

If the high-level control system of an exoskeleton has an insight into more complex human biomechanical models, it can exploit them and consequently optimize specific features during assistance (i.e., besides generally reducing the user's effort), thus making task execution more ergonomic. For example, limb and muscle manipulability properties change the ability of humans to produce endpoint forces and motions in different directions of Cartesian space. For example, a lower limb exoskeleton can use geometric limb manipulability to optimize for energy efficiency in walking (Kim et al., 2010). In another example, the muscle manipulability model can be used by an arm exoskeleton to keep the force and motion production capacity equal in all directions irrespectively of human arm configuration (Petrič et al., 2019).

More complex movements can be encoded by trajectories; however, they typically have to be preplanned or be of repetitive nature, like in rehabilitation scenarios (Tsagarakis and Caldwell, 2003; Riener et al., 2005). There are numerous ways to encode the trajectories, ranging from simple spline methods to more sophisticated methods involving nonlinear oscillators (Ronsse et al., 2011; Aguirre-Ollinger, 2013) or dynamical movement primitives (DMPs) (Huang et al., 2016; Peternel et al., 2016; Lauretti et al., 2018). These trajectories can be learned from EMG feedback (Peternel et al., 2016) or from recorded motion directly (Huang et al., 2016; Lauretti et al., 2018). However, unlike effort-amplification methods, methods that merely involve repetition of learned trajectories do not enable the desired online and real-time adaptation to the user.

It has been shown that the human neuromechanical system can adapt to exoskeletons (Galle et al., 2013; Gordon et al., 2013). Exoskeletons that to adapt to the human user can adaptively update their assistive behavior (e.g., joint torques or motion) to minimize the user's effort (e.g., measured by muscle activity or interaction force) (Kawamoto et al., 2003b; Krebs et al., 2003; Riener et al., 2005; Balasubramanian et al., 2008; Huang et al., 2016; Peternel et al., 2016; Teramae et al., 2018). For example, in Peternel et al. (2016), torque trajectories are increased as long as the user has to produce some effort to perform the given task (e.g., movement); therefore, the learned trajectories in a steady state compensate for all underlying dynamics (i.e., human, robot, and task). If the dynamics change, the user has to compensate for the change to perform the given task, and the changed human effort signal again reshapes the trajectories in real-time, thus achieving coadaptation between the user and the exoskeleton (Peternel et al., 2016). Alternatively, reinforcement learning can be applied to optimize learned trajectories based on some specific metrics while the user performs the task (Huang et al., 2016). The main advantages of

adaptive learning methods are that they can reshape complex trajectories in a real-time manner and, unlike direct effort amplification, do not require a constant user effort when the task movement or force is repetitive (Peternel et al., 2016).

As outlined above, several major fundamental principles for high-level control of wearable devices have been proposed in past decades. We predict that future research will mainly focus on improving upon existing fundamental principles, as has been already done in the recent years. Most of the fundamental human–robot coadaptation control methods still require various improvements in terms of processing speed and accuracy, control intelligence and adaptability, and design practicality in order to make them effective in realistic real-world scenarios. Nevertheless, some methods are also inherently reliant on research in complementary fields, especially sensing-related research to obtain meaningful and information-rich intention signals that can be used by the control system. For example, decoding brain activity (either through EEG or some other method) heavily relies on developments in the field of neuroscience, while EMG-based and model-based methods rely on developments in the fields of human motor control and biomechanical modeling. We believe that the most powerful and convenient way for the high-level control system to govern user-exoskeleton coadaptation would be to have a complete and detailed understanding of brain functions and a more practical means to measure it. However, as of now, we are still quite far from this point.

Benchmarking

Demonstrating that a device is working properly for a given application and communicating this information to end-users and stakeholders is an essential task when developing a new product. Very often, the performance of an exoskeleton prototype is considered “demonstrated” after the execution of a few experiments in a laboratory using ad hoc sets of metrics, which are usually established by the same researchers involved in the development process. This common practice should be seriously reconsidered if the wearable robot community wants to promote rigorous comparison across existing systems as a basis of optimized and efficient product development (Torricelli et al., 2020), and benchmarking may help in this process.

Benchmarking is a comparative evaluation process often used by large companies to measure and improve the quality of products, services or processes. By means of comparisons with a standard, or benchmark, it is possible to identify the gaps in the state of the art and find strategies to improve competitiveness. But how can we translate this concept to wearable robotics? What features should be the object of comparison? Which devices or behaviors should be considered as standard or reference? The main problem is that exoskeleton research is still at an early stage, with hundreds of solutions at low technology readiness levels and only a few commercial solutions with unclear performance levels. One of the main mistakes in comparing robotic systems is to think in terms of “this robot is better (or worse) than that robot.” This viewpoint, common in competitions, does not reflect the multifactorial nature of performance. For instance, a faster system is not necessarily better than a slower one, and similar principles apply to many variables and features that may be used to describe the performance of a system. Nonetheless, comparing a characteristic of a system with the same characteristic of another system is still a very valuable process in order to understand why that system is behaving in a certain way. And this is what benchmarking is about: comparing to understand.

Recent efforts in the wearable robotics community made apparent the difficulty of defining the performance of a system on simple scales, triggering interesting discussions on the terminology to be used to define a given aspect of performance (Torricelli et al., 2020). What we understood from these discussions is that the evaluation of performance is strongly dependent on the robotic technology under examination. For instance, “interaction” may assume very different meanings if applied to an industrial manipulator rather than to a wearable robot. This also applies to other system abilities, for example, “adaptability” and “decisional autonomy.” One of the first steps in benchmarking is to agree on the definition of these concepts with respect to the specific technology considered. The motor function and the application domain are two additional factors that influence the definition of performance. For

instance, would “stability” mean the same for rehabilitation and industrial robots? Would a unique definition apply to walking and reaching skills equally? Without clear definitions, it becomes impossible or even worse, incorrect, to compare two systems.

Another crucial aspect of benchmarking is categorizing the multiple facets of a given system into meaningful subgroups. Some years ago, Torricelli et al. (2015) proposed to categorize the “motion abilities” of a bipedal system into four classes: efficiency, stability, kinematics, and dynamics. Each of them was composed of four to five quantitative indicators obtainable by direct measurements. Whether this classification is appropriate or not is yet to be proven. However, the methodological approach of decomposing the performance of a system into measurable items seems promising. Future research efforts aiming at identifying quantitative indicators for all the missing system abilities would be strongly valuable for the wearable robotics community. In this respect, one largely overlooked aspect is human–robot interaction, which is particularly difficult to characterize since it is strongly dependent on the subjective perception and experience of the user (Beckerle et al., 2017). Another crucial research direction in robotic benchmarking is establishing causal relations between the specific actuator typology and the overall performance in a given application. This would greatly help developers in identifying how improvements in components would contribute to better system behavior.

The discourse on system abilities and their definitions should be taken to a practical level to bring a real impact to the community. It is useless to design a meticulous protocol able to test a specific performance aspect very rigorously if such a protocol cannot be easily reproduced in different labs and systems. If benchmarking is about comparison, the reality of experimental settings should be considered. For this reason, the concept of “reproducibility” as opposed to “replicability” is worth clarifying. As well explained by Plessner (2018), reproducibility allows comparability of results from experiments conducted using different equipment (e.g., different sensors or environment) and even different procedural steps (e.g., a different number of trials). Replicability, instead, requires two tests to be executed exactly in the same way, using identical experimental setups, to produce comparable results. For benchmarking to be extensively adopted, we support the direction of reproducibility, even when this implies renouncing accuracy to a certain level. Very recently, several researchers have started to identify reproducible metrics, protocols and testbeds to quantify performance of exoskeletons in a wide range of motor skills and application domains (Torricelli and Pons, 2019; Baltrush et al. 2021; Kozinc et al., 2020a; Madinei et al., 2020; Pinto-Fernandez et al., 2020; Pesenti et al., 2021). This notable increase in scientific interest is being accompanied by the intense work of several international committees for standards (e.g., ASTM F48, ISO TC299, and CEN/CENELEC), which contribute to fill two important gaps: the availability of standardized testing methods and rigorous quantitative data on exoskeleton performance.

Nevertheless, even in the presence of agreed terminology, reproducible protocols and testbeds, and data-driven evidence, a missing element to ensure wide adoption of any potential standard testing methodology is the data format. Anyone with a clear and reproducible protocol and a proper experimental setup can run an experiment, obtain data and calculate the desired metric. But, in order to benchmark such system, a comparison with a reference should be done. If results cannot be easily shared and put together, for example, using unified databases, individual efforts in producing datasets are likely to remain limited to local (single lab-based) comparisons, or relegated to simplistic comparisons using easy-to-understand but poor metrics (e.g., speed and power consumption). To exploit all the potential that a unified database may provide, it is imperative to first find a common format for data, and second, make the existing computational algorithms to be compliant with such standardized data format. This way, anyone can compare their own system with another one, or with normative data obtained from all similar previously tested systems. To reach this situation, some bottlenecks should be overcome, the most important one being the regulatory constraints on sharing data given that tests with wearable robots intrinsically include a human in the loop.

Conclusion

We have presented challenges related to seven important technological aspects of wearable robots that are currently limiting broader adoption of such robots in numerous everyday applications, from healthcare to

sports. In the following paragraphs, we summarize the most important challenges that should be considered by developers of exoskeletons and other wearable devices.

Regarding mechanical design, many traditional exoskeleton structures and layouts tend to be obtrusive for the wearer during use and general management of the device. To increase the usability of these devices, researchers are working on novel solutions that target easier management through innovative design concepts: for example, modular structures, lighter machines specifically optimized to the degree of assistance that needs to be delivered, exosuits, and textile elements.

Actuation represents a major bottleneck for wearable robots, as their torque density, energy efficiency and backdrivability do not yet meet expectations. The inherent limitations of electric motors and gearboxes are the root of the problem. Improvements in these conventional technologies are slow, and promising alternative technologies are not yet mature. Specialized actuator designs incorporating springs, locking mechanisms, nonlinear transmissions and/or multiple motors can provide solutions, but introduce new trade-offs in terms of versatility, system complexity, and compactness.

In the area of sensing, the greatest challenge is in high-level sensor fusion. Such fusion has achieved good accuracies in controlled laboratory conditions, but these accuracies do not generalize to more uncontrolled real-world environments, which require additional work on multilevel and adaptive sensor fusion algorithms. Furthermore, additional improvements may be achieved through broader adoption of less common sensor modalities, such as exteroceptive environment sensing.

Limitations related to user comfort and acceptance are a critical remaining barrier to widespread use of wearable robots. Interfaces should not limit the comfort of the user and should simultaneously allow efficient power transition, which is achieved by optimal placement of the attachments to the human, larger contact stiffness, and prevention of interface migration.

Low-level and mid-level control architectures constitute a solid basis for high-level decision-making blocks, such as human–robot coadaptation. A possible performance degradation in low- and mid-level controllers may be directly reflected to overall system behavior, causing it to underperform the expected tasks. Therefore, one must construct a hierarchical control architecture first by ensuring the high-fidelity tracking performance at the inner control loops before gradually synthesizing the outer high-level controllers.

It is imperative that wearable devices perform tasks in synergy with their human users and simultaneously adapt to each other. Therefore, a high-level control system is a necessary element to infer human intention and create/adapt exoskeleton actions that facilitate the execution of the desired task as intended by the user. The high-level controller can infer the user's intention based on various signals, each of them having pro and cons in terms of control.

Describing the performance of a wearable robot including its components is a multidimensional problem. Benchmarking is a methodology that can help researchers and developers to quantify the different facets of performance and allow truthful comparisons between systems. Several issues should be considered to convert benchmarking into a common practice in the wearable robotic community. The most relevant ones are: converging on agreed terminology, quantifying human–robot interaction, define reproducible protocols and testbeds, and identifying common and sharable data formats.

While we have focused on the main challenges and aspects that many wearable robot applications have in common, it remains important to be aware that eventual success of wearable robots will equally depend on their specific tailoring and adequacy to the specific application case, always demanding that all critical user needs associated with that specific use case have been met. Additionally, successful implementation of wearable robots in our everyday life also depends on equally or even more important ethical, legal, and societal questions that deserve attention of experts in the relevant fields.

Data Availability Statement. Data availability is not applicable to this article as no new data were created or analyzed in this study.

Authorship Contributions. All authors contributed equally to this work.

Funding Statement. This article is partially based on work from COST Action CA16116 (Wearable Robots for Augmentation, Assistance or Substitution of Human Motor Functions), supported by COST (European Cooperation in Science and Technology).

J.B. was partially supported by the Slovenian Research Agency Programme P2-0076. M.L. was supported by the Istituto Italiano per l'Assicurazione contro gli Infortuni sul Lavoro (INAIL) under grant agreement nos. PR19-RR-P1 and PR19-RR-P2. T.V. was supported as a postdoctoral fellow of the Research Foundation Flanders—Fonds voor Wetenschappelijk Onderzoek (FWO). D.N. was partially supported by the United States National Science Foundation under grant no. 1933409. B.U. was partially supported by the Scientific and Technological Research Council of Turkey (TUBITAK) under grant no. 215E138. D.T. was partially supported by EUROBENCH under grant agreement no. 779963.

Competing Interests. The authors declare no competing interests exist.

References

- Abdoli-E M and Stevenson JM** (2008) The effect of on-body lift assistive device on the lumbar 3D dynamic moments and EMG during asymmetric freestyle lifting. *Clinical Biomechanics* **23**(3), 372–380. <https://doi.org/10.1016/j.clinbiomech.2007.10.012>
- Abdoli-Eramaki M, Stevenson JM, Reid SA and Bryant TJ** (2007) Mathematical and empirical proof of principle for an on-body personal lift augmentation device (PLAD). *Journal of Biomechanics* **40**(8), 1694–1700. <https://doi.org/10.1016/j.jbiomech.2006.09.006>
- Aguirre-Ollinger G** (2013) Learning muscle activation patterns via nonlinear oscillators: Application to lower-limb assistance. *2013 IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 1182–1189. <https://doi.org/10.1109/IROS.2013.6696500>
- Alexander RM** (1990) Three uses for springs in legged locomotion. *The International Journal of Robotics Research* **9**(2), 53–61. <https://doi.org/10.1177/027836499000900205>
- Almenara M, Cempini M, Gómez C, Cortese M, Martín C, Medina J, Vitiello N and Opisso E** (2017) Usability test of a hand exoskeleton for activities of daily living: An example of user-centered design. *Disability and Rehabilitation: Assistive Technology* **12**(1), 84–96. <https://doi.org/10.3109/17483107.2015.1079653>
- AL-Quraishi M, Elamvazuthi I, Daud S, Parasuraman S and Borboni A** (2018) EEG-based control for upper and lower limb exoskeletons and prostheses: A systematic review. *Sensors* **18**(10), 3342. <https://doi.org/10.3390/s18103342>
- Andrikopoulos G, Nikolakopoulos G and Manesis S** (2015) Motion control of a novel robotic wrist exoskeleton via pneumatic muscle actuators. *2015 IEEE 20th Conference on Emerging Technologies & Factory Automation (ETFA)*, pp. 1–8. <https://doi.org/10.1109/ETFA.2015.7301464>
- Asbeck AT, De Rossi SMM, Holt KG and Walsh CJ** (2015) A biologically inspired soft exosuit for walking assistance. *The International Journal of Robotics Research* **34**(6), 744–762. <https://doi.org/10.1177/0278364914562476>
- Asbeck AT, Dyer RJ, Larusson AF and Walsh CJ** (2013) Biologically-inspired soft exosuit. *2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR)*, pp. 1–8. <https://doi.org/10.1109/ICORR.2013.6650455>
- Atalay A, Sanchez V, Atalay O, Vogt DM, Haufe F, Wood RJ and Walsh CJ** (2017) Batch fabrication of customizable silicone-textile composite capacitive strain sensors for human motion tracking. *Advanced Materials Technologies* **2**(9), 1700136. <https://doi.org/10.1002/admt.201700136>
- Awad LN, Bae J, O'Donnell K, De Rossi SMM, Hendron K, Sloat LH, Kudzia P, Allen S, Holt KG, Ellis TD and Walsh CJ** (2017) A soft robotic exosuit improves walking in patients after stroke. *Science Translational Medicine* **9**(400), eaai9084. <https://doi.org/10.1126/scitranslmed.aai9084>
- Babič J, Mombaur K, Lefeber D, van Dieën J, Graitmann B, Russold M, Šarabon N and Houdijk H** (2017) SPEXOR: Spinal exoskeletal robot for low back pain prevention and vocational reintegration. In González-Vargas J, Ibáñez J, Contreras-Vidal JL, van der Kooij H and Pons JL (eds), *Wearable Robotics: Challenges and Trends*. Cham: Springer International Publishing, pp. 311–315. https://doi.org/10.1007/978-3-319-46532-6_51
- Babič J, Petrič T, Mombaur K, Kingma I, Bornmann J, González-Vargas J, Baltrusch S, Šarabon N and Houdijk H** (2019) SPEXOR: Design and development of passive spinal exoskeletal robot for low back pain prevention and vocational reintegration. *SN Applied Sciences* **1**(3), 262. <https://doi.org/10.1007/s42452-019-0266-1>
- Bai S and Christensen S** (2017) Biomechanical HRI modeling and mechatronic design of exoskeletons for assistive applications. In *Human Modelling for Bio-Inspired Robotics*. Amsterdam: Elsevier, pp. 251–272. <https://doi.org/10.1016/B978-0-12-803137-7.00010-0>
- Balasubramanian S, Wei R, Perez M, Shepard B, Koenenman E, Koenenman J and He J** (2008) RUPERT: An exoskeleton robot for assisting rehabilitation of arm functions. *2008 Virtual Rehabilitation*, pp. 163–167. <https://doi.org/10.1109/ICVR.2008.4625154>
- Ball SJ, Brown IE and Scott SH** (2007) MEDARM: A rehabilitation robot with 5DOF at the shoulder complex. *2007 IEEE/ASME International Conference on Advanced Intelligent Mechatronics*, pp. 1–6. <https://doi.org/10.1109/AIM.2007.4412446>
- Baltrusch SJ, Houdijk H, van Dieën JH and Kruif JTCM** (2021) Passive trunk exoskeleton acceptability and effects on self-efficacy in employees with low-back pain: A mixed method approach. *Journal of Occupational Rehabilitation* **31**(1), 129–141. <https://doi.org/10.1007/s10926-020-09891-1>
- Baltrusch SJ, van Dieën JH, van Bennekom CAM and Houdijk H** (2018) The effect of a passive trunk exoskeleton on functional performance in healthy individuals. *Applied Ergonomics* **72**, 94–106. <https://doi.org/10.1016/j.apergo.2018.04.007>

- Banala SK, Kim SH, Agrawal SK and Scholz JP (2009) Robot assisted gait training with active leg exoskeleton (ALEX). *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 17(1), 2–8. <https://doi.org/10.1109/TNSRE.2008.2008280>
- Bao G, Pan L, Fang H, Wu X, Yu H, Cai S, Yu B and Wan Y (2019) Academic review and perspectives on robotic exoskeletons. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 27(11), 2294–2304. <https://doi.org/10.1109/TNSRE.2019.2944655>
- Beckerle P, Salvietti G, Unal R, Prattichizzo D, Rossi S, Castellini C, Hirche S, Endo S, Amor HB, Ciocarlie M, Mastrogiorganni F, Argall BD and Bianchi M (2017) A human–robot interaction perspective on assistive and rehabilitation robotics. *Frontiers in Neurorobotics* 11, 24. <https://doi.org/10.3389/fnbot.2017.00024>
- Beckerle P, Verstraten T, Mathijssen G, Furnemont R, Vanderborght B and Lefeber D (2016) Series and parallel elastic actuation: Influence of operating positions on design and control. *IEEE/ASME Transactions on Mechatronics* 22(1), 521–529. <https://doi.org/10.1109/TMECH.2016.2621062>
- Bergelin BJ and Voglewede PA (2012) Design of an active ankle-foot prosthesis utilizing a four-bar mechanism. *Journal of Mechanical Design* 134(6), 061004. <https://doi.org/10.1115/1.4006436>
- Bianchi M, Buonamici F, Furferi R and Vanni N (2016) Design and optimization of a flexion/extension mechanism for a hand exoskeleton system. Volume 1A: 36th Computers and Information in Engineering Conference. <https://doi.org/10.1115/DETC2016-59466>
- Bogue R (2018) Exoskeletons – A review of industrial applications. *Industrial Robot: An International Journal* 45(5), 585–590. <https://doi.org/10.1108/IR-05-2018-0109>
- Bolivar Nieto EA, Rezazadeh S and Gregg RD (2019) Minimizing energy consumption and peak power of series elastic actuators: A convex optimization framework for elastic element design. *IEEE/ASME Transactions on Mechatronics* 24(3), 1334–1345. <https://doi.org/10.1109/TMECH.2019.2906887>
- Bosch T, van Eck J, Knitel K and de Looze M (2016) The effects of a passive exoskeleton on muscle activity, discomfort and endurance time in forward bending work. *Applied Ergonomics* 54, 212–217. <https://doi.org/10.1016/j.apergo.2015.12.003>
- Brahmi B, Saad M, Rahman MH and Ochoa-Luna C (2019) Cartesian trajectory tracking of a 7-DOF exoskeleton robot based on human inverse kinematics. *IEEE Transactions on Systems, Man, and Cybernetics: Systems* 49(3), 600–611. <https://doi.org/10.1109/TSMC.2017.2695003>
- Brassitos E and Jalili N (2019) Dynamic model development and characterization of gear bearing transmission systems: Theory and experiments. *IEEE/ASME Transactions on Mechatronics* 24(4), 1651–1661. <https://doi.org/10.1109/TMECH.2019.2916333>
- Browning Raymond C, Modica Jesse R, Kram Rodger and Goswami Ambarish The Effects of Adding Mass to the Legs on the Energetics and Biomechanics of Walking, *Medicine & Science in Sports & Exercise*: March 2007 - Volume 39 - Issue 3 - p 515–525 doi: 10.1249/mss.0b013e31802b3562
- Buerger SP and Hogan N (2007) Complementary stability and loop shaping for improved human–robot interaction. *IEEE Transactions on Robotics* 23(2), 232–244. <https://doi.org/10.1109/TRO.2007.892229>
- Buongiorno D, Chiaradia D, Marcheschi S, Solazzi M and Frisoli A (2019) Multi-DoFs exoskeleton-based bilateral teleoperation with the time-domain passivity approach. *Robotica* 37(9), 1641–1662. <https://doi.org/10.1017/S0263574719000171>
- Caldwell DG, Tsagarakis NG, Kousidou S, Costa N and Sarakoglou I (2007) “Soft” exoskeletons for upper and lower body rehabilitation – Design, control and testing. *International Journal of Humanoid Robotics* 4(3), 549–573. <https://doi.org/10.1142/S0219843607001151>
- Chaparro-Cárdenas SL, Lozano-Guzmán AA, Ramirez-Bautista JA and Hernández-Zavala A (2018) A review in gait rehabilitation devices and applied control techniques. *Disability and Rehabilitation: Assistive Technology* 13(8), 819–834. <https://doi.org/10.1080/17483107.2018.1447611>
- Cherelle P, Grosu V, Matthys A, Vanderborght B and Lefeber D (2014) Design and validation of the ankle mimicking prosthetic (AMP-) foot 2.0. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 22(1), 138–148. <https://doi.org/10.1109/TNSRE.2013.2282416>
- Chiaradia D, Tiseni L, Xiloyannis M, Solazzi M, Masia L and Frisoli A (2021) An assistive soft wrist exosuit for flexion movements with an ergonomic reinforced glove. *Frontiers in Robotics and AI* 7, 595862. <https://doi.org/10.3389/frobt.2020.595862>
- Chiaradia D, Xiloyannis M, Antuvan CW, Frisoli A and Masia L (2018) Design and embedded control of a soft elbow exosuit. 2018 IEEE International Conference on Soft Robotics (RoboSoft), pp. 565–571. <https://doi.org/10.1109/ROBOSOFT.2018.8405386>
- Chiaradia D, Xiloyannis M, Solazzi M, Masia L and Frisoli A (2019) Comparison of a Soft Exosuit and a Rigid Exoskeleton in an Assistive Task. Cham: Springer, pp. 415–419. https://doi.org/10.1007/978-3-030-01887-0_80
- Choi H, Kang BB, Jung B-K and Cho K-J (2019) Exo-wrist: A soft tendon-driven wrist-wearable robot with active anchor for dart-throwing motion in hemiplegic patients. *IEEE Robotics and Automation Letters* 4(4), 4499–4506. <https://doi.org/10.1109/LRA.2019.2931607>
- Convens B. et al., “Modeling, Design and Test-Bench Validation of a Semi-Active Propulsive Ankle Prosthesis With a Clutched Series Elastic Actuator,” in *IEEE Robotics and Automation Letters*, vol. 4, no. 2, pp. 1823–1830, April 2019, doi: 10.1109/LRA.2019.2897993.
- Collins SH and Kuo AD (2010) Recycling energy to restore impaired ankle function during human walking. *PLoS One* 5(2), e9307. <https://doi.org/10.1371/journal.pone.0009307>

- Davis KG, Reid CR, Rempel DD and Treaster D** (2020) Introduction to the human factors special issue on user-centered design for exoskeleton. *Human Factors: The Journal of the Human Factors and Ergonomics Society* **62**(3), 333–336. <https://doi.org/10.1177/0018720820914312>
- de Looze MP, Bosch T, Krause F, Stadler KS and O’Sullivan LW** (2016) Exoskeletons for industrial application and their potential effects on physical work load. *Ergonomics* **59**(5), 671–681. <https://doi.org/10.1080/00140139.2015.1081988>
- de Looze MP, Krause F and O’Sullivan LW** (2017) *The Potential and Acceptance of Exoskeletons in Industry*. Cham: Springer, pp. 195–199. https://doi.org/10.1007/978-3-319-46532-6_32
- Di Natali C, Poliero T, Sposito M, Graf E, Bauer C, Pauli C, Bottenberg E, De Eyto A, O’Sullivan L, Hidalgo AF, Scherly D, Stadler KS, Caldwell DG and Ortiz J** (2019) Design and evaluation of a soft assistive lower limb exoskeleton. *Robotica* **37**(12), 2014–2034. <https://doi.org/10.1017/S0263574719000067>
- Ding Y and Park H-W** (2017) Design and experimental implementation of a quasi-direct-drive leg for optimized jumping. *2017 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*, pp. 300–305. <https://doi.org/10.1109/IROS.2017.8202172>
- Elery T, Rezazadeh S, Nesler C and Gregg RD** (2020) Design and validation of a powered knee–ankle prosthesis with high-torque, low-impedance actuators. *IEEE Transactions on Robotics* **36**(6), 1649–1668. <https://doi.org/10.1109/TRO.2020.3005533>
- Ergin MA and Patoglu V** (2012) ASSISTON-SE: A self-aligning shoulder-elbow exoskeleton. *2012 IEEE International Conference on Robotics and Automation*, pp. 2479–2485. <https://doi.org/10.1109/ICRA.2012.6225117>
- Esquenazi A, Talaty M, Packer A and Saulino M** (2012) The ReWalk powered exoskeleton to restore ambulatory function to individuals with thoracic-level motor-complete spinal cord injury. *American Journal of Physical Medicine & Rehabilitation* **91** (11), 911–921. <https://doi.org/10.1097/PHM.0b013e318269d9a3>
- Farris RJ, Quintero HA and Goldfarb M** (2011) Preliminary evaluation of a powered lower limb orthosis to aid walking in paraplegic individuals. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **19**(6), 652–659. <https://doi.org/10.1109/TNSRE.2011.2163083>
- Farris RJ, Quintero HA, Murray SA, Ha KH, Hartigan C and Goldfarb M** (2014) A preliminary assessment of legged mobility provided by a lower limb exoskeleton for persons with paraplegia. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **22**(3), 482–490. <https://doi.org/10.1109/TNSRE.2013.2268320>
- Feldman AG** (1980) Superposition of motor programs—I. Rhythmic forearm movements in man. *Neuroscience* **5**(1), 81–90. [https://doi.org/10.1016/0306-4522\(80\)90073-1](https://doi.org/10.1016/0306-4522(80)90073-1)
- Ferris DP, Gordon KE, Sawicki GS and Peethambaran A** (2006) An improved powered ankle–foot orthosis using proportional myoelectric control. *Gait & Posture* **23**(4), 425–428. <https://doi.org/10.1016/j.gaitpost.2005.05.004>
- Fitzsimons K, Acosta AM, Dewald JPA and Murphey TD** (2019) Ergodicity reveals assistance and learning from physical human–robot interaction. *Science Robotics* **4**(29), eaav6079. <https://doi.org/10.1126/scirobotics.aav6079>
- Fleischer C and Hommel G** (2008) A human–exoskeleton interface utilizing electromyography. *IEEE Transactions on Robotics* **24** (4), 872–882. <https://doi.org/10.1109/TRO.2008.926860>
- Frigo C, Crenna P and Jensen LM** (1996) Moment-angle relationship at lower limb joints during human walking at different velocities. *Journal of Electromyography and Kinesiology* **6**(3), 177–190. [https://doi.org/10.1016/1050-6411\(96\)00030-2](https://doi.org/10.1016/1050-6411(96)00030-2)
- Frisoli A, Rocchi F, Marcheschi S, Dettori A, Salsedo F and Bergamasco M** (2005) A new force-feedback arm exoskeleton for haptic interaction in virtual environments. *First Joint Eurohaptics Conference and Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems*, pp. 195–201. <https://doi.org/10.1109/WHC.2005.15>
- Fritz H, Patzer D and Galen SS** (2019) Robotic exoskeletons for reengaging in everyday activities: Promises, pitfalls, and opportunities. *Disability and Rehabilitation* **41**(5), 560–563. <https://doi.org/10.1080/09638288.2017.1398786>
- Galle S, Malcolm P, Derave W and De Clercq D** (2013) Adaptation to walking with an exoskeleton that assists ankle extension. *Gait & Posture* **38**(3), 495–499. <https://doi.org/10.1016/j.gaitpost.2013.01.029>
- Gao F, Liu Y and Liao W-H** (2018) Design of powered ankle-foot prosthesis with nonlinear parallel spring mechanism. *Journal of Mechanical Design* **140**(5), 055001. <https://doi.org/10.1115/1.4039385>
- García PL, Crispel S, Sacerens E, Verstraten T and Lefeber D** (2020) Compact gearboxes for modern robotics: A review. *Frontiers in Robotics and AI* **7**, 103. <https://doi.org/10.3389/frobt.2020.00103>
- Geeroms J, Flynn L, Jimenez-Fabian R, Vanderborght B and Lefeber D** (2017) Design and energetic evaluation of a prosthetic knee joint actuator with a lockable parallel spring. *Bioinspiration & Biomimetics* **12**(2), 026002. <https://doi.org/10.1088/1748-3190/aa575c>
- Gokcesu K, Ergeneci M, Ertan E and Gokcesu H** (2019) An adaptive algorithm for online interference cancellation in EMG sensors. *IEEE Sensors Journal* **19**(1), 214–223. <https://doi.org/10.1109/JSEN.2018.2874724>
- Gopura RARC, Bandara DSV, Kiguchi K and Mann GKI** (2016) Developments in hardware systems of active upper-limb exoskeleton robots: A review. *Robotics and Autonomous Systems* **75**, 203–220. <https://doi.org/10.1016/j.robot.2015.10.001>
- Gordon KE, Kinnaird CR and Ferris DP** (2013) Locomotor adaptation to a soleus EMG-controlled antagonistic exoskeleton. *Journal of Neurophysiology* **109**(7), 1804–1814. <https://doi.org/10.1152/jn.01128.2011>
- Gorgey AS** (2018) Robotic exoskeletons: The current pros and cons. *World Journal of Orthopedics* **9**(9), 112–119. <https://doi.org/10.5312/wjo.v9.i9.112>

- Goršič M, Regmi Y, Johnson AP, Dai B and Novak D (2020) A pilot study of varying thoracic and abdominal compression in a reconfigurable trunk exoskeleton during different activities. *IEEE Transactions on Biomedical Engineering* **67**(6), 1585–1594. <https://doi.org/10.1109/TBME.2019.2940431>
- Goršič M, Song Y, Dai B and Novak D (2021) Evaluation of the HeroWear Apex back-assist exosuit during multiple brief tasks. *Journal of Biomechanics* **126**, 110620. <https://doi.org/10.1016/j.jbiomech.2021.110620>
- Graven-Nielsen T and Arendt-Nielsen L (2010) Assessment of mechanisms in localized and widespread musculoskeletal pain. *Nature Reviews Rheumatology* **6**(10), 599–606. <https://doi.org/10.1038/nrrheum.2010.107>
- Guardian XO Exoskeleton Webpage (n.d.) Available at <https://www.sarcos.com/products/guardian-xo-powered-exoskeleton/>.
- Gui K, Liu H and Zhang D (2019) A practical and adaptive method to achieve EMG-based torque estimation for a robotic exoskeleton. *IEEE/ASME Transactions on Mechatronics* **24**(2), 483–494. <https://doi.org/10.1109/TMECH.2019.2893055>
- Guizzo E and Goldstein H (2005) The rise of the body bots [robotic exoskeletons]. *IEEE Spectrum* **42**(10), 50–56. <https://doi.org/10.1109/MSPEC.2005.1515961>
- Haddadin S, Albu-Schäffer A and Hirzinger G (2009) Requirements for safe robots: Measurements, analysis and new insights. *The International Journal of Robotics Research* **28**(11–12), 1507–1527. <https://doi.org/10.1177/0278364909343970>
- Haeufle DFB, Taylor MD, Schmitt S and Geyer H (2012) A clutched parallel elastic actuator concept: Towards energy efficient powered legs in prosthetics and robotics. *2012 4th IEEE RAS & EMBS International Conference on Biomedical Robotics and Biomechanics (BioRob)*, pp. 1614–1619. <https://doi.org/10.1109/BioRob.2012.6290722>
- Ham R, Sugar T, Vanderborght B, Hollander K and Lefeber D (2009) Compliant actuator designs. *IEEE Robotics & Automation Magazine* **16**(3), 81–94. <https://doi.org/10.1109/MRA.2009.933629>
- Haninger K, Asignacion A and Oh S (2020) Safe high impedance control of a series-elastic actuator with a disturbance observer. *2020 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 921–927. <https://doi.org/10.1109/ICRA40945.2020.9197402>
- Hasan Z (1986) Optimized movement trajectories and joint stiffness in unperturbed, inertially loaded movements. *Biological Cybernetics* **53**(6), 373–382. <https://doi.org/10.1007/BF00318203>
- Haufe FL, Kober AM, Schmidt K, Sancho-Puchades A, Duarte JE, Wolf P and Riener R (2019) User-driven walking assistance: First experimental results using the MyoSuit*. *2019 IEEE 16th International Conference on Rehabilitation Robotics (ICORR)*, pp. 944–949. <https://doi.org/10.1109/ICORR.2019.8779375>
- Hensel R and Keil M (2019) Subjective evaluation of a passive industrial exoskeleton for lower-back support: A field study in the automotive sector. *IIEE Transactions on Occupational Ergonomics and Human Factors* **7**(3–4), 213–221. <https://doi.org/10.1080/24725838.2019.1573770>
- Hill D, Holloway CS, Morgado Ramirez DZ, Smitham P and Pappas Y (2017) What are user PERSPECTIVES of exoskeleton technology? A literature review. *International Journal of Technology Assessment in Health Care* **33**(2), 160–167. <https://doi.org/10.1017/S0266462317000460>
- Hlucny SD and Novak D (2020) Characterizing human box-lifting behavior using wearable inertial motion sensors. *Sensors* **20**(8), 2323. <https://doi.org/10.3390/s20082323>
- Hogan N (1985) Impedance control: An approach to manipulation: Part I—Theory. *Journal of Dynamic Systems, Measurement, and Control* **107**(1), 1–7. <https://doi.org/10.1115/1.3140702>
- Hollander KW, Ilg R, Sugar TG and Herring D (2006) An efficient robotic tendon for gait assistance. *Journal of Biomechanical Engineering* **128**(5), 788–791. <https://doi.org/10.1115/1.2264391>
- Hollerbach JM, Hunter IW and Ballantyne J (1992) A comparative analysis of actuator technologies for robotics. In *The Robotics Review 2*. Cambridge, MA: MIT Press, pp. 299–342.
- Hollman John H, Robert H Deusinger, Linda R Van Dillen, Matthew J Matava, Knee Joint Movements in Subjects Without Knee Pathology and Subjects With Injured Anterior Cruciate Ligaments, *Physical Therapy*, Volume **82**, Issue 10, 1 October 2002, Pages 960–972, <https://doi.org/10.1093/ptj/82.10.960>
- Hu B, Simon AM and Hargrove L (2019) Deep generative models with data augmentation to learn robust representations of movement intention for powered leg prostheses. *IEEE Transactions on Medical Robotics and Bionics* **1**(4), 267–278. <https://doi.org/10.1109/TMRB.2019.2952148>
- Huang R, Cheng H, Guo H, Chen Q and Lin X (2016) Hierarchical interactive learning for a human-powered augmentation lower exoskeleton. *2016 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 257–263. <https://doi.org/10.1109/ICRA.2016.7487142>
- Huang R, Cheng H, Qiu J and Zhang J (2019) Learning physical human–robot interaction with coupled cooperative primitives for a lower exoskeleton. *IEEE Transactions on Automation Science and Engineering* **16**(4), 1566–1574. <https://doi.org/10.1109/TASE.2018.2886376>
- Huang H, Kuiken TA and Lipschutz RD (2009) A strategy for identifying locomotion modes using surface electromyography. *IEEE Transactions on Biomedical Engineering* **56**(1), 65–73. <https://doi.org/10.1109/TBME.2008.2003293>
- Huysamen K, Bosch T, de Looze M, Stadler KS, Graf E and O’Sullivan LW (2018) Evaluation of a passive exoskeleton for static upper limb activities. *Applied Ergonomics* **70**, 148–155. <https://doi.org/10.1016/j.apergo.2018.02.009>
- Hyun DJ, Park H, Ha T, Park S and Jung K (2017) Biomechanical design of an agile, electricity-powered lower-limb exoskeleton for weight-bearing assistance. *Robotics and Autonomous Systems* **95**, 181–195. <https://doi.org/10.1016/j.robot.2017.06.010>
- Jamšek M, Petrič T and Babič J (2020) Gaussian mixture models for control of quasi-passive spinal exoskeletons. *Sensors* **20**(9), 2705. <https://doi.org/10.3390/s20092705>

- Jeong M, Woo H and Kong K** (2020) A study on weight support and balance control method for assisting squat movement with a wearable robot, angel-suit. *International Journal of Control, Automation and Systems* **18**(1), 114–123. <https://doi.org/10.1007/s12555-019-0243-x>
- Jiménez-Fabián R and Verlinden O** (2012) Review of control algorithms for robotic ankle systems in lower-limb orthoses, prostheses, and exoskeletons. *Medical Engineering & Physics* **34**(4), 397–408. <https://doi.org/10.1016/j.medengphy.2011.11.018>
- Jin Y, Glover CM, Cho H, Araromi OA, Graule MA, Li N, Wood RJ and Walsh CJ** (2020) Soft sensing shirt for shoulder kinematics estimation. *2020 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 4863–4869. <https://doi.org/10.1109/ICRA40945.2020.9196586>
- Jyräkoski T, Merilampi S, Puustinen J and Kärki A** (2020) Over-ground robotic lower limb exoskeleton in neurological gait rehabilitation: User experiences and effects on walking ability. *Technology and Disability* **33**, 53–63. <https://doi.org/10.3233/TAD-200284>
- Kaneishi D, Matthew RP and Tomizuka M** (2018) Optimal control parameterization for active/passive exoskeleton with variable impedance actuator. *2018 7th IEEE International Conference on Biomedical Robotics and Biomechanics (Biorob)*, pp. 713–719. <https://doi.org/10.1109/BIOROB.2018.8487719>
- Kashiri N, Malzahn J and Tsagarakis NG** (2017) On the sensor design of torque controlled actuators: A comparison study of strain gauge and encoder-based principles. *IEEE Robotics and Automation Letters* **2**(2), 1186–1194. <https://doi.org/10.1109/LRA.2017.2662744>
- Kawamoto H, Kanbe S and Sankai Y** (2003a) Power assist method for HAL-3 estimating operator's intention based on motion information. *The 12th IEEE International Workshop on Robot and Human Interactive Communication, 2003. Proceedings. ROMAN 2003*, pp. 67–72. <https://doi.org/10.1109/ROMAN.2003.1251800>
- Kawamoto H, Lee S, Kanbe S and Sankai Y** (2003b) Power assist method for HAL-3 using EMG-based feedback controller. *SMC'03 Conference Proceedings. 2003 IEEE International Conference on Systems, Man and Cybernetics. Conference Theme - System Security and Assurance (Cat. No.03CH37483)*, 2, pp. 1648–1653. <https://doi.org/10.1109/ICSMC.2003.1244649>
- Kazerooni H** (1995) Dynamics and control of instrumented harmonic drives. *Journal of Dynamic Systems, Measurement, and Control* **117**(1), 15–19. <https://doi.org/10.1115/1.2798517>
- Kazerooni H** (2005) Exoskeletons for human power augmentation. *2005 IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 3459–3464. <https://doi.org/10.1109/IROS.2005.1545451>
- Kazerooni H and Guo J** (1993) Human extenders. *Journal of Dynamic Systems, Measurement, and Control* **115**(2B), 281–290. <https://doi.org/10.1115/1.2899068>
- Kazerooni H, Steger R and Huang L** (2006) Hybrid control of the Berkeley lower extremity exoskeleton (BLEEX). *The International Journal of Robotics Research* **25**(5–6), 561–573. <https://doi.org/10.1177/0278364906065505>
- Kermavnar T, de Vries AW, de Looze MP and O'Sullivan LW** (2021) Effects of industrial back-support exoskeletons on body loading and user experience: An updated systematic review. *Ergonomics* **64**, 685–711. <https://doi.org/10.1080/00140139.2020.1870162>
- Kermavnar T, Power V, de Eyto A and O'Sullivan LW** (2018) Computerized cuff pressure algometry as guidance for circumferential tissue compression for wearable soft robotic applications: A systematic review. *Soft Robotics* **5**(1), 1–16. <https://doi.org/10.1089/soro.2017.0046>
- Kiguchi K and Hayashi Y** (2012) An EMG-based control for an upper-limb power-assist exoskeleton robot. *IEEE Transactions on Systems, Man, and Cybernetics, Part B (Cybernetics)* **42**(4), 1064–1071. <https://doi.org/10.1109/TSMCB.2012.2185843>
- Kim B and Deshpande AD** (2017) An upper-body rehabilitation exoskeleton harmony with an anatomical shoulder mechanism: Design, modeling, control, and performance evaluation. *The International Journal of Robotics Research* **36**(4), 414–435. <https://doi.org/10.1177/0278364917706743>
- Kim YB, Kim U, Seok D-Y, So J, Lee YH and Choi HR** (2018a) Torque sensor embedded actuator module for robotic applications. *IEEE/ASME Transactions on Mechatronics* **23**(4), 1662–1672. <https://doi.org/10.1109/TMECH.2018.2850962>
- Kim D, Kwon J, Han S, Park Y-L and Jo S** (2019a) Deep full-body motion network for a soft wearable motion sensing suit. *IEEE/ASME Transactions on Mechatronics* **24**(1), 56–66. <https://doi.org/10.1109/TMECH.2018.2874647>
- Kim J, Lee G, Heimgartner R, Arumukhom Revi D, Karavas N, Nathanson D, Galiana I, Eckert-Erdheim A, Murphy P, Perry D, Menard N, Choe DK, Malcolm P and Walsh CJ** (2019b) Reducing the metabolic rate of walking and running with a versatile, portable exosuit. *Science* **365**(6454), 668–672. <https://doi.org/10.1126/science.aav7536>
- Kim T, Lee S, Hong T, Shin G, Kim T and Park Y-L** (2020) Heterogeneous sensing in a multifunctional soft sensor for human-robot interfaces. *Science robotics* **5**(49), eabc6878. <https://doi.org/10.1126/scirobotics.abc6878>
- Kim W, Lee S, Kang M, Han J and Han C** (2010) Energy-efficient gait pattern generation of the powered robotic exoskeleton using DME. *2010 IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 2475–2480. <https://doi.org/10.1109/IROS.2010.5649235>
- Kim S, Nussbaum MA, Mokhespour Esfahani MI, Alemi MM, Alabdulkarim S and Rashedi E** (2018b) Assessing the influence of a passive, upper extremity exoskeletal vest for tasks requiring arm elevation: Part I – “Expected” effects on discomfort, shoulder muscle activity, and work task performance. *Applied Ergonomics* **70**, 315–322. <https://doi.org/10.1016/j.apergo.2018.02.025>

- Koenig A, Novak D, Omlin X, Pulfer M, Perreault E, Zimmerli L, Mihelj M and Riener R (2011a) Real-time closed-loop control of cognitive load in neurological patients during robot-assisted gait training. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **19**(4), 453–464. <https://doi.org/10.1109/TNSRE.2011.2160460>
- Koenig A, Omlin X, Bergmann J, Zimmerli L, Bolliger M, Müller F and Riener R (2011b) Controlling patient participation during robot-assisted gait training. *Journal of Neuroengineering and Rehabilitation* **8**(1), 14. <https://doi.org/10.1186/1743-0003-8-14>
- Kong K and Jeon D (2006) Design and control of an exoskeleton for the elderly and patients. *IEEE/ASME Transactions on Mechatronics* **11**(4), 428–432. <https://doi.org/10.1109/TMECH.2006.878550>
- Kozinc Ž, Babič J and Šarabon N (2021) Human pressure tolerance and effects of different padding materials with implications for development of exoskeletons and similar devices. *Applied Ergonomics* **93**, 103379. <https://doi.org/10.1016/j.apergo.2021.103379>
- Kozinc Ž, Baltrusch S, Houdijk H and Šarabon N (2020a) Reliability of a battery of tests for functional evaluation of trunk exoskeletons. *Applied Ergonomics* **86**, 103117. <https://doi.org/10.1016/j.apergo.2020.103117>
- Kozinc Ž, Baltrusch S, Houdijk H and Šarabon N (2020b) Short-term effects of a passive spinal exoskeleton on functional performance, discomfort and user satisfaction in patients with low back pain. *Journal of Occupational Rehabilitation* **31**, 142–152. <https://doi.org/10.1007/s10926-020-09899-7>
- Krausz NE, Hu BH and Hargrove LJ (2019) Subject- and environment-based sensor variability for wearable lower-limb assistive devices. *Sensors* **19**(22), 4887. <https://doi.org/10.3390/s19224887>
- Krausz NE, Lenzi T and Hargrove LJ (2015) Depth sensing for improved control of lower limb prostheses. *IEEE Transactions on Biomedical Engineering* **62**(11), 2576–2587. <https://doi.org/10.1109/TBME.2015.2448457>
- Krebs HI, Palazzolo JJ, Dipietro L, Ferraro M, Krol J, Rannekleiv K, Volpe BT and Hogan N (2003) Rehabilitation robotics: Performance-based progressive robot-assisted therapy. *Autonomous Robots* **15**(1), 7–20. <https://doi.org/10.1023/A:1024494031121>
- Krimsky E and Collins SH (2020) Optimal control of an energy-recycling actuator for mobile robotics applications. *2020 IEEE International Conference on Robotics and Automation (ICRA)*, pp. 3559–3565. <https://doi.org/10.1109/ICRA40945.2020.9196870>
- Laffranchi M, Semprini M, Manzan E, Cerruti G, Vassallo C, De Giuseppe S, Maludrotto S, Succi A, Chiappalone M, Gruppioni E and De Michieli L (2018) A novel lower limb exoskeleton and new neurorhabilitation-based application scenarios. *40th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC'18)*. Available at https://embs.paperecept.net/conferences/conferences/EMBC18/program/EMBC18_ContentListWeb_3.html.
- Lamers EP, Yang AJ and Zelik KE (2018) Feasibility of a biomechanically-assistive garment to reduce low back loading during leaning and lifting. *IEEE Transactions on Bio-Medical Engineering* **65**(8), 1674–1680. <https://doi.org/10.1109/TBME.2017.2761455>
- Langlois K, Moltedo M, Bacek T, Rodriguez-Guerrero C, Vanderborght B and Lefeber D (2018) Design and development of customized physical interfaces to reduce relative motion between the user and a powered ankle foot exoskeleton. *2018 7th IEEE International Conference on Biomedical Robotics and Biomechanics (Biorob)*, pp. 1083–1088. <https://doi.org/10.1109/BIOROB.2018.8487706>
- Langlois K, Rodriguez-Cianca D, Serrien B, De Winter J, Verstraten T, Rodriguez-Guerrero C, Vanderborght B and Lefeber D (2020) Investigating the effects of strapping pressure on human–robot interface dynamics using a soft robotic cuff. *IEEE Transactions on Medical Robotics and Bionics* **3**(1), 146–155. <https://doi.org/10.1109/TMRB.2020.3042255>
- Laschowski B, McNally W, Wong A and McPhee J (2020) ExoNet database: Wearable camera images of human locomotion environments. *Frontiers in Robotics and AI* **7**, 188. <https://doi.org/10.3389/frobt.2020.562061>
- Lauretti C, Cordella F, Ciancio AL, Trigili E, Catalan JM, Badesa FJ, Crea S, Pagliara SM, Sterzi S, Vitiello N, Garcia Aracil N and Zollo L (2018) Learning by demonstration for motion planning of upper-limb exoskeletons. *Frontiers in Neurorobotics* **12**, 5. <https://doi.org/10.3389/fnbot.2018.00005>
- Lazzaroni M, Tabasi A, Toxiri S, Caldwell DG, De Momi E, van Dijk W, de Looze MP, Kingma I, van Dieën JH and Ortiz J (2020) Evaluation of an acceleration-based assistive strategy to control a back-support exoskeleton for manual material handling. *Wearable Technologies* **1**, e9. <https://doi.org/10.1017/wtc.2020.8>
- Lee H-D, Lee B-K, Kim W-S, Han J-S, Shin K-S and Han C-S (2014) Human–robot cooperation control based on a dynamic model of an upper limb exoskeleton for human power amplification. *Mechatronics* **24**(2), 168–176. <https://doi.org/10.1016/j.mechatronics.2014.01.007>
- Lenzi T, De Rossi SMM, Vitiello N and Carrozza MC (2012) Intention-based EMG control for powered exoskeletons. *IEEE Transactions on Biomedical Engineering* **59**(8), 2180–2190. <https://doi.org/10.1109/TBME.2012.2198821>
- Lessard S, Pansodtee P, Robbins A, Trombadore JM, Kurniawan S and Teodorescu M (2018) A soft exosuit for flexible upper-extremity rehabilitation. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **26**(8), 1604–1617. <https://doi.org/10.1109/TNSRE.2018.2854219>
- Li YD and Hsiao-Weckler ET (2013) Gait mode recognition and control for a portable-powered ankle-foot orthosis. *2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR)*, pp. 1–8. <https://doi.org/10.1109/ICORR.2013.6650373>
- Li N, Yang T, Yu P, Chang J, Zhao X, Zhao X, Elhajj IH, Xi N and Liu L (2018) Bio-inspired upper limb soft exoskeleton to reduce stroke-induced complications. *Bioinspiration & Biomimetics* **13**(6), 066001. <https://doi.org/10.1088/1748-3190/aad8d4>

- Li Z, Yuan Y, Luo L, Su W, Zhao K, Xu C, Huang J and Pi M (2019) Hybrid brain/muscle signals powered wearable walking exoskeleton enhancing motor ability in climbing stairs activity. *IEEE Transactions on Medical Robotics and Bionics* **1**(4), 218–227. <https://doi.org/10.1109/TMRB.2019.2949865>
- Lim B, Lee J, Jang J, Kim K, Park YJ, Seo K and Shim Y (2019) Delayed output feedback control for gait assistance with a robotic hip exoskeleton. *IEEE Transactions on Robotics* **35**(4), 1055–1062. <https://doi.org/10.1109/TRO.2019.2913318>
- Liu M, Wang D and Huang H (2016) Development of an environment-aware locomotion mode recognition system for powered lower limb prostheses. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **24**(4), 434–443. <https://doi.org/10.1109/TNSRE.2015.2420539>
- Lo HS and Xie SQ (2012) Exoskeleton robots for upper-limb rehabilitation: State of the art and future prospects. *Medical Engineering & Physics* **34**(3), 261–268. <https://doi.org/10.1016/j.medengphy.2011.10.004>
- Lotti N, Xiloyannis M, Durandau G, Galofaro E, Sanguineti V, Masia L and Sartori M (2020) Adaptive model-based myoelectric control for a soft wearable arm exosuit: A new generation of wearable robot control. *IEEE Robotics & Automation Magazine* **27**(1), 43–53. <https://doi.org/10.1109/MRA.2019.2955669>
- Luger T, Bär M, Seibt R, Rieger MA and Steinhilber B (2021) Using a back exoskeleton during industrial and functional tasks—Effects on muscle activity, posture, performance, usability, and wearer discomfort in a laboratory trial. *Human Factors: The Journal of the Human Factors and Ergonomics Society*, 001872082110072. <https://doi.org/10.1177/00187208211007267>
- Luger T, Seibt R, Cobb TJ, Rieger MA and Steinhilber B (2019) Influence of a passive lower-limb exoskeleton during simulated industrial work tasks on physical load, upper body posture, postural control and discomfort. *Applied Ergonomics* **80**, 152–160. <https://doi.org/10.1016/j.apergo.2019.05.018>
- Lv G, Zhu H and Gregg RD (2018) On the design and control of highly backdrivable lower-limb exoskeletons: A discussion of past and ongoing work. *IEEE Control Systems* **38**(6), 88–113. <https://doi.org/10.1109/MCS.2018.2866605>
- Madden JD (2007) Mobile robots: Motor challenges and materials solutions. *Science* **318**(5853), 1094–1097. <https://doi.org/10.1126/science.1146351>
- Madinei S, Alemi MM, Kim S, Srinivasan D and Nussbaum MA (2020) Biomechanical evaluation of passive back-support exoskeletons in a precision manual assembly task: “Expected” effects on trunk muscle activity, perceived exertion, and task performance. *Human Factors: The Journal of the Human Factors and Ergonomics Society* **62**(3), 441–457. <https://doi.org/10.1177/0018720819890966>
- Mallwitz M, Will N, Teiwes J and Kirchner E (2015) *The CAPIO active upper body exoskeleton and its application for teleoperation*.
- Mao Y and Agrawal SK (2012) Design of a cable-driven arm exoskeleton (CAREX) for neural rehabilitation. *IEEE Transactions on Robotics* **28**(4), 922–931. <https://doi.org/10.1109/TRO.2012.2189496>
- Marchal-Crespo L and Reinkensmeyer DJ (2009) Review of control strategies for robotic movement training after neurologic injury. *Journal of Neuroengineering and Rehabilitation* **6**(1), 20. <https://doi.org/10.1186/1743-0003-6-20>
- Marcheschi S, Salsedo F, Fontana M and Bergamasco M (2011) Body extender: Whole body exoskeleton for human power augmentation. *2011 IEEE International Conference on Robotics and Automation*, pp. 611–616. <https://doi.org/10.1109/ICRA.2011.5980132>
- Mathijssen G, Lefeber D and Vanderborght B (2015) Variable recruitment of parallel elastic elements: Series–parallel elastic actuators (SPEA) with dephased mutilated gears. *IEEE/ASME Transactions on Mechatronics* **20**(2), 594–602. <https://doi.org/10.1109/TMECH.2014.2307122>
- Matsuki H, Nagano K and Fujimoto Y (2019) Bilateral drive gear—A highly backdrivable reduction gearbox for robotic actuators. *IEEE/ASME Transactions on Mechatronics* **24**(6), 2661–2673. <https://doi.org/10.1109/TMECH.2019.2946403>
- Mehrholz J, Hädrich A, Platz T, Kugler J and Pohl M (2012) Electromechanical and robot-assisted arm training for improving generic activities of daily living, arm function, and arm muscle strength after stroke. In Mehrholz J (ed.), *Cochrane Database of Systematic Reviews*. Hoboken, NJ: John Wiley & Sons. <https://doi.org/10.1002/14651858.CD006876.pub3>
- Meijneke C, van Dijk W and van der Kooij H (2014) Achilles: An autonomous lightweight ankle exoskeleton to provide push-off power. *5th IEEE RAS/EMBS International Conference on Biomedical Robotics and Biomechanics*, pp. 918–923. <https://doi.org/10.1109/BIOROB.2014.6913898>
- Meyer JT, Schrade SO, Lamercy O and Gassert R (2019) User-centered design and evaluation of physical interfaces for an exoskeleton for paraplegic users. *2019 IEEE 16th International Conference on Rehabilitation Robotics (ICORR)*, pp. 1159–1166. <https://doi.org/10.1109/ICORR.2019.8779527>
- Milia P, De Salvo F, Caserio M, Cope T, Weber P, Santella C, Fiorini S, Baldoni G, Bruschi R, Bigazzi B, Cencetti S, Da Campo M, Bigazzi P and Bigazzi M (2016) Neurorehabilitation in paraplegic patients with an active powered exoskeleton (Ekso). *Digital Medicine* **2**(4), 163. https://doi.org/10.4103/digm.digm_51_16
- Molteni F, Gasperini G, Cannaviello G and Guanziroli E (2018) Exoskeleton and end-effector robots for upper and lower limbs rehabilitation: Narrative review. *PM&R* **10**, S174–S188. <https://doi.org/10.1016/j.pmrj.2018.06.005>
- Mooney L and Herr H (2013) Continuously-variable series-elastic actuator. *2013 IEEE 13th International Conference on Rehabilitation Robotics (ICORR)*, pp. 1–6. <https://doi.org/10.1109/ICORR.2013.6650402>
- Mooney LM and Herr HM (2016) Biomechanical walking mechanisms underlying the metabolic reduction caused by an autonomous exoskeleton. *Journal of Neuroengineering and Rehabilitation* **13**(1), 4. <https://doi.org/10.1186/s12984-016-0111-3>

- Mortenson WB, Pysklywec A, Chau L, Prescott M and Townson A (2020) Therapists' experience of training and implementing an exoskeleton in a rehabilitation centre. *Disability and Rehabilitation*, 1–7. <https://doi.org/10.1080/09638288.2020.1789765>
- Mouchoux J, Carisi S, Dosen S, Farina D, Schilling AF and Markovic M (2021) Artificial perception and semiautonomous control in myoelectric hand prostheses increases performance and decreases effort. *IEEE Transactions on Robotics* **37**, 1–15. <https://doi.org/10.1109/TRO.2020.3047013>
- Näf MB, Junius K, Rossini M, Rodriguez-Guerrero C, Vanderborght B and Lefeber D (2018) Misalignment compensation for full human-exoskeleton kinematic compatibility: State of the art and evaluation. *Applied Mechanics Reviews* **70**(5), 050802. <https://doi.org/10.1115/1.4042523>
- Näf MB, Koopman AS, Baltrusch S, Rodriguez-Guerrero C, Vanderborght B and Lefeber D (2018) Passive back support exoskeleton improves range of motion using flexible beams. *Frontiers in Robotics and AI* **5**, 72. <https://doi.org/10.3389/frobt.2018.00072>
- Nagai K, Nakanishi I, Hanafusa H, Kawamura S, Makikawa M and Tejima N (1998) Development of an 8 DOF robotic orthosis for assisting human upper limb motion. *Proceedings. 1998 IEEE International Conference on Robotics and Automation (Cat. No. 98CH36146)*, 4, pp. 3486–3491. <https://doi.org/10.1109/ROBOT.1998.680977>
- Naseri A, Grimmer M, Seyfarth A and Ahmad Sharbafi M (2020) Neuromechanical force-based control of a powered prosthetic foot. *Wearable Technologies* **1**, e6. <https://doi.org/10.1017/wtc.2020.6>
- National Research Council (2004) *Meeting the Energy Needs of Future Warriors*. Washington, DC: National Academies Press. <https://doi.org/10.17226/11065>
- Nazarahari M and Rouhani H (2021) 40 years of sensor fusion for orientation tracking via magnetic and inertial measurement units: Methods, lessons learned, and future challenges. *Information Fusion* **68**, 67–84. <https://doi.org/10.1016/j.inffus.2020.10.018>
- Nef T and Lum P (2009) Improving backdrivability in geared rehabilitation robots. *Medical & Biological Engineering & Computing* **47**(4), 441–447. <https://doi.org/10.1007/s11517-009-0437-0>
- Nef T, Mihelj M, Kiefer G, Perndl C, Muller R and Riener R (2007) ARMin - Exoskeleton for arm therapy in stroke patients. *2007 IEEE 10th International Conference on Rehabilitation Robotics*, pp. 68–74. <https://doi.org/10.1109/ICORR.2007.4428408>
- Nie H, Arendt-Nielsen L, Andersen H and Graven-Nielsen T (2005) Temporal summation of pain evoked by mechanical stimulation in deep and superficial tissue. *The Journal of Pain* **6**(6), 348–355. <https://doi.org/10.1016/j.jpain.2005.01.352>
- Noda T, Teramae T, Ugurlu B and Morimoto J (2014) Development of an upper limb exoskeleton powered via pneumatic electric hybrid actuators with bowden cable. *2014 IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 3573–3578. <https://doi.org/10.1109/IROS.2014.6943062>
- Novak D and Riener R (2015) A survey of sensor fusion methods in wearable robotics. *Robotics and Autonomous Systems* **73**, 155–170. <https://doi.org/10.1016/j.robot.2014.08.012>
- Ohnishi K, Shibata M and Murakami T (1996) Motion control for advanced mechatronics. *IEEE/ASME Transactions on Mechatronics* **1**(1), 56–67. <https://doi.org/10.1109/3516.491410>
- Otten A, Voort C, Stienen A, Aarts R, van Asseldonk E and van der Kooij H (2015) LIMPACT: A hydraulically powered self-aligning upper limb exoskeleton. *IEEE/ASME Transactions on Mechatronics* **20**(5), 2285–2298. <https://doi.org/10.1109/TMECH.2014.2375272>
- Ozcinar E and Ugurlu B (2021) *Torque control strategies for a torsion based series elastic actuator*. <https://doi.org/10.24433/CO.7392529.v1>
- Pacifico I, Scano A, Guanziroli E, Moise M, Morelli L, Chiavenna A, Romo D, Spada S, Colombina G, Molteni F, Giovacchini F, Vitiello N and Crea S (2020) An experimental evaluation of the proto-MATE: A novel ergonomic upper-limb exoskeleton to reduce workers' physical strain. *IEEE Robotics & Automation Magazine* **27**(1), 54–65. <https://doi.org/10.1109/MRA.2019.2954105>
- Paluska D and Herr H (2006) The effect of series elasticity on actuator power and work output: Implications for robotic and prosthetic joint design. *Robotics and Autonomous Systems* **54**(8), 667–673. <https://doi.org/10.1016/j.robot.2006.02.013>
- Perry JC, Powell JM and Rosen J (2009) Isotropy of an upper limb exoskeleton and the kinematics and dynamics of the human arm. *Applied Bionics and Biomechanics* **6**(2), 175–191. <https://doi.org/10.1080/11762320902920575>
- Pesenti M, Antonietti A, Gandolla M and Pedrocchi A (2021) Towards a functional performance validation standard for industrial low-back exoskeletons. *State of the Art Review. Sensors* **21**(3), 808. <https://doi.org/10.3390/s21030808>
- Peternel L, Noda T, Petrič T, Ude A, Morimoto J and Babič J (2016) Adaptive control of exoskeleton robots for periodic assistive behaviours based on EMG feedback minimisation. *PLoS One* **11**(2), e0148942. <https://doi.org/10.1371/journal.pone.0148942>
- Petrič T, Peternel L, Morimoto J and Babič J (2019) Assistive arm-exoskeleton control based on human muscular manipulability. *Frontiers in Neurobotics* **13**, 30. <https://doi.org/10.3389/fnbot.2019.00030>
- Pinto-Fernandez D, Torricelli D, Sanchez-Villamanan MDC, Aller F, Mombaur K, Conti R, Vitiello N, Moreno JC and Pons JL (2020) Performance evaluation of lower limb exoskeletons: A systematic review. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **28**(7), 1573–1583. <https://doi.org/10.1109/TNSRE.2020.2989481>
- Plessner HE (2018) Reproducibility vs. replicability: A brief history of a confused terminology. *Frontiers in Neuroinformatics* **11**, 76. <https://doi.org/10.3389/fninf.2017.00076>
- Plouff M, Wolfslag W and Wisse M (2017) Clutched elastic actuators. *IEEE/ASME Transactions on Mechatronics* **22**(2), 739–750. <https://doi.org/10.1109/TMECH.2017.2657224>

- Pratt JE, Krupp BT, Morse CJ and Collins SH (2004) The RoboKnee: An exoskeleton for enhancing strength and endurance during walking. *IEEE International Conference on Robotics and Automation, 2004. Proceedings. ICRA '04. 2004* **3**, 2430–2435. <https://doi.org/10.1109/ROBOT.2004.1307425>
- Pratt GA and Williamson MM (1995) Series elastic actuators. *Proceedings 1995 IEEE/RSJ International Conference on Intelligent Robots and Systems. Human Robot Interaction and Cooperative Robots* **1**, 399–406. <https://doi.org/10.1109/IROS.1995.525827>
- Pratt GA, Willisson P, Bolton C and Hofman A (2004) Late motor processing in low-impedance robots: Impedance control of series-elastic actuators. *Proceedings of the 2004 American Control Conference* **4**, 3245–3251. <https://doi.org/10.23919/ACC.2004.1384410>
- Proietti T, Crocher V, Roby-Brami A and Jarrasse N (2016) Upper-limb robotic exoskeletons for Neurorehabilitation: A review on control strategies. *IEEE Reviews in Biomedical Engineering* **9**, 4–14. <https://doi.org/10.1109/RBME.2016.2552201>
- Realmuto J, Klute G and Devasia S (2015) Nonlinear passive Cam-based springs for powered ankle prostheses. *Journal of Medical Devices* **9**(1), 011007. <https://doi.org/10.1115/1.4028653>
- Rehmat N, Zuo J, Meng W, Liu Q, Xie SQ and Liang H (2018) Upper limb rehabilitation using robotic exoskeleton systems: A systematic review. *International Journal of Intelligent Robotics and Applications* **2**(3), 283–295. <https://doi.org/10.1007/s41315-018-0064-8>
- Reinkensmeyer DJ, Wolbrecht ET, Chan V, Chou C, Cramer SC and Bobrow JE (2012) Comparison of three-dimensional, assist-as-needed robotic arm/hand movement training provided with Pneu-WREX to conventional Tabletop therapy after chronic stroke. *American Journal of Physical Medicine & Rehabilitation* **91**(11), S232–S241. <https://doi.org/10.1097/PHM.0b013e31826bce79>
- Riener R, Lunenburger L, Jezernik S, Anderschitz M, Colombo G and Dietz V (2005) Patient-cooperative strategies for robot-aided treadmill training: First experimental results. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **13**(3), 380–394. <https://doi.org/10.1109/TNSRE.2005.848628>
- Ronsse R, Vitiello N, Lenzi T, van den Kieboom J, Carrozza MC and Ijspeert AJ (2011) Human–robot synchrony: Flexible assistance using adaptive oscillators. *IEEE Transactions on Biomedical Engineering* **58**(4), 1001–1012. <https://doi.org/10.1109/TBME.2010.2089629>
- Rose CG and O'Malley MK (2019) Hybrid rigid-soft hand exoskeleton to assist functional dexterity. *IEEE Robotics and Automation Letters* **4**(1), 73–80. <https://doi.org/10.1109/LRA.2018.2878931>
- Rosen J, Brand M, Fuchs MB and Arcan M (2001) A myosignal-based powered exoskeleton system. *IEEE Transactions on Systems, Man, and Cybernetics - Part A: Systems and Humans* **31**(3), 210–222. <https://doi.org/10.1109/3468.925661>
- Rouse EJ, Mooney LM and Herr HM (2014) Clutchable series-elastic actuator: Implications for prosthetic knee design. *The International Journal of Robotics Research* **33**(13), 1611–1625. <https://doi.org/10.1177/0278364914545673>
- Royer TD and Martin PE (2005) Manipulations of leg mass and moment of inertia: Effects on energy cost of walking. *Medicine & Science in Sports & Exercise* **37**(4), 649–656. <https://doi.org/10.1249/01.MSS.0000159007.56083.96>
- Sanchez V, Walsh CJ and Wood RJ (2021) Textile technology for soft robotic and autonomous garments. *Advanced Functional Materials* **31**(6), 2008278. <https://doi.org/10.1002/adfm.202008278>
- Sanchez-Villamañan MDC, Gonzalez-Vargas J, Torricelli D, Moreno JC and Pons JL (2019) Compliant lower limb exoskeletons: A comprehensive review on mechanical design principles. *Journal of Neuroengineering and Rehabilitation* **16**(1), 55. <https://doi.org/10.1186/s12984-019-0517-9>
- Sariyildiz E, Mutlu R and Yu H (2020) A sliding mode force and position controller synthesis for series elastic actuators. *Robotica* **38**(1), 15–28. <https://doi.org/10.1017/S0263574719000420>
- Schmidt K, Duarte JE, Grimmer M, Sancho-Puchades A, Wei H, Easthope CS and Riener R (2017) The Myosuit: Bi-articular anti-gravity exosuit that reduces hip extensor activity in sitting transfers. *Frontiers in Neurobotics* **11**, 57. <https://doi.org/10.3389/fnbot.2017.00057>
- Schürmann T, Mohler BJ, Peters J and Beckerle P (2019) How Cognitive Models of Human Body Experience Might Push Robotics. *Front. Neurobot.* **13**:14. doi: 10.3389/fnbot.2019.00014
- Sensingier JW, Clark SD and Schorsch JF (2011) Exterior vs. interior rotors in robotic brushless motors. *2011 IEEE International Conference on Robotics and Automation*, pp. 2764–2770. <https://doi.org/10.1109/ICRA.2011.5979940>
- Seok S, Wang A, Chuah MY, Hyun DJ, Lee J, Otten DM, Lang JH and Kim S (2015) Design principles for energy-efficient legged locomotion and implementation on the MIT cheetah robot. *IEEE/ASME Transactions on Mechatronics* **20**(3), 1117–1129. <https://doi.org/10.1109/TMECH.2014.2339013>
- Seyfarth A, Geyer H, Blickhan R, Lipfert S, Rummel J, Minekawa Y and Iida F (2006) Running and walking with compliant legs. In *Fast Motions in Biomechanics and Robotics*. Berlin: Springer, pp. 383–401. https://doi.org/10.1007/978-3-540-36119-0_18
- Shepherd MK and Rouse EJ (2017) The VSPA foot: A quasi-passive ankle-foot prosthesis with continuously variable stiffness. *IEEE Transactions on Neural Systems and Rehabilitation Engineering : A Publication of the IEEE Engineering in Medicine and Biology Society* **25**(12), 2375–2386. <https://doi.org/10.1109/TNSRE.2017.2750113>
- Shi C, Yang D, Zhao J and Liu H (2020) Computer vision-based grasp pattern recognition with application to myoelectric control of dexterous hand prosthesis. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **28**(9), 2090–2099. <https://doi.org/10.1109/TNSRE.2020.3007625>

- Singer R, Maufroy C and Schneider U** (2020) Automatic support control of an upper body exoskeleton – Method and validation using the Stuttgart Exo-jacket. *Wearable Technologies* **1**, e2. <https://doi.org/10.1017/wtc.2020.1>
- Sirawattanakul S and Sanngoen W** (2020) Review of upper limb exoskeleton for rehabilitation and assistive application. *International Journal of Mechanical Engineering and Robotics Research* **9**, 752–758. <https://doi.org/10.18178/ijmerr.9.5.752-758>
- Slajpah S, Kamnik R and Munih M** (2017) Compensation for magnetic disturbances in motion estimation to provide feedback to wearable robotic systems. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **25**(12), 2398–2406. <https://doi.org/10.1109/TNSRE.2017.2760356>
- Soekadar SR, Witkowski M, Vitiello N and Birbaumer N** (2015) An EEG/EOG-based hybrid brain-neural computer interaction (BNCI) system to control an exoskeleton for the paralyzed hand. *Biomedical Engineering / Biomedizinische Technik* **60**(3), 199–205. <https://doi.org/10.1515/bmt-2014-0126>
- Spada S, Ghibaud L, Gilotta S, Gastaldi L and Cavatorta MP** (2017) Investigation into the applicability of a passive upper-limb exoskeleton in automotive industry. *Procedia Manufacturing* **11**, 1255–1262. <https://doi.org/10.1016/j.promfg.2017.07.252>
- Spanias JA, Simon AM, Finucane SB, Perreault EJ and Hargrove LJ** (2018) Online adaptive neural control of a robotic lower limb prosthesis. *Journal of Neural Engineering* **15**(1), 016015. <https://doi.org/10.1088/1741-2552/aa92a8>
- Sposito M, Di Natali C, Toxiri S, Caldwell DG, De Momi E and Ortiz J** (2020) Exoskeleton kinematic design robustness: An assessment method to account for human variability. *Wearable Technologies* **1**, e7. <https://doi.org/10.1017/wtc.2020.7>
- Stadler KS, Altenburger R, Schmidhauser E, Scherly D, Ortiz J, Toxiri S, Mateos L and Masood J** (2016) Robo-Mate an exoskeleton for industrial use: Concept and mechanical design. In *Advances in Cooperative Robotics*. Singapore: World Scientific, pp. 806–813. https://doi.org/10.1142/9789813149137_0094
- Sterling M** (2011) Pressure algometry: What does it really tell us? *Journal of Orthopaedic & Sports Physical Therapy* **41**(9), 623–624. <https://doi.org/10.2519/jospt.2011.0106>
- Stienen AHA, Hekman EEG, van der Helm FCT and van der Kooij H** (2009) Self-aligning exoskeleton axes through decoupling of joint rotations and translations. *IEEE Transactions on Robotics* **25**(3), 628–633. <https://doi.org/10.1109/TRO.2009.2019147>
- Sugar TG and Holgate M** (2013). *Compliant mechanisms for robotic ankles. Volume 6A: 37th Mechanisms and Robotics Conference*. <https://doi.org/10.1115/DETC2013-12167>
- Teramae T, Noda T and Morimoto J** (2018) EMG-based model predictive control for physical human–robot interaction: Application for assist-as-needed control. *IEEE Robotics and Automation Letters* **3**(1), 210–217. <https://doi.org/10.1109/LRA.2017.2737478>
- Theurel J and Desbrosses K** (2019) Occupational exoskeletons: Overview of their benefits and limitations in preventing work-related musculoskeletal disorders. *IIEE Transactions on Occupational Ergonomics and Human Factors* **7**(3–4), 264–280. <https://doi.org/10.1080/24725838.2019.1638331>
- Tiseni L. et al.**, “On the edge between soft and rigid: an assistive shoulder exoskeleton with hyper-redundant kinematics,” 2019 IEEE 16th International Conference on Rehabilitation Robotics (ICORR), 2019, pp. 618–624, doi: 10.1109/ICORR.2019.8779546.
- Torricelli D, Gonzalez-Vargas J, Veneman JF, Mombaur K, Tsagarakis N, Del-Ama AJ, Gil-Agudo A, Moreno JC and Pons JL** (2015) Benchmarking bipedal locomotion: A unified scheme for humanoids, wearable robots, and humans. *IEEE Robotics & Automation Magazine* **22**(3), 103–115. <https://doi.org/10.1109/MRA.2015.2448278>
- Torricelli D and Pons JL** (2019) *EUROBENCH: Preparing Robots for the Real World*. Cham: Springer, pp. 375–378. https://doi.org/10.1007/978-3-030-01887-0_72
- Torricelli D, Rodriguez-Guerrero C, Veneman JF, Crea S, Briem K, Lenggenhager B and Beckerle P** (2020) Benchmarking wearable robots: Challenges and recommendations from functional, user experience, and methodological perspectives. *Frontiers in Robotics and AI* **7**, 168. <https://doi.org/10.3389/frobt.2020.561774>
- Toxiri S, Ortiz J, Masood J, Fernandez J, Mateos LA and Caldwell DG** (2015) A wearable device for reducing spinal loads during lifting tasks: Biomechanics and design concepts. *2015 IEEE International Conference on Robotics and Biomimetics (ROBIO)*, pp. 2295–2300. <https://doi.org/10.1109/ROBIO.2015.7419116>
- Tran M, Gabert L, Cempini M and Lenzi T** (2019) A lightweight, efficient fully powered knee prosthesis with actively variable transmission. *IEEE Robotics and Automation Letters* **4**(2), 1186–1193. <https://doi.org/10.1109/LRA.2019.2892204>
- Tsagarakis NG and Caldwell DG** (2003) Development and control of a ‘soft-actuated’ exoskeleton for use in physiotherapy and training. *Autonomous Robots* **15**(1), 21–33. <https://doi.org/10.1023/A:1024484615192>
- Tucker MR, Olivier J, Pagel A, Bleuler H, Bouri M, Lamercy O, Millán JR, Riener R, Vallery H and Gassert R** (2015) Control strategies for active lower extremity prosthetics and orthotics: A review. *Journal of Neuroengineering and Rehabilitation* **12**(1), 1. <https://doi.org/10.1186/1743-0003-12-1>
- Ugurlyu B, Forni P, Doppmann C, Sariyildiz E and Morimoto J** (2019) Stable control of force, position, and stiffness for robot joints powered via pneumatic muscles. *IEEE Transactions on Industrial Informatics* **15**(12), 6270–6279. <https://doi.org/10.1109/TII.2019.2916228>
- Ugurlyu B, Nishimura M, Hyodo K, Kawanishi M and Narikiyo T** (2015) Proof of concept for robot-aided upper limb rehabilitation using disturbance observers. *IEEE Transactions on Human-Machine Systems* **45**(1), 110–118. <https://doi.org/10.1109/THMS.2014.2362816>

- Ugurlu B, Oshima H, Sariyildiz E, Narikiyo T and Babic J (2020) Active compliance control reduces upper body effort in exoskeleton-supported walking. *IEEE Transactions on Human-Machine Systems* **50**(2), 144–153. <https://doi.org/10.1109/THMS.2019.2961969>
- Ugurlu Barkan, Emre Sariyildiz, Ahmet Talha Kansizoglu, Erim Can Ozcinar, and Sinan Coruk, “Benchmarking Torque Control Strategies for a Torsion based Series Elastic Actuator”, *IEEE Robotics & Automation Magazine*, Accepted, 2021, 10.1109/MRA.2021.3124154.
- Ugurlu Barkan, Corinne Doppmann, Masashi Hamaya, Paolo Forni, Tatsuya Teramae, Tomoyuki Noda, and Jun Morimoto, Variable Ankle Stiffness Improves Balance Control: Experiments on a Bipedal Exoskeleton, *IEEE/ASME Transactions on Mechatronics*, vol. **21**, no. 1, February 2016, pp. 79–87. DOI: 10.1109/TMECH.2015.2448932
- Vallery H, Ekkelenkamp R, van der Kooij H and Buss M (2007) Passive and accurate torque control of series elastic actuators. 2007 *IEEE/RSJ International Conference on Intelligent Robots and Systems*, pp. 3534–3538. <https://doi.org/10.1109/IROS.2007.4399172>
- Vallery H, Veneman J, van Asseldonk E, Ekkelenkamp R, Buss M and van Der Kooij H (2008) Compliant actuation of rehabilitation robots. *IEEE Robotics & Automation Magazine* **15**(3), 60–69. <https://doi.org/10.1109/MRA.2008.927689>
- van der Kooij H, van Asseldonk E, van Oort G, Sluiter V, Emmens A, Witteveen H, Tagliamonte NL, Tamburella F, Pisotta I, Masciullo M, Arquilla M, Molinari M, Wu A, Ijspeert A, Dzeladini FF, Thorsteinsson F, Arami A, Burdet E, Huang H-Y., Gregoor W and Meijneke C (2019) *Symbiotron: Symbiotic Man-Machine Interactions in Wearable Exoskeletons to Enhance Mobility for Paraplegics*, Carrozza MC, Micera S and Pons JL (eds). Cham: Springer International Publishing, pp. 361–364. https://doi.org/10.1007/978-3-030-01887-0_69
- van der Kooij H, Veneman JF and Ekkelenkamp R (2006) Compliant actuation of exoskeletons. In *Mobile Robots: Towards New Applications*. Wein: I-Tech Education and Publishing. <https://doi.org/10.5772/4688>
- Van Engelenhoven L, Poon N, Kazerooni H, Barr A, Rempel D and Harris-Adamson C (2018) Evaluation of an adjustable support shoulder exoskeleton on static and dynamic overhead tasks. *Proceedings of the Human Factors and Ergonomics Society Annual Meeting* **62**(1), 804–808. <https://doi.org/10.1177/1541931218621184>
- Vanderborght B, Albu-Schaeffer A, Bicchi A, Burdet E, Caldwell DG, Carloni R, Catalano M, Eiberger O, Friedl W, Ganesh G, Garabini M, Grebenstein M, Grioli G, Haddadin S, Hoppner H, Jafari A, Laffranchi M, Lefeber D, Petit F, Stramigioli S, Tsagarakis N, Van Damme M, Van Ham R, Visser LC and Wolf S (2013) Variable impedance actuators: A review. *Robotics and Autonomous Systems* **61**(12), 1601–1614. <https://doi.org/10.1016/j.robot.2013.06.009>
- Varghese RJ, Mukherjee G, King R, Keller S and Deshpande AD (2018) Designing variable stiffness profiles to optimize the physical human robot interface of hand exoskeletons. 2018 *7th IEEE International Conference on Biomedical Robotics and Biomechatronics (Biorob)*, pp. 1101–1108. <https://doi.org/10.1109/BIOROB.2018.8487862>
- Varol HA, Sup F and Goldfarb M (2010) Multiclass real-time intent recognition of a powered lower limb prosthesis. *IEEE Transactions on Biomedical Engineering* **57**(3), 542–551. <https://doi.org/10.1109/TBME.2009.2034734>
- Vassallo C, De Giuseppe S, Piezzo C, Maludrottu S, Cerruti G, D’Angelo ML, Gruppioni E, Marchese C, Castellano S, Guanziroli E, Molteni F, Laffranchi M and De Michieli L (2020) Gait patterns generation based on basis functions interpolation for the TWIN lower-limb exoskeleton. 2020 *IEEE International Conference on Robotics and Automation (ICRA)*, pp. 1778–1784. <https://doi.org/10.1109/ICRA40945.2020.9197250>
- Veale AJ and Xie SQ (2016) Towards compliant and wearable robotic orthoses: A review of current and emerging actuator technologies. *Medical Engineering & Physics* **38**(4), 317–325. <https://doi.org/10.1016/j.medengphy.2016.01.010>
- Veneman JF, Ekkelenkamp R, Kruidhof R, van der Helm FCT and van der Kooij H (2006) A series elastic- and Bowden-cable-based actuation system for use as torque actuator in exoskeleton-type robots. *The International Journal of Robotics Research* **25** (3), 261–281. <https://doi.org/10.1177/02783649060603829>
- Veneman JF, Kruidhof R, Hekman EEG, Ekkelenkamp R, Van Asseldonk EHF and van der Kooij H (2007) Design and evaluation of the LOPES exoskeleton robot for interactive gait rehabilitation. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **15**(3), 379–386. <https://doi.org/10.1109/TNSRE.2007.903919>
- Verstraten T, Flynn L, Geeroms J, Vanderborght B and Lefeber D (2018) On the electrical energy consumption of active ankle prostheses with series and parallel elastic elements. 2018 *7th IEEE International Conference on Biomedical Robotics and Biomechatronics (Biorob)*, pp. 720–725. <https://doi.org/10.1109/BIOROB.2018.8487656>
- Verstraten T, Furnémont R, López-García P, Crispel S, Vanderborght B and Lefeber D (2019) *A Series Elastic Dual-Motor Actuator Concept for Wearable Robotics*. Cham: Springer, pp. 165–169. https://doi.org/10.1007/978-3-030-01887-0_32
- Verstraten T, Geeroms J, Mathijssen G, Convens B, Vanderborght B and Lefeber D (2017) Optimizing the power and energy consumption of powered prosthetic ankles with series and parallel elasticity. *Mechanism and Machine Theory* **116**, 419–432. <https://doi.org/10.1016/j.mechmachtheory.2017.06.004>
- Veeger H.E.J., The position of the rotation center of the glenohumeral joint, *Journal of Biomechanics*, Volume **33**, Issue 12, 2000, Pages 1711–1715, ISSN 0021-9290, [https://doi.org/10.1016/S0021-9290\(00\)00141-X](https://doi.org/10.1016/S0021-9290(00)00141-X).
- Vitiello N, Lenzi T, Roccella S, De Rossi SMM, Cattin E, Giovacchini F, Vecchi F and Carrozza MC (2013) NEUROExos: A powered elbow exoskeleton for physical rehabilitation. *IEEE Transactions on Robotics* **29**(1), 220–235. <https://doi.org/10.1109/TRO.2012.2211492>
- Vouga T, Baud R, Fasola J, Bouri M and Bleuler H (2017) TWICE – A lightweight lower-limb exoskeleton for complete paraplegics. 2017 *International Conference on Rehabilitation Robotics (ICORR)*, pp. 1639–1645. <https://doi.org/10.1109/ICORR.2017.8009483>

- Wang J, Li X, Huang T-H, Yu S, Li Y, Chen T, Carriero A, Oh-Park M and Su H (2018) Comfort-centered design of a lightweight and backdrivable knee exoskeleton. *IEEE Robotics and Automation Letters* 3(4), 4265–4272. <https://doi.org/10.1109/LRA.2018.2864352>
- Wang S, Wang L, Meijneke C, van Asseldonk E, Hoellinger T, Cheron G, Ivanenko Y, La Scaleia V, Sylos-Labini F, Molinari M, Tamburella F, Pisotta I, Thorsteinnsson F, Ilzkovitz M, Gancet J, Nevatia Y, Hauffe R, Zanow F and van der Kooij H (2015) Design and control of the MINDWALKER exoskeleton. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 23(2), 277–286. <https://doi.org/10.1109/TNSRE.2014.2365697>
- Wang Z, Wu X, Zhang Y, Chen C, Liu S, Liu Y, Peng A and Ma Y (2021) A semi-active exoskeleton based on EMGs reduces muscle fatigue when squatting. *Frontiers in Neurobotics* 15, 625479. <https://doi.org/10.3389/fnbot.2021.625479>
- Washabaugh EP, Guo J, Chang C-K, Remy CD and Krishnan C (2019) A portable passive rehabilitation robot for upper-extremity functional resistance training. *IEEE Transactions on Biomedical Engineering* 66(2), 496–508. <https://doi.org/10.1109/TBME.2018.2849580>
- Wehner M, Quinlivan B, Aubin PM, Martínez-Villalpando E, Baumann M, Stirling L, Holt K, Wood R and Walsh C (2013) A lightweight soft exosuit for gait assistance. *2013 IEEE International Conference on Robotics and Automation*, pp. 3362–3369. <https://doi.org/10.1109/ICRA.2013.6631046>
- Wehner M, Rempel D and Kazerooni H (2009) Lower extremity exoskeleton reduces Back forces in lifting. *ASME 2009 Dynamic Systems and Control Conference* 2, 49–56. <https://doi.org/10.1115/DSCC2009-2644>
- Wendong W, Hanhao L, Menghan X, Yang C, Xiaoqing Y, Xing M and Bing Z (2020) Design and verification of a human–robot interaction system for upper limb exoskeleton rehabilitation. *Medical Engineering & Physics* 79, 19–25. <https://doi.org/10.1016/j.medengphy.2020.01.016>
- Weston EB, Alizadeh M, Knapik GG, Wang X and Marras WS (2018) Biomechanical evaluation of exoskeleton use on loading of the lumbar spine. *Applied Ergonomics* 68, 101–108. <https://doi.org/10.1016/j.apergo.2017.11.006>
- Windrich M, Grimmer M, Christ O, Rinderknecht S and Beckerle P (2016) Active lower limb prosthetics: A systematic review of design issues and solutions. *Biomedical Engineering Online* 15(S3), 140. <https://doi.org/10.1186/s12938-016-0284-9>
- Wittmann F, Lamberg O and Gassert R (2019) Magnetometer-based drift correction during rest in IMU arm motion tracking. *Sensors* 19(6), 1312. <https://doi.org/10.3390/s19061312>
- Writers C (2018) Rise of the exoskeletons: Latest news in built environment research and innovation. *Construction Research and Innovation* 9(4), 86–90. <https://doi.org/10.1080/20450249.2018.1556515>
- Xiao ZG and Menon C (2019) A review of force Myography Research and Development. *Sensors* 19(20), 4557. <https://doi.org/10.3390/s19204557>
- Xiloyannis M, Chiaradia D, Frisoli A and Masia L (2019) Physiological and kinematic effects of a soft exosuit on arm movements. *Journal of Neuroengineering and Rehabilitation* 16(1), 29. <https://doi.org/10.1186/s12984-019-0495-y>
- Yan T, Cempini M, Oddo CM and Vitiello N (2015) Review of assistive strategies in powered lower-limb orthoses and exoskeletons. *Robotics and Autonomous Systems* 64, 120–136. <https://doi.org/10.1016/j.robot.2014.09.032>
- Yandell MB, Quinlivan BT, Popov D, Walsh C and Zelik KE (2017) Physical interface dynamics alter how robotic exosuits augment human movement: Implications for optimizing wearable assistive devices. *Journal of Neuroengineering and Rehabilitation* 14(1), 40. <https://doi.org/10.1186/s12984-017-0247-9>
- Yandell MB, Ziemnicki DM, McDonald KA and Zelik KE (2020) Characterizing the comfort limits of forces applied to the shoulders, thigh and shank to inform exosuit design. *PLoS One* 15(2), e0228536. <https://doi.org/10.1371/journal.pone.0228536>
- Yang X, Huang T-H, Hu H, Yu S, Zhang S, Zhou X, Carriero A, Yue G and Su H (2019) Spine-inspired continuum soft exoskeleton for stoop lifting assistance. *IEEE Robotics and Automation Letters* 4(4), 4547–4554. <https://doi.org/10.1109/LRA.2019.2935351>
- Yildirim MC, Kansizoglu AT, Sendur P and Ugurlu B (2019) *High Power Series Elastic Actuator Development for Torque-Controlled Exoskeletons*. Cham: Springer, pp. 70–74. https://doi.org/10.1007/978-3-030-01887-0_14
- Yildirim MC, Sendur P, Kansizoglu AT, Uras U, Bilgin O, Emre S, Yapici GG, Arik M and Ugurlu B (2021) Design and development of a durable series elastic actuator with an optimized spring topology. *Proceedings of the Institution of Mechanical Engineers, Part C: Journal of Mechanical Engineering Science* (accepted).
- Yu S, Huang T-H, Wang D, Lynn B, Sayd D, Silivanov V, Park YS, Tian Y and Su H (2019) Design and control of a high-torque and highly Backdrivable hybrid soft exoskeleton for knee injury prevention during squatting. *IEEE Robotics and Automation Letters* 4(4), 4579–4586. <https://doi.org/10.1109/LRA.2019.2931427>
- Yu S, Huang T-H, Yang X, Jiao C, Yang J, Chen Y, Yi J and Su H (2020) Quasi-direct drive actuation for a lightweight hip exoskeleton with high Backdrivability and high bandwidth. *IEEE/ASME Transactions on Mechatronics* 25(4), 1794–1802. <https://doi.org/10.1109/TMECH.2020.2995134>
- Yun Y, Na Y, Esmatloo P, Dancausse S, Serrato A, Agarwal P and Deshpande AD (2020) Improvement of hand functions of spinal cord injury patients with electromyography-driven hand exoskeleton: A feasibility study. *Wearable Technologies* 1, e8. <https://doi.org/10.1017/wtc.2020.9>
- Zhang F, D'Andrea SE, Nunnery MJ, Kay SM and Huang H (2011) Towards design of a stumble detection system for artificial legs. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* 19(5), 567–577. <https://doi.org/10.1109/TNSRE.2011.2161888>
- Zhang T and Huang H (2018) A lower-back robotic exoskeleton: Industrial handling augmentation used to provide spinal support. *IEEE Robotics & Automation Magazine* 25(2), 95–106. <https://doi.org/10.1109/MRA.2018.2815083>

- Zhang T, Tran M and Huang H** (2019) Admittance shaping-based assistive control of SEA-driven robotic hip exoskeleton. *IEEE/ASME Transactions on Mechatronics* **24**(4), 1508–1519. <https://doi.org/10.1109/TMECH.2019.2916546>
- Zhang K, Xiong C, Zhang W, Liu H, Lai D, Rong Y and Fu C** (2019) Environmental features recognition for lower limb prostheses toward predictive walking. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **27**(3), 465–476. <https://doi.org/10.1109/TNSRE.2019.2895221>
- Zheng E, Wang Q and Qiao H** (2019) Locomotion mode recognition with robotic transtibial prosthesis in inter-session and inter-day applications. *IEEE Transactions on Neural Systems and Rehabilitation Engineering* **27**(9), 1836–1845. <https://doi.org/10.1109/TNSRE.2019.2934525>
- Zhu H, Nesler C, Divekar N, Peddinti V and Gregg R** (2021) Design principles for compact, backdrivable actuation in partial-assist powered knee orthoses. *IEEE/ASME Transactions on Mechatronics*, p. 1. <https://doi.org/10.1109/TMECH.2021.3053226>
- Zimmermann Y, Forino A, Rienen R and Hutter M** (2019) ANYexo: A versatile and dynamic upper-limb rehabilitation robot. *IEEE Robotics and Automation Letters* **4**(4), 3649–3656. <https://doi.org/10.1109/LRA.2019.2926958>
- Zinn M, Khatib O and Roth B** (2004) A new actuation approach for human friendly robot design. *IEEE International Conference on Robotics and Automation, 2004. Proceedings. ICRA '04*, Vol. 1, pp. 249–254. <https://doi.org/10.1109/ROBOT.2004.1307159>



Jan Babič received his PhD degree from Faculty of Electrical Engineering, University in Ljubljana, Slovenia, examining the role of biarticular muscles in human locomotion. During the years 2006/2007, he was a visiting researcher at ATR Computational Neuroscience Laboratories in Japan. In November 2014, he was a visiting professor at The Institute for Intelligent Systems and Robotics, University of Pierre and Marie Curie in France. Currently, he is the head of Laboratory for Neuromechanics and Biorobotics at Jožef Stefan Institute in Slovenia. His current research is concerned with robustness and adaptations of human movement and with the design of biologically plausible solutions for a broad spectrum of robotic systems such as industrial robots, humanoids, exoskeletons, and rehabilitation devices.



Matteo Laffranchi is the R&D Coordinator of the Rehab Technologies Lab. His expertise lies in the development of novel mechatronic systems in the robotics field. He received his master's degree in Mechatronics Engineering at Polytechnic of Turin in 2006 and a PhD degree in Robotics from the University of Sheffield in 2011. After a brief experience in the automation industry at OSAI A.S., from 2008 to 2011, he was a research fellow at the Italian Institute of Technology and a postdoctoral researcher at the same institute from 2011 to 2014. Since then, he is actively involved into technology transfer and product development activities. During his PhD and postdoc, he spent 6 years developing compliant actuators and robots for safe human robot collaboration. Currently, he is the head of research and product development activities at the Rehab Technologies Lab, with focus on robotic medical devices.



Federico Tessari received his MSc degree in Mechatronic Engineering at Politecnico di Torino, Italy, in 2016, then he received his PhD degree in Mechanical Engineering at Politecnico di Torino in collaboration with the Italian Institute of Technology (IIT) in 2021. He is currently working as postdoctoral researcher in the Rehab Technologies Laboratories of the IIT. His research interests include novel actuating solutions for powered wearable robots and bionic limbs, multibody kinematic and dynamic analysis of rehabilitation robotic devices as well as innovative control strategies to enhance human robot interaction in upper limb exoskeletons.



Tom Verstraten was born in Brussels, Belgium. He received his MSc degree in Electromechanical Engineering from the Vrije Universiteit Brussel (VUB) in 2012 and his PhD degree from the same institution in 2018. He was appointed as Assistant Professor at the Robotics & Multibody Mechanics Research Group (R&MM) of the VUB in 2020. He was awarded fellowships of the Research Foundation—Flanders (FWO) for both his doctoral and postdoctoral research, as well as a Fulbright grant for visiting scholars for a research stay at the University of Tulsa (United States) in 2018–2019. He also worked in industry as a R&D Engineer at Aquasystems International (2012–2013) and as a visiting researcher at TU Darmstadt (Germany) in 2017. His main research focus is the study and development of energy-efficient actuation systems for robotic prostheses, exoskeletons, and collaborative robots. Research interests include elastic actuators and redundant actuation.



Domen Novak received their diploma and PhD degrees in Electrical Engineering from the University of Ljubljana, Slovenia, in 2008 and 2011, respectively. They were a postdoctoral fellow at ETH Zurich, Switzerland, from 2012 to 2014. They joined the University of Wyoming in 2014 as an Assistant Professor and were promoted to Associate Professor in 2020. Their research spans diverse areas of human–machine interaction, including wearable robotics, rehabilitation engineering, serious games, and affective computing.




Nejc Šarabon received his BSc degree in Sports Education as well as Physical Therapy, in Slovenia in 2000 and 2001, respectively, and received his PhD in 2004. Today, he is the dean of Faculty of Health Sciences (University of Primorska), where he also works as a lecturer and researcher, and the head of Human Health Research Group at Innorenew CoE, Izola. His scientific focus is motor abilities and neuromuscular control in the context of sport performance, health, and ergonomics. He is closely involved in bringing scientific advances to society and has authored over 15 patents. He is regularly involved with various national and international sports and health organization to translate scientific knowledge into practice.



Barkan Ugurlu received the PhD degree in Electrical and Computer Engineering from Yokohama National University, Yokohama, Japan, in 2010, with the Monbukagakusho scholarship granted by the Ministry of Education, Culture, Sports, and Technology of Japan. From May 2010 to March 2013, he was a postdoctoral researcher, with the Istituto Italiano di Tecnologia, Genova, Italy, and Toyota Technological Institute, Nagoya, Japan. Between March 2013 and February 2015, he was a Research Scientist with the Computational Neuroscience Laboratories, Advanced Telecommunications Research Institute International (ATR), Kyoto, Japan. He currently holds an Assistant Professor position with the Department of Mechanical Engineering, Ozyegin University, Istanbul, Turkey. His research interests include biological sensorimotor control and motor recovery, active orthoses and exoskeletons, robot-aided rehabilitation, humanoid/quadruped locomotion control, and multibody dynamics, force, and compliance control.



Luka Peternel received a PhD degree in robotics from University of Ljubljana, Slovenia in 2015. He conducted PhD studies at the Department for Automation, Biocybernetics and Robotics, Jožef Stefan Institute in Ljubljana from 2011 to 2015, and at the Department of Brain–Robot Interface, ATR Computational Neuroscience Laboratories in Kyoto, Japan in 2013 and 2014. He was with the Human–Robot Interfaces and Physical Interaction Lab of Italian Institute of Technology in Genoa, Italy from 2015 to 2018. From 2019, he is with the Department of Cognitive Robotics, Delft University of Technology in the Netherlands.



Diego Torricelli received the MS degree in Mechanical Engineering in 2004, and the PhD degree in Biomedical Engineering in 2009, both from the University of Roma TRE. Since 2010, he is with the Neural Rehabilitation Group of the Spanish National Research Council (CSIC). His research line spans the understanding of motor control principles in healthy and impaired people, and the development of benchmarking techniques for human and robot performance evaluation. He is currently the coordinator of the H2020 project EUROBENCH (www.eurobench2020.eu).



Jan F. Veneman received his doctoral degree in 2007 from the Faculty of Engineering Technology, University of Twente, the Netherlands, on the development and evaluation of the LOPES exoskeleton robot for interactive gait rehabilitation. From 2008 to 2018, he continued his research on rehabilitation technology with Tecnalia Research and Innovation, a private research institute in the Basque country. Since 2018, he has been employed as Technical Project Lead in Hocoma AG, Switzerland, the world-leading manufacturer of robotic devices for neuro-rehabilitation, coordinating research and development projects in the field of lower extremity rehabilitation (gait and balance). He is currently Scientific and Management Committee Chair of the European wide COST Action network on Wearable Robots (CA16116—wearablerobots.eu). He has also been closely involved as a national representative in the development of the IEC 80601-2-78 safety standard for rehabilitation robots.