Automatic attenuation map estimation from SPECT data only for brain perfusion scans using convolutional neural networks

Chen, Yuan; Goorden, Marlies C.; Beekman, Freek J.

DOI

[10.1088/1361-6560/abe557](https://doi.org/10.1088/1361-6560/abe557)

Publication date

2021

Document Version

Final published version

Published in

Physics in Medicine and Biology

Citation (APA)

Chen, Y., Goorden, M. C., & Beekman, F. J. (2021). Automatic attenuation map estimation from SPECT data only for brain perfusion scans using convolutional neural networks. *Physics in Medicine and Biology*, 66(6), [065006]. <https://doi.org/10.1088/1361-6560/abe557>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

PAPER • OPEN ACCESS

Automatic attenuation map estimation from SPECT data only for brain perfusion scans using convolutional neural networks

To cite this article: Yuan Chen *et al* 2021 *Phys. Med. Biol.* **66** 065006

View the [article online](#) for updates and enhancements.

IPEM | IOP

Series in Physics and Engineering in Medicine and Biology

Your publishing choice in medical physics,
biomedical engineering and related subjects.

Start exploring the collection—download the
first chapter of every title for free.

PAPER

OPEN ACCESS

RECEIVED
13 May 2020REVISED
25 January 2021ACCEPTED FOR PUBLICATION
11 February 2021PUBLISHED
3 March 2021

Original content from this work may be used under the terms of the [Creative Commons Attribution 4.0 licence](#).

Any further distribution of this work must maintain attribution to the author(s) and the title of the work, journal citation and DOI.

Automatic attenuation map estimation from SPECT data only for brain perfusion scans using convolutional neural networks

Yuan Chen¹, Marlies C Goorden¹ and Freek J Beekman^{1,2,3}¹ Section Biomedical Imaging, Department of Radiation, Science and Technology, Delft University of Technology, Delft, The Netherlands² MILabs B.V., Utrecht, The Netherlands³ Department of Translational Neuroscience, Brain Center Rudolf Magnus, University Medical Center Utrecht, The NetherlandsE-mail: y.chen-4@tudelft.nl**Keywords:** CNN, attenuation correction, quantitative SPECT, multi-pinhole SPECT, deep learning, Monte Carlo simulation

Abstract

In clinical brain SPECT, correction for photon attenuation in the patient is essential to obtain images which provide quantitative information on the regional activity concentration per unit volume ($\text{kBq} \cdot \text{ml}^{-1}$). This correction generally requires an attenuation map (μ map) denoting the attenuation coefficient at each voxel which is often derived from a CT or MRI scan. However, such an additional scan is not always available and the method may suffer from registration errors. Therefore, we propose a SPECT-only-based strategy for μ map estimation that we apply to a stationary multi-pinhole clinical SPECT system (G-SPECT-I) for $^{99\text{m}}\text{Tc}$ -HMPAO brain perfusion imaging. The method is based on the use of a convolutional neural network (CNN) and was validated with Monte Carlo simulated scans. Data acquired in list mode was used to employ the energy information of both primary and scattered photons to obtain information about the tissue attenuation as much as possible. Multiple SPECT reconstructions were performed from different energy windows over a large energy range. Locally extracted 4D SPECT patches (three spatial plus one energy dimension) were used as input for the CNN which was trained to predict the attenuation coefficient of the corresponding central voxel of the patch. Results show that Attenuation Correction using the Ground Truth μ maps (GT-AC) or using the CNN estimated μ maps (CNN-AC) achieve comparable accuracy. This was confirmed by a visual assessment as well as a quantitative comparison; the mean deviation from the GT-AC when using the CNN-AC is within 1.8% for the standardized uptake values in all brain regions. Therefore, our results indicate that a CNN-based method can be an automatic and accurate tool for SPECT attenuation correction that is independent of attenuation data from other imaging modalities or human interpretations about head contours.

1. Introduction

In SPECT, attenuation of photons in tissue hampers quantitative analysis of regional tracer uptake and may lead to image artefacts. Attenuation correction (AC) is thus required to improve the diagnostic value and quantitative accuracy of reconstructed images. Besides, quantitative SPECT (i.e. SPECT that provides a precise estimate of the activity level in $\text{kBq} \cdot \text{ml}^{-1}$) is also important for dosimetry planning (Bailey and Willowson 2013).

Correction for photon attenuation is commonly based on a 3D map (attenuation map or μ map) that quantifies the amount of attenuation in each voxel within the patient at the given photon energy. Today, these μ maps are often derived from an additional CT or MRI scan. However, such an additional scan may not be available, can add radiation dose in case of CT, and is prone to registration errors (King *et al* 2004, Goetze *et al* 2007, Berker and Li 2016). Besides the use of additional CT or MRI scans, simple methods based solely on emission data that delineate the object contour and assume a uniform attenuation within the contour are used. Currently, manual placement of an ellipse approximating the head contour is still the most widely implemented approach for brain studies with commercial SPECT systems (Zaidi and Hasegawa 2003). Such a method

however can be highly subjective and suffer from limited accuracy due to the operator dependency, coarse approximation of the skull contour and lack of internal head anatomy.

In clinical SPECT, photon-interactions in biological tissues are dominated by Compton scatter while photoelectric absorption is almost negligible. For example, for photons at 140 keV traveling through brain, only 1% (a linear coefficient of 0.0015/cm) of the total attenuation (0.1461/cm) is due to photoelectric effects. Therefore, one might expect that the detected scattered photons contain essential information about attenuation maps. Conventionally, the use of SPECT images reconstructed from scattered photons or projections in scatter windows for μ map generation has been investigated in the 90s for brain (Macey *et al* 1988) and body (cardiac and liver) SPECT (Macey *et al* 1988, Wallis *et al* 1995, Pan *et al* 1996, Younes *et al* 1998), yet in most cases solely a contour was estimated from a single scatter window while a uniform attenuation coefficient was assigned to the volume within the contour. This approach is seldom applied clinically, possibly due to the increased complexity in light of the limited improvement of accuracy compared to the ellipse method. With the advance of SPECT instruments, list mode acquisition—in which case the estimated interaction position as well as the energy are recorded simultaneously for every detected event—is gaining popularity. In the work of Jha *et al* (2013), the authors studied the information content in SPECT list mode data and proposed a MLEM approach using scattered photons from the observed list mode data to jointly reconstruct the attenuation and activity map. This, however, remains a theoretical study due to the complexity of the approach. Similarly, jointly reconstructing the attenuation and activity map using both the photopeak and scattered projections was also proposed in a newly published work for PET scans without time of flight information (Brusaferrì *et al* 2020).

Recently, deep learning with convolutional neural networks (CNNs) has been widely investigated in medical image restoration and analysis, e.g. for tissue segmentation, image de-noising, and image transformation (e.g. MRI to CT) (Beekman 1993, Gong *et al* 2018, Leynes *et al* 2018, Mok *et al* 2018, Wang *et al* 2019, Shi *et al* 2020, Shiri *et al* 2020). These networks can capture relevant information inherent in data and establish highly nonlinear mapping from input to output. This enables the extraction of energy-spatial information from the scatter- and primary-window reconstructed SPECT images for μ map estimation. Successful implementations of this deep learning based strategy have been demonstrated in Shi *et al* (2020) for ^{99m}Tc -tetrofosmin SPECT scans, and in Liu *et al* (2018a), Reimold *et al* (2019) for ^{18}F -FDG brain PET scans. Specifically in Shi *et al* (2020), the authors performed image reconstructions from two energy windows (primary and one scatter window) of ^{99m}Tc -tetrofosmin myocardial SPECT; a generative adversarial network was used to transform 4D SPECT image patches (3D SPECT over two energy windows) to 3D attenuation map patches. Results were validated on clinical myocardial scans acquired from a GE dual-head parallel-hole SPECT/CT 850 system.

The aim of the present work is to develop and validate a CNN approach for estimating μ maps of ^{99m}Tc -HMPAO brain perfusion scanning with a full ring stationary SPECT system (G-SPECT-I, (Beekman *et al* 2015)). The full emission data including both primary and scattered photons over a broad energy range was utilized via multiple image reconstructions at different energy windows. A patch-based CNN approach was implemented in the present work. Such an approach is often applied in medical image restoration and analysis (Beekman 1993, Guo *et al* 2018, Leynes *et al* 2018, Liu *et al* 2018b) due to the reduced number of parameters and increased amount of training data that is required compared to the full-image to full-image approaches e.g. U-Net (Ronneberger *et al* 2015). In the present study, a large number of 4D (i.e. XYZ dimensions) SPECT patches (sub-volumes of $21 \times 21 \times 21$ voxels $\times$ 5 energies) were used as input for the CNN to estimate the attenuation coefficient of the corresponding central voxel of the patch. Our proposed method was tested on Monte Carlo (MC) simulated ^{99m}Tc -HMPAO SPECT scans based on the G-SPECT-I geometry.

2. Methods

2.1. G-SPECT-I system

The G-SPECT-I is a multi-pinhole system with stationary detectors that demonstrates excellent resolution down to 2.5 mm (resolved rod size for Derenzo hot rod phantom) and a central sensitivity of 415 cps/MBq when using a dedicated brain collimator with 3 mm $\varnothing$ pinholes (Beekman *et al* 2015). This system (see figure 1) consists of nine large-area NaI detectors, an interchangeable collimator and a precisely controlled *xyz*-stage used for bed translation. The collimator assumed in this paper was developed for high sensitivity brain and pediatric imaging and has a total of 54 pinholes with a pinhole diameter of 4.5 mm. The G-SPECT-I has a focused geometry design, similar to that of its preclinical predecessors, i.e. various versions of U-SPECT/CT and VECTor/CT scanners that are now in use by many labs worldwide (Beekman *et al* 2005, van der Have *et al* 2009, Goorden *et al* 2013, Ivashchenko *et al* 2014). Such focused geometries entail that all pinholes simultaneously ‘view’ a central volume in which a very high sensitivity and complete data (without bed movement) is obtained. This central volume is termed the complete data volume (CDV). This CDV has a transaxial diameter and axial length of 100 mm and 60 mm respectively (see figure 1). For a scan of an object larger than the CDV, the bed is translated to extend the

volume in which sufficient sampling is obtained for optimal image reconstruction. In the present paper, for a whole brain perfusion scan, 18 bed translations with overlapping CDVs are used for sufficient sampling of the entire scan volume based on findings in (Chen *et al* 2020). All pinhole projections from all bed positions together are used simultaneously for image reconstruction using the so called scanning focus method (Vastenhout and Beekman 2007). Other details concerning G-SPECT-I are described in Chen *et al* (2018).

2.2. SPECT data simulations

The CNN estimation of the μ maps is based entirely on MC simulated SPECT list-mode data for the G-SPECT-I geometry using head phantoms. As the presence of noise in the MC simulated realistic brain images may hamper visualization of AC effects when evaluating the μ maps, we additionally performed a voxelized ray tracing (VRT) simulation (Goorden *et al* 2016, Wang *et al* 2017) to generate noise-free images, such that the accuracy of AC and thus the quality of the μ maps could be studied on noisy as well as on noiseless images. These simulations are described in detail below (see also figure 2).

2.2.1. Digital phantoms for simulation

The publicly available Brainweb database with digital phantoms generated based on normal subjects (Aubert-Broche *et al* 2006) were used to simulate ^{99m}Tc -HMPAO brain perfusion scans. This type of scan was chosen given the wide application of perfusion SPECT in the diagnosis of cerebrovascular (e.g. stroke), neurological (e.g. epilepsy) and psychiatric disorders (e.g. post-traumatic stress disorder) (Catafau 2001, Juni *et al* 2009, Amen 2015). A total of six phantoms (the first six phantoms when ordered by subject number) were used in this work, with five used for training and one for testing in a leave-one-out cross validation manner.

For each phantom, the activity map was generated by assigning tracer concentrations to grey matter, white matter and background regions (e.g. skin, skeletal muscle) with a mean ratio of 80:20:5 (Glick and Soares 1997, Stodilka *et al* 2000, Pato *et al* 2015) to mimic a realistic blood flow. This ratio was introduced with a random variation (normally distributed) with a standard deviation of 10% for each number to make the activity distribution variable among phantoms. Besides, a total activity of 50 MBq in average was set in the head (resembling an injected dose of 25 mCi as in (Laere *et al* 2000, Nobili *et al* 2002, Bowen *et al* 2011)). This number was induced with a standard deviation of 10% (normally distributed) for each phantom. Simulations assumed a scan time of 30 min.

The attenuation map was obtained by segmenting the phantom into different regions. For MC simulations, the maps with each region assigned with a material (e.g. water) were used (see figure 2(a)), while for VRT simulations, the maps with each region assigned with an attenuation coefficient were used (subject to the different requirements of both simulators). A total of eight different regions were segmented, i.e. skull, skin, blood, muscle, brain, water, fat and air, with a corresponding coefficient of 0.248, 0.155, 0.149, 0.147, 0.146,

0.142, 0.128 and 0 cm^{-1} respectively. These values were calculated based on the chemical component of each tissue and the mass attenuation coefficient given in NIST (National Institute of Standards & Technology) for photons at 140 keV. The map with given attenuation coefficient is also the ground truth (GT) μ map that was used in the cross-validation step for training and for evaluation of the network predicted μ maps. All phantoms for simulations were down-sampled with trilinear interpolation to have a voxel size of $1 \times 1 \times 1 \text{ mm}^3$ from its original voxel size of $0.5 \times 0.5 \times 0.5 \text{ mm}^3$.

2.2.2. Noisy MC simulations

The software used for MC simulations was the Geant4 Application for Tomographic Emission (GATE) (Jan *et al* 2004), with Geant4 v10 and GATE v8.0 installed on a CentOS 6.6 cluster. The long scanning time (30 min) simulation was divided into multiple simulations with shorter time intervals (randomly seeded) that were executed in parallel on a computer cluster.

The system geometry in GATE was designed to closely represent the actual G-SPECT-I design, namely, the computer-aided design (CAD) drawing of the G-SPECT-I collimator was put into GATE. Additionally, nine NaI-scintillators were created natively by defining a $497 \times 410 \times 9.5 \text{ mm}^3$ box and replicating it with a ring repeater with a collimator center to detector distance of 757 mm (same as in the G-SPECT-I prototype). The full emission spectrum of ^{99m}Tc was used in the simulations. Physics processes modelled were photoelectric effect, Compton and Rayleigh scattering for gamma photons; bremsstrahlung and multiple scattering for electrons. Within the scintillators, the interaction time, total deposited energy, and energy-weighted average interaction location for each gamma photon were recorded. The uncertainty of the scintillation process and light collection is not fully modelled in the MC simulations, but rather accounted for by taking random samples from a Gaussian distribution in both the spatial and energy domain for each photon recorded on the detector. This is to accelerate the simulation process. The full width at half max (FWHM) of the Gaussians were set according to findings in (Nguyen *et al* 2019): for photons at 140 keV, the respective spatial and energy blurring were 3.5 mm and 10% FWHM (based on measurements of detectors at our institute); for photons at other energies these two values were calculated based on models from literature (Nguyen *et al* 2019).

Background counts due to cosmic radiation were emulated to make the simulation more realistic. This was done by acquiring a long void scan (10 h scan without radioactivity) with the G-SPECT-I prototype, followed by count scaling (by dividing the count numbers by 360 to emulate a 100 s scan at one bed position) and Poisson statistics generation. This process was performed for all 18 bed positions to emulate a total acquisition time of 30 min. Finally, the background counts at each bed position were added to the corresponding projections obtained with the MC simulation.

2.2.3. Noise-free VRT simulations

The VRT simulator takes the system geometry (i.e. the precise pinhole and detector positions and detector orientations) as input and models the collimator and detector crystal penetration but ignores scattering (Goorden *et al* 2016, Wang *et al* 2017). For these noise-free VRT scan simulations, the same scanner geometry and the same acquisition parameters (e.g. the same 18 bed positions) as in the MC simulation were assumed. Cosmic radiation counts were not added in the VRT simulations. Effects of patient attenuation were included in the VRT simulated projection data.

2.3. Image reconstruction

For the MC simulated projection data, five SPECT reconstructions were performed (see figure 2(a)); one used the photons detected in the photopeak window combined with a triple energy window (TEW) scatter correction (28 keV width centered at 140 keV for the photopeak window and 5.6 keV width at each side for scatter correction) and four additional reconstructions were done from different scatter windows (28 keV width centered at 120, 100, 80 and 60 keV respectively). For the VRT simulated projection data, only primary photons were simulated and could thus be reconstructed, resulting in noise- and scatter-free SPECT scans (see figure 2(b)).

All image reconstructions were performed on a 1.5 mm grid, larger than the voxel size of the digital phantoms, to mimic a continuous activity distribution reconstructed on a discrete grid. The system matrix for reconstruction was calculated using the VRT simulator for photons at 140 keV and excludes effects of object scatter. This system matrix was used in image reconstruction for all energy windows. No attenuation correction was performed during reconstruction. Similarity regulated OSEM (Vaissier *et al* 2016) with 8 subsets and 10 iterations was implemented for image reconstruction.

2.4. Attenuation map (μ map) estimation

2.4.1. SPECT image preprocessing

Before being used as input to the CNN to estimate μ maps, the reconstructed MC simulated SPECT images were preprocessed (see also figure 3). Firstly, the reconstructed noisy MC simulated SPECT images (from all energy windows) were masked to remove artefacts outside the head (figure 3(a)). The mask used was a cylinder with a relatively large diameter of 240 mm to safely preserve the brain/head structures. Secondly, intensity normalization was performed to compensate for variance of reconstructed image intensity among phantoms (figure 3(a)). Here for each phantom, the maximum intensity from the photopeak reconstructed image was firstly calculated. To reduce the effects of noise and/or any strong edge artefacts (which might appear at the edge slices of the reconstructed volume), this maximum value was obtained from slightly filtered SPECT scans (with a 6 mm FWHM Gaussian) and from the central 36-mm-thick slices of the reconstructed volume (the reconstructable length even with one bed position in the axial direction of G-SPECT-I, see figure 1). Subsequently, all five reconstructions from different energy windows were normalized by division by this maximum value such that the images from different phantoms that are used as input to CNN have a similar dynamic range.

The MC SPECT images were interpolated (tri-linearly) to a voxel size of $3 \times 3 \times 3 \text{ mm}^3$ from an original voxel size of $1.5 \times 1.5 \times 1.5 \text{ mm}^3$ before being fed into the neural network. This is to speed up the training process with a relatively larger image voxel size.

2.4.2. CNN regression

The neural network takes 4D (XYZ and energy) patches centered at each voxel as input and was trained to predict the attenuation coefficient of the corresponding central voxel in object space (figure 3(b)). The patch size used was $21 \times 21 \times 21$ voxels $\times$ 5 energies. The voxel size of the patch images is $3 \times 3 \times 3 \text{ mm}^3$. We used a typical CNN architecture that consists of multiple stages of convolution, batch normalization, ReLU activation and max pooling, followed by a fully connected layer with a *sigmoid* for voxel regression (see figure 3(b)). The network that takes in total five energy windows as input is termed CNN5E. In addition, we included two networks using less energy windows to investigate the effect of energy range of scattered photons on μ map estimation. These two networks are termed CNN3E and CNN1E respectively, with three energy windows (one photopeak window and the two high-energy scatter windows, see figure 2(a)) and only the photopeak window involved respectively. Convolution was performed with a kernel size of $3 \times 3 \times 3$. The number of the filters is 32, 64, 128, 256 and 512 as indicated in the figure. Pooling was done with a grid of $2 \times 2 \times 2$.

Leave-one-out cross validation was performed using five phantoms in training and one for testing (see figure 3(c)). Each network was trained with a balanced set containing 5k samples (4D patches) randomly selected from each of the three main tissue classes ($0 \leq \mu < 0.07/\text{cm}$ for air voxels, $0.07 \leq \mu < 0.20/\text{cm}$ for water-like voxels and $0.20 \leq \mu < 0.25/\text{cm}$ for bone voxels; thus 15k samples in total). Meanwhile, for each epoch during

training, a new selection of 15k samples was made in order to feed the network with as many samples as possible, similar as in Moeskops *et al* (2016). The network used in this study was trained to minimize the mean square error between the predicted attenuation coefficient μ and the GT μ map (see description in section 2.2.1 for definition).

The proposed network was implemented using TensorFlow. The Adam optimizer (Kingma and Jimmy 2014) with default settings (learning rate = 0.001, $\beta_1 = 0.9$, $\beta_2 = 0.999$) was used to train the network. A mini-batch of 15 4D patches was used. No validation set was used to determine the optimal epoch due to the limited number of phantoms in this study. The model was trained for 200 epochs for convergence.

2.5. Attenuation correction with estimated maps

As the aim is to use the μ maps to realize quantitative SPECT, attenuation correction was implemented using the estimated maps. Correction was done using an adapted first-order Chang’s method (Chang 1978). In Chang’s method with traditional parallel-hole collimation, the transmission along every projection line from a given voxel is calculated, and the transmission fraction (TF) for that voxel is defined as the average transmission value among all projection lines (equation (1)).

$$TF = \frac{1}{M} \sum_{m=1}^M \exp\left(-\int_{L_m} \mu(l) dl\right). \quad (1)$$

Here M is the number of projections involved in data acquisition for a certain voxel, L_m is the m th projection path of gamma photons and μ is the attenuation coefficient along the projection line L_m .

Due to the (multi-pinhole multi-bed-position) characteristic of G-SPECT-I, for a given voxel at a given bed position, only photons that travel toward some distinct directions will be captured by one of the pinholes (see figure 4). Thus, here we implemented an adapted multi-pinhole Chang's method. This was done by first checking—at every bed position—if a voxel is seen by a pinhole. If yes, the transmission along this projection line (from that voxel center to the pinhole center) is counted, and weighted by the pinhole's sensitivity for that voxel (see equations (2)–(4)). The parameters used in the adapted multi-pinhole Chang's method (equations (2)–(4)) are summarized in table 1.

$$TF = \frac{1}{\sum_{n=1}^N \sum_{p=1}^P S_{n,p}} \sum_{n=1}^N \sum_{p=1}^P \left(S_{n,p} \exp\left(-\int_{L_{n,p}} \mu(l) dl\right) \right) \quad (2)$$

$$S_{np} = \frac{d_e^2}{16 r_{n,p}^2} \cos \vartheta_{n,p} \quad (3)$$

$$d_e = \sqrt{d^2 + \frac{2}{\mu_0} d \tan\left(\frac{\alpha}{2}\right) + \frac{2}{\mu_0^2} d \tan^2\left(\frac{\alpha}{2}\right)}, \quad (4)$$

2.6. Evaluation

2.6.1. Visual inspection

The network estimated μ maps were compared to the GT μ maps. Besides, attenuation corrected SPECT images using the three CNN estimated μ maps (CNN5E-AC, CNN3E-AC and CNN1E-AC) were compared to the ground-truth-AC (GT-AC) images which use the GT μ maps for correction.

2.6.2. Quantitative analysis

2.6.2.1. Standard uptake value (SUV)

As often used in absolute quantification for PET and SPECT, the standardized uptake values SUVs were calculated in a number of regions of interest (ROIs). The SUV (in units of g ml^{-1}) is the mean concentration in a region C_{ROI} (Bq ml^{-1}) normalized by the injected dose per patient weight (Bq g^{-1}), as shown in equation (5). In this work a clinically injected dose of 25 mCi and a body weight of 70 kg were assumed for brain perfusion scans for all phantoms. To define the ROIs, a template phantom based on the 'Colin 27 Average Brain' (Collins *et al* 1998) provided by the Brainweb database was registered to subject phantoms. ROIs were generated by warping

Table 1. Parameters used in the adapted multi-pinhole Chang's method.

Symbol	Description
N	Number of bed positions in data acquisition
P	Number of pinholes 'seeing' the voxel
$S_{n,p}$	Sensitivity of the given voxel corresponding to pinhole p at bed position n (Paix 1967)
μ	Attenuation coefficient along the line $L_{n,p}$ from voxel to pinhole p center at bed position n
$r_{n,p}$	Voxel-to-pinhole distance, given pinhole p and bed position n
$\theta_{n,p}$	Angle between the line of voxel to pinhole center and the line of collimator center to pinhole center, given pinhole p and bed position n (see figure 4)
d_e	Equivalent pinhole diameter
d	Actual pinhole diameter (4.5 mm here)
μ_0	Attenuation coefficient (24.3 cm^{-1} , measured experimentally) for pinholes made of hard lead (antimonial lead, containing a mixture of Pb and Sn)
α	Pinhole opening angle (27° here)

the automated anatomical labeling (AAL v3) template (Tzourio-Mazoyer *et al* 2002) to subject space using the same transformation. A total of 166 regions were defined in the AAL template. These localized regions have a size in the 0.07–41.15 ml range with a median value of 3.75 ml (mean of 6.40 ml) respectively (across all phantoms).

The capability to achieve quantitative SPECT was assessed by comparing the SUVs calculated from the attenuation corrected SPECT images to those of the digital phantom image. This direct comparison of the SUVs was performed in eight big structures as in Gong *et al* (2018), Yang *et al* (2019), including four big lobes (temporal, occipital, parietal and frontal lobe), three subcortical structures (thalamus, putamen and caudate nucleus) and the cerebellum. Here in our work these eight structures were further separated into 16 ROIs by hemispheres (e.g. thalamus in the left hemisphere and in the right).

$$SUV (\text{g ml}^{-1}) = \frac{C_{ROI}}{\text{Dose/Weight}}. \quad (5)$$

2.6.2.2. Difference from GT-AC

Differences in units of SUV (*DIFF*) when using the CNN estimated μ maps for attenuation correction compared to that of the GT μ maps were evaluated (see equation (6)). The distribution of the differences were presented in Bland–Altman plots as in Yang *et al* (2019). Differences (in percentage) from the GT-AC images, which is termed deviations (*DEV*) here, were calculated as in equation (7). Statistical significance of the deviations was assessed using paired *t* tests ($p < 0.05$ is considered as statistically significant) as in Liu *et al* (2018a), Yang *et al* (2019).

$$DIFF(SUV) = SUV - SUV_{GT-AC} \quad (6)$$

$$DEV_{SUV}(\%) = \left| \frac{SUV - SUV_{GT-AC}}{SUV_{GT-AC}} \right| \times 100\%. \quad (7)$$

For all SPECT images shown in this paper, a 3D Gaussian post filter with 6 mm FWHM was applied. For quantitative analysis, measurements were performed on the unfiltered SPECT images to avoid any bias from filtering.

3. Results

3.1. Visual inspection

Figure 5 shows a comparison of the CNN estimated and the GT μ maps. Five slices obtained from one of the phantoms (indicated by the solid black lines) are displayed. A full comparison of results for all six phantoms is included in the appendix (figure A1). Figure 5 shows that compared to the GT, μ maps can be well estimated with CNN5E and CNN3E in which cases scatter windows (besides the photopeak window) were involved as input to the CNN; in these cases, the shape and size of the heads are well retrieved. The improvement by including two additional low-energy scatter windows from CNN3E to CNN5E is small.

CNN1E has inferior performance compared to CNN5E and CNN3E; the head size can be inaccurately predicted with CNN1E as in the example shown in figure 5 where the predicted μ maps appear incorrectly large. Besides, an insufficient estimation of the air and bone structures was observed for CNN1E as demonstrated in figure A1 (phantom number 5). This might be due to the fact that input to CNN1E includes only the photopeak window which represents solely the activity distribution. Information to correctly determine the air and bone voxels where there is barely any tracer accumulation is thus insufficient.

Figure 6 shows a comparison of the SPECT images as well as the image profiles when using the GT or CNN μ maps for attenuation correction. SPECT images presented in this figure are the noiseless scans from VRT simulations to better visualize the effects of attenuation correction. Results of corrections performed on the realistic MC simulated scans are included in the appendix (figure A3). Figure 6 shows that the image and image profile differences from the ground-truth-AC are small for CNN5E-AC and CNN3E-AC images, while CNN1E-AC images deviate more from the ground-truth-AC.

3.2. Quantitative analysis

3.2.1. SUV

The absolute quantification results when using the ground-truth-AC and CNN-AC are shown in figure 7. The SUVs in the 16 merged ROIs (8 structures $\times$ 2 hemispheres) are plotted. Figure 7 shows that attenuation correction is essential to achieve accurate quantification; without correction, an underestimation of about 70% for the SUVs is observed. Compared to the digital phantom, the ground-truth-AC, CNN5E-AC and CNN3E-AC images suffer from slight underestimation of the SUVs, which could be due to the partial volume effects given the finite system resolution (~ 3.5 mm) or the imperfect attenuation correction method.

3.2.2. Difference from GT-AC

In the Bland–Altman plot (figure 8), the distribution of the SUV differences is assessed. Figure 8(a) plots the distribution for measurements taken from the merged 16×6 regions while those from the localized 166×6 regions over the entire brain are given in figure 8(b). In figure 8, differences from the ground-truth-AC are small (close to the zero line) for both the CNN5E-AC and CNN3E-AC. Among these two methods, CNN3E-AC shows a slightly more diverging distribution and thus a larger difference. A deviation from the ground-truth-AC of

10% is highlighted by the semi-transparent grey region. For CNN5E-AC, all measurements are within 10% deviation across all regions.

The mean value of the deviations for measurements among all ROIs is summarized in table 2 for different CNN-AC methods. For CNN5E-AC and CNN3E-AC, the mean deviations across 16×6 regions are within 1.60% and are statistically insignificant ($p > 0.05$), while for assessments across 166×6 localized regions, the mean deviations are within 1.82% ($p < 0.05$).

3.3. Counts from each energy window

Table 3 provides the number of counts detected in each energy window. An example of the energy spectrum from one phantom is shown in figure 2(a). Table 3 shows that (i) the detected counts from tracer emission decreases for energy windows going from E1 to E5, and (ii) contribution of cosmic radiation to the total increases (from 5% to 15%) for energy window going from E1 to E5.

Figure 7. Comparison of the SUV values when using various AC methods in 16 regions (8 structures $\times$ 2 hemispheres). The mean and standard deviation from all 6 phantoms are displayed for each region.

Figure 8. Bland-Altman plots when using different CNN-AC methods. Measurements taken from the merged 16×6 regions are plotted in (a) while those from the localized 166×6 regions are given in (b).

4. Discussion

The current work shows the feasibility of estimating attenuation maps by using SPECT data only without additional (radiation) scans or the need to draw head contours. This could facilitate SPECT imaging with

Table 2. SUV deviation (mean $\pm$ standard deviation) from the ground-truth-AC across 16 merged ROIs and 166 small localized ROIs for all six phantoms. $P < 0.05$ is defined as the significance level.

16 $\times$ 6 regions	CNN5E-AC	CNN3E-AC	CNN1E-AC	No-AC
DEV_{SUV} (%)	1.60 $\pm$ 1.53	1.59 $\pm$ 1.68	7.35 $\pm$ 8.91	67.50 $\pm$ 5.30
P	0.368	0.616	0.002	<0.001
166 $\times$ 6 regions	CNN5E-AC	CNN3E-AC	CNN1E-AC	No-AC
DEV_{SUV} (%)	1.63 $\pm$ 1.66	1.82 $\pm$ 1.90	7.53 $\pm$ 8.80	69.25 $\pm$ 4.40
P	<0.001	0.016	<0.001	<0.001

Table 3. Comparison of the count number from different energy windows (keV). The mean counts across all phantoms are given. ‘M’ stands for the unit million. The first row gives the mean counts detected from tracer emission based on Monte Carlo simulations. The second row gives the background counts from cosmic radiation based on a long void scan (10 h scan without radioactivity) with the G-SPECT-I prototype. Note that these two types of counts were added in projections for image reconstruction.

Mean counts	Photopeak E1 (126–154)	E2 (106–134)	E3 (86–114)	E4 (66–94)	E5 (46–74)
From tracer	7.50 M	4.20 M	3.59 M	3.29 M	2.19 M
From cosmic	0.45 M	0.44 M	0.45 M	0.54 M	0.35 M
Cosmic/tracer	6.0%	10.5%	12.5%	16.4%	16.0%

minimal ionizing radiation or user interactions, enabling quantitative brain perfusion SPECT to be independent from data from other scanners, thereby avoiding registration issues between modalities as well.

Figures 5–7 and A1 show that despite of the inadequate μ map estimation of CNN1E, a reasonable quantification accuracy could still be obtained with this method. This might explain the broad use of an ellipsoidal region for attenuation correction in clinical routine given its benefit of simplicity and decent effectiveness (i.e. the improved accuracy compared to that of not performing a correction at all). While various contour-based methods have been proposed to attain a uniform μ map, devising one here that is suitable for our work is beyond the scope of this paper. Alternatively, a ground-truth-uniform (GT-uniform) μ map, which can be regarded as the best achievable μ map using the contour-based method, was generated by replacing all tissues in the GT μ map with the attenuation coefficient of brain tissue ($\mu = 0.146/\text{cm}$). This part of the results is included in the appendix (figure A2 and table A1). GT-uniform-AC suffers from underestimation of the SUVs compared to the GT-AC since all bone voxels are replaced by brain. The Bland–Altman plot of GT-uniform-AC shows a more diverging distribution of the SUV differences compared to those of the CNN5E-AC and CNN3E-AC. The mean deviation from GT-AC is within 3.78%, which is also larger than CNN5E-AC (1.63%) and CNN3E-AC (1.82%). Note that the GT-uniform μ map is the ideal μ map one could get using the contour-based method. In reality, obtaining such a contour generally requires image processing steps, which involve a set of parameters (e.g. smoothing and threshold) that are sensitive to tracer distribution and noise in image. On the other hand, with the proposed CNN method, an optimal mapping with a large number of parameters is generated automatically. Here our results show that, even though the tracer distribution and the amount of activity were variable for each phantom, CNN could still accurately estimate the attenuation map when trained on a small set with five phantoms.

Three neural networks (CNN5E, CNN3E and CNN1E) were tested in this work, where we found that compared to the use of only the photopeak window (CNN1E), the involvement of scatter windows as done in CNN3E and CNN5E improves the performance. The use of five energy windows (CNN5E) instead of three (CNN3E) only has a limited effect on results. This might be due to (i) the intrinsic weak correlation between the multi-order scattered photons and the attenuation coefficients and (ii) the relatively low scatter signal compared to the level of noise in the additional two low-energy scatter windows of CNN5E (see table 3). The results also indicate that investigating the utilization of photons from high-energy scatter windows close to the photopeak might be beneficial. In this paper, energy windows used all have an equal width of 28 keV and are already slightly overlapping (see figure 2(a)), thus adding additional high-energy scatter windows may not lead to improvements. Given better energy resolution, the incorporation of more energy windows may be possible by narrowing the width of each window. This could provide the network with photons with more precise energy-spatial information. However, the exact effect on CNN estimation of μ maps requires additional studies.

In this paper, evaluation of attenuation correction with the CNN estimated μ maps was performed on the noise-free VRT simulated scans (for better visualization of the AC effects when using different μ maps). Results of correction performed on the realistic MC simulated scans are included in the appendix (figure A3). These results show that visually the differences of using GT-AC, GT-uniform-AC, CNN5E-AC and CNN3E-AC are small on the MC simulated scans. Besides, comparing the realistic MC simulated SPECT images to the noiseless VRT scans (figure 6(a)), one could see that the effects on image quality due to, e.g. noise and imperfect scatter correction as in the former scans can be larger than effects caused by using different CNN-AC methods (e.g. CNN5E-AC and CNN3E-AC). This is not surprising as noise generally plays a critical role in SPECT. While we aim to estimate the attenuation map using deep learning for automatic and accurate attenuation correction, future development that could achieve SPECT denoising can be beneficial, and using deep learning techniques for this task is also being actively investigated (Ramon *et al* 2018, Reyman *et al* 2019, Zhang *et al* 2019).

As a feasibility study based on Monte Carlo simulated data, limitations of this work include a validation using experimental scans which is required to bring the method closer to clinical applicability. For implementation on experimental data, we expect that the network may be trained on physical scans with CT GT attenuation maps acquired from only a few patients, as it was shown that the proposed CNN needed only a limited number of training samples. Besides, the effect of subtle differences of attenuation correction when using a CNN instead of a CT μ map needs to be evaluated on real patient data by human readers with a specific clinical task to ultimately prove the merit of the proposed CNN approach. Additionally, we used discretized maps consisting of a limited number of tissue classes (eight classes) as the GT attenuation map. For brain scans with relatively simple attenuation maps, this approximation could be acceptable. For other applications especially those involving fine structures, e.g. myocardium imaging, more complex attenuation maps with more tissue classes or with continuous attenuation coefficients would be preferable. Furthermore, the proposed method was tested on a G-SPECT-I geometry as it is an ultra-high resolution system we are developing at our institute. Given the validity of the proposed approach on G-SPECT-I, it would be worthy to try it for other standalone SPECT devices as well. Moreover, the proposed methodology was validated only for brain perfusions scans with ^{99m}Tc HMPAO in the present work. We expect that the approach may be applicable to other types of scans when the CNN is trained on that specific scan, yet the accuracy and effectivity have to be validated.

5. Conclusion

We have implemented and validated a neural network approach to generate attenuation maps solely from SPECT emission data for ^{99m}Tc -HMPAO brain perfusion scans. This could enable quantitative SPECT imaging with minimal ionizing radiation and make SPECT independent of data from other imaging modalities, while the fully automated approach could reduce the subjectivity due to intra- or inter-observer variability.

Acknowledgments

Financial disclosures of authors: FB is employee and shareholder of MILabs BV. This work is conducted with financial support of the Netherlands Organization for Scientific Research (NWO-I), Physics Valorization Prize 2015 ‘Ultra-fast, ultra-sensitive and ultra-high resolution SPECT’ co-financed by MILabs BV.

Appendix

Comparison of the GT and CNN estimated μ maps for all six phantoms are given in figure A1. Five slices with an equal separation of 25.5 mm are displayed for each phantom. Figure A2 and table A1 provide the quantitative analysis results of GT-uniform-AC. The results from CNN5E-AC are shown together here as a comparison. Figure A3 displays the MC simulated realistic SPECT scans after attenuation correction with different μ maps. Quantitative analysis on these MC simulated scans of the deviations from the GT-AC for various CNN-AC methods are not included. These deviations are the same as those given in table 2 where assessments are performed on the noiseless VRT scans as the deviations depend only on the transmission fraction of the μ maps (see equation (2)).

Figure A1. Comparison of the ground truth GT and the CNN estimated μ maps for all six phantoms. Five slices from the top to the bottom of the head (equally distributed with 25.5 mm separation) are displayed.

Figure A2. Bland–Altman plots for SUVs calculated from the noiseless VRT simulated SPECT images when using ground-truth-uniform-AC. The results from CNN5E-AC are also plotted as a comparison. The difference for ground-truth-uniform-AC are negative since all bone voxels are replaced by the attenuation coefficient of brain tissue.

Table A1. SUV deviation (mean $\pm$ standard deviation) from the ground-truth-AC for ground-truth-uniform-AC when assessed on the VRT simulated SPECT images. The results from CNN5E-AC are also given here as a comparison. $P < 0.05$ is defined as the significance level.

	16 $\times$ 6 regions		166 $\times$ 6 regions	
	CNN5E-AC	Ground-truth-uniform-AC	CNN5E-AC	Ground-truth-uniform-AC
DEV_{SUV} (%)	1.60 $\pm$ 1.53	3.72 $\pm$ 1.58	1.63 $\pm$ 1.66	3.78 $\pm$ 1.65
p	0.368	<0.001	<0.001	<0.001

References

- Amen D G 2015 *Change Your Brain, Change Your Life: Revised and Expanded Edition: The Breakthrough Programme for Conquering Anxiety, Depression, Anger and Obsessiveness* (New York: Times Books)
- Aubert-Broche B, Griffin M, Pike G B, Evans A C and Collins D L 2006 Twenty new digital brain phantoms for creation of validation image data bases *IEEE Trans. Med. Imaging* **25** 1410–6
- Bailey D L and Willows K P 2013 An evidence-based review of quantitative SPECT imaging and potential clinical applications *J. Nucl. Med.* **54** 83–9

- Beekman F J 1993 *Restoration of computer-tomographic images with neural networks* WO/1993/006560
- Beekman F J, van der Have F, Vastenhout B, van der Linden A J A, van Rijk P P, Burbach J P H and Smidt M P 2005 U-SPECT-I: a novel system for submillimeter-resolution tomography with radiolabeled molecules in mice *J. Nucl. Med.* **46** 1194–200
- Beekman F J, van der Have F, Goorden M C, Vaissier P E B, van Roosmalen J, During H and Vastenhout B 2015 G-SPECT-I: a full ring high sensitivity and ultra-fast digital molecular imaging system with <3 mm resolution *Eur. J. Nucl. Med. Mol. Imaging* **42** S209
- Berker Y and Li Y 2016 Attenuation correction in emission tomography using the emission data—a review *Med. Phys.* **43** 807–32
- Bowen J D, Huang Q, Gullberg G T and Seo Y 2011 Phantom measurements and simulations of cardiac and brain studies using a multipinhole collimator with 20 apertures *IEEE Nucl. Sci. Symp. Med. Imaging Conf. Rec.* pp 3417–21
- Brusaferri L et al 2020 Joint activity and attenuation reconstruction from multiple energy window data with photopeak scatter re-estimation in non-TOF 3D PET *IEEE Trans. Radiat. Plasma Med. Sci.* **4** 410–21
- Catafau A M 2001 Brain SPECT in clinical practice: I. Perfusion *J. Nucl. Med.* **42** 259–71
- Chang L 1978 A method for attenuation correction in radionuclide computed tomography *IEEE Trans. Nucl. Sci.* **25** 638–43
- Chen Y, Vastenhout B, Wu C, Goorden M C and Beekman F J 2018 Optimized image acquisition for dopamine transporter imaging with ultra-high resolution clinical pinhole SPECT *Phys. Med. Biol.* **63** 225002
- Chen Y, Goorden M C, Vastenhout B and Beekman F J 2020 Optimized sampling for high resolution multi-pinhole brain SPECT with stationary detectors *Phys. Med. Biol.* **65** 015002
- Collins D L, Zijdenbos A P, Kollokian V, Sled J G, Kabani N J, Holmes C J and Evans A C 1998 Design and construction of a realistic digital brain phantom *IEEE Trans. Med. Imaging* **17** 463–8
- Glick S J and Soares E J 1997 Noise characteristics of SPECT iterative reconstruction with a mis-matched projector-backprojector pair *IEEE Trans. Nucl. Sci.* **2** 1515–18
- Goetze S, Brown T L, Lavery W C, Zhang Z and Bengel F M 2007 Attenuation correction in myocardial perfusion SPECT/CT: effects of misregistration and value of reregistration *J. Nucl. Med.* **48** 1090–5
- Gong K, Yang J, Kim K, El Fakhri G, Seo Y and Li Q 2018 Attenuation correction for brain PET imaging using deep neural network based on Dixon and ZTE MR images *Phys. Med. Biol.* **63** 125011
- Goorden M C, van Roosmalen J, van der Have F and Beekman F J 2016 Optimizing modelling in iterative image reconstruction for preclinical pinhole PET *Phys. Med. Biol.* **61** 3712–33
- Goorden M C, van der Have F, Kreuger R, Ramakers R M, Vastenhout B, Burbach J P H, Booij J, Molthoff C F M and Beekman F J 2013 VECToR: A Preclinical Imaging System for Simultaneous Submillimeter SPECT and PET *J. Nucl. Med.* **54** 306–12
- Guo Z, Li X, Huang H, Quo N and Li Q 2018 Medical image segmentation based on multi-modal convolutional neural network: study on image fusion schemes *IEEE 15th Int. Symp. Biomed. Imaging (ISBI 2018)* vol 2018, pp 903–7
- Ivashchenko O, van der Have F, Villena J L, Groen H C, Ramakers R M, Weinans H H and Beekman F J 2014 Quarter-millimeter-resolution molecular mouse imaging with U-SPECT+ *Mol. Imaging* **14** 7290–2014
- Jan S et al 2004 GATE: a simulation toolkit for PET and SPECT *Phys. Med. Biol.* **49** 4543
- Jha A K, Clarkson E, Kupinski M A and Barrett H H 2013 Joint reconstruction of activity and attenuation map using LM SPECT emission data *Med. Imaging* **8668** 86681W
- Juni J E, Waxman A D, Devous M D, Tikofsky R S, Ichise M, Van Heertum R L, Carretta R F and Chen C C 2009 Procedure Guideline for Brain Perfusion SPECT Using 99mTc Radiopharmaceuticals 3.0 *J. Nucl. Med. Technol.* **37** 191–5
- King M A, Glick S J, Pretorius P H, Wells R G, Gifford H C, Narayanan M V and Farncombe T 2004 Attenuation, scatter, and spatial resolution compensation in SPECT *Emission tomography: The fundamentals of PET and SPECT* (Cambridge, MA: Academic Press) pp 473–98
- Kingma D P and Jimmy B 2014 Adam: a method for stochastic optimization arXiv:1412.6980
- Laere K V, Koole M, Kauppinen T, Monsieurs M, Bouwens L and Dierck R 2000 Nonuniform transmission in brain SPECT using 201 Tl, 153 Gd, and 99 m Tc static line sources : anthropomorphic dosimetry studies and influence on brain quantification *J. Nucl. Med.* **41** 2051–63
- Leynes A P, Yang J, Wiesinger F, Kaushik S S, Shanbhag D D, Seo Y, Hope T A and Larson P E Z 2018 Zero-echo-time and dixon deep pseudo-CT (ZeDD CT): direct generation of pseudo-CT images for Pelvic PET/MRI attenuation correction using deep convolutional neural networks with multiparametric MRI *J. Nucl. Med.* **59** 852–8
- Liu F, Jang H, Kijowski R, Zhao G, Bradshaw T and Mcmillan A B 2018a A deep learning approach for 18 F-FDG PET attenuation correction *EJNMMI Phys.* **5** 24
- Liu M, Cheng D, Wang K and Wang Y 2018b Multi-modality cascaded convolutional neural networks for Alzheimer's disease diagnosis *Neuroinformatics* **16** 295–308
- Macey D J, Denardo G L and Denardo S J 1988 Comparison of three boundary detection methods for SPECT using compton scattered photons *J. Nucl. Med.* **29** 203–7
- Moeskops P, Viergever M A, Mendrik A M, De Vries L S, Benders M J N L and Isgum I 2016 Automatic segmentation of MR brain images with a convolutional neural network *IEEE Trans. Med. Imaging* **35** 1252–61
- Mok G S P, Zhang Q, Cun X, Zhang D, Pretorius P H and King M A 2018 Initial investigation of using a generative adversarial network for denoising in dual gating myocardial perfusion SPECT 2018 *IEEE Nucl. Sci. Symp. Med. Imaging Conf. NSS/MIC 2018—Proc.* pp 1–3 National Institute of Standards & Technology (Accessed September 2020) (<https://physics.nist.gov/PhysRefData/Xcom/html/xcom1.html>)
- Nguyen M P, Goorden M C, Kamphuis C and Beekman F J 2019 Evaluation of pinhole collimator materials for micron-resolution *ex vivo* Evaluation of pinhole collimator materials for micron-resolution *ex vivo* SPECT *Phys. Med. Biol.* **64** 105017
- Nobili F, Koulilaly M, Vitali P, Migneco O, Mariani G, Ebmeier K, Alberto P, Robert P, Rodriguez G and Darcourt J 2002 Brain perfusion follow-up in Alzheimer's patients during treatment with acetylcholinesterase inhibitors *J. Nucl. Med.* **43** 983–91
- Paix D 1967 pinhole imaging of gamma rays *Phys. Med. Biol.* **12** 489–500
- Pan T S, King M A, De Vries D J and Ljungberg M 1996 Segmentation of the body and lungs from compton scatter and photopeak window data in SPECT: a Monte-Carlo investigation *IEEE Trans. Med. Imaging* **15** 13–24
- Pato L R V, Vandenbergh S, Zedda T and Van Holen R 2015 Parallel-hole collimator concept for stationary SPECT imaging *Phys. Med. Biol.* **60** 8791–807
- Ramon A J, Yang Y, Pretorius P H, Johnson K L, King M A and Wernick M N 2018 Initial Investigation of Low-Dose SPECT-MPI via deep learning *IEEE NSS/MIC—Proc.* pp 1–3
- Reimold M, Nikolaou K, Christian La Fougère M D and Gatidis S 2019 Independent brain 18 F-FDG PET attenuation correction using a deep learning approach with Generative Adversarial Networks *Hell. J. Nucl. Med.* **22** 179–86
- Reymann M P, Wurfl T, Ritt P, Stimpel B, Cachovan M, Vija A H and Maier A 2019 U-Net for SPECT image denoising *IEEE NSS/MIC* pp 1–2

- Ronneberger O, Fischer P and Brox T 2015 U-net: convolutional networks for biomedical image segmentation *International Conference on Medical Image Computing and Computer-Assisted Intervention* (Cham: Springer) pp 234–41
- Shi L, Onofrey J A, Liu H, Liu Y H and Liu C 2020 Deep learning-based attenuation map generation for myocardial perfusion SPECT *Eur. J. Nucl. Med. Mol. Imaging* **47** 2383–95
- Shiri I, Arabi H, Geramifar P, Hajianfar G, Ghafarian P, Rahmim A, Ay M R and Zaidi H 2020 Deep-JASC: joint attenuation and scatter correction in whole-body 18F-FDG PET using a deep residual network *Eur. J. Nucl. Med. Mol. Imaging* **47** 2533–48
- Stodilka R Z, Kemp B J, Prato F S, Kertesz A, Kuhl D and Nicholson R L 2000 Scatter and attenuation correction for brain SPECT using attenuation distributions inferred from a head atlas *J. Nucl. Med.* **41** 1569–78
- Tzourio-Mazoyer N, Landeau B, Papathanassiou D, Crivello F, Etard O, Delcroix N, Mazoyer B and Joliot M 2002 Automated anatomical labeling of activations in SPM using a macroscopic anatomical parcellation of the MNI MRI single-subject brain *Neuroimage* **15** 273–89
- Vaissier P E B, Beekman F J and Goorden M C 2016 Similarity-regulation of OS-EM for accelerated SPECT reconstruction *Phys. Med. Biol.* **61** 4300–15
- van der Have F, Vastenhouw B, Ramakers R M, Branderhorst W, Krah J O, Ji C, Staelens S G and Beekman F J 2009 U-SPECT-II: an ultra-high-resolution device for molecular small-animal imaging *J. Nucl. Med.* **50** 599–605
- Vastenhouw B and Beekman F 2007 Submillimeter total-body murine imaging with U-SPECT-I *J. Nucl. Med.* **48** 487–93
- Wallis J W, Miller T R and Koppel P 1995 Attenuation correction in cardiac SPECT without a transmission measurement *J. Nucl. Med.* **36** 506–12
- Wang B, Van Roosmalen J, Piët L, VanSchie M A, Beekman F J and Goorden M C 2017 Voxelized ray-tracing simulation dedicated to multi-pinhole molecular breast tomosynthesis *Biomed. Phys. Eng. Express* **3** 045021
- Wang G, Li W, Ourselin S and Vercauteren T 2019 Automatic brain tumor segmentation based on cascaded convolutional neural networks with uncertainty estimation *Front. Comput. Neurosci.* **13** 1–13
- Yang J, Park D, Gullberg G T and Seo Y 2019 Joint correction of attenuation and scatter in image space using deep convolutional neural networks for dedicated brain 18F-FDG PET *Phys. Med. Biol.* **64** 075019
- Younes R B, Mas J and Bidet R 1998 A fully automated contour decision algorithm the preliminary step for scatter and attenuation compensation in SPECT *Eur. J. Nucl. Med.* **14** 586–9
- Zaidi H and Hasegawa B 2003 Determination of the attenuation map in emission tomography *J. Nucl. Med.* **44** 291–315
- Zhang Q, Sun J and Mok G S P 2019 Low dose SPECT image denoising using a generative adversarial network arXiv:1907.11944 pp 1–4