Focused ultrasound neuromodulation on a multiwell MEA

Saccher, M.; Kawasaki, S.; Proietti Onori, Martina; van Woerden, Geeske M. ; Giagka, Vasiliki; Dekker, R.

DOI

[10.1186/s42234-021-00083-7](https://doi.org/10.1186/s42234-021-00083-7)

Publication date

2022

Document Version

Final published version

Published in

Bioelectronic Medicine

Citation (APA)

Saccher, M., Kawasaki, S., Proietti Onori, M., van Woerden, G. M., Giagka, V., & Dekker, R. (2022). Focused ultrasound neuromodulation on a multiwell MEA. *Bioelectronic Medicine*.
<https://doi.org/10.1186/s42234-021-00083-7>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

METHODOLOGY

Open Access

Focused ultrasound neuromodulation on a multiwell MEA

Marta Saccher^{1*} , Shinnosuke Kawasaki¹, Martina Proietti Onori², Geeske M. van Woerden^{2,3}, Vasiliki Giagka^{1,4} and Ronald Dekker^{1,5}

Abstract

Background: Microelectrode arrays (MEA) enable the measurement and stimulation of the electrical activity of cultured cells. The integration of other neuromodulation methods will significantly enhance the application range of MEAs to study their effects on neurons. A neuromodulation method that is recently gaining more attention is focused ultrasound neuromodulation (FUS), which has the potential to treat neurological disorders reversibly and precisely.

Methods: In this work, we present the integration of a focused ultrasound delivery system with a multiwell MEA plate.

Results: The ultrasound delivery system was characterised by ultrasound pressure measurements, and the integration with the MEA plate was modelled with finite-element simulations of acoustic field parameters. The results of the simulations were validated with experimental visualisation of the ultrasound field with Schlieren imaging. In addition, the system was tested on a murine primary hippocampal neuron culture, showing that ultrasound can influence the activity of the neurons.

Conclusions: Our system was demonstrated to be suitable for studying the effect of focused ultrasound on neuronal cultures. The system allows reproducible experiments across the wells due to its robustness and simplicity of operation.

Keywords: Ultrasound, Neuromodulation, Bioelectronics, Multielectrode arrays, Focused ultrasound

Background

In the past decades, several methods for treating neurological disorders such as epilepsy, Parkinson's disease and motor neuron disorders have been developed as an alternative to drugs. The most common method is to use electrical stimulation, however, other stimulation methods are recently considered. Among them, Low Intensity Focused Ultrasound (LIFU) is gaining particular attention because it can be focused in the submillimetre range, and its effects are reversible (Naor et al. 2016). Although its neuromodulation capabilities were observed almost one century ago (Harvey 1929), the scientific community has not agreed yet on the exact mechanisms responsible for the excitation or inhibition of the neurons with

LIFU, and it is considered most likely that different mechanisms dominate depending on the sonication frequency (Kamimura et al. 2020). A broad range of ultrasound (US) parameters have been tested, however a gap in the research on the effects of high frequency neuromodulation exist, especially concerning which combination of parameters exactly causes excitation or inhibition of the neurons (Blackmore et al. 2019). Most of the studies on ultrasound neuromodulation have been done in vitro and on small animals, while only a few studies have been done on humans. Experiments on in vivo animals are complex due to the handling of the animals and the influence of anaesthesia on the response to ultrasound (Kamimura et al. 2020; Jerusalem et al. 2019). A simpler system to study the effect of ultrasound neuromodulation is to use cultured neurons on a microelectrode array (MEA). This makes it possible to study the effect of different ultrasound parameters on the activity of neurons, as well as using

*Correspondence: M.Saccher@tudelft.nl

¹ Department of Microelectronics, Delft University of Technology, Delft, Netherlands

Full list of author information is available at the end of the article

© The Author(s). 2022 **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

drugs to characterize the mechanism underlying these effects. Furthermore, significant advancements in human stem cell technology have enabled the quick and efficient generation of human neurons in a dish that could be used to test potential ultrasound treatments to find the optimal solution for a specific patient (Keller and Frega 2019).

Current MEA systems allow both electrical stimulation and recording of the cell activity, and integrating other neuromodulation methods will greatly increase the versatility of the MEA platform. Although other studies have tested the effects of ultrasound on in vitro neuronal cultures (Rinaldi et al. 1991; Tyler et al. 2008; Lin et al. 2018; Prieto et al. 2018; Kim et al. 2016; Hee-Sok et al. 2014), most of them used frequencies below 1 MHz. Here, we describe the integration of a 24-well MEA system with an ultrasound delivery system that allows the testing of different wells during the same experiment.

Methods

Mice

FvB/NHsD females were crossed with FvB/NHsD males (ordered at 8–10 weeks old from Envigo) for the neuronal cultures. Mice were kept group-housed in IVC cages (Sealsafe 1145T, Tecniplast) with bedding material (Lignocel BK 8/15 from Rettenmayer) on a 12/12 h light/dark cycle in $21^{\circ}\text{C}(\pm 1)^{\circ}\text{C}$, humidity at 40–70%. Food pellets (801727CRM(P) from Special Dietary Service) and water were available ad libitum. All animal experiments were conducted in accordance with the European Commission Council Directive 2010/63/EU (CCD approval AVD101002017893).

Primary hippocampal culture preparation and plating on the mEA

Primary hippocampal cultures were prepared from FvB/NHsD wild-type mice according to the procedure described by Goslin and Banker (Banker and Goslin 1991). Hippocampi were isolated from brains of E16.5 embryos and collected in 10 ml of neurobasal medium

(NB, Gibco) on ice. After two washings with NB, the samples were incubated in pre-warmed trypsin/EDTA solution (Invitrogen) at 37°C for 20 min. After two washings in pre-warmed NB, the cells were resuspended in 500 μL NB medium supplemented with 2% B27, 1% penicillin/streptomycin, and 1% glutamax (Invitrogen), and dissociated using a 1 ml pipette. Following dissociation, neurons were plated in a small drop in the centre of each well of a 24-wells MEA plate precoated with poly-d-lysine (25mg/ml, Sigma) at a density of 35000 cells per well. The plates were stored at $37^{\circ}\text{C}/5\% \text{CO}_2$ for 45 minutes and 500 μL of pre-warmed supplemented NB medium was later added to each well. One-third of the medium was consecutively refreshed every three days.

MEA and acquisition system

The MEA plates used in this work have 24 wells, each with 12 gold recording electrodes and 4 reference electrodes coated with PEDOT (poly3,4-ethylene-dioxythiophene). Each recording electrode has a diameter of 30 μm , and it is arranged on a 4×4 grid at a pitch of 300 μm (MultiChannelSystem, Germany). The electrodes layout is illustrated in Fig. 1A. The MEA plates are coupled to the MEA read-out headstage (MultiChannelSystem, Germany), which connects the recording electrodes to a computer running the acquisition software. The extracellular signals were recorded at a sampling frequency of 10 kHz and filtered using a 4th-order low-pass Butterworth filter with a cut-off frequency of 2 kHz and a 2nd-order high-pass Butterworth filter with cut-off frequency of 100 Hz. The baseline activity of the neurons was recorded for 5 minutes previous to the start of the ultrasound exposure for each well used in the experiment.

System design

A custom lid was designed and fabricated to cover the MEA headstage to ensure constant ambient conditions over the MEA plate of 5% CO_2 and 95% O_2 using an external oxygenation system interfaced to the lid via a small

tube. Ultrasound was generated by a focused piezoelectric ultrasound transducer (V310, Olympus Panametrics NTD). The transducer is a single element curved transducer with a centre frequency of 5 MHz. It is spherically focused at 10 mm, with a focal length of 11 mm and a beam diameter at the focus point of approximately 600 μm (-6dB). The ultrasound transducer has a cylindrical shape, therefore circular holes of the same diameter of the transducer were cut in the lid concentric to each MEA well to allow the US transducer to enter the well. A rigid mechanical support structure for the US transducers was designed to ensure a constant distance and alignment between the bottom of the transducer and the surface of the well plate. Alignment holes were engraved onto the lid to allow the transducer support structure to be easily and precisely positioned with respect to the lid and the

MEA well (Fig. 2). In addition, few drops of water were placed on the tip of the transducer and then sealed with Parafilm[®] film. This ensured the coupling of the transducer to the cell medium and prevented contamination of the transducer with biological material. In addition, the transducer was electrically isolated from the recording electrodes. Holes corresponding to unused wells were capped to maintain the ambient conditions.

Program for generation of the sonication protocol

A RedPitaya board (RedPitaya) was used to generate the sinusoidal signal for the ultrasound transducer. This small board contains two programmable function generators that can be programmed to generate arbitrary waveform patterns. The board was connected to a power amplifier (240L, Electronics & Innovation Ltd.) The output of the

Fig. 2 Experimental setup. **A**) 3D model of the experimental setup. The black box at the bottom is the MEA headstage, at the top of which the custom-made lid is placed. The circular holes in the lid are aligned to the underlying MEA plate. The ultrasound transducer is inserted in the transducer holder structure which can then be positioned in the engraved alignment holes for precise centering with respect to the MEA well. **B**) Cross-section view of the experimental setup showing the relative position of the transducer with respect to the neuronal culture. **C**) Detail picture of the experimental setup showing the transducer in the support structure on the left and on the right, the wells with the coloured cap are used to record the trigger signal. **D**) Schematic view of the complete experimental setup showing the connections between the various elements

power amplifier was, in turn, connected to the ultrasound transducer. The RedPitaya board also generated a trigger signal to indicate the ON- and OFF-time of the ultrasound signal. The electrodes of one well of the MEA system were used to record the trigger signal to provide the time stamp needed to correlate the neuron response to the US stimulation. The LED indicators on the RedPitaya board were also programmed to indicate that the ultrasound was ON and to signal the time left to the end of the trial. The complete experimental setup is depicted in Fig. 2D.

Sonication protocol

Two different sonication protocols were chosen for the validation of the FUS system on the multiwell-MEA. Continuous ultrasound at a frequency of 5 MHz and exposure times of 15 and 60 seconds were used. These treatment durations were chosen based on the range of values found in the literature for continuous wave neuromodulation (Blackmore et al. 2019). Each treatment was repeated three times with an observation period, namely the post period, of 5 minutes between each treatment to test the consistency of the response to the ultrasound neuromodulation and the reversibility of this potential treatment method. The experiment was performed twice for each sonication protocol (Fig. 1B).

Data analysis

Extracellular signals recordings were obtained using the Multiwell-Screen software application and analysed using the Multiwell-Analyzer software, both from MultiChannelSystem. The threshold for spike detection was set to $\pm 15\mu\text{V}$. Bursts were identified when at least 5 consecutive spikes were detected at an interval of less than 50 ms, defined as the inter-spike interval in burst (ISI in burst). The minimum interval between consecutive bursts was set to 100 ms, defined as inter-burst interval (IBI) (Chiappalone et al. 2005) (Fig. 1C). Data from the electrodes in each well was averaged to consider each well as a whole, in each different phase of the experiment. Data was then normalised with respect to the baseline activity, recorded previously to the exposure to ultrasound. Parameter calculation and analysis were done using custom MATLAB scripts, and results from different wells, in which the same ultrasound treatment duration was used, were then compared.

Results

System characterisation

Pressure measurements

Direct measurement of the acoustic pressure at the bottom of the well is not possible in this setup. Therefore, a separate measurement was performed in a water tank setup using a calibrated hydrophone needle (Precision Acoustic Ltd.) mounted on a high precision x-y-z scanner.

The input voltage of the ultrasound transducer was the same as the one used during the cell experiments. The hydrophone needle was aligned to the centre of the transducer and scanned in a plane parallel to the face of the transducer, at the same distance as between the bottom of the well and the transducer in the cell experiments. An area equivalent to the bottom of the MEA well was mapped, and for each location, the peak-to-peak pressure values were calculated by dividing the peak-to-peak values of the recorded signal by the sensitivity of the needle at the corresponding frequency.

Modelling of ultrasound field

To estimate the effect of the MEA plate on the ultrasound field, a finite element model was computed in COMSOL (Multiphysics). A simulation consisting of the ultrasound transducer immersed in an infinite water medium was computed and matched to the hydrophone pressure measurements. The structure of the MEA well was then included in the model, which was analysed using pressure acoustic and solid mechanics models. The bottom of the MEA well was assumed to be entirely made of quartz glass and the smaller structures, like electrodes, were not included in the model because their size is one tenth of the wavelength, therefore negligible. The thickness of the quartz glass was 600 μm , and it was modelled as a perfectly reflecting surface with respect to the ultrasound waves. The culture medium was approximated as water at 37°C, and the primary hippocampal neurons layer was assumed to be an incompressible and hyperplastic material with a thickness of 150 μm (Menz et al. 2019). The distance between the ultrasound transducer face and the bottom of the plate was 6 mm. A symmetrical boundary condition was applied due to the radial symmetry of the geometry. A frequency-domain study was chosen because the system is actuated at a harmonic frequency of 5 MHz. The acoustic properties of the primary hippocampal neurons were retrieved from the literature (IT'IS 2021; Lay et al. 2019) and are summarised in Table 1, while the COMSOL default values for the properties of the other materials in the model were used. To solve the model, a free triangular mesh with a size of $\lambda/8$ was chosen, and the default solver was used. In Fig. 3B, the results of the pressure acoustic study show the presence of an interference pattern generated by the reflections of the acoustic waves at the bottom of the MEA well. In addition, due to the presence of the glass layer at the bottom of the MEA well, the maximum pressure at the level of the neurons was simulated to be 1.05MPa.

Validation of the computed ultrasound field

To validate the results of the COMSOL simulation of pressure field, a pulse laser Schlieren system was used (Fig. 3A). The Schlieren system is an optic system that

Table 1 COMSOL parameters used for the hippocampal neurons

Parameter	Value
Speed of sound	1500 [m/s]
Density	1045 [kg/m ³]
Thermal conductivity	0.55 [W/m/K]
Heat capacity	3696 [J/kg/K]
Shear modulus	2 [kPa]

allows the ultrasound field to be visualised, being able to image the change in the refractive index of a transparent medium (Neumann and Ermert 2006). For this work, the ultrasound transducer was immersed in a glass bath filled with water. The distance between the tip of the transducer and the bottom of the water bath was kept the same as the distance between the transducer and the bottom of the MEA well. The camera of the Schlieren system visualises the ultrasound waves as they are moving through the water. In this way, it is also possible to follow the temporal evolution of the ultrasound field. The ultrasound transducer was driven with the same input voltage used during the cells experiment, and in Fig. 3C, the interference pattern generated by the reflections of the ultrasound beam on the bottom of the water bath in which the transducer is

submerged is shown. Comparing Fig. 3B and C, the similarity between the COMSOL simulated field and the one pictured with the Schlieren setup is evident. Nevertheless, as reported in (Andrews 2014), the exact interpretation of Schlieren images is rather difficult and should be considered as a validation and illustration of the results from the FEM simulations.

Validation of the temperature increase

To assess the increase in temperature at the location of the cells during the experiments, a K type thermocouple with a sensitivity of $\pm 0.1^\circ\text{C}$ was placed at the bottom of one of the wells of the multiwell MEA plate. The well was filled with water and placed on a temperature-controlled surface, at 36°C . The ultrasound transducer was then placed inside the well and turned ON for 60 seconds with the same input voltage as during the cell experiments. This resulted in a maximum temperature increase of 0.8°C . The temperature returned to the starting value within 5 minutes after the ultrasound was turned OFF. In addition, the Schlieren system can also be used to visualise heat profiles since temperature differences reflect in different densities of the medium, hence different refractive indexes. The ultrasound transducer was placed in the same Schlieren setup configuration used to

Fig. 3 Schlieren setup. **A)** View of the complete setup. **B)** Result of ultrasound field simulation in COMSOL. **C)** Interference pattern generated by the reflection of the ultrasound field on a hard surface. The distance between the ultrasound transducer and the reflecting surface is the same as the distance between the transducer and the bottom of the MEA well (6 mm). **D)** Heat profile generated by the ultrasound transducer at intensity 4 times higher than the experimental conditions. The shape of the thermocouple is also visible

validate the ultrasound field, and the same thermocouple as the one used for the first experiment was placed close to the tip of the transducer. The ultrasound transducer was then turned ON for 60 seconds with the same input voltage as during the cell experiments. However, due to the small temperature increase, the heat profile was difficult to visualise due to the resolution limits of the Schlieren system, therefore a third experiment was performed driving the ultrasound transducer with an input voltage 4.5 times higher than the one used during the cell experiments, to better visualise the heat profile. Figure 3D depicts the resulting heat profile at 60 seconds. The maximum increase in temperature was of 3.3°C. In the last two experiments, the objective was to assess the heat profile and not the actual temperature rise, since the liquid volume of the bath used in the Schlieren system is much larger than the one of the MEA well. Moreover, it can be

noticed that the heat profile is shallow to the tip of the transducer and the water surface, suggesting that, even at a much higher intensity, the heat does not reach the neuronal culture and is dispersed at the liquid surface rather than towards the neurons. The video recordings associated with these experiments are publicly available at (Saccher et al. 2021).

Validation of the fUS system for neuromodulation on multiwell MEA

The focused ultrasound system was tested on mouse primary hippocampal neurons. Preceding the exposure to ultrasound, activity of the neuronal culture was recorded for 5 minutes as a baseline measurement in seven MEA wells. During the periods of US-ON (15 second or 60 second period), neurons initially showed an increased firing rate, as shown in Fig. 4A and B, immediately followed by

Fig. 4 Ultrasound neuromodulation results. A) Example of raster plot of one of the wells used in the experiment, for a treatment duration of 15 seconds. Dark green shaded area indicates the US-ON period, while stripes-patterned area indicates the US-OFF period, namely the post period. B) Same as A) but for a treatment duration of 60 seconds. Dark green shaded area indicates the first 15 seconds of US-ON period and light green area indicates the remaining 45 seconds. Stripes-patterned area indicates the US-OFF period. C) Spike rate trend corresponding to the raster plot in A), calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the US-ON period and stripe-patterned area indicates the US-OFF period. D) Spike rate trend corresponding to the raster plot in B), calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the first 15 seconds of US-ON and light green area the remaining 45 seconds. Stripe-patterned area indicates the period of US-OFF

a reduction or even a pause in firing, after which the neurons resumed their baseline firing pattern. However, the response to the stimulus showed some differences in the two conditions. In the 15 second condition, the increased firing pattern was observed during the total stimulation time, followed by a pause in firing. In the 60 second condition, only in the first 15 seconds of the 60 second stimulus, the firing frequency was increased. In the subsequent 45 seconds stimulus, the firing frequency already showed a strong reduction (Fig. 4D). The duration of the post-stimulus pause appeared to last longer in the 60 second than in the 15 second condition (Fig. 4C and D), however a precise calculation of this variable was not possible due to the low n and the variability in the activity recorded by the electrodes. The responses of the other wells are presented in Supplemental Figure 1 & 2. These results indicate that the ultrasound stimulation can affect neuronal activity.

Discussion

In this work, the use of a multiwell plate in combination with an ultrasound delivery system to study the influence of focused ultrasound on the neural activity was shown. Other studies performed in vitro ultrasound neuromodulation yet using a wide variety of experimental setups, mostly ignoring the presence of reflections and standing waves patterns. In FUS neuromodulation applications, reflections and standing waves are relevant because the actual pressure amplitude or intensity level to which the neurons are exposed is different from what can be measured in an acoustic measurement setup. This is because measuring the actual pressure field in the real experimental conditions is very difficult, if not impossible, since the tools used to perform these measurements cannot be used in vivo or easily integrated in in-vitro experiments. Mediums with different densities or mechanical properties such as bones, skin or other organs cause standing waves and reflections due to their different acoustic impedances. For example in transcranial FUS, reflections are created at the boundaries between the transducer and the skull, and the skull and the brain, especially at frequencies above 1 MHz (Pinton et al. 2011). In addition, ultrasound waves reflect off the opposite side of the skull creating standing waves and pressure hotspots in the brain at locations other than the target point (Younan et al. 2013). As a consequence, the actual ultrasound field at the target location is different from what could be measured in an experimental setup. The uncontrolled presence of standing waves in an experiment makes the comparison of the results from different studies unreliable, hence the need to more accurately measure and analyse the real experimental conditions, which, in this work, was done with finite element methods simulations and the use of Schlieren imaging.

The MEA plate and the ultrasound delivery system were integrated and characterised with pressure measure-

ments and simulations of the experimental conditions. The simulations highlighted the presence of standing waves between the transducer and the bottom of the MEA well. This is due to the impedance mismatch between the cell medium and the two hard surfaces, namely the quartz glass and the ultrasound transducer, making the acoustic wave reflect back and forth between them. The resulting pattern depends on the distance between the two elements. Locations at $\lambda/2$ experience constructive interference (nodes) while locations at λ experience destructive interference (antinodes), with λ the acoustic wavelength. Standing waves increased the pressure inside the MEA well, compared to the pressure measured with the hydrophone at the same location and in a water tank. This was also observed in another study, (Secomski et al. 2017) and it could be potentially used to reduce the intensity generated by the ultrasound transducer.

Acoustic pressure is converted into radiation pressure due to absorption. This generates a pattern of stretching and compression at the location of the nodes and antinodes, which can influence some ionic channels or the membrane capacitance, with an impact on the neural activity. In addition to mechanical effects, heat is also a possible neuromodulation mechanism that is able to affect the spiking rate, amplitude and duration of action potentials. However, previous studies suggest that for temperature variations of less than a degree, the effect on neuromodulation is negligible (Tyler et al. 2008; Kim et al. 2019; Khraiche et al. 2008; Tufail et al. 2010), while a temperature increase of a few degrees was found to cause a decrease in spiking activity (Schiff and Somjen 1985; Gulick et al. 2017). In our experiments, the results of the measurements with a thermocouple indicate that the temperature increase was 0.8°C, and that the temperature resettled to the initial value within 5 minutes, hence during the recovery time of the neuromodulation treatment. In addition, the measurement of the heat profile with the Schlieren system suggests that most of the heat does not reach the level of the neuronal culture excluding thermal damage and thermal effects as potential mechanisms for neuromodulation.

The functioning of the newly designed system was validated with a proof-of-concept experiment on primary hippocampal neurons, showing that ultrasound affected the activity of the neurons. It was observed that the neuronal spiking activity increased during the period of US-ON. These results are comparable to what was observed by (Khraiche et al. 2008) using ultrasound at 7.75 MHz at an intensity higher than 50 W/cm². A period of reduced spiking activity followed each ultrasound treatment, which could be due to a feedback mechanism from interneurons, also present in the cultures. This phenomenon (of reduced firing) can also be observed after electrical stimulation of the neurons, although the dura-

tion of the electrical stimuli is usually 3 orders of magnitude lower than the ones used for ultrasound neuromodulation in this experiment (Hales et al. 2010). The response of the hippocampal neurons was consistent across three consecutive treatments and, in the recovery period following each ultrasound treatment, the spiking activity tended to values similar to the baseline, as observed in other studies (Hee-Sok et al. 2014; Choi et al. 2013). However, the strength of the response differed when looking at different wells (Supplemental Figure 1 & 2). What exactly causes this difference in response is difficult to say based on the low n . However, the baseline firing pattern of the neurons differed across the wells, which could be due to, amongst others, the density of the neurons plated in the dish and/or the number of dead cells. This could potentially be an explanation for the variation in responses to FUS. Future experiments will investigate how differences in cell densities and other biological factors influence the responses seen with FUS.

Previous studies showed that FUS at 43 MHz can activate the mechanosensitive Piezo1, but not the voltage-gated sodium channel $\text{Na}_V1.2$ (Prieto et al. 2018). Additionally, using the same FUS frequency on hippocampal slices, an effect was found on the resting membrane potential and the shape of the action potential of CA1 pyramidal neurons (Prieto et al. 2020), suggesting that FUS indeed is capable of altering neuronal characteristics. The setup presented in this work can be used in future studies to gain more insight into whether the effects of different stimulation frequencies and ultrasound parameters can differentially affect neuronal activity. Primary hippocampal cultures of different knock-out or knock-in mouse models can be used, or neurons can be transduced with different shRNAs to knock down specific channels or proteins involved in neuronal signalling. Finally, specific inhibitors or enhancers of specific channels can be added to disentangle the mechanism underlying the effect of FUS on neuronal firing.

The system was proven to be easy to assemble and operate, given the simplicity of the positioning of the ultrasound transducer in each well, which leaves small margins for alignment errors and ensures the reproducibility of the experimental conditions across the different wells. The system was also made free from biological contamination due to the presence of the Parafilm^M film that isolates the ultrasound transducer from the cell medium yet not influencing the ultrasound field. The Parafilm^M layer also electrically isolates the recording electrodes from the ultrasound transducer, ensuring that the response of the neurons is influenced only by the ultrasound field.

Conclusions

In this work, the integration of a focused ultrasound delivery system with a multiwell MEA plate was presented.

The system was characterised by pressure measurements, Schlieren imaging and finite element methods simulations. This method could be used to correct and estimate the effects of a certain setup geometry and experimental conditions to find the optimal conditions for an experiment. A proof-of-concept experiment was performed on a primary hippocampal neurons culture, demonstrating the potential of this system for investigations of focused ultrasound neuromodulation. Future work will focus on studying the effects of ultrasound on neurons using different sets of parameters and higher ultrasound frequencies. The focusing effect of ultrasound on the response of the neurons will also be investigated, and future experiments will also explore the effects on different types of cells.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s42234-021-00083-7>.

Additional file 1: Additional results of ultrasound neuromodulation. A) Raster plot of a well with a treatment duration of 15 seconds. Dark green shaded area indicates the US-ON period, while stripes-patterned area indicates the US-OFF period, namely the post period. B) Spike rate trend corresponding to the raster plot in A), calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the US-ON period and stripe-patterned area indicates the US-OFF period. C,E) Same as A) but for a treatment duration of 60 seconds. Dark green shaded area indicates the first 15 seconds of US-ON period and light green area indicates the remaining 45 seconds. Stripes-patterned area indicates the US-OFF period. D,F) Spike rate trend corresponding to the raster plot in C) and E) respectively, calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the first 15 seconds of US-ON and light green area the remaining 45 seconds. Stripe-patterned area indicates the period of US-OFF.

Additional file 2: Additional results of ultrasound neuromodulation. A) Raster plot of a well with a treatment duration of 15 seconds. Dark green shaded area indicates the US-ON period, while stripes-patterned area indicates the US-OFF period, namely the post period. B) Spike rate trend corresponding to the raster plot in A), calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the US-ON period and stripe-patterned area indicates the US-OFF period. C) Same as A) but for a treatment duration of 60 seconds. Dark green shaded area indicates the first 15 seconds of US-ON period and light green area indicates the remaining 45 seconds. Stripes-patterned area indicates the US-OFF period. D) Spike rate trend corresponding to the raster plot in C), calculated for 15 seconds bins over the duration of the experiment. Dark green shaded area indicates the first 15 seconds of US-ON and light green area the remaining 45 seconds. Stripe-patterned area indicates the period of US-OFF.

Acknowledgements

The authors would like to thank the researchers working at Philips for helping in the performing the pressure profile measurements and fabricating some parts of the experimental setup.

Authors' contributions

MS and SK designed and performed the experiments. MPO and GvW supervised the experimental work involving the neuronal culture. MS performed the simulations and analysed all the experimental data. All authors contributed to writing the manuscript. The authors read and approved the final manuscript.

Funding

This work is funded by the Moore4Medical project. Moore4Medical received funding from the ECSEL JU, under the grant agreement

H2020-ECSEL-2019-IA-876190. GMvW is funded by NWO-VIDI (016.Vidi.188.014).

Availability of data and materials

The dataset used during the current study is available from the corresponding author on reasonable request. The video recordings associated with the experiments are available in the 4TU.ResearchData repository, <https://doi.org/10.4121/17055329.v2>.

Declarations

Ethics approval and consent to participate

All animal experiments were approved by the Erasmus MC Institutional Animal Care and Ethical Committee, in accordance with European and Institutional Animal Care and Use Committee guidelines.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹Department of Microelectronics, Delft University of Technology, Delft, Netherlands. ²Department of Neuroscience, Erasmus MC, Rotterdam, Netherlands. ³Department of Clinical Genetics, Erasmus MC, Rotterdam, Netherlands. ⁴Fraunhofer Institute for Reliability and Microintegration IZM, Berlin, Germany. ⁵Philips Research, Eindhoven, Netherlands.

Received: 27 October 2021 Accepted: 6 December 2021

Published online: 27 January 2022

References

- Andrews D. Modelling of ultrasonic transducers and ultrasonic wave propagation for commercial applications using finite elements with experimental visualization of waves for validation. In: 2014 COMSOL Conference in Cambridge. Stockholm: COMSOL Multiphysics; 2014.
- Banker G, Goslin K. Culturing Nerve Cells. Bradford Book. Cambridge: The MIT Press; 1991.
- Blackmore J, Shrivastava S, Sallet J, R. Butler C, Cleveland RO. Ultrasound neuromodulation: A review of results, mechanisms and safety. *Ultrasound Med Biol*. 2019;45(7):1509–36. <https://doi.org/10.1016/j.ultrasmedbio.2018.12.015>.
- Chiappalone M, Novellino A, Vajda I, Vato A, Martinoia S, van Pelt J. Burst detection algorithms for the analysis of spatio-temporal patterns in cortical networks of neurons. *Neurocomputing*. 2005;65-66:653–62. <https://doi.org/10.1016/j.neucom.2004.10.094>.
- Choi JB, Lim SH, Cho KW, Kim DH, Jang DP, Kim IY. The effect of focused ultrasonic stimulation on the activity of hippocampal neurons in multi-channel electrode. In: 2013 6th International IEEE/EMBS Conference on Neural Engineering (NER); 2013. p. 731–34. <https://doi.org/10.1109/NER.2013.6696038>.
- Gulick DW, Li T, Kleim JA, Towe BC. Comparison of electrical and ultrasound neurostimulation in rat motor cortex. *Ultrasound Med Biol*. 2017;43(12):2824–33. <https://doi.org/10.1016/j.ultrasmedbio.2017.08.937>.
- Hales CM, Rolston JD, Potter SM. How to culture, record and stimulate neuronal networks on micro-electrode arrays (meas). *J Visualized Exp*. 2010;39: <https://doi.org/10.3791/2056>.
- Harvey EN. The effect of high frequency sound waves on heart muscle and other irritable tissues. *Am J Physiol-Legacy Content*. 1929;91(1):284–90. <https://doi.org/10.1152/ajplegacy.1929.91.1.284>.
- Hee-Sok H, Hwang S, Akram F, Jeon H, Nam S, Jun SB. Neural activity modulation via ultrasound stimulation measured on multi-channel electrodes. In: World Congress on Engineering, WCE 2014. Lecture Notes in Engineering and Computer Science, vol. 1. Hong Kong: Newswood Limited; 2014. p. 604–07.
- IT'IS F. Acoustic Properties. 2021. <https://itis.swiss/virtual-population/tissue-properties/overview/>. Accessed 07 Jan 2021.
- Jerusalem A, Al-Rekabi Z, Chen H, Ercole A, Malboubi M, Tamayo-Elizalde M, Verhagen L, Contera S. Electrophysiological-mechanical coupling in the neuronal membrane and its role in ultrasound neuromodulation and general anaesthesia. *Acta Biomater*. 2019;97:116–40. <https://doi.org/10.1016/j.actbio.2019.07.041>.
- Kamimura HAS, Conti A, Toschi N, Konofagou EE. Ultrasound neuromodulation: Mechanisms and the potential of multimodal stimulation for neuronal function assessment. *Front Phys*. 2020;8: <https://doi.org/10.3389/fphy.2020.00150>.
- Keller JM, Frega M. Past, present, and future of neuronal models in vitro. In: *Advances in Neurobiology*. Springer; 2019. p. 3–17. https://doi.org/10.1007/978-3-030-11135-9_1.
- Khraiche ML, Phillips WB, Jackson N, Muthuswamy J. Ultrasound induced increase in excitability of single neurons. In: 2008 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society. IEEE; 2008. p. 4246–49. <https://doi.org/10.1109/iembs.2008.4650147>.
- Kim H, Kim S, Sim NS, Pasquinelli C, Thielscher A, Lee JH, Lee HJ. Miniature ultrasound ring array transducers for transcranial ultrasound neuromodulation of freely-moving small animals. *Brain Stimulation*. 2019;12(2):251–55. <https://doi.org/10.1016/j.brs.2018.11.007>.
- Kim H-B, Swanberg KM, Han H-S, Kim J-C, Kim J-W, Lee S, Lee CJ, Maeng S, Kim T-S, Park J-H. Prolonged stimulation with low-intensity ultrasound induces delayed increases in spontaneous hippocampal culture spiking activity. *J Neurosci Res*. 2016;95(3):885–96. <https://doi.org/10.1002/jnr.23845>.
- Lay F-Y, Chen P-Y, Cheng H-F, Kuo Y-M, Huang C-C. Ex vivo evaluation of mouse brain elasticity using high-frequency ultrasound elastography. *IEEE Trans Biomed Eng*. 2019;66(12):3426–35. <https://doi.org/10.1109/tbme.2019.2905551>.
- Lin Z, Zhou W, Huang X, Wang K, Tang J, Niu L, Meng L, Zheng H. On-chip ultrasound modulation of pyramidal neuronal activity in hippocampal slices. *Adv Biosyst*. 2018;2(8):1800041. <https://doi.org/10.1002/adbi.201800041>.
- Menz MD, Ye P, Firouzi K, Nikoozadeh A, Pauly KB, Khuri-Yakub P, Baccus S. Radiation force as a physical mechanism for ultrasonic neurostimulation of the ex vivo retina. *J Neurosci*. 2019;39(32):6251–64. <https://doi.org/10.1523/jneurosci.2394-18.2019>.
- Multiphysics C. Understand, Predict, and Optimize Physics-Based Designs and Processes with COMSOL Multiphysics. <https://www.comsol.nl/comsol-multiphysics>. Accessed 07 Jan 2021.
- Naor O, Krupa S, Shoham S. Ultrasonic neuromodulation. *J Neural Eng*. 2016;13(3):031003. <https://doi.org/10.1088/1741-2560/13/3/031003>.
- Neumann T, Ermet H. 4c-4 a new designed schlieren system for the visualization of ultrasonic pulsed wave fields with high spatial and temporal resolution. In: 2006 IEEE Ultrasonics Symposium. IEEE; 2006. p. 244–47. <https://doi.org/10.1109/ULTSYM.2006.74>.
- Pinton G, Aubry J-F, Fink M, Tanter M. Numerical prediction of frequency dependent 3d maps of mechanical index thresholds in ultrasonic brain therapy. *Med Phys*. 2011;39(1):455–67. <https://doi.org/10.1118/1.3670376>.
- Prieto ML, Firouzi K, Khuri-Yakub BT, Madison DV, Maduke M. Spike-frequency dependent inhibition and excitation of neural activity by high-frequency ultrasound. *bioRxiv*. 2020. <https://doi.org/10.1101/2020.06.01.128710>. <https://www.biorxiv.org/content/early/2020/08/18/2020.06.01.128710.full.pdf>.
- Prieto ML, Firouzi K, Khuri-Yakub BT, Maduke M. Activation of piezo1 but not na_v1.2 channels by ultrasound at 43 mhz. *Ultrasound Med Biol*. 2018;44(6):1217–32.
- RedPitaya. Swiss army knife for engineers. <https://www.redpitaya.com/>. Accessed 07 Jan 2021.
- Rinaldi PC, Jones JP, Reines F, Price LR. Modification by focused ultrasound pulses of electrically evoked responses from an in vitro hippocampal preparation. *Brain Res*. 1991;558(1):36–42. [https://doi.org/10.1016/0006-8993\(91\)90711-4](https://doi.org/10.1016/0006-8993(91)90711-4).
- Saccher M, Kawasaki S, Proietti Onori M, van Woerden GM, Giagka V, Dekker R. Video recordings associated with the publication: Focused ultrasound neuromodulation on MEA. 4TU.ResearchData. 2021. <https://doi.org/10.4121/17055329.v2>. https://data.4tu.nl/articles/dataset/Video_recordings_associated_with_the_publication_Focused_ultrasound_neuromodulation_on_MEA/17055329/2.
- Schiff SJ, Somjen GG. The effects of temperature on synaptic transmission in hippocampal tissue slices. *Brain Res*. 1985;345(2):279–84. [https://doi.org/10.1016/0006-8993\(85\)91004-2](https://doi.org/10.1016/0006-8993(85)91004-2).
- Secomski W, Bilmin K, Kujawska T, Nowicki A, Grieb P, Lewin PA. In vitro ultrasound experiments: Standing wave and multiple reflections influence

on the outcome. *Ultrasonics*. 2017;77:203–13. <https://doi.org/10.1016/j.ultras.2017.02.008>.

Tufail Y, Matyushov A, Baldwin N, Tauchmann ML, Georges J, Yoshihiro A, Tillery SIH, Tyler WJ. Transcranial pulsed ultrasound stimulates intact brain circuits. *Neuron*. 2010;66(5):681–94. <https://doi.org/10.1016/j.neuron.2010.05.008>.

Tyler WJ, Tufail Y, Finsterwald M, Tauchmann ML, Olson EJ, Majestic C. Remote excitation of neuronal circuits using low-intensity, low-frequency ultrasound. *PLoS ONE*. 2008;3(10):3511. <https://doi.org/10.1371/journal.pone.0003511>.

Younan Y, Deffieux T, Larrat B, Fink M, Tanter M, Aubry J-F. Influence of the pressure field distribution in transcranial ultrasonic neurostimulation. *Med Phys*. 2013;40(8):082902. <https://doi.org/10.1118/1.4812423>.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions