

Long distance measurement with a femtosecond laser based frequency comb

Bhattacharya, N.; Cui, M.; Zeitouny, M. G.; Urbach, H. P.; Van Den Berg, S. A.

DOI

[10.1117/12.2309260](https://doi.org/10.1117/12.2309260)

Publication date

2018

Document Version

Final published version

Published in

International Conference on Space Optics, ICSO 2012

Citation (APA)

Bhattacharya, N., Cui, M., Zeitouny, M. G., Urbach, H. P., & Van Den Berg, S. A. (2018). Long distance measurement with a femtosecond laser based frequency comb. In E. Armandillo, B. Cugny, & N. Karafolas (Eds.), *International Conference on Space Optics, ICSO 2012* (Vol. 10564). [105643F] SPIE. <https://doi.org/10.1117/12.2309260>

Important note

To cite this publication, please use the final published version (if applicable).
Please check the document version above.

Copyright

Other than for strictly personal use, it is not permitted to download, forward or distribute the text or part of it, without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license such as Creative Commons.

Takedown policy

Please contact us and provide details if you believe this document breaches copyrights.
We will remove access to the work immediately and investigate your claim.

International Conference on Space Optics—ICSO 2012

Ajaccio, Corse

9–12 October 2012

Edited by Bruno Cugny, Errico Armandillo, and Nikos Karafolas



Long distance measurement with a femtosecond laser based frequency comb

N. Bhattacharya

M. Cui

M. G. Zeitouny

H. P. Urbach

et al.



Long distance measurement with a femtosecond laser based frequency comb.

N. Bhattacharya (n.bhattacharya@tudelft.nl),
M. Cui (m.cui@tudelft.nl),
M. G. Zeitouny (m.zeitouny@tudelft.nl)
H. P. Urbach (h.p.urbach@tudelft.nl)

Technische Universiteit Delft, Lorentzweg 1,
2628 CJ Delft, The Netherlands.

S. A. van den Berg (SvdBerg@vsl.nl)

VSL, Thijsseweg 11, 2629 JA Delft, The Netherlands.

Recent advances in the field of ultra-short pulse lasers have led to the development of reliable sources of carrier envelope phase stabilized femtosecond pulses. The pulse train generated by such a source has a frequency spectrum that consists of discrete, regularly spaced lines known as a frequency comb. In this case both the frequency repetition and the carrier-envelope-offset frequency are referenced to a frequency standard, like an atomic clock. As a result the accuracy of the frequency standard is transferred to the optical domain, with the frequency comb as transfer oscillator. These unique properties allow the frequency comb to be applied as a versatile tool, not only for time and frequency metrology, but also in fundamental physics, high-precision spectroscopy, and laser noise characterization. The pulse-to-pulse phase relationship of the light emitted by the frequency comb has opened up new directions for long range highly accurate distance measurement.

For the case of distance metrology a frequency comb can also be viewed as a ruler, which is based on the fact that the vacuum distance between subsequent pulses is known with the accuracy of the used time standard. We have recently demonstrated absolute distance measurements using a frequency comb laser applying a cross-correlation technique, which was supported by a theoretical and a numerical study on the formation of cross-correlation in dispersive media. We have also studied a distance measurement scheme based on dispersive (spectral) interferometry using the frequency comb. For the measurements of distances up to 50 meter, we achieve accuracy of 1 μm .

In this work we will present two experimental schemes to demonstrate distance measurements of up to 50 metres in air. The first experiment is done by analyzing the cross correlations measured between pulses emitted from a stabilized frequency comb source which are fed into a Michelson interferometer. To understand our results we implemented a model of pulse propagation in air to account for the effect of dispersion due to air on the measured cross-correlation functions. The measurement results obtained with the frequency comb and a conventional counting laser interferometer are compared. In a following experiment we apply frequency-comb based dispersive interferometry for measuring the distance of 50 m (100 m propagation) in air. During both the measurements, the temperature, the pressure, and the humidity have been monitored. The variation and the uncertainty on the measurement of these environmental parameters limit the accuracy in measuring the absolute distances. The measured distances in air agree with reference measurements to within 1 μm .

The frequency comb has the advantage of a large non-ambiguity range and depending on the line width of the particular comb source, coherence lengths of kilometers or more are feasible. Therefore making this technique a good candidate for measuring very long distances, especially in space where the dispersion of air is absent. One promising application is absolute distance metrology between satellites or for formation flying in space.

M. Cui, M. G. Zeitouny, N. Bhattacharya, S. A. van den Berg, H. P. Urbach and J. J. M. Braat,
Opt. Lett. **34**, 1982 (2009).

M. G. Zeitouny, M. Cui, N. Bhattacharya, S.A. van den Berg, A. J. E. M. Janssen and H. P. Urbach,
Phys. Rev. A **82**. 023808 (2010).

M. Cui, M. G. Zeitouny, N. Bhattacharya, S. A. van den Berg and H. P. Urbach,
Opt. Express **19**, 6549 (2011)